



**CORRECTION**



# Correction: Anterior segment optical coherence tomography in ocular surface tumours and simulating lesions

Ahmet Kaan Gündüz , Ibadulla Mirzayev , Aylin Okcu Heper, Işınsu Kuzu, Zarifakhanım Gahramanli, Cevriye Cansız Ersöz, Ömür Özlenen Gündüz and Ömür Ataoğlu

© The Author(s), under exclusive licence to The Royal College of Ophthalmologists 2023

*Eye* (2023) 37:3517–3518; <https://doi.org/10.1038/s41433-023-02541-9>

Correction to: *Eye* <https://doi.org/10.1038/s41433-022-02339-1>, published online 13 December 2022

Due to a typesetting mistake, the column headings of Table 1 have originally appeared incorrectly. Additionally, the table citation has been corrected from “Table adapted from Medina

CA, Plesec T, Singh AD. Optical coherence tomography imaging of ocular and periocular tumours. *Br J Ophthalmol.* 2014;98 Suppl 2:ii40–ii46” to “Table adapted from Ang M, Baskaran M, Werkmeister RM, Chua J, Schmidl D, Aranha Dos Santos V, et al. Anterior segment optical coherence tomography. *Prog Retin Eye Res* 2018;66:132–156”. The original article has been corrected.

**Table 1.** Summary of features of various commercially available OCT systems used for anterior segment examination<sup>†</sup>.

OCT type	Manufacturer	Measurement principle	Optical source	Axial resolution (µm)	Transverse resolution (µm)	Scan speed (A-scans per second)	Scan depth (mm)	Maximum scan width (mm)
<b>Visante OCT*</b>	Carl Zeiss Meditec, Dublin, CA	Time-domain	SLD 1310 nm	18	60	2000	6	16
<b>Slit Lamp OCT*</b>	Heidelberg Engineering, Heidelberg, Germany	Time-domain	SLD 1310 nm	<25	20–100	200	7	15
<b>Cirrus OCT<sup>‡</sup></b>	Carl Zeiss Meditec, Dublin, CA	Spectral-domain	SLD 840 nm	5 <sup>‡</sup>	15 <sup>‡</sup>	27,000	2	6
<b>Stratus OCT<sup>‡</sup></b>	Carl Zeiss Meditec	Time-domain	SLD 820 nm	10 <sup>‡</sup>	20 <sup>‡</sup>	400	2	N/A
<b>Spectralis OCT<sup>§</sup></b>	Heidelberg Engineering, Heidelberg, Germany	Spectral-domain	SLD 820 nm	7 <sup>‡</sup>	20 <sup>‡</sup>	40,000	2	6
<b>OCT SLO<sup>§</sup></b>	Optos	Spectral-domain	SLD 830 nm	<6 <sup>‡</sup>	20 <sup>‡</sup>	27,000	2	N/A
<b>3D OCT 2000<sup>§</sup></b>	Topcon Corporation, Tokyo, Japan	Spectral-domain	SLD 840 nm	5–6 <sup>‡</sup>	20 <sup>‡</sup>	50,000	2.3	6
<b>SOCT Copernicus<sup>§</sup></b>	Optopol	Spectral-domain	SLD 850 nm	3 <sup>‡</sup>	12–18 <sup>‡</sup>	52,000	2	10
<b>Optovue iVue<sup>§</sup></b>	Optovue, Inc, Fremont, CA	Spectral-domain	SLD 840 nm	5 <sup>‡</sup>	15 <sup>‡</sup>	26,000–80,000**	2–2.3	13
<b>Nidek RS 3000<sup>§</sup></b>	Nidek, Gamagori, Japan	Spectral-domain	SLD 880 nm	7 <sup>‡</sup>	15 <sup>‡</sup>	53,000	2	8
<b>Revo NX<sup>§</sup></b>	Optopol, Zawiercie, Poland	Spectral-domain	SLD 830 nm	5 <sup>‡</sup>	18 <sup>‡</sup>	110,000	2.4	16
<b>Anterior*</b>	Heidelberg Engineering, Heidelberg, Germany	Swept-source	Swept-source laser 1300 nm	<10	<30	50,000	14	16.5
<b>CASIA SS-1000 OCT*</b>	Tomey Corporation, Nagoya, Japan	Swept-source	Swept-source laser 1310 nm	10	30	30,000	6	16
<b>Triton OCT<sup>§</sup></b>	Topcon Corporation, Tokyo, Japan	Swept-source	Swept-source laser 1050 nm	8 <sup>‡</sup>	20 <sup>‡</sup>	100,000	3	3–12

OCT Optical coherence tomography. SLD Superluminescent light emitting diode. N/A Not available.

<sup>†</sup>Table adapted in part from Ang M, Baskaran M, Werkmeister RM, Chua J, Schmidl D, Aranha Dos Santos V, et al. Anterior segment optical coherence tomography. Prog Retin Eye Res 2018;66:132–156.

<sup>‡</sup>The quoted resolution values refer to posterior segment examination.

<sup>§</sup>Hybrid OCT device: for both anterior and posterior segment examination.

\*For anterior segment examination only.

\*\*The range refers to earlier and latest versions of the OCT equipment.