www.bjcancer.com

A simple, reproducible method for monitoring the treatment of tumours using dynamic contrast-enhanced MR imaging

B Morgan^{*,1}, JF Utting², A Higginson³, AL Thomas¹, WP Steward¹ and MA Horsfield⁴

¹Department of Cancer Studies and Molecular Medicine, University of Leicester, Leicester Royal Infirmary, Leicester LE1 5WW, UK; ²Department of Medical Physics, University Hospitals of Leicester NHS Trust, Leicester LE I 5WW, UK. ³Department of Radiology, University Hospitals of Leicester NHS Trust, Leicester LE1 5WW, UK; ⁴Department of Cardiovascular Sciences, University of Leicester, Leicester Royal Infirmary, Leicester LE1 5WW, UK

Dynamic contrast-enhanced MR imaging (DCE-MRI) may act as a biomarker for successful cancer therapy. Simple, reproducible techniques may widen this application. This paper demonstrates a single slice imaging technique. The image acquisition is performed in less than 500 ms making it relatively insensitive to respiratory motion. Data from phantom studies and a reproducibility study in solid human tumours are presented. The reproducibility study showed a coefficient of variation (CoV) of 19.1% for K^{trans} and 15.8% for the initial area under the contrast enhancement curve (IAUC). This was improved to 16 and 13.9% if tumours of diameter less than 3 cm were excluded. The individual repeatability (the range within which individual measurements are expected to fall) was 30.6% for K^{trans} and 26.5% for IAUC for tumours greater than 3 cm diameter. This approach to DCE-MRI image acquisition can be performed with standard clinical scanners, and data analysis is straightforward. For treatment trials with 10 patients in a cohort, the CoV implies that the method would be sensitive to a treatment effect of greater than 18%. The individual repeatability is well inside the 40% change shown to be important in clinical studies using this DCE-MRI technique. British Journal of Cancer (2006) 94, 1420–1427. doi:10.1038/sj.bjc.6603140 www.bjcancer.com

Published online 2 May 2006 © 2006 Cancer Research UK

Keywords: dynamic contrast-enhanced magnetic resonance imaging; DCE-MRI, biomarker; clinical trial

Dynamic contrast-enhanced magnetic resonance imaging (DCE-MRI) is the continuous acquisition of MR images over a period of between 1 and 20 min after the injection of a contrast agent into a peripheral vein, typically in the antecubital fossa. Normally, this involves a gadolinium chelate, which causes an increase in signal intensity on T_1 -weighted MR images. The increase in signal intensity (enhancement) is caused by the change in the nuclear spin relaxation rate R_1 ($R_1 = 1/T_1$), which is proportional to the concentration of the contrast agent. Enhancement, as measured by DCE-MRI, may be used to study the pathophysiology of tumours (Taylor et al, 1999; Yuh, 1999; Morgan et al, 2004). Parameters of interest include tumour vascularity, vascular endothelial permeability and the volume fraction of the extravascular, extracellular space (Knopp et al, 1999). There is an increasing interest in the use of these techniques as biomarkers in the development of novel cancer therapies (Jayson et al, 2002; Galbraith et al, 2003; Morgan et al, 2003; Leach et al, 2005; Rehman and Jayson, 2005; Thomas et al, 2005). Different approaches to the acquisition and analysis of DCE-MRI studies have been reviewed elsewhere (Morgan et al, 2004; Buckley and Parker, 2005; Parker and Buckley, 2005).

Ideally, DCE-MRI data acquisition should have high spatial resolution allowing mapping of enhancement parameters for the tumour in a pixel-by-pixel manner. This is important for gaining

*Correspondence: Dr B Morgan; E-mail: bm11@le.ac.uk

prognostic information about a tumour where small areas of increased tumour activity could otherwise be missed (Degani et al, 1997; Parker et al, 1997). This approach typically involves image acquisition times of between 6 and 30s for each image in the dynamic series. As a large percentage of metastatic disease treated by chemotherapy is in parts of the body that cannot be easily immobilised, such as the liver and lung, pixel-by-pixel data analysis is complicated by the need for sophisticated registration of the tumour in consecutive images. Long imaging times also involve multiple breath holds, which may be difficult for a patient with advanced cancer. This suggests an approach with short image acquisition times that can freeze motion and an analysis based on the whole tumour as the region of interest. A rapid imaging approach may also allow better depiction of the early part of the contrast enhancement curve. Successful systemic treatment may be expected to have an effect detectable by averaging the signal from the whole region of interest (ROI), but changes that are restricted to a small part of the tumour, or to poorly enhancing pixels, could be missed.

One problem when analysing tumour microcirculation is that signal intensities obtained from the MRI scanner are arbitrary and need to be 'standardised' if comparisons between different patients or patient visits are required. One possible method is by dividing the signal change due to gadolinium enhancement by the precontrast signal of the tumour (Mayr et al, 1999; Reddick et al, 1999). This has the disadvantage that the precontrast R_1 can change during therapy, for example with increasing tumour

Received 8 August 2005; revised 3 April 2006; accepted 3 April 2006; published online 2 May 2006

 (M_0) is given by:

(1) simplifies to

$$\frac{M_z}{M_0} = C^{(m-1)} \left\{ 1 - \text{EI} \left\{ 1 + \frac{\left[1 \text{ED} \left(1 - C^{Nk} (1 - \text{EI}) \right) + \frac{\text{ED} (\text{ER} - 1) \left(C^{Nk} - 1 \right)}{C - 1} \right]}{1 + \text{ED} \times C^{Nk} \times \text{EI}} \right\} \right\}$$
$$+ (1 - \text{ER}) \left(\frac{C^{(m-1)} - 1}{C - 1} \right)$$
(1)

where $\text{ER} = e^{-\text{TR}/T_1}$, $\text{EI} = e^{-\text{TI}/T_1}$, $\text{ED} = e^{-\text{TD}/T_1}$ and $\text{C} = \text{ER} \cos \alpha$. The sum of TI, TD and $(\text{TR} \times Nk)$ is TR_0 , the time between successive inversion pulses. As $\text{TR}_0 \rightarrow \infty$ and $\alpha \rightarrow 0$, then equation

$$\frac{M_z}{M_0} = 1 - 2e^{-TI_{\rm eff}/T_1}$$
(2)

where TI_{eff} is the time from the inversion pulse to the centre of k space. The signal intensity is given by $S = M_z \sin \alpha$.

In order to provide data that are comparable between patient visits, it is necessary to convert signal intensities to quantitative data. One method is to calculate the relaxation rate R_1 , which can be done by acquiring an image without the inversion pulse applied, after the contrast enhanced run, removing a major component of T_1 weighting. The ratio of signal intensities with (S) and without (S₀) the inversion pulse is then related to the R_1 of the tissue by equations (1) and (2), from which R_1 may be determined.

Phantom study

This approach to R_1 estimation was tested in a phantom study. The phantom consisted of vials of water doped with different concentrations of Gd-DTPA giving a range of R1 values from 0.4 to 10 s^{-1} . The R_1 values were accurately measured using an IRprepared turbo spin echo (TSE) sequence with TE = 10 ms, TR = 5000 ms, echo train length = 5, $128 \times 128 \text{ matrix}$, 1.56 mm in plane resolution, 5 mm slice thickness and inversion times of 23, 123, 323, 723 and 1520 ms. Phantom work was carried out using a Siemens symphony 1.5 T system (Siemens, Erlangen, Germany). Based on equation (1) this sequence effectively nulls signal at an R_1 of 0.53 s^{-1} . Where R_1 was less than 0.53 s^{-1} , the positive signal of the magnitude images was inverted. In a DCE-MRI study, rectified signal can be recognised if the signal reduces in the first few images of contrast enhancement. The estimated R_1 values, using the chosen DCE-MRI sequence and equations. (1) and (2), were compared with the accurately measured R_1 values.

Clinical study

Eleven patients with advanced cancers, including colorectal cancers with liver metastases, were recruited as part of two 'phase I' trials. Patients had two DCE-MRI scans 1 week apart, without treatment. All patients were informed about the investigational nature of the study according to institutional and regional guidelines, and subsequently gave informed consent before start of the study. Permission of local ethics regulatory bodies was obtained.

All MR images were acquired using a 1.5 T whole body magnet equipped with 25 mTm⁻¹ gradient coils and phased array surface coils (Siemens Magnetom Vision, Erlangen, Germany). All patients underwent standard transverse breath hold gradient recalled echo (GRE) T_1 and TSE T_2 -weighted imaging of the tumour region (GRE, typically TR = 150 ms, TE = 4 ms, $\alpha = 70^{\circ}$ and matrix 256 × 256; TSE typically TR = 5000 ms, TE = 90, $\alpha = 90^{\circ}$, echo train length = 23). In regions where respiratory motion was a factor, the T_1 -weighted multislice GRE examination was repeated during gentle respiration to improve slice positioning for the

oedema, causing an apparent change in enhancement (Evelhoch, 1999). A mechanism to standardise signal intensities is therefore required that is not sensitive to changes in pre-contrast R_1 .

We have shown that high temporal resolution DCE-MRI, using a single slice approach with a single ROI evaluating the whole tumour cross-section, performed without breath holding, is feasible and can act as a biomarker for the beneficial effects of an anti-angiogenesis agent (Morgan *et al*, 2003; Thomas *et al*, 2005). However, in order to plan such trials it is important to know the inherent variability of the results from the DCE-MRI approach in order to assess the significance of individual changes, and to use power statistics to determine the cohort sizes required to give statistical significance for a desired effect.

When planning a multicentre clinical trial for patients with advanced cancer, a DCE-MRI technique that is quick, does not involve multiple breath holds and is applicable in all areas of the body has obvious advantages. A validated, straightforward way of analysing the data from DCE-MRI would allow wider application of the technique. This study shows the rationale for developing a single slice DCE-MRI technique and investigates different approaches to standardising the measured signal intensities, and analysing the time course of signal intensities from an ROI that covers the whole tumour cross-section. The chosen imaging sequence was performed in cancer patients on two occasions 1 week apart, without treatment, in order to assess the reproducibility of several common approaches used for analysing DCE-MRI data.

The results of this reproducibility analysis were then put in a clinical context by re-analysing and comparing the data obtained in a previous published clinical trial that successfully used this technique as a biomarker (Thomas *et al*, 2005). In this dose escalating phase I study, the pharmacodynamic effect of PTK787/ZK22584 (PTK/ZK), an orally active inhibitor of vascular endothelial growth factor (VEGF) receptor tyrosine kinases, which inhibits VEGF-mediated angiogenesis, was evaluated.

MATERIALS AND METHODS

Pulse sequence

To achieve high temporal resolution and avoid the need for breath holding, we used scan parameters with individual image acquisition times of less than 500 ms, effectively freezing motion owing to respiration. To achieve this, a repetition time (TR) of 5 ms or less is necessary for 100 phase encoding steps. T_1 weighting can then be created by applying a magnetisation preparation pulse (Snapshot FLASH) (Haase, 1990). We used an inversion pulse for magnetisation preparation; this pulse was nonslice selective (i.e. affecting the whole bore of the magnet) to minimise the effects of through-plane motion and blood inflow. As respiratory motion in the abdomen and lower part of the lung is largely in the cranio-caudal direction, a coronal or sagittal oblique plane was used to keep the tumour within the imaged slice during the dynamic run. The parameters of the sequence used were: TR = 3.3 ms, echo time (TE) = 1.4 ms, flip angle (α) = 8°, k space matrix 100 phase encoding steps \times 128 points in the readout direction, inversion time (TI) = 655 ms, and time between successive inversion pulses $(TR_0) = 3000 \text{ ms.}$

Quantification of R_1

Assuming minimal transverse relaxation between the excitation pulses and signal acquisition, the contrast obtained from an inversion recovery snapshot FLASH sequence is affected by TR, TI, α , the number of lines of k space (Nk) and the delay between the end of image acquisition and the next inversion pulse (TD). From previous work (Jivan *et al*, 1997), the z magnetisation after m phase encoding steps relative to the equilibrium magnetisation **Clinical Studies**

Method for monitoring tumour treatment using DCE-MRI

B Morgan et al

dynamic series, which was also performed during gentle respiration. A representative disease site was selected (target lesion) and a single coronal oblique slice was planned to bisect the midline of the tumour and a major blood vessel (usually aorta). This became the imaged tumour for all subsequent scans. The DCE-MRI sequence was then run with 100 measurements, 3s apart $(T\hat{R}_0 = 3000 \text{ ms})$. Low molecular weight gadolinium-chelate 0.1 mmol kg^{-1} (Magnevist, Schering AG or Omniscan, Nycomed) was injected as a rapid bolus through an arm vein in less than 5 s by hand injection (a power injector was not available at the time of the study). This injection speed was adhered to for all patients. Both contrast agents display comparable relaxivity and distribution characteristics and therefore would be expected to provide similar enhancement parameters (Sze et al, 1991). The same contrast agent was always used for individual patients. Injection commenced after the first four measurements to allow magnetisation to reach a steady state. At the end of the dynamic run, the sequence was repeated with the inversion pulse switched off to acquire the S₀ image. For subsequent scans, patient positioning and slice orientation was replicated using the previous image set.

Signal intensities were taken from regions of interest drawn around the tumour and a major blood vessel (e.g. aorta) separately using an image analysis package, Analyze[™] (Mayo Clinic, Rochester, MN, USA). For the tumour, the ROI was drawn based on the precontrast images after review of the contrast-enhanced images so that large blood vessels and non-tumour tissues were avoided. For the major blood vessel, the ROI was drawn based on the peak arterial enhancement phase of imaging. The positions of the ROIs were corrected on a time point by time point basis for any movement during the dynamic run. The shapes of the ROIs were not changed throughout the dynamic run. The maximum diameter and appearance on T_1 - and T_2 -weighted images of the target lesion was recorded. The slice position for the dynamic run was checked on both T_1 - and T_2 -weighted images to ensure consistency between scans and care was taken to ensure that the ROIs had the same anatomical positions and sizes for both scans.

In clinical DCE-MRI, measurement of S_0 at every time point would result in unacceptably low temporal resolution. Therefore, we performed a single measurement of S_0 at the end of the dynamic run, by acquiring an image without the inversion preparation pulse. Equation (1) shows that measured S_0 depends on R_1 as there is still some T_1 weighting, owing to spin relaxation during the TR intervals, even when $TI \rightarrow \infty$ (the equivalent of turning the inversion pulse off). A single reading of S_0 therefore produces inaccurate estimates of R_1 when R_1 differs from its value at the time when S_0 is measured. An estimate of the maximum error introduced by the assumption of a constant S₀ was made using estimates of R_1 precontrast and at maximum enhancement. Equation (1) was used to predict the true S_0 at each of these time points. R_1 calculations were then performed using equation (2), firstly with the single measured S_0 value and second using the true S_0 values (corrected for R_1). This error analysis was performed for typical enhancement curves seen in the clinical study.

The signal intensities were preprocessed in three forms as summarised in Table 1. First, the raw signal intensities were used unchanged. In this form, any variability caused by different coil sensitivity or receiver gain on the two visits, or by different contrast bolus characteristics remained uncorrected. Secondly, the signal intensities were divided by the initial area under the arterial enhancement curve (IAUC-A) for a defined period of time. The IAUC-A was calculated by first subtracting the average precontrast signal intensities in the artery from the arterial time series, then using the trapezium rule to calculate the total area under the resulting change in enhancement out to times of 60 and 180 s (IAUC-A(60) and IAUC-A(180)) after contrast injection. Thirdly, the signal intensities were converted to R_1 values using equation 2) and the S_0 image collected after the end of the dynamic run.
 Table I
 Summary of parameters derived from two DCE-MRI scans, performed I week apart, without therapeutic intervention

Mean chang Parameter (%)		CoV (%)	Repeatability (%)		
Unscaled signal intensity data					
PE [′]	21.2	48.6	99.5		
IAUC(60)	15.4	50. I	98.0		
IAUC(180)	17.0	50.0	98.7		
Signal intensity related to measured arterial input function					
PE/IALIC_A(180)	12.0	35.7	70.4		
IAUC(60)/	6.6	37.2	70.3		
IAUC(180)/	8.3	37.8	72.1		
K ^{trans}	4.6	34	75.0		
R ₁ data					
PE	1.0	15.9	29.5		
IAUC(60)	1.7	15.8	29.5		
IAUC(180)	0.9	16.1	29.9		
K ^{trans}	3.1	19.1	36.1		

PE = peak enhancement; IAUC(t) = initial area under the *tumour* contrast enhancement curve for first *t* seconds; IAUC-A(t) = Initial area under the *arterial* contrast enhancement curve for first *t* seconds.

These three data sets were then used to calculate peak enhancement and the initial area under the enhancement curve (calculated as above) for 60 and 180 s (IAUC(60) and IAUC(180)). The bidirectional transfer constant K^{trans} was also calculated using a two compartment model (Tofts *et al*, 1999). As the imaging parameters are optimised for tumour rather than arterial enhancement, it was not possible to calculate R_1 accurately for the arterial ROI. K^{trans} was therefore measured in two ways: first using the raw signal intensities from the tumour, and a standard data set for aortic input, similar to previous published data (Figure 1) (Tofts and Kermode, 1991; Fritz–Hansen *et al*, 1996). The timing of the start of this standard arterial input function (AIF) was measured from the artery imaged in the study.

The data obtained in a previous published clinical trial that successfully used this technique as a biomarker (Thomas *et al*, 2005) were re-evaluated. Dynamic contrast-enhanced MR imaging was performed pre-treatment, on day 2, and at the end of the 28-day cycle to show changes in the contrast enhancement parameters of tumours in 35 patients with advanced cancer. *K*^{trans}, IAUC(60) and IAUC(180) were calculated. These data are presented as per cent change from the pretreatment value after treatment and are then compared with pharmacokinetic end points.

Statistical analysis

Dynamic contrast-enhanced MR imaging is applied to assess changes caused by treatment, with patients acting as their own controls. Previous experience shows a wide interpatient coefficient of variation (CoV) of 61% in pretreatment values for K_i (Morgan *et al*, 2003), equivalent to K^{trans} (Tofts *et al*, 1999). This considerable variability in the pre-treatment enhancement of tumours makes the percentage rather than absolute changes of enhancement parameters more amenable to analysis. Statistical analysis was therefore performed on the percentage changes of the observed parameter between the two scans. The percentage change data were tested for normality of distribution by the Shapiro–Wilk

Method for monitoring tumour treatment using DCE-MRI B Morgan et al



Figure I Change in R₁ against time used to represent a standard AIF for a 0.1 mmol kg⁻ gadolinium chelate injected as a bolus. The continuous line represents the AIF calculated using a formula from a model (Tofts and Kermode, 1991). The dotted line is our standard AIF created from locally acquired data, and modified based on published data (Fritz-Hansen et al, 1996) for high concentrations obtained in the early part of the enhancement curve owing to the contrast bolus.

test. The mean change and CoV were calculated and expressed as percentages (Bland and Altman, 1996a). Assuming that the posttreatment variability is similar, the CoV can be used to assess the statistical power of studies, or anticipate the patient numbers required for a study to demonstrate a given degree of treatment effect. The repeatability value was also evaluated (Bland and Altman, 1996a), which is defined as the range within which 95% of measurements will fall, assuming no treatment effect, and is therefore helpful in assessing the significance of individual patient results.

RESULTS

Pulse sequence, quantification of R_1 and phantom study

Figure 2 shows simulated signal intensities for the IR snapshot FLASH sequence, using equation (1) and the simplified equation (2) for the sequence described above. Although there is considerable difference between the signal intensities predicted by the two expressions, Figure 3 shows that R_1 can be estimated accurately for R_1 values between 1 and $5 s^{-1}$ using either equations (1) and (2). The marked error occurring at R_1 values above $5 \, \text{s}^{-1}$ is due to almost complete relaxation during the TI interval when S approaches S_0 , as shown in Figure 2. Thus, with this sequence, selection of sequence parameters to accurately measure relatively short R_1 compromises the ability to measure long R_1 values accurately. The chosen TI_{eff} is therefore a compromise based on the pre-contrast and range of R_1 seen in tumours during DCE-MRI.

Clinical study

Of the 11 patients recruited, one patient was excluded owing to incorrect positioning of the imaging slice on the second scan. Despite breathing motion, it was possible to maintain the size and shape of the tumour and arterial ROI in all remaining cases. Figure 4 demonstrates an example of the images obtained for a liver metastasis. Although all slices included a major artery (aorta or iliac artery), ghosting from motion artefact in the phase-



Figure 2 Signal intensities for the IR snapshot FLASH sequence predicted using equation (1) and the simplified equation (2). TR = 3.3 ms, $TE = 1.4 \text{ ms}, \alpha = 8^{\circ}, TI = 655 \text{ ms}, TR_0 = 3000 \text{ ms}.$



Figure 3 Comparison of R_1 estimated from the DCE-MRI sequence using equations (1) and (2) and the 'true' value measured by IR-prepared turbo spin echo sequence in a phantom with $R_1 = 0.5 - 10 \text{ s}^{-1}$ $0.5 - 2 \text{ s}^{-1}$ (**B**). (**A**) and

encoded direction was not a significant problem, presumably due to rapid image acquisition and the nonselective inversion pulse. The minimum observed precontrast R_1 , averaged over the whole

1423

B Morgan et al



Figure 4 Image set showing a T_1 (**A**) and T_2 (**B**)-weighted image, demonstrating the tumour and the planned dynamic slice position (white line). Image (**C**) is the post contrast S_0 image. The dynamic contrast enhancement series is represented by a precontrast (**D**), and early and late contrast enhanced images (**E** and **F**). The inset image in (**D**) shows how the ROI is drawn for analysis.



Figure 5 The R_1 time courses for four DCE–MRI series, covering the range of contrast enhancement and tumour size. The highest variation in R_1 between data points is seen for tumour A, which is the smallest measured tumour. Although the smoothness of the enhancement curve is related to tumour size, all cases are amenable to analysis.

ROI, was 0.95 s^{-1} ($T_1 = 1.05 \text{ s}$). The maximum R_1 was 3.4 s^{-1} ($T_1 = 0.3 \text{ s}$).

From Figure 5, it can be seen that the smoothness of the curves improves with increasing tumour size. As S_0 is not continuously measured as R_1 changes, errors are introduced in the dynamic measurement of R_1 , which can be estimated as described earlier. For lesion A (precontrast $R_1 = 1 \text{ s}^{-1}$, maximum $R_1 = 3.4 \text{ s}^{-1}$, and R_1 at S_0 signal acquisition = 3 s^{-1}), the error in the maximum change of estimated R_1 (2.4 s^{-1}) is 0.13 s^{-1} or 5.4%. For lesion B, there is gradually increasing enhancement throughout (precontrast $R_1 = 1 \text{ s}^{-1}$, maximum $R_1 \sim 2 \text{ s}^{-1}$) and the error in the maximum change of estimated R_1 (1 s^{-1}) is approximately 0.009 s^{-1} , or less than 1%. For lesions C and D, the maximum estimated error is also less than 1%. As tissue contrast enhancement changes slowly at the end of the dynamic sequence (Figure 5), and the S_0 sequence is only mildly sensitive to these changes, the signal intensity in S_0

image will change very little in the period from the end of the dynamic image sequence to the acquisition of the S_0 image.

In this study, there was correlation between the difference in the enhancement parameters of repeated measurements in a single patient and their mean value for that patient (for $K^{\text{trans}} r = 0.42$, Pearson's correlation coefficient). This causes a skew of the 'normal' distribution and a transformation of the data should be performed in these circumstances (Bland and Altman, 1996b, c). As stated in the Materials and Methods section, the data were transformed to percentage difference from the first scan to the second. There was no significant correlation between the percentage difference and the mean measured parameters and therefore no further data transformation was required. For all listed parameters expressed as percentage change, there was no significant evidence against a normal distribution (Shapiro–Wilk test). The mean difference, CoV and repeatability are shown for all measured parameters in Table 1.

In this study, use of raw signal intensities for calculating peak enhancement and IAUC resulted in high CoV, which is improved by dividing by the arterial IAUC. Even though the peak enhancement is taken from just a single data value along the enhancement curves, rather than an integration of many more data points as is the case with IAUC, the reproducibility is very similar to IAUC. This is a reflection of the smoothness of the tumour enhancement curves. The K^{trans} calculated from tumour and arterial signal intensities showed a higher CoV (34%) than the K^{trans} from tumour R_1 and a standard AIF (19.1%). This is probably a reflection of the fact that any real variation in the AIF between the two scans is outweighed by the fact that the AIF is measured inaccurately with this sequence.

The individual patient data for two commonly used parameters, K^{trans} and IAUC[60], calculated from R_1 values, are given in Table 2. Although no correlation was seen between T_2 signal intensity and enhancement parameters, the second case in Table 2 had very high T_2 compared with the other cases, consistent with a cystic nature of the metastasis. Guidelines from a recent US national cancer institute workshop on DCE-MRI state that tumours in a fixed superficial location should be at least 2 cm in diameter (2004). This study shows a tendency for greater variability with reducing

Site	Primary	Size (cm)	K^{trans} (min ⁻¹)			IAUC(60)		
			Scan I	Scan 2	% change	Scan I	Scan 2	% change
Liver	Colorectal	7	0.116	0.111	-4.3	10.1	9.7	-3.7
Liver	Colorectal	2.8	0.018	0.014	-22.2	1.2	1.1	-15.2
Liver	Colorectal	4.7	0.116	0.100	- 4.0	9.3	8.2	-11.6
Liver	Colorectal	6	0.186	0.192	3.2	16.6	15.1	-9.1
Liver	Colorectal	10.4	0.037	0.044	18.9	3.6	4.2	15.6
Liver	Colorectal	17	0.085	0.081	-4.7	6.7	6.7	-0.5
Liver	Lung	4.6	0.039	0.05	30.8	3.5	4.3	24.1
Lung	Lung	2.5	0.231	0.298	29.0	17.4	21.0	20.8
Lung	Lung	7	0.152	0.171	12.5	12.6	14.5	15.0
Lymph node	Melanoma	2.4	0.49	0.40	-18.5	35.0	28.7	-18.0
			K ^{trans}		IAUC(60)			
All cases	Mean change (%)		3.1		1.7			
	CoV (%)		19.1		15.8			
	Repeatability (%)		36.1		29.5			
Size > 3 cm	Mean change (%)		6.1		4.3			
	CoV (%)		15.5		13.9			
	Repeatability (%)		30.6		26.5			

Table 2 Individual patient data showing tumour size, mean difference, coefficient of variation (CoV) and repeatability for K^{trans} and IAUC(60) for two scans, I week apart, without therapeutic intervention

IUAC = initial area under the contrast enhancement curve. Parameters were calculated using R1 values and a standardised arterial input function for K^{trans}.

size, and excluding lesions less that 3 cm in diameter reduced CoV. The colorectal liver metastases group also had lower CoV and repeatability values (K^{trans} 14.2 and 26.5% and IAUC(60) 11 and 21.3%, respectively), although this may be related to the fact that this group had relatively larger tumours.

Review of our previous study using DCE-MRI as a biomarker for the effect of antiangiogenesis therapy in advanced cancer (Thomas et al, 2005) showed that for 43 recruited patients with advanced cancer, in four out of 39 available patients the DCE-MRI was unsuccessful. In three cases, the ROI could not be accurately followed through the image series (all these tumours were 3 cm or smaller) and in the other case the imaging slice was planned using a breath hold image rather than during gentle respiration, causing misregistration of a 3.1-cm tumour. For the analysed cases, the difference in K^{trans} correlated with the pretreatment K^{trans} value, but, similar to this study, there was no correlation of the percentage difference to the pre-treatment K^{trans} value. There was good correlation between percent changes on treatment seen in IAUC and K^{trans} (Figure 6) calculated from R_1 values. After 1 day (3 doses) of treatment with PTK/ZK, there were similar significant reductions in K^{trans} of 46±6.3% (mean±s.e. of the mean (s.e.m.)) and IAUC(60) of 43.5 \pm 6.5%. For colorectal liver metastases receiving 1000 mg or more of PTK/ZK per day (n = 14), the effect was increased with a reduction in K^{trans} of 55.8±5.6% and IAUC(60) of $55.1 \pm 4.9\%$, again both showing a similar treatment effect.

DISCUSSION

Dynamic contrast-enhanced MR imaging using an inversion recovery snapshot FLASH sequence with a relatively long TI_{eff} of 815 ms and an interval of 3 s between subsequent images, performed in a coronal oblique plane provides a rapid, reasonably accurate measure of R_1 for values between 1 and 5 s⁻¹, and therefore of Gd-DTPA concentration during contrast enhancement. The R_1 values seen in the clinical arm of this study were within this range. As S_0 is not continuously measured as R_1 changes, there are errors of up to 6% in the estimated R_1 . Systematic errors also occur in estimating dynamic parameters based on signal intensity, as MRI signal intensity is not



Figure 6 A comparison of the percentage change of the DCE-MRI parameters IAUC and K^{trans} after treatment with PTK/ZK, as calculated from R_1 values.

proportional to Gd-DTPA concentration. These errors do not necessarily affect the reproducibility of contrast enhancement parameters but may affect the assessment of change in enhancement owing to treatment.

Using a long TI_{eff} is helpful as it provides good contrast to noise for low levels of enhancement such as those that may occur in necrotic tumours. However, this has the disadvantage that arterial contrast concentration cannot be accurately measured owing to rapid relaxation during the long TI_{eff} period. Sequences designed to measure blood concentration of contrast agent (Fritz-Hansen *et al*, 1998) have a very short TI_{eff} (150 ms), which may show little or no enhancement in necrotic tumours as they often have a relatively short precontrast R_1 .

Measurement of arterial Gd-DTPA concentration (AIF) is considered important when assessing absolute values of tumour

microcirculation (Evelhoch, 1999). However, the AIF may not be so important for measuring a treatment effect in patients, unless there is a change in cardiac function or haemodynamics on therapy. Techniques that allow accurate, reproducible measurement of the AIF as well as tissue enhancement may well improve the reproducibility of these parameters, but this study shows that suboptimal measurement of AIF makes the reproducibility of the study worse than if a standardised AIF is used.

The simplified equation (equation (2)) used to convert signal intensities into R_1 values does not take into account the signal saturation due to successive inversion pulses. If the TR₀ were to be shortened, the estimated R_1 would become progressively more inaccurate. Imaging of tumours with a short precontrast R_1 can be a problem with this sequence as in some cases enhancement may be negative in the first few images of the dynamic series. In a DCE-MRI study, rectified signal can be recognised if the signal reduces in the first few images of contrast enhancement. In practice, this is rarely a problem in liver metastases and this correction was not required in the clinical arm of this study as precontrast R_1 was always greater than 0.53 s^{-1} . This problem could be overcome by using saturation recovery rather than inversion recovery, at the expense of lower contrast at shorter R_1 values.

Although respiratory motion causes the tumour to move throughout the series of image acquisitions, the use of the coronal oblique plane minimises through plane motion and repositioning the ROI for each image accounts for in-plane motion. The fact that there was no need to change the size or shape of the ROI suggests that the effect of through plane motion was indeed small. The use of a single region of interest for the whole tumour does not allow discrimination of regions within the tumour, but it does give a good signal-to-noise ratio by averaging the signal from a large volume of tissue and minimises partial volume effects. Acquiring multiple slices, or three-dimensional images, through the tumour and then averaging the tumour ROI for each slice would allow even better signal-to-noise but this would reduce the temporal resolution of the sequence.

The CoV of the repeated measures depends on the preprocessing (use of raw signal intensities or conversion to R_1), analysis method (peak enhancement, IAUC or K^{trans}) and patient cohort used. Peak enhancement is a simple measure, shown to have similar reproducibility to IAUC. However, peak enhancement is much more a reflection of the extravascular, extracellular space volume fraction (v_e) rather than K^{trans} , as changes in K^{trans} shift the position of the peak of the enhancement curve more than its amplitude, while peak enhancement is, in the limit of large K^{trans} , proportional to v_e . The validity of using peak enhancement as a biomarker for angiogenesis would therefore seem questionable.

The study shows it is important to use some method for 'standardising' the signal intensities from the dynamic image series. The results using calculated R_1 values for IAUC(60) and IAUC(180) and R_1 values with a standard AIF for K^{trans} , show a CoV ranging from 11 to 19.1%. Assuming the post-treatment CoV is similar, this value can be used to assess the statistical power of studies or anticipate the patient numbers required for a study. Interestingly, our previous published data (Morgan *et al*, 2003) show a 58% mean reduction of K_i (equivalent to K^{trans}) with an s.e.m. of 5.2% for 15 patients with colorectal liver metastases treated with PTK/ZK with 1000 mg or more. This is consistent with a CoV of 20% (the CoV is equivalent to the standard deviation (s.d.) of percent changes and the s.e.m. is the s.d. divided by square

root of the number of cases). Reanalysis of our other published work (Thomas *et al*, 2005) for patients with liver metastases treated with PTK/ZK with 1000 mg or more (n = 14) showed a 56% mean reduction in K^{trans} and a CoV of 21.1%. A CoV of ~20%, for both the reproducibility study and treatment effect, implies a cohort of 10 patients would be expected to show a 25% treatment effect with statistical significance (power 0.8). If the CoV is reduced to 14%, as for IAUC(60) in tumours greater than 3 cm diameter, then an 18% treatment effect is expected then only 3–4 patients would be required to expect a statistically significant result to P < 0.05.

Evelhoch et al (2004) measured the median IAUC parameter (similar to our IAUC(60)) in 19 human tumours with a 7.2-s image acquisition time. They demonstrated pretreatment tumour interpatient CoV of 64% and intrapatient CoV, in repeated measurements without treatment, of 18%, similar to this study. The high CoV for interpatient tumour measurement supports the notion of assessing percentage rather than absolute changes. Galbraith et al (2002) assessed reproducibility in 16 patients with tumours 3 cm in diameter or greater. They use an 11 s image acquisition time. Their data are presented in a slightly different manner and uses both pixel-by-pixel and ROI analysis. For ROI analysis, the data can be summarised to show that for a cohort of 16 patients, IAUC can measure greater than 12% changes and K^{trans} can measure 14–17% changes. Similarly, our data extrapolated for 16 patients and tumours 3 cm or greater, (IAUC(60) CoV = 14% and K^{trans} CoV = 16%) would be sensitive to 14 and 16% changes, respectively. Both studies use similar methodology and do not measure AIF, but our study has an image acquisition time of less than 500 ms as opposed to 7.2 and 11 s, dropping the requirement for multiple breath holds and increasing temporal resolution, but at the expense of signal to noise of any given image.

The repeatability varied from 26.5% for IAUC(60) (tumours of diameter greater than 3 cm) to 36.1% for K^{trans} (whole group). This is a measure of the significance of an individual result. From our previously published data (Morgan *et al*, 2003; Thomas *et al*, 2005), a 40% change in enhancement parameters is considered to be clinically significant (the change required to predict a tumour response in colorectal liver metastases). A 40% change in an individual patient can therefore be considered both a statistically and a clinically significant finding. Both K^{trans} and IAUC are shown to give similar results in the clinical application of this technique and the improved reproducibility of IAUC in this study suggests it is a valuable, straightforward method of evaluating contrast dynamics from DCE-MRI.

In this study, DCE – MRI failed in one patient owing to incorrect positioning of the slice. The incorrect placement was demonstrated by studying the reference slice on both T_1 - and T_2 -weighted images but was more apparent on T_2 -weighted imaging as central tumour necrosis could be seen. When selecting the target lesion, we suggest avoiding metastases with very high T_2 -weighted signal intensity to avoid purely necrotic/cystic tumours and to select metastatic deposits with a diameter of greater than 3 cm.

In summary, this technique provides a rapid, straightforward, robust method of measuring tumour enhancement to monitor therapy. All stages of analysis are simple to perform if equation (2) is used to calculate R_1 and IAUC is used to assess tumour enhancement. The speed of image acquisition freezes motion, allowing a wide variety of tumour applications. Also, as multiple breath holds are not required, the scanning protocol is easier both for patients and scanning technicians.

REFERENCES

Anonymous (2004) DCE-MRI group consensus recommendations, NCI CIP MR Workshop on Translational Research in Cancer – Tumor Response. http://imaging.cancer.gov/reportsandpublications/ReportsandPresentations/MagneticResonance



Clinical Studies

- Bland JM, Altman DG (1996a) Statistics notes: measurement error proportional to the mean 313, 106. Br Med J **313**: 744
- Bland JM, Altman DG (1996c) Transforming data. BMJ 312: 770
- Bland JM, Altman DG (1996b) Statistics notes: Measurement error proportional to the mean. 23. Br Med J 313: 106
- Buckley DL, Parker GJM (2005) Measuring contrast agent concentration in T1 weighted dynamic contrast-enhanced MRI. In *Dynamic Contrast-Enhanced Magnetic Resonance Imaging in Oncology*, Jackson A, Buckley DL, Parker GJM (eds), pp 69-79. New York: Springer
- Degani H, Gusis V, Weinstein D, Fields S, Strano S (1997) Mapping pathophysiological features of breast tumors by MRI at high spatial resolution. Nat Med 3: 780-782
- Evelhoch JL (1999) Key factors in the acquisition of contrast kinetic data for oncology. J Magn Reson Imaging 10: 254-259
- Evelhoch JL, LoRusso PM, He ZQ, DelProposto Z, Polin L, Corbett TH, Langmuir P, Wheeler C, Stone A, Leadbetter J, Ryan AJ, Blakey DC, Waterton JC (2004) Magnetic resonance imaging measurements of the response of murine and human tumors to the vascular-targeting agent ZD6126. *Clin Cancer Res* **10**: 3650-3657
- Fritz-Hansen T, Rostrup E, Larsson HBW, Sondergaard L, Ring P, Henriksen O (1996) Measurement of the arterial concentration of Gd-DTPA using MRI: a step toward quantitative perfusion imaging. *Magn Reson Med* 36: 225-231
- Fritz-Hansen T, Rostrup E, Ring PB, Larsson HBW (1998) Quantification of gadolinium-DTPA concentrations for different inversion times using an IR-turbo flash pulse sequence: a study on optimizing multislice perfusion imaging. *Magn Reson Imaging* **16:** 893–899
- Galbraith SM, Lodge MA, Taylor NJ, Rustin GJS, Bentzen S, Stirling JJ, Padhani AR (2002) Reproducibility of dynamic contrast-enhanced MRI in human muscle and tumours: comparison of quantitative and semiquantitative analysis. *NMR Biomed* **15**: 132-142
- Galbraith SM, Maxwell RJ, Lodge MA, Tozer GM, Wilson J, Taylor NJ, Stirling JJ, Sena L, Padhani AR, Rustin GJS (2003) Combretastatin A4 phosphate has tumor antivascular activity in rat and man as demonstrated by dynamic magnetic resonance imaging. J Clin Oncol 21: 2831-2842
- Haase A (1990) Snapshot flash MRI applications to T1, T2, and chemicalshift imaging. *Magn Reson Med* **13:** 77–89
- Jayson GC, Zweit J, Jackson A, Mulatero C, Julyan P, Ranson M, Broughton L, Wagstaff J, Hakannson L, Groenewegen G, Bailey J, Smith N, Hastings D, Lawrance J, Haroon H, Ward T, Mcgown AT, Tang M, Levitt D, Marreaud S, Lehmann FF, Herold M, Zwierzina H (2002) Molecular imaging and biological evaluation of HuMV833 anti-VEGF antibody: implications for trial design of antiangiogenic antibodies. J Natl Cancer Inst 94: 1484-1493
- Jivan A, Horsfield MA, Moody AR, Cherryman GR (1997) Dynamic T-1 measurement using snapshot-FLASH MRI. J Magn Res 127: 65-72
- Knopp MV, Weiss E, Sinn HP, Mattern J, Junkermann H, Radeleff J, Magener A, Brix G, Delorme S, Zuna I, van Kaick G (1999) Pathophysiologic basis of contrast enhancement in breast tumors. J Magn Reson Imaging 10: 260-266
- Leach MO, Brindle KM, Evelhoch JL, Griffiths JR, Horsman MR, Jackson A, Jayson GC, Judson IR, Knopp MV, Maxwell RJ, McIntyre D, Padhani AR, Price P, Rathbone R, Rustin GJ, Tofts PS, Tozer GM, Vennart W,

Waterton JC, Williams SR, Workmanw P (2005) The assessment of antiangiogenic and antivascular therapies in early-stage clinical trials using magnetic resonance imaging: issues and recommendations. *Br J Cancer* 92: 1599–1610

- Mayr NA, Hawighorst H, Yuh WTC, Essig M, Magnotta VA, Knopp MV (1999) MR microcirculation assessment in cervical cancer: Correlations with histomorphological tumor markers and clinical outcome. J Magn Reson Imaging 10: 267-276
- Morgan B, Horsfield MA, Steward WP (2004) The role of imaging in the clinical development of antiangiogenic agents. *Hematol-Oncol Clin North* Am 18: 1183 1206
- Morgan B, Thomas AL, Drevs J, Hennig J, Buchert M, Jivan A, Horsfield MA, Mross K, Ball HA, Lee L, Mietlowski W, Fuxius S, Unger C, O'Byrne K, Henry A, Cherryman GR, Laurent D, Dugan M, Marme D, Steward WP (2003) Dynamic contrast-enhanced magnetic resonance imaging as a biomarker for the pharmacological response of PTK787/ZK 222584, an inhibitor of the vascular endothelial growth factor receptor tyrosine kinases, in patients with advanced colorectal cancer and liver metastases: results from two phase I studies. *JClin Oncol* **21:** 3955–3964
- Parker GJM, Buckley DL (2005) Tracer kinetic modelling for T1 weighted DCE-MRI. In *Dynamic Contrast-Enhanced Magnetic Resonance Imaging in Oncology*, Jackson A, Buckley DL, Parker GJM (eds), pp 81–92. New York: Springer
- Parker GJM, Suckling J, Tanner SF, Padhani AR, Revell PB, Husband JE, Leach MO (1997) Probing tumor microvascularity by measurement, analysis and display of contrast agent uptake kinetics. J Magn Reson Imaging 7: 564-574
- Reddick WE, Taylor JS, Fletcher BD (1999) Dynamic MR imaging (DEMRI) of microcirculation in bone sarcoma. J Magn Reson Imaging 10: 277-285
- Rehman S, Jayson GC (2005) Molecular imaging of antiangiogenic agents. Oncologist 10: 92-103
- Sze G, Brant-Zawadzki M, Haughton VM, Maravilla KR, McNamara MT, Kumar AJ, Aisen AM, Dreisbach JN, Bradley Jr WG, Weinreb JC (1991) Multicenter study of gadodiamide injection as a contrast agent in MR imaging of the brain and spine. *Radiology* **181**: 693–699
- Taylor JS, Tofts PS, Port R, Evelhoch JL, Knopp M, Reddick WE, Runge VM, Mayr N (1999) MR imaging of tumor microcirculation: promise for the new millenium. J Magn Reson Imaging 10: 903-907
- Thomas AL, Morgan B, Horsfield MA, Higginson A, Kay A, Lee L, Masson E, Puccio-Pick M, Laurent D, Steward WP (2005) Phase I study of the safety, tolerability, pharmacokinetics, and pharmacodynamics of PTK787/ZK 222584 administered twice daily in patients with advanced cancer. *J Clin Oncol* **23**: 4162-4171
- Tofts PS, Brix G, Buckley DL, Evelhoch JL, Henderson E, Knopp M, Larsson HBW, Lee TY, Mayr NA, Parker GJM, Port RE, Taylor J, Weisskoff RM (1999) Estimating kinetic parameters from dynamic contrast-enhanced T- 1-weighted MRI of a diffusable tracer: standardized quantities and symbols. J Magn Reson Imaging 10: 223-232
- Tofts PS, Kermode AG (1991) Measurement of the blood—brain barrier permeability and leakage space using dynamic Mr imaging. 1. Fundamental-concepts. *Magn Reson Med* **17:** 357-367
- Yuh WTC (1999) An exciting and challenging role for the advanced contrast MR imaging. J Magn Reson Imaging 10: 221-222