Sir,

Salomon *et al.*¹ cite a publication of ours² as advocating 'administration of aspirin for prevention of a second event of nonarteritic anterior ischemic optic neuropathy (NAION).' This is a miscitation. On the contrary, we did not conclude, based on our data, that aspirin was beneficial. Rather we concluded that 'our findings suggest a possible short-term (1 to 2 years) benefit but little or no long-term (5 years) benefit in using aspirin to reduce the risk of NAION in the fellow eye.' We also advised the reader that our results be viewed with caution because our study was not conducted to prospectively assess the benefit of aspirin with a standard controlled protocol.

The authors' finding of a lower incidence of second eye NAION in aspirin-treated compared with non-treated patients is almost certainly due to chance. This apparent association is being driven by an implausibly high second-eye rate (50%) in their small number of non-treated patients ($n = 16$). We believe, based on our data from 431 patients with NAION, that the incidence of second-eye NAION within 5 years, without aspirin therapy, is likely to be only about 10–15%. In fact, the authors' life-table estimate of 20% for the 5-year incidence of second eye NAION in the aspirin-treated patients (325 mg/day) actually exceeds our estimate for non-aspirin-treated patients.

Although we share the authors' opinion that a prospective study is warranted, we noted in our article that the expected low incidence of second-eye NAION limits the feasibility of such a study.

References

1. Salomon O, Huna-Baron R, Steinberg DM, Kurtz S, Seligsohn U. Role of aspirin in reducing the frequency of second eye involvement in patients with non-arteritic anterior ischaemic optic neuropathy. *Eye* 1999;13:357–9.
2. Beck RW, Hayreh SS, Podhajsky PA, Tan ES, Moke PS. Aspirin therapy in nonarteritic anterior ischemic optic neuropathy. *Am J Ophthalmol* 1997;123:212–7.

Roy W. Beck, MD, PhD
Jaeb Center for Health Research
3010 East 138th Avenue
Suite 9
Tampa

FL 33613, USA
Tel: +1 (813) 975 8690
Fax: +1 (813) 975 8761

Sohan S. Hayreh MD, PhD
Iowa City, Iowa, USA

Sir,

In their letter Beck and Hayreh pointed out that the incidence of second eye NAION in our study¹ was higher than in their study² (50% versus 20%, respectively). However, another study by Hayreh's group³ reported a cumulative incidence of second eye NAION of 34.6% at 5 years and 56.7% at 10 years. Thus, the incidences of second eye involvement seem to vary from one study to another.

In our study the lower incidence of second eye involvement among those patients taking 325 mg/day aspirin was not statistically significant as we carefully pointed out ($p = 0.358$). Nonetheless, in our view the estimated 250% increase in mean time to second eye involvement in patients on aspirin (325 mg/day) appeared to be of clinical significance. An interesting contrast between our results and those of Beck *et al.*² is that we found that aspirin provided the greatest reduction in risk of second eye NAION from > 2 years after the first event, whereas they found the greatest benefit of aspirin treatment during the first 2 years after the first event. Obviously, only a large multicentre prospective study of patients with NAION with and without aspirin may resolve this important issue.

References

1. Salomon O, Huna-Baron R, Steinberg DM, Kurtz S, Seligsohn U. Role of aspirin in reducing the frequency of second eye involvement in patients with non-arteritic anterior ischemic optic neuropathy. *Eye* 1999;13:357–9.
2. Beck RW, Hayreh SS, Podhajsky PA, Tan ES, Moke PS. Aspirin therapy in nonarteritic anterior ischemic optic neuropathy. *Am J Ophthalmol* 1997;123:212–7.
3. Beri M, Klugman MR, Kohler JA, Hayreh SS. Anterior ischemic optic neuropathy. VII. Incidence of bilaterality and various influencing factors. *Ophthalmology* 1987;94:1020–8.

Ophira Salomon
Ruth Huna-Baron
David Steinberg
Shimon Kurtz

Uri Seligsohn
Chaim Sheba Medical Center
Tel-Hashomer, Israel

Uri Seligsohn, MD
Department of Hematology
Institute of Thrombosis & Hemostasis
Chaim Sheba Medical Center
Tel-Hashomer 52621, Israel
Tel: +972 3 5302104
Fax: +972 3 5351565

Sir,

The article by Murray *et al.* on biometry of the silicone oil-filled eye¹ both complicates and oversimplifies the issue of axial length estimation. The length of an object measured ultrasonically is the speed of sound through it multiplied by the time the sound wave takes to traverse it. Before biometry became as critical as it is now many machines used an average sound speed through an 'average' eye, usually 1550 metres per second (not 1532 m/s as the authors state). But this immediately fails once one is measuring other than an average eye. Thus in a long or a short eye, or in an eye with a thick or a thin lens, using that speed could not give a correct measurement. For example, in a long eye approximately 84% of the time the sound is travelling at 1532 m/s (sound speed in aqueous and vitreous) while in a short eye it is travelling at this speed for only 72% of the time. Therefore the 'average' speed in these two types of eye is very different. To calculate any distance where the speed varies we need to know for how long sound travelled at each speed and then calculate how far it travelled at each speed. The sum of all those little distances is the total distance travelled.

Many recently manufactured A-scan machines calculate axial length in this way. If one looks at Fig. 1 one can see that the axial length is the sum of the anterior chamber depth (ACD) plus the lens thickness (LT) plus the distance from the back of the lens to the retina. The machine has measured the time taken for the echo to travel to and from the front of the lens, divided it by 2 and multiplied it by 1532 giving a length of 2.76 mm. Similarly the same calculation for the lens at a sound speed of 1641 m/s gives a thickness of 4.75 mm. Finally the vitreous is measured at 14.36 mm in the same way. The axial length is therefore the sum of the three: 21.67 mm. However if the vitreous cavity is filled with silicone oil then the sound wave slows down to 987 m/s (in this

PHAKIC
 AVERAGE VALUE : 21.72 mm SD : 0.09
 AXIAL LENGTH : 21.67 mm
 1532m/s 1641m/s 1532m/s
 AC = 2.76 L = 4.55 V = 14.36

Fig. 1. Sound speed of vitreous internal set at 1532 m/s.

example). Fig. 2 shows the same scan now with the calculation for the vitreous cavity set at 987 m/s. The ACD and lens thickness are still the same but the vitreous cavity is now measured at 9.25 mm (obviously incorrect as in this eye the normal vitreous was present). If this eye were filled with silicone oil the correct vitreous interval would have been calculated. In Murray *et al.*'s example illustrated in fig. 1 of their article,¹ simply recalculating the vitreous length by dividing it by 1550 to get back the length of time it took the sound wave to traverse it and then multiplying it by 987 we get 14.84 mm (or 15.02 mm if we use 1532 m/s), which is very close to what they got after the oil was removed. However, there is an unexplained discrepancy in their measurements of the ACD and lens thickness in their example. These should be the same before and after silicone oil removal. In their pre-operative scan (fig. 1a) the ACD is 2.06 mm while it is 4.13 mm in their post-operative scan (fig. 1b). Similarly the lens is 6.15 mm in their pre-operative scan and 3.91 mm in their post-operative scan. These discrepancies are confirmed in their table 4, which shows large variations in what should be static measurements of ACD and LT. Unless the authors can explain these different measurements in the same eyes, doubt is cast on all measurements used in the paper. Furthermore, because of the large variations in the total length of ACD

PHAKIC
 AVERAGE VALUE : 16.59 mm SD : 0.08
 AXIAL LENGTH : 16.56 mm
 1532m/s 1641m/s 987 m/s
 AC = 2.76 L = 4.55 V = 9.25

Fig. 2. Sound speed of vitreous interval set at 987 m/s.

plus LT in these four patients (from 0.13 to 2.65 mm, average 0.94 mm), calculation of any ratio for the vitreous cannot be reliable. Even if the measurement of the vitreous cavity were accurate the authors have the potential for an average 4 D error in the intraocular lens (IOL) calculation in the anterior segment alone.

The authors also calculate IOL power using the two sets of measurements derived from their method. As the conversion factor is derived from the same set of figures, i.e. the apparent change in axial length following oil removal, the two sets of figures are dependent on each other. Therefore to point to the similarity of the results of IOL power calculation is no proof of the accuracy of their work.

Moving to the issue of IOL calculation in these cases, I believe it is important to take account of whether the oil is to be removed or not. If it is to be left behind then IOL calculation is even more difficult as the oil has a refractive index very different from vitreous and therefore the IOL power must be modified to take account of this. Holladay² suggests that one may need to add up to 3 D to the IOL power in such cases. Additionally the contour of the back surface of the IOL affects what refractive effect the IOL/oil interface will have, so that the calculation will change for different lenses.

Patients who require IOL power calculation in an eye with silicone oil in the vitreous cavity should have their eye measured by an experienced biometrist using a machine in which sound speeds can be varied and the gates marking the intraocular anatomy moved. The biometrist must be able to identify each acoustic interface and adjust the gates manually if required. Even then the surgeon must be aware that a successful refractive outcome may well entail lens exchange in the first week following surgery and must be prepared for this. Unfortunately not all the maculae in such patients who have needed silicone oil to reposition their retinae are capable of good visual acuity.

Biometry is set to assume its rightful place in cataract surgery. This group of patients present special difficulties which push current technology and expertise to its limit.

References

1. Murray DC, Potamitis T, Good P, Kirkby GR, Benson MT. Biometry of the silicone oil-filled eye. *Eye* 1999;13:319-24.
2. Holladay. Advanced IOL power calculations. ASCRS course, 29 April 1999, Seattle.

John Bolger
 BUPA Hospital
 Ambrose Lane

Harpenden
 Herts AL5 4BP, UK
 Fax: +44 (0)1582 462301

Sir,

We thank Mr Bolger for his comments on our paper.¹ The aim of the study was to determine whether, in the silicone oil-filled eye (viscosity 1300 centistokes), measured axial length (AL) varied from true AL by a constant factor. We found a remarkably consistent variation in this ratio, ranging from 0.70 to 0.73 in seven eyes (mean 0.71, SD 0.01). This could enable eyes undergoing combined phacoemulsification of cataract, lens implantation and removal of silicone oil to have desired lens power determined, since true AL can be estimated by simply multiplying measured AL of the silicone oil-filled eye by 0.71.

We accept that there are limitations with this method, not least of which is the fact that true AL may be underestimated in very long eyes. We are of the opinion, however, that possible errors are likely to be less than those incurred by using measurements based on the fellow eye, or by implanting an 'average' power or 'best guess' intraocular lens, as is common practice at present. Table 5 in our paper¹ shows that true AL and AL calculated using the conversion factor of 0.71 are quite similar, differences ranging from -0.61 to +0.39 mm, with mean percentage error being only -0.57%.

We appreciate the concern expressed about the differences in anterior chamber depth (ACD) and lens thickness (LT) in four patients prior to and following the removal of silicone oil. However, measurements were taken several weeks apart and this may account for the variation in ACD and LT in these four patients. There is inevitably inter-test variability due to variation in biometry parameters. This is especially true of ACD measurements, which can vary with the extent of corneal compression from forward or backward movement of the patient. ACD was found to have increased in three of four patients following removal of silicone oil. Backward displacement of the lens-iris diaphragm with secondary deepening of the anterior chamber in the vitrectomised eye may have contributed to this increase. In addition, rapid-onset nuclear sclerosis is a well-recognised complication of silicone oil use.² This may well account for the increase in LT in three of four patients. This, in combination with intra- and inter-observer variability, most likely explains the difference in ACD and LT prior to and following removal of silicone oil and further validates the use of an overall constant which encompasses ACD, LT and vitreous cavity depth (VCD) measurements. Mr Bolger