SUMMARY

A new method of calculating the specific airway resistance in a whole-body plethysmograph is proposed. The only data needed are the simultaneous pneumotachograph flow and plethysmograph box volume fluctuations, without the knowledge of alveolar pressure.

This will often be the only objective datum that can be obtained with a whole-body plethysmograph in nonsedated children under the age of 6 or 7 years.

REFERENCES AND NOTES

1. Bouhuys, A.: Breathing. In: Physiology, Environment and Lung Disease, p. 160 (Grune & Stratton, New York, 1974).
2. Briscoe, W. A., and Du Bois, A. B.: The relationship between airway resistance, airway conductance and lung volume in subjects of different age and body size. *J. Clin. Invest.*, 37: 1279, (1958).
3. Brough, F. K., Du Wayne Schmidt, C., Dickman, M., and Jackson, B.: Effect of two instructional procedures on the performance of the spirometry test in children five through seven years of age. *Amer. Rev. Resp. Dis.*, 106: 604, (1972).
4. Doershuk, C. F., Downs, T. D., Matthews, L. W., and Lough, M. D.: A method for ventilatory measurements in subjects 1 month-5 years of age: Normal results and observations in disease. *Pediat. Res.*, 4: 165, (1970).
5. Doershuk, C. F., Fisher, B. J., and Matthews, L. W.: Specific airway resistance from perinatal period into adulthood. *Amer. Rev. Resp. Dis.*, 109: 452 (1974).
6. Du Bois, A. B., Botelho, S. Y., Bedell, G. N., Marshall, R., and Comroe, J. H., Jr.: A rapid plethysmographic method for measuring thoracic gas volume: A comparison with a nitrogen washout method for measuring functional residual capacity in normal subjects. *J. Clin. Invest.*, 35: 322 (1956).
7. Du Bois, A. B., Botelho, S. Y., and Comroe, J. H., Jr.: A new method for measuring airway resistance in man using a body-plethysmograph: Values in normal subjects and in patients with respiratory disease. *J. Clin. Invest.*, 35: 327, (1956).
8. Lloyd, T. C., Jr., and Wright, G. W.: Evaluation of methods used in detecting changes of airway resistance in man. *Amer. Rev. Resp. Dis.*, 87: 529, (1963).
9. Matthys, H.: Lungenfunktionsdiagnostik mittels Ganzkörperplethysmographie, pp. 2-7 (F. K. Schattauer Verlag, Stuttgart, 1972).
10. Nolte, D., Reif, E., and Ulmer, W. T.: Die Ganzkörperplethysmographie. *Respiration*, 25: 14 (1968).
11. Payne, C. B., Jr., Chester, E. H., and Hsi, B. P.: Airway responsiveness in chronic obstructive pulmonary disease. *Amer. J. Med.*, 42: 554 (1965).
12. Radford, M.: Measurement of airway resistance and thoracic gas volume in infancy. *Arch. Dis. Childhood*, 49: 611 (1974).
13. Fenyves & Gut, Basel, Switzerland.
14. Sony three-head stereo recorder, Sony Instrument Co., Tokyo, Japan.
15. Hewlett-Packard 7041A, Hewlett-Packard, Avondale, Pa.
16. Compucorp 325 Scientist, Compucorp Corp., Los Angeles, Calif.
17. The writers thank Mrs. Monballiu for the careful assistance with the respiratory function tests.
18. These data were presented at the VIIIth Interasma-Congress, Vlissingen-De Haan (Netherlands and Belgium), October 1975.
19. This research was supported by Grant no. 3.0015.75, National Fund for Medical and Scientific Research in Belgium.
20. Requests for reprints should be addressed to: I. Dab, M.D., Universitair Ziekenhuis St. Pieter. Dienst Kindergeneeskunde, Hoogstraat 320, 1000 Brussels (Belgium).
21. Accepted for publication June 11, 1976.

Copyright © 1976 International Pediatric Research Foundation, Inc.

Printed in U.S.A.

Pediat. Res. 10: 999-1000 (1976)

Letter to the Editor

Q. T. SMITH,^{1,2} M. J. HAMILTON, AND B. L. SHAPIRO

Division of Oral Biology, School of Dentistry, University of Minnesota, Minneapolis, Minnesota, USA

Research on cystic fibrosis (CF), early detection of CF, and genetic counseling relative to CF would all benefit greatly from availability of a simple technique for detection of CF heterozygotes and homozygotes. Consequently, various investigators have devoted much effort in attempts to identify differences in body fluids, especially serum, from CF heterozygotes and homozygotes. The objective of several of these studies has been detection of an abnormal component, for example the "CF factor," which could be used for identification of affected individuals and carriers and which would be useful for the laboratory study of the pathophysiology of CF. Thus the recent reports of serum protein differences in CF (5) and the claim recently published in *Pediatric Research* (4) to have "developed a standardized biophysical assay for the rapid detection of individuals homozygous and heterozygous for C/F" were particularly significant. According to these papers a protein band with an isoelectric point of 8.4-8.5 was present following thin layer isoelectric focusing in polyacrylamide gel from serum of CF homozygotes and heterozygotes but not from control subjects. This protein was reported to be dissociated from IgG by 4 M urea and to have a molecular weight between 1,000 and 10,000. Because of our experience with isoelectric focusing and polyacryl-

amide gel electrophoresis and our interest in CF we have attempted to confirm the presence of the pH 8.4-8.5 isoelectric point protein in CF homozygotes. The results of these efforts are summarized below.

We examined serum proteins from 18 CF subjects and 14 control individuals according to the first procedure described by Wilson, Jahn, and Fonseca in 1973 (5) and as later modified and published in *Pediatric Research* in 1975 by Wilson, Fudenberg, and Jahn (4). The only difference in our technique of which we are aware is use of an LKB Multiphor electrophoresis apparatus rather than the equipment described by Williamson (3). This difference should not be significant since a linear and reproducible pH gradient was obtained with our apparatus and since large amounts of protein were present at pH 8.0-9.0 in the gels. We were unable to detect any consistent differences between the stained serum protein patterns of the CF and the control subjects. Figure 1 gives an example of the isoelectric focusing patterns in polyacrylamide gel of CF and control serum proteins and of the isoelectric focusing pH gradients obtained in our laboratory. These patterns show greater resolution of proteins than those previously published (4, 5), but, in contrast, no consistent differences between CF and

Fig. 1. Isoelectric focusing (pH 5–10 gradient) of serum proteins stained with Coomassie blue in a thin layer polyacrylamide gel for three control subjects (C1, C2, and C3) and from three cystic fibrosis homozygotes (CF1, CF2, and CF3). The pH gradient in the gel at 4° after isoelectric focusing is given on the right margin of the figure. The serum samples were applied to the gel on filter paper strips in the area devoid of distinct bands on the lower portion of the gel.

control subjects. Individual bands were evident in the heavily stained area at approximately pH 8.5 when the fresh gels were examined on a light box.

In addition, we have modified the published methods with procedures which would enhance the likelihood of electrophoretic detection of an abnormal protein with the reported properties. These modifications included: (1) increased quantities of serum; (2) pretreatment of samples with urea before isoelectric focusing; (3) concentration of the IgG fraction by Amicon ultrafiltration before isoelectric focusing; (4) complete processing of samples in plastic ware; and (5) subsection of the pH 7.0–9.0 isoelectric focusing

fraction to electrophoresis in a second dimension according to the procedure given by Altland *et al.* (1). A portion of the samples analyzed by the published procedures (4, 5) also were analyzed by each of these modifications. None of the modified procedures showed consistent differences between serum proteins from the CF and the control subjects. Thomas, Merritt, and Hodes (2) also found no differences between CF subjects and controls after application of isoelectric focusing procedures based on the reports of Wilson *et al.* (4, 5).

No explanation is readily apparent for the difference between results of Thomas *et al.* (2) and our own results on the one hand and the previously reported results on the other (4, 5). We attempted to rigorously reproduce the isoelectric focusing techniques given in the literature (4, 5). However, isoelectric focusing in polyacrylamide gel can be affected by small technical differences and, thus, factors essential to reproducing the techniques (4, 5) may not have been evident in the published procedure. Genetic heterogeneity of the disease and variations in patient treatment are possible but unlikely explanations for the differences between our findings and the earlier conclusions (4, 5).

We hope that other laboratories also will carefully evaluate the isoelectric focusing methods and results reported by Wilson *et al.* (4, 5) for the detection of CF homozygotes and heterozygotes. However, we believe that those who are considering use of this procedure should be aware of the problems others have experienced since significant effort is required to develop technical competency in isoelectric focusing and thin layer polyacrylamide gel techniques.

REFERENCES AND NOTES

1. Altland, K., Schmidt, S. R., Kaiser, G., and Knoche, W.: Demonstration of a factor in the serum of homozygotes and heterozygotes for cystic fibrosis by a non-biological technique. *Humangenetik*, 28: 207 (1975).
2. Thomas, J. M., Merritt, D., and Hodes, M. E.: Electrophoretic analysis of serum proteins in cystic fibrosis. In: *Cystic Fibrosis Club Abstracts*, p. 19 (Cystic Fibrosis Foundation, Atlanta, 1976).
3. Williamson, A. R.: Isoelectric focusing of immunoglobulins. In: D. M. Weir: *Handbook of Experimental Immunology*, Ed. 2, Chapt. 8 (Blackwell Scientific Publications, Oxford, 1973).
4. Wilson, G. B., Fudenberg, H. H., and Jahn, T. L.: Studies on cystic fibrosis using isoelectric focusing. I. An assay for detection of cystic fibrosis homozygotes and heterozygote carriers from serum. *Pediat. Res.* 9: 635 (1975).
5. Wilson, G. B., Jahn, T. L., and Fonseca, J. R.: Demonstration of serum protein differences in cystic fibrosis by isoelectric focusing in thin-layer polyacrylamide gels. *Clin. Chim. Acta*, 49: 79 (1973).
6. We acknowledge the cooperation and interest in these studies by Dr. Warren J. Warwick, Director of Cystic Fibrosis Clinic, University of Minnesota.
7. Requests for reprints should be addressed to: Q. T. Smith, Ph.D., Division of Oral Biology, School of Dentistry, University of Minnesota, Minneapolis, Minn. 55455 (USA).