Continuous Limb Blood Flow Estimation in the Newborn Using Electrical Impedance Plethysmography

KATE COSTELLOE¹ AND P. ROLFE

Department of Paediatrics, John Radcliffe Hospital, Headington, Oxford, England

Summary

This study aims to define a method for acquiring limb blood flow data from newborn babies on a continuous basis which would be suitable for automatic processing. Pulsatile electrical impedance and strain gauge signals were analyzed in 3 different ways. The results were compared with estimates of flow derived from venous occlusion plethysmography with a mercury-in-rubber strain gauge. Measurements were made from 33 babies. The impedance technique was easier to perform than the strain gauge technique. The reproducibility of the various estimates of flow is shown in Table 3. The beat-to-beat reproducibility of simultaneous pulsatile measurements and of repeated measurements on different subjects was greater from the impedance recordings than from the strain gauge recordings. The results of linear regression analyses between the different flow estimates are shown in Table 4. The best correlation with flow measured using venous occlusion plethysmography was with flow calculated from the impedance pulse height, $F(Z_{\max})$ ($r = 0.76$; $n = 30$; $P < 0.001$). The regression line between these 2 variables was not statistically significantly different from the line of identity, ($P < 0.98$), and the intercept is close to zero.

The measurement of the impedance pulse height, $\Delta Z_{\max}$, and the subsequent calculation of $F(Z_{\max})$ could be automated and provide continuous beat-to-beat blood flow estimations.

Speculation

A method has been defined which is suitable for continuous estimation of limb blood flow in newborn babies under intensive care. It is believed that this simple technique will help the development of full understanding of the circulatory changes associated with events in the neonatal period and thus help to lead towards further improvements in the treatment of such infants.

It is likely that the changes in cardiac output and its distribution which may be associated with apnoeic attacks and changes in blood gas tensions during the neonatal period are involved in the pathogenesis of permanent brain damage. To improve further the treatment of infants requiring intensive care, it is important that such circulatory changes are more fully understood, and to achieve this, there is a need for the development of simple, noninvasive methods for the continuous estimation of blood flow.

Venous occlusion plethysmography, using both air-filled (21, 23) and water-filled cuffs (5) or mercury in rubber strain gauges (9, 18, 19) has been used to estimate lower limb blood flow in the newborn. However, this method does not allow continuous measurements to be made, and furthermore, the essentially mechanical techniques used are cumbersome and often difficult to apply. This is particularly the case when long-term studies are required on babies undergoing intensive care who require minimal handling and who are likely already to be encumbered by the attachment of physiologic monitoring apparatus. There is an additional ob-

jection to venous occlusion plethysmography in that the repeated inflation and deflation of the occluding cuff may influence blood flow when multiple measurements are made rapidly.

Electrical impedance plethysmography has been used in the adult to estimate limb blood flow and by comparison with mechanical techniques has been found to be simpler to use and unencumbering to the subject (8). Estimates of limb blood flow have been derived both from venous occlusion plethysmography (12) and from the analysis of individual arterial pulses (2, 4, 15, 20). Such analyses include the straightforward measurement of pulse height (20) as well as more complex methods (2, 4, 15) which aim to correct for inaccuracies introduced by venous and arterial runoff. In addition to providing quantitative estimates of limb blood flow, the impedance technique has been used for making qualitative assessments of peripheral arterial disease (1, 6, 14).

The use of pulse analysis clearly offers the possibility to derive estimates of blood flow on a beat by beat basis and, as such, has an obvious advantage compared with venous occlusion plethysmography. The aim of this study was to assess the feasibility and performance of electrical impedance plethysmography for the estimation of limb blood flow in ill newborn infants. Estimations of flow were made using various forms of pulse analysis and these were compared with estimates of flow derived from venous occlusion plethysmography using a mercury in rubber strain gauge. The ultimate goal of the study was to develop a method of analysis which would be suitable for automatic processing to derive blood flow data on a continuous basis.

MATERIALS AND METHODS

ELECTRICAL IMPEDANCE MEASUREMENTS

A four-electrode system was used for impedance measurements. With this arrangement, a constant current is supplied to two outer current electrodes and the measurement of the voltage appearing between two inner voltage electrodes then provides an indication of electrical impedance. The instrument used in this study (P.M.S. Ltd., Slough, Berks, United Kingdom) produced a constant amplitude sinusoidal current of 2 mA rms at 100 kHz and had an output impedance of 13 kohm at a load of 150 ohms.

The electrical impedance of biologic tissue in which there is pulsatile blood flow consists of two components: a basal impedance, Z_0 ohm, with a superimposed pulsatile component, ΔZ ohm, which corresponds to pulsatile blood flow within the tissue. ΔZ is small compared with Z_0 , and for recording purposes Z_0 must be subtracted from the composite signal to allow adequate amplification of ΔZ . The instrument used in this study had an automatic balancing circuit to achieve this. Z_0 was displayed in digital form on the instrument, and electrical outputs were available for recording ΔZ and its first derivative, dZ/dt /ohms.

Measurements of the electrical impedance of infants' legs were made using disposable band electrodes. These consisted of self-adhesive tape having a strip of aluminum running along the center

(3M electrode tape). One current electrode was placed around the foot, and the second was placed across the lower abdomen. The voltage sensing electrodes were positioned around the calf (Fig. 1). The adhesive tape on the voltage electrodes was trimmed to leave a sufficiently wide area of the calf for attachment of a mercury-in-rubber strain gauge. All electrodes were applied using electrode paste (Hewlett-Packard Redux cream), and no measurements were made until the electrodes had been applied for at least 10 min. The importance of electrode position was investigated in a preliminary series of experiments by making recordings of ΔZ . The ratio of $\Delta Z_{\max}/Z_0$ was determined for different values of voltage sensing electrode spacing, L (Fig. 1).

MERCURY-IN-RUBBER STRAIN GAUGE MEASUREMENTS

Strain gauge measurements (22) were made using gauges constructed from mercury-filled silicone rubber tubing, internal diameter 0.5 mm, external diameter 1 mm. (Dow-Corning, Ltd.), electrical contact with the mercury column being made with copper pins. The gauge formed one arm of a Wheatstone bridge, which was operated at constant current. The 95% response time of the gauge and bridge together was found to be 16.5 msec, and the system will faithfully reproduce signals up to a frequency of 23 Hz.

Fig. 1. Diagram to show positions of impedance electrodes and strain gauge on infant's leg. A, sphygmomanometer cuff; B, strain gauge; I_1 , I_2 , impedance current electrodes; E_1 , E_2 , impedance recording electrodes; L , distance between impedance recording electrodes.

The gauge was passed once around the circumference of the calf, between the voltage sensing electrodes (Fig. 1), and was secured by a small Perspex mount. The tension in the gauge was maintained between 14 and 16 g. The response of the gauge and bridge was found to be linear at the tensions used. Electrical calibration of the gauge was performed *in situ* by switching either 1/200th or 1/1000th increase in resistance into the Wheatstone bridge (7).

RECORDINGS

A four-channel Hewlett-Packard direct writing recorder, type 7404A, was used to collect permanent records of electrical impedance and strain gauge signals. Pulsatile recordings were made at a paper speed of 10 cm/sec and venous occlusion recordings at a speed of 0.5 or 1 cm/sec.

Simultaneous recordings were made of strain gauge pulses, ΔZ , and dZ/dt (Fig. 2).

Venous occlusion plethysmography was carried out by inflating a thigh encircling pressure cuff to 30 mm Hg. Between 5 and 8 occlusion slopes were obtained from such subject, approximately 30 sec being allowed between each occlusion.

ANALYSIS OF RECORDINGS

ELECTRICAL IMPEDANCE PULSE ANALYSIS

The electrical impedance pulses were analysed in three different ways. The first two were attempts to quantitate limb blood flow and were based upon the formula developed simultaneously by Nyboer (15, 16) and by Bonjer (3) which is commonly used for impedance flow calculations.

$$V = \rho \cdot (L/Z_0)^2 \cdot Z \quad (1)$$

where V is the pulse volume, ml, ρ is the electrical resistivity of blood (ohm cm), L is the voltage electrode spacing (cm), Z_0 is the mean basal impedance (ohms) and ΔZ is the pulsatile impedance change (ohm).

The resistivity may be written as:

$$\rho = Z_0 A/L$$

where A is the cross-sectional area of the limb segment. Substituting for ρ in equation 1 yields:

$$\Delta V = (L/Z_0) \cdot A \cdot \Delta Z \quad (2)$$

Thus ΔV per 100 ml of tissue = $(\Delta Z/Z_0) \cdot 100$.

Fig. 2. Trace of arterial pulsations from the calf of a newborn baby recorded simultaneously using electrical impedance, ΔZ and dZ/dt , and mercury-in-rubber strain gauge. The measurement of the maximum rate of impedance change $dZ/dt_{\max}$ is shown.

The first method used for calculating flow simply uses the maximum pulsatile impedance height, $\Delta Z_{\max}$ (Fig. 3), for the value of ΔZ in equation 2:

$$F(Z_{\max}) = (\Delta Z_{\max}/Z_0) \cdot \dot{H} \cdot 100 \text{ ml} \cdot 100 \text{ ml}^{-1} \cdot \text{min}^{-1}$$

where $\dot{H}$ is the heart rate in beats per min.

The second method for calculating flow uses Nyboer and Bonjer's formula but uses the end-diastolic impedance slope, ΔZ_1 (Fig. 3), as suggested by Brown *et al.* (4). The calculation is based on the assumption that there is no arterial flow at the end of the cardiac cycle, and that calf venous flow is nonpulsatile. Thus, the change of impedance at the end of the cardiac cycle relates to mean venous flow alone. Because over the complete cardiac cycle inflow and outflow are equal, this slope averaged over the cycle will give mean arterial flow:

$$F(Z_1) = (\Delta Z_1/Z_0) \cdot 6000 \text{ ml} \cdot 100 \text{ ml}^{-1} \cdot \text{min}^{-1}$$

where ΔZ_1 is the impedance slope at the end of the pulse cycle, in ohm/sec.

The third variable measured from the impedance signals was the maximum rate of impedance change occurring during the inflow phase of the pulse cycle, $dZ/dt_{\max}$ (Fig. 2). This value was estimated because the corresponding slope from transthoracic impedance pulses has been shown to be closely correlated with stroke volume (11). This measurement was made from the records of the first differential of ΔZ and was expressed in terms of Z_0 .

Ten successive beats were analyzed from each impedance recording. The mean values of $F(Z_{\max})$, $F(ZD)$, and $\frac{dZ/dt_{\max}}{Z_0}$ were then used for comparison with strain gauge measurements.

STRAIN GAUGE FLOW CALCULATIONS

The pulsatile recordings from the strain gauge were analyzed in a similar fashion to the impedance recordings. Calculations were made using the equation derived from the work of Whitney (22):

$$\Delta V = \Delta R/R \cdot 100 \quad (3)$$

where ΔV is the volume change per 100 ml tissue, ΔR is the change in strain gauge resistance (ohm), and R is the resting strain gauge resistance (ohm). The use of electrical calibration as described above means that the actual value of R need not be known.

The two estimations of flow from the strain gauge recordings were then made from the following:

$$F(SG_{\max}) = \Delta SG_{\max}/R \cdot \dot{H} \cdot 100 \text{ ml} \cdot 100 \text{ ml}^{-1} \cdot \text{min}^{-1}$$

where $\dot{H}$ = heart rate (beats/min),
and

$$F(SGD) = \Delta SGD/R \cdot 6000 \text{ ml} \cdot 100 \text{ ml}^{-1} \cdot \text{min}^{-1}$$

The terms $\Delta SG_{\max}$ and ΔSGD relate to the maximum height of the strain gauge pulse and its end diastolic slope and are equivalent to $\Delta Z_{\max}$ and ΔZ_1 in Figure 3.

Fig. 3. The measurement of maximum impedance pulse height, $\Delta Z_{\max}$ and slope at the end of the cardiac cycle, ΔZD from arterial impedance pulses.

The maximum change in strain gauge resistance seen during the inflow phase of the pulse cycle was expressed as $\frac{dR/dt_{\max}}{R}$.

The same ten successive beats from the strain gauge recording were analyzed as had been from the impedance recording.

Flow during venous occlusion was calculated from the strain gauge recording using the formula:

$$F(SC_{\max}) = \Delta R/R \cdot 6000 \text{ ml} \cdot 100 \text{ ml}^{-1} \cdot \text{min}^{-1}$$

where ΔR is the change in strain gauge resistance per sec during venous occlusion.

REPRODUCIBILITY OF FLOW ESTIMATIONS

The reproducibility of the three variables measured from the simultaneous pulsatile impedance and strain gauge recordings was investigated by calculating the coefficient of variation (CV):

$$CV = (\text{standard deviation/mean}) \cdot 100\%$$

The CV was calculated for each variable from the ten successive beats analyzed. In addition, six babies had ten impedance recordings made at 30-sec intervals. From these recordings, the CV was calculated for heart rate, Z_0 , mean $F(Z_{\max})$, the mean pulse volume calculated using $Z_{\max}$.

The reproducibility of blood flow calculations made using the strain gauge during venous occlusion was estimated from data collected from a further six babies who each had 10 measurements made at 30-sec intervals.

COMPARISON OF IMPEDANCE AND STRAIN GAUGE FLOW ESTIMATIONS

To determine which of the three methods of impedance pulse analysis could best provide a reliable estimate of beat-to-beat limb blood flow, linear regression analysis was performed between each of the three estimates and flow calculated from strain gauge venous occlusion measurements. Linear regression analysis was also performed between the corresponding estimates of simultaneously recorded pulsatile impedance and strain gauge signals. In each case, when two flow estimates were being compared, a test was performed to determine whether or not the regression line differed statistically significantly from the line of identity.

SUBJECTS

All the babies studied were patients on the Special Care Baby Unit at the John Radcliffe Hospital. Clinical details of these babies are given in Table 1. There was no selection in terms of maturity or size, and babies both with and without clinical problems were studied. In no case was the infant's routine disturbed, and only brief handling was made as part of the measurement procedure. Informed parental consent was obtained before each study.

RESULTS

EFFECT OF ELECTRODE SEPARATION

In five babies, impedance recordings were made with four or five different recording electrode positions on the leg. The ratio $\Delta Z_{\max}/Z_0$ was calculated for each value of electrode separation (L) (Fig. 1). The results are shown in Figure 4, and it can be seen that, with the exception of one baby (●) who was restless throughout the study, the ratio is constant.

COMPARISON OF IMPEDANCE AND STRAIN GAUGE FLOW ESTIMATIONS

The results of the comparative impedance and strain gauge calculations collected in 33 babies are shown in Table 2. As a general comment, the impedance technique was found to be easier to perform than the strain gauge technique. Pulsatile strain gauge

Table 1. *Clinical details of babies*

Subject	Sex	GA ¹	B. Wt. (g)	Age (days)	Wt (g)	Nursing	Clinical comments
1	F	38	2060	14	2150	C _{or}	Microcephalic
2	F	36	1630	7	1620	C _{or}	Well ²
3	M	34	2000	5	1960	C _{or}	Well
4	M	30	1300	34	1730	C _{or}	Well
5	M	37	3180	2	2950	IMI	TTN, pneumothorax, headbox oxygen
6	M	38	2930	2	2870	IMI	TTN, UAC in Rt. a., leg blanched and catheter withdrawn prior to study
7	M	34	2240	4	2155	Inc	Well
8	F	31	1130	13	1140	Inc	Twin, well
9	F	30	1180	63	1900	C _{or}	Twin, BPD following severe HMD
10	M	35	2350	16 hr		Inc	Mild HMD, headbox O ₂
11	M	39	2280	2 hr		IMI	Severe B. A. irritable
12	M	35	2200	15	2170	C _{or}	Well
13	F	34	1350	9	1350	Inc	Well
14	M	32	1350	4	1170	IMI	Severe HMD, IPPV
15	M	30	1355	2	1370	IMI	Severe HMD, IVH, fits
16	M	36	2835	2	2840	IMI	Severe HMD, bilat. pneumothorax, IPPV
17	M	30	1400	25	1420	Inc	Well
18	F	35	1930	3	1870	Inc	Well
19	M	30	1250	7	1200	IMI	Severe HMD, headbox O ₂
20	F	34	1800	11	1865	Inc	Well
21	M	30	1640	1	1600	Inc	Mild HMD, headbox O ₂
22	M	27	1030	1		IMI	B. A. Severe HMD, thrombosis L. leg, R. leg studied, + d.2
23	M	26	870	5	760	IMI	Severe HMD, IPPV
24	M	34	2520	2	2430	Inc	Mild HMD, headbox O ₂
25	F	32	1840	4	1710	Inc	Well
26	M	33	1820	3	1740	Inc	Well
27	M	33	2000	8 hr		Inc	TTN, headbox O ₂
28	F	33	1560	6	1430	Inc	Well
29	M	35	2360	5	2180	Inc	Well
30	M	30	1240	3	1260	Inc	Mild HMD, headbox O ₂
31	F	36	1920	2	1880	Inc	Well
32	M	32	1260	2	1250	IMI	Severe HMD, IPPV
33	F	30	1350	19 hr		IMI	Severe HMD, IPPV

¹ GA, gestational age, completed wk; B.Wt., Birth weight; IMI, open incubator, using radiant heater; TTN, transient tachypnea of the newborn; UAC, umbilical arterial catheter; Inc, closed incubator; BPD, bronchopulmonary dysplasia; HMD, hyaline membrane disease; BA, birth asphyxia; IPPV, receiving intermittent positive pressure ventilation; IVH, intraventricular hemorrhage.

² An infant described as well is well at the time of study, but might have had earlier illness.

Fig. 4. The relationship between the ratio of maximum pulsatile impedance change to mean limb impedance and the interrecording electrode distance, L cm. (Fig. 1), studied in 5 separate babies.

recordings were not made from subjects 1 to 4 because amplification of the signal was not adequate at that time. The failure of strain gauge occlusion measurements in subject 7 was due to electrical problems, and all other failures of strain gauge measurements were due to the difficulty of obtaining satisfactory records in dyspneic babies. Satisfactory pulsatile impedance recordings were obtained from all babies, although in the case of subject 22 the left leg was studied initially, and no pulse was obtained. This leg was dusky and clinically pulseless following the insertion of an umbilical arterial catheter which had been rapidly withdrawn. Good recordings were made from the right leg of this baby. At post-mortem examination following death from severe hyaline membrane disease on the second day of life, it was found that the left common iliac artery was completely occluded by antemortem thrombus. This was the only occasion on which no impedance recording could be made. Good recordings were made from subjects 6 and 11, despite the fact that they had no pulses palpable below the femoral arteries.

The impedance and strain gauge pulses were found to be strikingly different in form (see Fig. 2). The impedance pulse waveform has a more rapid rise than that obtained with the strain gauge. In most babies, especially the more dyspneic, a respiratory-related component of both impedance and strain gauge signals was present, but this was much more apparent in the strain gauge recordings.

Table 2. Results of comparative strain gauge and electrical impedance measurements

Subject	Strain gauge pulses					Impedance pulses				
	F(SG _{osc}) ¹ (ml. 100 ml ⁻¹ min ⁻¹)	H (beats/min)	dSG/dt _{max} Rs ⁻¹	F(SGD) (ml. 100 ml/min)	F(SG _{max}) (ml. 100 ml/min)	Z ₀ (ohm)	dZ/dt _{max} (ohms ⁻¹)	dZ/dt _{max} (Z ₀ S ⁻¹)	F(ZD) (ml. 100 ml/min)	F(Z _{max}) (ml. 100 ml/min)
1	13.4	128				142	4.34	0.0306	10.9	10.8
2	14.6	141				139	4.66	0.0335	11.0	16.3
3	11.1	126				129	3.58	0.0277	13.7	14.8
4	10.7	167				132	2.69	0.0204	13.3	10.5
5	5.9	116				63	0.65	0.0103	9.0	5.7
6	3.8	123				78	1.10	0.0141	6.5	4.3
7		126	0.0034	5.7	7.6	78	1.05	0.0135	4.6	8.0
8	9.4	145	0.0050	12.6	7.0	137	3.43	0.0250	15.8	14.3
9	19.2	164	0.0085	23.3	11.5	106	2.53	0.0239	20.0	14.6
10	6.3	129	0.0056	10.3	6.7	94	0.77	0.0082	3.4	5.1
11		153	0.0134	19.6	16.7	88	2.78	0.0316	11.9	17.3
12	10.4	158	0.0074	12.4	7.1	89	2.99	0.0336	9.1	17.4
13	19.7	140	0.0029	8.1	4.4	126	2.84	0.0225	13.5	14.7
14	10.7	130				125	1.67	0.0134	14.6	9.3
15	10.0	143	0.0048	9.4	5.3	84	1.17	0.0139	8.5	8.6
16	18.4	125				63	1.54	0.0244	6.2	14.1
17	27.2	153	0.0107	22.8	13.5	94	3.08	0.0328	21.7	18.3
18	7.5	129				79	0.83	0.0105	3.5	6.2
19	8.0	133	0.0099	20.4	8.9	110	2.63	0.0239	8.9	9.9
20	6.7	129	0.0073	8.4	7.1	74	1.88	0.0254	4.5	9.5
21	8.5	159	0.0032	7.8	4.0	115	2.58	0.0224	14.7	15.0
22	7.6	159	0.0076	22.9	10.0	136	1.88	0.0138	10.9	9.7
23	8.1	153	0.0041	7.8	5.1	158	3.24	0.0205	14.9	16.6
24	8.8	141	0.0144	20.0	10.3	95	2.53	0.0266	12.9	11.4
25	30.1	133	0.0131	23.9	15.7	104	3.82	0.0367	24.8	23.3
26	19.5	127	0.0136	23.9	17.6	103	3.31	0.0321	5.0	12.4
27	6.7	115	0.0036	6.1	3.5	70	0.73	0.0104	3.0	5.9
28	16.6	145	0.0028	10.1	10.9	107	2.86	0.0267	12.1	15.5
29		123	0.0044	10.8	4.7	112	3.54	0.0316	10.2	14.2
30	4.8	120	0.0033	11.2	4.9	99	1.67	0.0169	9.5	7.8
31	5.9	123	0.0021	3.9	2.5	117	1.15	0.0099	3.5	4.8
32	14.1	186	0.0185	35.3	21.3	88	2.37	0.0269	12.3	15.1
33	10.4	143	0.0064	14.2	7.9	42	1.59	0.0379	10.0	7.4

¹ $F(SG_{\text{osc}})$, blood flow calculated using venous occlusion plethysmography strain gauge; H, heart rate; $\frac{dSG/dt_{\text{max}}}{R}$, variable calculated from the upslope of the pulse recorded using the strain gauge; $F(SGD)$ and $F(SG_{\text{max}})$, blood flow calculated from the end of the pulse cycle and the maximum pulse height recorded using the strain gauge (see text); Z_0 , mean baseline electrical impedance; dZ/dt_{max} , maximum change of impedance seen during the inflow phase of the pulse cycle (Fig. 2); $\frac{dZ/dt_{\text{max}}}{Z_0}$, maximum impedance change in terms of mean impedance; $F(ZD)$ and $F(Z_{\text{max}})$, blood flow calculated from the end of the pulse cycle and the maximum impedance pulse height.

The reproducibility of the various estimates of flow is shown in Table 3. It can be seen that the beat-to-beat reproducibility of simultaneous measurements from the upslopes and from the maximum pulse heights was greater from the impedance recordings than from strain gauge recordings. The reproducibility of the impedance estimations in subjects 5, 6, 14, 16, and 18 (Table 1) was less good than in the other babies because of the presence of respiratory related components; it is emphasized, however, that the strain gauge recordings from these babies were so much more distorted as to be rendered useless for analysis. The reproducibility of flow estimations from the slopes at the end of the pulse cycle [$F(Z_0)$ and $F(SGD)$] was poor for both methods.

The mean CV for 10 repeated measurements in six different subjects was greater for the strain gauge than for impedance measurements. Unfortunately, only one baby appears in both groups of six. In that subject, who was clinically well, the CV of successive $F(Z_{\text{max}})$ estimations was 7.8% and that of $F(SC_{\text{osc}})$ estimations was 17.9%. Comparing the two groups of six babies, two of the impedance group and three in the strain gauge group were receiving intermittent positive pressure ventilation, the re-

maining three babies in the latter group being perfectly well, while two of the remaining four in the impedance group were severely tachypneic.

The results of linear regression analyses are shown in Table 4. It can be seen that each of the three estimates of flow from the pulsatile impedance signals correlates statistically significantly with the flow estimations made by strain gauge venous occlusion. The best correlation ($r = 0.76$) was between $F(Z_{\text{max}})$ and $F(SC_{\text{osc}})$.

Because the aim of the comparisons was to assess how well estimations based on impedance pulse analysis might correlate with flow measurements made by strain gauge venous occlusion plethysmography, the latter was initially considered as the independent variable, x , in the regression analysis. For such analysis, the regression lines between $F(Z_{\text{max}})$ and $F(SG_{\text{osc}})$ and between $F(Z_0)$ and $F(SC_{\text{osc}})$ were found to be statistically significantly different from the line of identity ($P < 0.001$). However, because the reproducibility of the variable $F(Z_{\text{max}})$ was significantly better than that of $F(SG_{\text{osc}})$, it is valid to consider the impedance estimates of flow as the independent variables. When the analysis was performed in this way, the regression line between $F(SG_{\text{osc}})$

Table 3. *Reproducibility of observations*

Pulsatile data, 10 successive beats simultaneously recorded (<i>n</i> = 24)						
	Strain gauge			Impedance		
	$\frac{dR/dt_{max}}{R}$	F(SGD) ¹	F(SG _{max})	$\frac{dZ/dt_{max}}{Z_0}$	F(Z ₀)	F(Z _{max})
CV%						
Mean	21.0	38.1	17.4	9.0	36.8	7.8
Range	(6.3-49.8)	(6.3-71.2)	(5.9-45.3)	(4.3-16.9)	(11.3-93.1)	(2.1-16.9)
No. of subjects with CV ≥10%	21	22	18	8	24	5
0.3,0.3						
10 impedance recordings at 30-sec intervals (<i>n</i> = 6)						
	F(Z _{max})	Pulse volume		Z ₀	H	
CV%						
Mean	9.8	9.1		1.2	3.1	
Range	(6.1-13.5)	(3.7-14.0)		(0.0-1.9)	(0.5-4.5)	
10 measurements of F(SG _{oc}) at 30-sec intervals (<i>n</i> = 6)						
CV%						
Mean		16.1				
Range		(5.7-27.6)				

¹ $\frac{dR/dt_{max}}{R}$ and $\frac{dZ/dt_{max}}{Z_0}$ = variables measured from the upslope of strain gauge and impedance pulsatile recordings (see Fig. 2 and text); $F(SGD)$, $F(SG_{max})$, $F(SGD)$, and $F(SG_{max})$; blood flow calculations from the end of the pulse cycle and maximum pulse heights from strain gauge and impedance recordings (see Fig. 3 and text); CV, coefficient of variation (see text); $F(SG_{oc})$, blood flow calculated using the strain gauge during venous occlusion; $\dot{H}$, heart rate.

Table 4

Variables				Slope of regression line compared with line of identity (P)			
	x	y	n	Regression equation	r	Statistical significance of r ($P <$)	
$F(SG_{oc})$ compared with variables from pulsatile traces	$F(SG_{oc})$	$F(Z_{max})$	30	$y = 5.14 + 0.56x$	0.76	0.001	<0.001
	$F(Z_{max})$	$F(SG_{oc})$	30	$y = 0.33 + 1.03x$	0.76	0.001	>0.980
	$F(SG_{oc})$	FZD	30	$y = 4.51 + 0.54x$	0.64	0.001	>0.001
	$F(SG_{oc})$	$\frac{dZ/dt_{max}}{Z_0}$	30	$y = 0.0124 + 0.0009x$	0.63	0.001	
	$\frac{dZ/dt_{max}}{Z_0}$	$F(SG_{oc})$	30	$y = 1.28 + 467x$	0.63	0.001	
	$F(SG_{oc})$	$F(SG_{max})$	21	$y = 3.55 + 0.43x$	0.62	0.01	<0.001
	$F(SG_{oc})$	$F(SGD)$	21	$y = 7.97 + 0.57x$	0.50	0.05	N.S.
	$F(SG_{oc})$	$\frac{dR/dt_{max}}{R}$	21	$y = 0.0040 + 0.0003x$	0.43	N.S.	
Comparison of simultaneous pulsatile traces	$F(SG_{oc})$	$F(Z_{max})$	24	$y = 8.11 + 0.48x$	0.50	0.01	<0.01
	$F(Z_{max})$	$F(SG_{oc})$	24	$y = 2.50 + 0.52x$	0.50	0.01	<0.025
	$F(SGD)$	$F(Z_0)$	24	$y = 6.11 + 0.34x$	0.46	0.05	<0.001
	$\frac{dR/dt_{max}}{R}$	$\frac{dZ/dt_{max}}{Z_0}$	24	$y = 2.50 + 414x$	0.77	0.001	
	$\frac{dZ/dt_{max}}{Z_0}$	$\frac{dR/dt_{max}}{R}$	24	$y = 0.00097 + 0.270x$	0.77	0.001	

¹ r, linear correlation coefficient; P = probability; $F(SG_{oc})$, blood flow calculated using the strain gauge during venous occlusion; $\frac{dZ/dt_{max}}{Z_0}$, and $\frac{dR/dt_{max}}{R}$, variables measured from the upslope strain gauge and impedance pulsatile recordings (see Fig. 2 and text); $F(SGD)$, $F(SG_{max})$, $F(SGD)$, and $F(SG_{max})$, blood flow calculations from the end of the pulse cycle and maximum pulse heights from strain gauge and impedance recordings (see Fig. 3 and text).

and $F(Z_{\max})$ was not statistically significantly different from the line of identity ($P < 0.98$), and the intercept of the line is very close to zero (Fig. 5).

The correlation coefficient between $F(SG_{\max})$ and the variables $F(SG_{\max})$ and $F(SG_{15})$ reached statistical significance ($P < 0.01$ and < 0.05 , respectively), but these coefficients were lower than those obtained with the variables from the impedance pulse. Furthermore, the regression lines differed statistically significantly from the line of identity. The relationship between the value $\frac{(dR/dt_{\max})}{R}$ and $F(SG_{\max})$ did not reach statistical significance.

The linear correlations between the simultaneously recorded impedance and strain gauge pulses reached statistical significance, but, with the exception of the upslope correlation, i.e., $\frac{(dR/dt_{\max})}{R}$ with $\frac{(dZ/dt_{\max})}{Z_0}$, the coefficients were comparatively low, and the regression lines differed significantly from the lines of identity.

DISCUSSION

The results show that of the methods of analysis performed the estimation of limb blood flow using the maximum pulsatile impedance change, $F(Z_{\max})$, gave the best correlation with strain gauge venous occlusion estimations, $F(SG_{\max})$, and that it was the most reproducible beat-to-beat estimation. This is in accordance with experience in adults, and animals where the maximum height of the impedance pulse, $\Delta Z_{\max}$, has been used for the estimation of blood flow, where it has been found to correlate well with simultaneous measurements using electromagnetic flowmeters (20).

Estimation of flow based on the downslope of the impedance pulse at the end of the pulse cycle is based upon the assumption that venous flow is nonpulsatile and that there is no forward flow in the artery at the end of the pulse cycle (4). We have confirmed that venous flow is indeed nonpulsatile in the newborn infant with the use of a Doppler ultrasound probe positioned over the femoral vein. In addition, it has been found that while there is generally no forward flow in the femoral or popliteal arteries at the end of the pulse cycle there may in fact be reverse flow which could invalidate the calculation. Neither is it possible to say that these

findings in the great vessels may be extrapolated to predict what might occur in the smaller vessels. The basis for this form of estimation therefore appears to be unreliable, although it does represent an attempt to overcome the difficulties of correcting for venous and arterial runoff which is necessary for the accurate analysis of pulsatile recordings.

The third method of pulse analysis, based on a measurement of the maximum rate of change of impedance at the beginning of the pulse cycle, $dZ/dt_{\max}$, was considered in this study because of the success of a similar analysis in estimating stroke volume (11). $dZ/dt_{\max}$ was expressed as a fraction of Z_0 so that it should be directly comparable with the corresponding slope of the strain gauge pulse, $(dR/dt_{\max})/R$. It was also necessary to express the results in this way to arrive at the correct dimensions for the flow estimation.

Other methods of pulse analysis have been described using extrapolation of the slopes (2) and integration of the pulse form (15, 16). The basis for these forms of analysis appears not to have been justified theoretically, and when results obtained in the present study were analyzed using these methods, no improvement in beat-to-beat reproducibility or correlation between pulsatile and venous occlusion flow estimations was gained.

Some of the values calculated for limb flow in this study using both the impedance and strain gauge techniques are higher than previously reported values from the newborn except in experimentally induced reactive hyperaemia (5, 18) or in a very warm environment (9). This is likely to be due to the variety of circumstances under which babies were studied some of which would lead to increased peripheral blood flow, such as the presence of patent ductus arteriosus, the use of phototherapy, and nursing under radiant heaters.

Although the results of this study indicate that at least one technique of impedance pulse analysis can provide limb blood flow estimations which correlate with the established method of strain gauge venous occlusion plethysmography, they also lead one to raise objections to the use of the strain gauge method as a standard. However, for measurements in the newborn, no better alternative standard technique exists. The major problems encountered in this study with the strain gauge method were concerned both with the practical application of the gauge and with the reproducibility of the measurements made. Although the very fine silicone rubber tubing used for the gauges was found to be robust, there was a constant need to reposition the gauge on the leg because there was a tendency for it to roll up or down the calf as the baby moved. This would make long-term use of the method satisfactory. Movement-induced artefact was found to be a problem with the strain gauge method both when making pulsatile recordings and during attempts to make venous occlusion estimations. Despite the fact that satisfactory tracings were eventually obtained with the strain gauge from 30 of the 33 babies, success was often not achieved at the first attempt because of movement related artefacts occurring both before and during the occlusion which made reliable measurement of the occlusion slope difficult and sometimes impossible. For this reason, pulse and venous occlusion estimations in the same baby were often separated in time by as much as 10 min. Previous reports of studies using the mercury-in-rubber strain gauge method in the newborn have not emphasized this problem which was possibly accentuated in this work because of the heterogeneous nature of the group of babies studied, several of whom were very ill at the time of study.

In contrast to the problems associated with the strain gauge method, the electrical impedance recordings were comparatively easy to obtain; in only one infant, in whom the common iliac artery was shown to be completely occluded, was it impossible to analyze the electrical impedance pulses. The automatic balancing device of the impedance instrument was used during all recordings, but the signal was so stable that it was not needed during the comparatively short periods of time when recordings were being made. However, the auto-balance facility was useful for eliminating large variations in the baseline which occurred between recordings when collecting intermittent data as was the case when

Fig. 5. Relationship between blood flow measured using the strain gauge during venous occlusion, $F(SG_{\max})$ and from the maximum impedance pulse height, $F(Z_{\max})$. The regression line is statistically significantly similar to the line of identity, $P > 0.98$.

estimating the reproducibility of the technique. Furthermore, the auto-balance feature would be essential for long-term data collection.

The calculations of reproducibility of the various estimations (Table 3) must include physiological beat-to-beat variation. In particular, respiration can be expected to produce beat-to-beat variation of pulse volume, probably of the order of $\pm 10\%$. This level of reproducibility was only achieved for $dZ/dt_{\max} \cdot Z_0$ and $F(Z_{\max})$. The particularly poor reproducibility of $F(SG_{15})$ and $F(Z_{15})$ is because these estimations are derived from the slope of a comparatively low-frequency portion of the pulsatile signal which is more susceptible to base line variations. For this reason, it is considered that the estimation of flow from the end of the downslope is not suitable as a method for long-term beat-to-beat flow assessment. The poorer reproducibility of the strain gauge estimations compared with impedance estimations fully justifies considering the electrical impedance estimations as the independent variable in the regression analysis.

There are fundamental reasons why one might expect to obtain different results with the electrical impedance and strain gauge techniques. The latter responds solely to mechanical phenomena, whereas the impedance method will be influenced by both mechanical and electrical phenomena. The difference in the form of the pulses obtained with the two techniques has already been mentioned. The rapid upslope of the impedance pulse is probably due to the fact that as soon as blood enters the large arteries in the limb segment an electrical impedance change will be seen, but its arrival will not be detected immediately by the strain gauge until the blood has been distributed through the tissues and produced a change in limb circumference. Furthermore, a change in the electrical impedance of the limb can be observed simply due to a change in the alignment of red blood cells (17) and in the redistribution of a fixed volume of blood from large arteries to capillaries. Although both methods should be capable of providing the same information, this will only be achieved if the methods of analysis take account of these fundamental differences.

The equation used as the basis for the impedance calculations (equation 2) has been used by various authors (8, 10, 13), but its derivation deserves comment. The electrical resistivity of blood is eliminated from the equation by making the substitution $\rho = Z_0 A/L$. However, the relationship used to make this substitution is in fact an expression for the mean resistivity of the tissue segment, ρ_t . Thus, the derivation of the equation assumes that blood resistivity, ρ_b , is the same as ρ_t . Blood resistivity varies with temperature, blood motion, and haematocrit. Measurements made at 37°C on fresh samples of whole blood from newborn infants have shown an exponential relationship between ρ_b and haematocrit; $\rho = 67.919 \exp(0.02433Hct)$ (13).

At a haematocrit of 40%, the resistivity is 180 ohm cm and at 55% is 259 ohm cm. Kubicek *et al.* (10) have measured ρ_t in 8 adult subjects and found a mean value of 211 ohm cm. 12 of the babies in the present study had careful measurements made of the mean interelectrode distance (L cm) and the mean calf circumference, C cm, and ρ_t , calculated from the relationship $\rho_t = C^2 Z_0/4$, was found to have a mean value of 200 ohm cm (range, 173 to 246). Thus, some error must be introduced by the assumption that $\rho_b = \rho_t$. However, it is felt that this error is no larger than that which would be introduced by attempting to make measurements of haematocrit, interelectrode distance, and leg circumference because the latter two measurements are associated with a significant error in the small leg of a preterm infant.

The measurement of $\Delta Z_{\max}$ and the subsequent calculation of $F(Z_{\max})$ could be automated and thus provide continuous beat-to-

beat blood flow estimations. Although the accuracy of these estimations in absolute terms is likely to be only moderate, the reproducibility is good, and the technique simple and safe to apply. We feel that at present there is no better technique for making noninvasive estimations of peripheral blood flow in the newborn, and we believe that this method could provide valuable information from acutely ill babies.

REFERENCES AND NOTES

- Allison, R. D.: The effects of aging, vascular disease and hypertension on the peripheral pulse volume. *J. Am. Geriatr. Soc.*, **16**: 39 (1968).
- Bloom, W. L., Kaufman, S. S., Nims, L. E., and Nyboer, J.: The effect of positive pressure breathing at sea level on peripheral blood flow and cardiac output. *Natl. Res. Coun. Comm. Aviation Med. Report 246* (1944).
- Bonjer, F. H.: *Circulatieonderzoek door Impedantiemeting*. Groningen, Drukkerij L. Oppenheim, N. V., 1950.
- Brown, B. H., Pryce, W. I. J., Baumber, D., and Clarke, R. G.: Impedance plethysmography: can it measure changes in limb blood flow. *Med. Biol. Eng.*, **13**: 674 (1975).
- Celander, O.: Blood flow in the foot and calf of the newborn. A plethysmographic study. *Acta Paediatr.*, **49**: 488 (1960).
- Couch, N. D., Van de Water, J. M., and Dmochowski, M. D.: Noninvasive measurement of peripheral arterial flow. *Arch. Surg.*, **102**: 435 (1971).
- Hallbook, T., Masson, B., and Nilsen, R.: A strain gauge plethysmograph with electrical calibration. *Scand. J. Clin. Lab. Invest.*, **25**: 413 (1970).
- Hill, D. W., and Lowe, H. J.: The use of the electrical impedance technique for the monitoring of cardiac output and limb blood flow during anaesthesia. *Med. Biol. Eng.*, **11**: 534 (1973).
- Kidd, L., Levison, H., Gemmel, P., Aharon, A., and Swyer, P. R.: Limb blood flow in the normal and sick newborn. *Am. J. Dis. Child.*, **112**: 402 (1966).
- Kubicek, W. G., Witsoe, D. A., Patterson, R. P., and From, A. H. L.: Development and evaluation of an impedance cardiographic system to measure cardiac output and other cardiac parameters. Final Progress Report, Contract NAS 9-4500, NASA Manned Space Craft Centre, Houston, Texas (1970).
- Mohapatra, S. N.: Noninvasive cardiac monitoring by electrical impedance plethysmography. Ph.D. Thesis, London University, 1977.
- Mohapatra, S. N., and Aronson, H. M.: The measurement of peripheral blood flow by the electrical impedance technique. *J. Med. Eng. Tech.*, **3**: 132 (1979).
- Mohapatra, S. N., Costeloe, K., and Hill, D. W.: Blood resistivity and its implications for the calculation of cardiac output by the thoracic electrical impedance technique. *Eur. J. Intensive Care Med.*, **1**: 153 (1976).
- Mullick, S. C., Wheeler, H. B., and Songster, G. E.: Diagnosis of deep vein thrombosis by measurement of electrical impedance. *Am. J. Surg.*, **119**: 417 (1970).
- Nyboer, J.: Electrical impedance plethysmography—A physical and physiologic approach to peripheral vascular study. *Circulation*, **2**: 811 (1950).
- Nyboer, J.: *Electrical Impedance Plethysmography*, 2nd ed. pp. 41-44 (Charles C Thomas, Springfield, Ill., 1970).
- Peura, R. A., Wheeler, W. H., Penney, B. C., and Arcuri, J.: Impedance plethysmography: relative contribution of blood volume and red cell velocity changes on the signal strength. Digest of the 11th International Conference of Medical and Biological Engineering, Ottawa, Canada, p. 240 (1976).
- Powers, W. E., and Swyer, P. R.: Limb blood flow following umbilical artery catheterisation. *Pediatrics*, **55**: 248 (1975).
- Storrs, C. N.: Cardiovascular effects of apnoea in preterm infants. *Arch. Dis. Child.*, **52**: 534 (1977).
- Van de Water, J. M., Dmochowski, J. R., Dove, G. B., and Couch, N. P.: Evaluation of an impedance flowmeter in arterial surgery. *Surgery*, **70**: 954 (1971).
- Wallgren, G., and Lind, J.: Quantitative studies on the human neonatal circulation. IV. Observations on the newborn infant's peripheral circulation and plasma expansion during moderate hypovolaemia. *Acta Paediatr. Scand. Suppl.*, **179**: 55 (1967).
- Whitney, R. J.: The measurement of volume changes in human limbs. *J. Physiol.*, **121**: 1 (1953).
- Yao, A. C., Wallgren, G., Sachchida, N. S., and Lind, J.: Peripheral circulatory response to feeding in the newborn infant. *Pediatrics*, **47**: 378 (1971).
- Dr. Costeloe is supported by a grant from the Spastics Society.
- The authors would like to express their gratitude to Professor J. P. M. Tizard for help with the preparation of this manuscript.
- Requests for reprints should be addressed to: Dr. Kate Costeloe, Department of Paediatrics, John Radcliffe Hospital, Headington, Oxford, England.
- Received for publication September 26, 1979.
- Accepted for publication December 13, 1979.