Quantitative Near Infrared Spectroscopy Measurement of Cerebral Hemodynamics in Newborn Piglets

DEREK W. BROWN, PAUL A. PICOT, JAFAR GHARAVI NAEINI, ROGER SPRINGETT,
DAVID T. DELPY, AND TING-YIM LEE

Imaging Division, Lawson Health Research Institute, London, Ontario, Canada, N6A 4V2 [D.W.B., P.A.P., J.G.N., T.-Y.L.]; Imaging Research Laboratories, Robarts Research Institute, London, Ontario, Canada, N6A 5K8 [D.W.B., T.-Y.L.]; Department of Medical Biophysics, University of Western Ontario, London, Ontario, Canada, N6A 4B8 [D.W.B., P.A.P., J.G.N., T.-Y.L.]; Department of Medical Physics and Bioengineering, University College London, London, U.K., WC1E 6JA [R.S., D.T.D.]

ABSTRACT

Severely premature infants are often at increased risk of cerebral hemorrhage and/or ischemic injury caused by immature autoregulatory control of blood flow to the brain. If blood flow is too high, the infant is at risk of hemorrhage, whereas too little blood flow can result in ischemic injury. The development of a noninvasive, bedside means of measuring cerebral hemodynamics would greatly facilitate both diagnosis and monitoring of afflicted individuals. It is to this end that we have developed a near infrared spectroscopy (NIRS) system that allows for quantitative, bedside measurement of cerebral blood flow (CBF), cerebral blood volume (CBV), and mean transit time (MTT). The technique requires an i.v. injection of the near infrared chromophore indocyanine green. Six newborn piglets, median age of 18 h (range 6–54 h), median weight of 1.75 kg (range 1.5–2.1 kg), were studied. Measurements of CBF, CBV, and MTT were made at normocapnia, hypocapnia, and hypercapnia to test the technique over a range of hemodynamic conditions. The accuracy of our new approach has been determined by direct com-

parison with measurements made using a previously validated computed tomography technique. Paired *t* tests showed no significant difference between computed tomography and NIRS measurements of CBF, CBV, and MTT, and mean biases between the two methods were $-2.05 \text{ mL} \cdot \text{min}^{-1} \cdot 100 \text{ g}^{-1}$, $-0.18 \text{ mL} \cdot 100 \text{ g}^{-1}$, and 0.43 s , respectively. The precision of NIRS CBF, CBV, and MTT measurements, as determined by repeated-measures ANOVA, was 9.71%, 13.05%, and 7.57%, respectively. (*Pediatr Res* 51: 564–570, 2002)

Abbreviations:

CBF, cerebral blood flow
CBV, cerebral blood volume
MTT, mean transit time
CT, computed tomography
NIRS, near infrared spectroscopy
ICG, indocyanine green
Paco₂, partial arterial CO₂ tension

Since the first publication by Jobsis in 1977 (1), NIRS has been used in a variety of studies to investigate cerebral hemodynamics (2, 3). The underlying principles behind the use of NIRS to probe biologic media are relatively simple and have been described in detail elsewhere (1, 4–8). There exist in biologic tissue four endogenous near infrared (NIR) light absorbers—oxy-Hb (HbO₂), deoxy-Hb (Hb), cytochrome oxidase (Cyt), and water. Because HbO₂ and Hb are generally present at relatively low concentrations in tissue, NIR light is able to penetrate tissue to a greater extent than other low-energy forms of light, in some cases up to distances of 8–9 cm (6).

As NIR light enters tissue, it is multiply scattered. The result of this scatter is that the total path length traveled by the NIR light is greater than the physical distance between the points of emission and detection. This extra distance can be accounted for using the differential path length factor (DPF), first described by Delpy *et al.* (9). With accurate knowledge of the DPF, a modified version of the Beer-Lambert law can be used to determine absolute changes in concentrations of NIR absorbers within tissue:

$$\Delta c = \frac{\Delta A}{\alpha \cdot L \cdot B} \quad (1)$$

where Δc is the change in concentration, ΔA is the change in attenuation, α is the extinction coefficient, L is the physical distance between emission and detection of NIR light, and B is the DPF. Measurement of concentration changes over time can

Received February 8, 2001; accepted August 8, 2001.
Correspondence: Ting-Yim Lee, Ph.D., Imaging Division, Lawson Health Research Institute, 268 Grosvenor Street, London, Ontario, Canada, N6A 4V2; e-mail: tlee@irus.rii.on.ca
Supported by Photonics Research Ontario, the Lawson Health Research Institute, Robarts Research Institute, and the Child Health Research Institute.

then be used to quantify cerebral hemodynamics, namely CBF (10, 11).

Previous attempts to investigate cerebral hemodynamics using NIRS have achieved somewhat limited success. This has been, in part, due to the use of discrete wavelength spectrometers that acquire intensity measurements at several distinct wavelengths. CBF, CBV, and MTT calculations made using these spectrometers are subject to error resulting from an inability to accurately measure the individual DPF. Also, previous algorithms used to calculate CBF, CBV, and MTT have been unable to account for loss of tracer from the brain resulting from venous outflow, greatly limiting the length of relevant data acquisition that can be used in the calculation of CBF (10). We used a cooled CCD (charged coupled device) spectrometer to acquire high signal-to-noise NIR spectra from 600 to 980 nm with a spectral sampling width of 0.395 ± 0.001 nm. This spectrometer enables the individual DPF to be estimated accurately using the second derivative technique, described by Matcher *et al.* (12). In addition, outflow of tracer from the brain was explicitly considered in our CBF, CBV, and MTT calculations, thus removing another confounder of the accuracy of the measurement.

In this study, we have measured CBF, CBV, and MTT in six newborn piglets using a cooled CCD NIRS system. The accuracy of our new approach has been determined by direct comparison with measurements made using a previously validated CT technique (13). To test the accuracy of our new approach over a range of hemodynamic conditions, measurements were made at various P_{aCO_2} levels: hypocapnia (P_{aCO_2} 30 ± 4 mm Hg), normocapnia (P_{aCO_2} 40 ± 4 mm Hg), and hypercapnia (P_{aCO_2} 50 ± 4 mm Hg).

P_{aCO_2} is now generally considered to be the most influential regulator of the cerebral circulation (14), and, although the exact mechanism of the response is not known, the physiologic basis for it is well understood. CO_2 is one of the major products of cellular metabolism. Thus, an increase in CO_2 levels indicates an increase in cerebral cellular activity. This increased metabolism demands increased oxygen delivery, which is achieved through increased CBF and CBV and decreased MTT. Higher CO_2 levels would give rise to both increased CBF and CBV and decreased MTT. Similarly, lower CO_2 levels would give rise to decreased CBF and CBV and increased MTT. Cerebral CO_2 reactivity is defined as the change in CBF per unit change in P_{aCO_2} and has been shown to be linear in normal individuals for P_{aCO_2} values ranging from 25 to 65 mm Hg (15).

Crucial to the comparison of NIRS with CT measurements is the mapping of the region through which detected NIR light has traveled. Without this detailed knowledge, a comparison between NIRS and CT is not possible. Previous studies, using computer modeling, have shown that photon paths through the brain can be traced and used to determine the most likely NIR light path (16). Thus, the volume for CBF, CBV, and MTT measurements by CT and NIRS can be matched, eliminating any potential sampling errors. These studies show that, in a homogeneous medium, the detected NIR light travels in a banana-shaped path through the head. The maximum penetration depth of the path is given approximately by the square root

of the physical distance between the points of emission and detection.

Theory. The CBF, CBV, and MTT calculations are based on a generalized application of the Fick principle, which states

$$Q(t) = F \int_0^t [C_a(u) - C_v(u)] du \quad (2)$$

where $Q(t)$, $C_a(t)$, and $C_v(t)$ are the tissue, arterial, and venous concentrations of the NIR chromophore, respectively, and F is the cerebral blood flow. Previous studies have made use of the assumption that there is no venous outflow of the chromophore during the time interval 0 to t , in which case equation 2 can be simplified to

$$Q(t) = F \int_0^t C_a(u) du \quad (3)$$

Blood flow can thus be estimated at each time within the time interval (0 to t) by calculating the slope of the tissue concentration curve *versus* the integral of the arterial concentration curve (11). Although useful, this technique carries with it a serious drawback. To assume no venous outflow, measurements must be made within the minimum transit time of the organ, typically only 3–5 s in the brain. This limitation in the amount of usable data results in variability in the blood flow estimation.

We have developed an alternative means of measuring CBF in which the outflow of the tracer from the brain is explicitly considered in the calculation. If we consider a network of capillaries in a certain mass of brain tissue, then CBF into the network is F ($\text{mL} \cdot \text{min}^{-1} \cdot 100 \text{ g}^{-1}$) and is carrying with it ICG at a concentration of $C_a(t)$ ($\mu\text{mol} \cdot \text{mL}^{-1}$). The tissue concentration function, or the tissue residue function, $Q(t)$ ($\mu\text{mol} \cdot \text{mL}^{-1}$), can be measured using NIRS. In the special case when $F \cdot C_a$ is a delta function such that a unit mass of ICG is deposited in the tissue instantaneously at time 0, then the tissue residue function becomes the impulse residue function (IRF) or $R(t)$ (17). For a tracer that remains intravascular, such as ICG, the IRF is of the general form shown in Figure 1. The length of the initial plateau at unity height corresponds to the minimum time required for the blood to traverse the network from the arterial inlet to the venous outlet, or the minimum transit time.

When ICG is injected intravenously in a peripheral vein, the rate of delivery of the tracer to the capillary network is $F \cdot C_a(t)$. If the mass of ICG in the network is linear with respect to the arterial (input) concentration and F is constant in time, then, by linear superimposition, it can be shown that

$$Q(t) = F \cdot C_a(t) * R(t) \quad (4)$$

where $*$ is the convolution operator. $Q(t)$ and $C_a(t)$ can be measured and deconvolution between the two curves then yields $F \cdot R(t)$ (13), the initial height of which corresponds to CBF and the

Figure 1. Schematic representation of the IRF for a vascular network in which the blood flow tracer (ICG) remains intravascular.

area under the curve to CBV (18). From the central volume principle (19), MTT can then be calculated as follows:

$$MTT = \frac{CBV}{CBF} \quad (5)$$

METHODS

Subjects and studies. Six newborn piglets, median age 18 h (range 6–54 h) and median weight 1.75 kg (range 1.5–2.1 kg), were studied. The study was approved by the Council on Animal Care—Animal Use Subcommittee at the University of Western Ontario. The piglets were anesthetized using 1% isoflurane with vecuronium, intubated, and ventilated. Physiologic parameters, including partial arterial oxygen tension (P_{aO_2}), bicarbonate levels (HCO_3^-), pH, and mean arterial blood pressure were monitored throughout the experiment.

The piglets were placed in a Lightspeed QXi multislice CT scanner (General Electric Medical Systems, Milwaukee, WI, U.S.A.). Because of the portable nature of the NIRS system, both NIRS and CT measurements could be made with the piglet in the

CT scanner. Table 1 shows the number of CT and NIRS measurements made at each P_{aCO_2} level for individual piglets. In some piglets, we were able to perform CT and NIRS measurements only once at each P_{aCO_2} level because of time constraints or technical problems in maintaining the piglet at stable conditions. In other piglets, we were able to perform two NIRS measurements for each CT measurement. CT and NIRS measurements were performed at only two P_{aCO_2} levels in two of the six piglets and at all three P_{aCO_2} levels in the remaining four piglets. The comparison of NIRS and CT data were analyzed by two-way ANOVA with unequal repeated measures (20).

NIRS tissue ICG concentration measurement. The NIRS system is comprised of a tungsten halogen lightsource, two fiber-optic cables, and a spectrometer. The spectrometer consists of a holographic grating housed in a custom-designed, light-tight container and a cooled CCD camera (Wright Instruments, Enfield, Middlesex, U.K.). For the NIRS measurements, two fiberoptic optodes were placed 3.0 cm apart on the head of the piglet. One of the optodes was used to transmit and the other to collect NIR light in the range of 600–980 nm. Multiply scattered light collected from the head of the piglet was channelled to the holographic grating, where it was dispersed across the cooled CCD chip (cooled to -70°C to reduce electronic dark noise). The piglets received a 1.0-mL injection of ICG solution at a concentration of 0.1 mg/mL into the ear vein. A modified version of the Beer-Lambert law, which uses the DPF to account for the scatter, was used to calculate absolute change in ICG concentration within the illuminated tissue region with a temporal resolution of 200 ms. Because the concentration of ICG in tissue is 0 before injection, the absolute change in ICG is therefore the absolute concentration of ICG in tissue.

Arterial ICG concentration measurement. Arterial ICG concentration was measured noninvasively on a hindfoot of the piglet using a dye densitogram unit (model DDG-2001 A/K, Nihon Kohden, Tokyo, Japan). The dye densitogram probe was held on the skin of the piglet using a spring-loaded clip, much the same as is used for a regular pulse oximeter probe. This system measured absolute ICG concentration in the arteries of the piglet, making one measurement with every heartbeat. The normal piglet heart rate is between 110 and 130 beats/min, giving approximately one measurement every half second.

CT measurement. For CT measurements, each piglet received a 1.0 mL/kg injection of Omnipaque 300 (Nycomed Imaging AS, Oslo, Norway), a nonionic, radio-opaque iodine solution, at a rate of 1.0 mL/s into the umbilical vein. CT

Table 1. Number of CT and NIRS measurements at each P_{aCO_2} level

Piglet no.	CT			NIRS		
	Hypocapnia	Normocapnia	Hypercapnia	Hypocapnia	Normocapnia	Hypercapnia
1	1	1	X	2	1	X
2	2	1	2	2	2	2
3	1	1	1	2	2	1
4	2	1	X	2	2	X
5	1	1	1	1	2	2
6	1	1	1	2	2	2

X = no measurement was made.

Figure 2. Representative tissue and arterial ICG concentration curves measured with NIRS and the dye densitogram unit, respectively. Curves are measured simultaneously over a period of 40 s after injection of 1.0 mL of ICG solution at a concentration of 0.1 mg/mL. The peak arterial concentration is approximately 20 times greater than that of tissue. This is because blood volume in tissue is approximately 5% of the total tissue volume.

scans were then performed with a Lightspeed QXi multislice scanner. Each scan provided four simultaneously obtained slices, once every second, for a period of 40 s. The four slices were spaced 5.0 mm apart, coronally on the head of the piglet, with a field of view (12 cm) set to encompass the entire head and front legs of the piglet, in a prone position. From this series of images, it was possible to determine both arterial and tissue concentrations of the contrast agent with a temporal resolution of 1.0 s (13). Functional maps of CBF, CBV, and MTT across the four slices of the piglet head were calculated using the CT perfusion software package from General Electric Medical Systems, which was developed and validated in our lab (13).

Deconvolution algorithm. For the NIRS studies, the measured $Q(t)$ and $C_a(t)$ were deconvolved, according to equation 4, to yield the blood flow scaled IRF, or $FR(t)$. In the implementation of the deconvolution algorithm, equation 4 was first discretized, resulting in a system of linear equations. Because negative values for $R(t)$ are not allowed, the system of equations was solved by the non-negative least squares algorithm of Lawson and Hanson (21).

Accuracy and precision. The accuracy of hemodynamic measurements made using the NIRS system was determined via direct comparison with CT results. Because the NIRS system is portable, both sets of measurements were made with the animal in the CT scanner. This allowed for direct comparison of hemodynamic measurements obtained by the two modalities at various $Paco_2$ levels. The NIRS measurements are averaged values over the entire detected NIR light path. To

compare NIRS with CT measurements, averages in the CT functional images over the area corresponding to the detected NIR light path were calculated.

RESULTS

The NIRS system employs an emission/detection scheme consisting of two optodes placed 3 cm apart on the head of the piglet. Calculations of CBF, CBV, and MTT are averages over the entire tissue region illuminated by the detected NIR light. Tissue and arterial concentration *versus* time graphs, as measured with the NIRS and dye densitogram unit systems, are presented in Figure 2. The peak arterial concentration is approximately 20 times greater than that of tissue. This is because blood volume in tissue is approximately 5% of total tissue volume. The result of deconvolution between these two curves, the blood flow scaled IRF, is presented in Figure 3, and a graph of the NIRS-measured CBF $Paco_2$ reactivity is presented in Figure 4. Linear regression was used to fit a line to the CBF $Paco_2$ reactivity data set as a whole. This line is shown in Figure 4 as a dashed line. The slope of the regression line is $1.36 \text{ mL} \cdot \text{min}^{-1} \cdot 100 \text{ g}^{-1} / \text{mm Hg}$. Thus, for every mm Hg increase in $Paco_2$, CBF increases by $1.36 \text{ mL} \cdot \text{min}^{-1} \cdot 100 \text{ g}^{-1}$. Similar graphs have been produced for CBV and MTT but are omitted here in the interest of space.

Table 2 presents the measured physiologic parameters for repeated NIRS measurements at various $Paco_2$ levels on each piglet. Paired t tests showed no significant difference between physiologic parameters for repeated measurements ($p > 0.05$).

Figure 5A shows two of a set of four cross-sectional CT images at peak contrast enhancement. The brain of the piglet was centered in the field of view of the CT scanner. The two front legs of the

Figure 3. Calculated impulse residue function. CBF scaled impulse residue function $[F \cdot R(t)]$ from a piglet at normocapnia. Height of initial plateau yields CBF, area under the curve yields CBV, and MTT is equal to the area (CBV) divided by the height (CBF).

Figure 4. NIRS-measured PaCO_2 reactivity. Mean cerebral blood flow at various PaCO_2 levels for six piglets. In four piglets, repeat CBF measurements were made at each of three PaCO_2 levels, and means at each level are presented (open circles). In the remaining two piglets, repeat CBF measurements were made at two PaCO_2 levels and means at each level are presented (filled circles). The dashed line shows the linear regression line for the CBF PaCO_2 reactivity data set as a whole. The slope of the regression line is $1.36 \text{ mL} \cdot \text{min}^{-1} \cdot 100 \text{ g}^{-1} / \text{mm Hg}$.

piglet are visible in the bottom corners of the image. The piglet was held in a frame, visible above and to the side of the animal. Two holes used to hold the NIR probes on the head of the piglet are visible in the top section of the frame. Figure 5, B–D, shows CBF, CBV, and MTT cross-sectional maps, at normocapnia, corresponding to the images presented in Figure 5A.

Figure 6 presents the Bland-Altman plot comparing CT and NIRS measurements of CBF (22). The mean difference between the two methods is $-2.05 \text{ mL} \cdot \text{min}^{-1} \cdot 100 \text{ g}^{-1}$. The limits of agreement, the region in which 95% of the differences lie, are -14.48 and $10.39 \text{ mL} \cdot \text{min}^{-1} \cdot 100 \text{ g}^{-1}$. The 95% confidence interval of the mean difference, the region in which we expect (with 95% confidence) the bias lies, is -4.21 to $0.12 \text{ mL} \cdot \text{min}^{-1} \cdot 100 \text{ g}^{-1}$. Similar Bland-Altman plots for CBV and MTT measurements have also been produced and show mean differences of $-0.18 \text{ mL} \cdot 100 \text{ g}^{-1}$ and 0.43 s , respectively. Two-way ANOVA showed no significant differences between CBF, CBV, and MTT measurements made using the CT and NIRS techniques ($p > 0.1$). Regression plots showed relationships of $\text{CBF}_{\text{CT}} = -1.98 + 1.00 \cdot \text{CBF}_{\text{NIRS}}$ ($r^2 = 0.86$) for CBF, $\text{CBV}_{\text{CT}} = 0.29 + 0.84 \cdot \text{CBV}_{\text{NIRS}}$ ($r^2 = 0.75$) for CBV, and $\text{MTT}_{\text{CT}} = 0.11 + 1.07 \cdot \text{MTT}_{\text{NIRS}}$ ($r^2 = 0.80$) for MTT. The precision of NIRS CBF, CBV, and MTT measurements, determined using repeated-measures ANOVA, are 9.71%, 13.05%, and 7.57%, respectively.

Figure 5. CT images and corresponding CBF, CBV, and MTT functional maps. (A) Two of a set of four coronal CT images at peak contrast enhancement. The two front legs of the piglet are visible in the bottom corners of the images (black arrows), the brain of the piglet is centered in the image (gray arrow), the lateral ventricles are visible in the mid-brain (transparent arrows), and the arteries are visible as bright spots (white arrows). (B) Corresponding blood flow maps. (C) Corresponding blood volume maps. (D) Corresponding MTT maps. Maps are color coded: CBF blue = 0 mL/min/100 g , orange = 100 mL/min/100 g ; CBV blue = 0 mL/100 g , orange = 10 mL/100 g ; MTT blue = 0 s , orange = 10 s . Structures in the CBF and CBV maps that appear orange, flow or volume exceeds 100 mL/min/100 g or 10 mL/100 g , respectively, are arterial and venous structures.

DISCUSSION

The NIRS results correlate well with CT measurements over a range of hemodynamic conditions. Two-way

Table 2. *Physiological parameters for repeated NIRS measurements*

Piglet no.	Measurement	pH	Paco ₂ (mm Hg)	Pao ₂ (mm Hg)	HCO ₃ (mmol/L)	MAP (mm Hg)
1	Hypocapnia	7.587	32.1	381.3	30.7	44
		7.581	32.4	378.8	30.8	45
2	Hypocapnia	7.53	30.2	396.4	25.5	55
		7.541	30.9	408.3	26.8	57
	Normocapnia	7.45	39.3	420.3	27.3	57
		7.425	38.4	373.1	25.1	57
	Hypercapnia	7.35	53.1	364.2	28.7	55
		7.322	54.5	369.5	28.9	56
3	Hypocapnia	7.58	29.3	358.9	27.6	50
		7.589	29.6	340.6	28.5	54
	Normocapnia	7.486	39.2	381.8	29.2	51
		7.466	39.5	390	28.1	50
4	Hypocapnia	7.566	30.9	83.4	28	53
		7.568	31.1	110.1	28.3	52
	Normocapnia	7.455	42	187	29	61
		7.452	42.4	108.1	29.2	56
5	Normocapnia	7.502	38.5	207	30.2	56
		7.474	37.9	234.1	27.7	56
	Hypercapnia	7.357	57.5	181.6	31.5	57
		7.337	56.2	174.7	29.2	57
6	Hypocapnia	7.548	32.3	204.4	28.8	52
		7.555	31.4	199.9	28.4	48
	Normocapnia	7.459	41.4	204.8	29.8	52
		7.457	42.1	208.7	30.2	50
	Hypercapnia	7.342	54	177	28.8	51
		7.357	52.3	209.5	29	50

MAP, mean arterial pressure.

ANOVA showed no significant differences between CT and NIRS measurements of CBF, CBV, and MTT. The estimated bias, using the Bland-Altman method (22), was small relative to the respective mean values in each case. In some instances, the disagreement between NIRS- and CT-measured CBF values was quite large. This, coupled with the observation that linear regression fits to the individual NIRS-measured CBF Paco₂ data were approximately parallel, suggests that the NIRS technique may be better suited for comparison of CBF values measured from the same piglet rather than for comparisons between piglets. These results show that the NIRS system is capable of accurate, reliable, bedside measurement of cerebral hemodynamics in newborn piglets. There are, however, several factors that should be addressed before clinical trials of the NIRS system can be considered. With respect to skull composition and thickness, the piglet head mimics that of a newborn infant. However, the two are significantly different in terms of size and dimension. In this study, an optode separation of 3.0 cm was used; this yields a maximum light path penetration depth of 1.73 cm. It has been shown that in newborn infants, the tissues overlying the brain are <5 mm thick and do not interfere with NIR monitoring of the brain for emitter-detector separation of at least 3 cm (23). An increase in optode separation provides increased penetration depth, but it also introduces noise into the detected NIR signal. This would eventually affect the ability of the NIRS system to reliably measure concentration and thus would have a deleterious effect on the precision of NIRS cerebral hemodynamic measurements. In our piglet studies, physical lim-

itations prevented use of a greater optode separation. The minimal level of noise present in the concentration data presented here suggests that small increases in optode separation would not greatly affect the overall performance of the system. In addition, the system allows measurement in a single region only, meaning that an investigation of the entire cerebral cortex would require repositioning of optodes. Although repositioning the optodes is not difficult, it does detract from the convenience and ease of use of the system. Addition of further detection optodes around the emission optode would allow sampling of multiple regions, thereby providing measurements across the entire brain (cortex) and eliminating the need to reposition optodes.

A final concern with the NIRS method is that it requires an injection of ICG. Although the use of an exogenous blood flow tracer is somewhat less appealing than using one of the endogenous NIR absorbers (*e.g.* HbO₂ or Hb), it is often of less risk to an already unstable patient than changing the arterial oxygen saturation (Sao₂) of the subject and can be more convenient in the clinical setting. Furthermore, ICG results in much higher signal-to-noise ratio data than does manipulation of Sao₂. ICG has been used to measure various physiologic parameters such as cerebral blood flow (11), hepatic function (24), plasma volume (25), cardiac output (26), and left-to-right shunts (27) in infants since 1965. One study, involving more than 3000 human subjects, showed that i.v. injection of ICG resulted in significant allergic reaction in only one subject (28).

In current clinical practice, identification of neonates at risk of ischemic injury relies heavily on measurements made using ultrasonography (29). This modality is not able to

Figure 6. Bland-Altman plot comparing CT and NIRS measurements of CBF. The mean difference between the two methods is $-2.05 \text{ mL} \cdot \text{min}^{-1} \cdot 100 \text{ g}^{-1}$. The limits of agreement, the region in which 95% of the differences lie, are -14.73 and $10.64 \text{ mL} \cdot \text{min}^{-1} \cdot 100 \text{ g}^{-1}$. The 95% confidence interval of the mean difference, the region in which we expect (with 95% confidence) the bias lies, is -4.21 to $0.12 \text{ mL} \cdot \text{min}^{-1} \cdot 100 \text{ g}^{-1}$. The *dashed line* shows the mean difference, the *dash-dot-dash lines* show the limits of agreement, and the *solid lines* show the 95% confidence intervals of the mean difference.

provide absolute values for CBF because it measures only blood flow velocities in the vessels studied, generally the middle cerebral arteries. Because the middle cerebral artery carries a large portion of the total blood flow to the brain, it is assumed that its flow is representative of the bulk hemispheric blood flow. Although detection of high velocities can lead to relatively accurate assumptions about the state of global hemispheric blood flow, the method is not able to detect minor, but potentially damaging, changes in CBF, nor is it able to provide any information about the location of possible infarct or injury.

The results presented here clearly demonstrate the ability of the NIRS system to overcome some of the limitations of ultrasonography. It is able to accurately measure cerebral hemodynamics at the bedside and thus would be an invaluable tool for monitoring and diagnosing both hemorrhage and ischemic injury in severely premature infants. In the case of ischemic injury, blood flow is often low for an extended period of time before symptoms of injury are apparent. The ability to detect diminished blood flow at an earlier stage would significantly aid earlier intervention and greatly reduce mortality and morbidity from ischemic injury. It has also been suggested that loss of Paco_2 reactivity is associated with poor neurodevelopmental outcome and/or hypoxic-ischemic encephalopathy and that the loss of Paco_2 reactivity is a candidate for predicting early severe brain damage in preterm infants (30, 31). Because the NIRS system presented in this study is capable of providing reliable, noninvasive, repeatable measurements, it can be used to detect loss of Paco_2 reactivity, thereby greatly improving diagnosis. The portable nature of the system also provides a distinct advantage over conventional CT/MRI scanners for those patients whose condition renders transport unfavorable.

Acknowledgments. The authors thank animal health technician Jennifer Hadway for her help in the experiments.

REFERENCES

1. Jobsis FF 1977 Non-invasive infrared monitoring of cerebral and myocardial oxygen sufficiency and circulatory parameters. *Science* 198:1264–1267
2. Brazy JE, Lewis DV 1986 Changes in cerebral blood volume and cytochrome aa3 during hypertensive peaks in preterm infants. *Pediatrics* 108:983–987
3. Ferrari M, De Marchis C, Giannini I, DiNicola A, Agostino R, Nodari S, Bucci G 1986 Cerebral blood volume and haemoglobin oxygen saturation monitoring in neonatal brain by near infrared spectroscopy. *Adv Exp Med Biol* 200:203–211
4. Chance B, Legallais V, Sorge J, Graham N 1975 A versatile time-sharing multichannel spectrophotometer, reflectometer, and fluorometer. *Anal Biochem* 66:498–514
5. Wyatt JS, Cope M, Delpy DT, Wray S, Reynolds EO 1986 Quantification of cerebral oxygenation and haemodynamics in sick newborn infants by near infrared spectrophotometry. *Lancet* 2:1063–1066
6. Cope M, Delpy DT 1988 System for long-term measurement of cerebral blood and tissue oxygenation on newborn infants by near infra-red transillumination. *Med Biol Eng Comput* 26:289–294
7. Yoxall CW, Weindling AM, Dawani NH, Peart I 1995 Measurement of cerebral venous oxyhemoglobin saturation in children by near-infrared spectroscopy and partial jugular venous occlusion. *Pediatr Res* 38:319–323
8. Reynolds EO, Wyatt JS, Azzopardi D, Delpy DT, Cady EB, Cope M, Wray S 1998 New non-invasive methods for assessing brain oxygenation and haemodynamics. *Br Med Bull* 44:1052–1075
9. Delpy DT, Cope M, van der Zee P, Arridge S, Wray S, Wyatt JS 1988 Estimation of optical pathlength through tissue from direct time of flight measurement. *Phys Med Biol* 33:1433–1442
10. Elwell CE, Cope M, Edwards AD, Wyatt JS, Delpy DT, Reynolds EO 1994 Quantification of adult cerebral hemodynamics by near-infrared spectroscopy. *J Appl Physiol* 77:2753–2760
11. Patel J, Marks K, Roberts I, Azzopardi D, Edwards AD 1998 Measurement of cerebral blood flow in newborn infants using near infrared spectroscopy with indocyanine green. *Pediatr Res* 43:34–39
12. Matcher SJ, Cope M, Delpy DT 1994 Use of the water absorption spectrum to quantify tissue chromophore concentration changes in near-infrared spectroscopy. *Phys Med Biol* 39:177–196
13. Cenic A, Nabavi DG, Craen RA, Gelb AW, Lee TY 1999 Dynamic CT measurement of cerebral blood flow: a validation study. *AJNR Am J Neuroradiol* 20:63–73
14. van Bel F, van de Bor M, Baan J, Ruys JH 1988 The influence of abnormal blood gases on cerebral blood flow velocity in the preterm newborn. *Neuropediatrics* 19:27–32
15. Madden JA 1993 The effect of carbon dioxide on cerebral arteries. *Pharmacol Ther* 59:229–250
16. Bonner RF, Nossal R, Havlin S, Weiss GH 1987 Model for photon migration in turbid biological media. *J Opt Soc Am A Opt Image Sci Vis* 4:423–432
17. Bassingthwaite JB, Chinard FP, Crone C, Lassen NA, Perl W 1970 Definitions and terminology for indicator dilution methods. In: Crone C, Lassen NA (eds) *Capillary Permeability*. Munksgaard, Copenhagen, pp 665–669
18. Zierler KL 1965 Equations for measuring blood flow by external monitoring of radioisotopes. *Circ Res* 16:309–321
19. Meier PP, Zierler KL 1954 On the theory of the indicator-dilution method for measurement of blood flow and volume. *J Appl Physiol* 6:731–744
20. Snedecor GW, Cochran WG 1967 *Statistical Methods*. The Iowa State University Press, Ames, IA, pp 472–477
21. Lawson CE, Hanson RJ 1974 *Solving Least Squares Problems*. Prentice-Hall, Englewood Cliffs, NJ, pp 159–173
22. Bland JM, Altman DG 1986 Statistical methods for assessing agreement between two methods of clinical measurement. *Lancet* 1:307–310
23. Benaron DA, Kurth CD, Steven JM, Delivoria-Papadopoulos M, Chance B 1995 Transcranial optical path length in infants by near-infrared phase-shift spectroscopy. *J Clin Monit Comput* 11:109–117
24. Evans WE, Relling MV, de Graaf S, Rodman JH, Pieper JA, Christensen ML, Crom WR 1989 Hepatic drug clearance in children: studies with indocyanine green as a model substrate. *J Pharm Sci* 78:452–456
25. Anthony MY, Goodall SR, Papouli M, Levene MI 1992 Measurement of plasma volume in neonates. *Arch Dis Child* 67:36–40
26. O'Connell AJ, Tibbals J, Coulthard M 1991 Improving agreement between thoracic bioimpedance and dye dilution cardiac output estimation in children. *Anaesth Intensive Care* 19:434–440
27. Krovetz LJ, Gessner IH 1965 A new method utilizing indicator-dilution techniques for estimation of left-to-right shunts in infants. *Circulation* 32:772–777
28. Hope-Ross M, Yannuzzi LA, Gragoudas ES, Guyer DR, Slakter JS, Sorenson JA, Krupsky S, Orlock DA, Puliafito CA 1994 Adverse reactions due to indocyanine green. *Ophthalmology* 101:529–533
29. Blankenberg FG, Loh NN, Bracci P, D'Arceuil HE, Rhine WD, Norbush AM, Lane B, Berg A, Person B, Coutant M, Enzmann DR 2000 Sonography, CT, and MR imaging: a prospective comparison of neonates with suspected intracranial ischemia and hemorrhage. *AJNR Am J Neuroradiol* 21:213–218
30. Blankenberg FG, Loh NN, Norbush AM, Craychee JA, Spielman DM, Person BL, Berg CA, Enzmann DR 1997 Impaired cerebrovascular autoregulation after hypoxic-ischemic injury in extremely low-birth-weight neonates: detection with power and pulsed wave Doppler US. *Radiology* 205:563–568
31. Muller AM, Morales C, Briner J, Baenziger O, Duc G, Bucher HU 1997 Loss of CO_2 reactivity of cerebral blood flow is associated with severe brain damage in mechanically ventilated very low birth weight infants. *Europ J Paediatr Neurol* 1:157–163