ARTICLE

Open Access

Optical conductivity-based ultrasensitive mid-infrared biosensing on a hybrid metasurface

Yibo Zhu¹, Zhaoyi Li², Zhuang Hao¹, Christopher DiMarco¹, Panita Maturavongsadit³, Yufeng Hao⁴, Ming Lu⁵, Aaron Stein⁵, Qian Wang³, James Hone¹, Nanfang Yu² and Qiao Lin¹

Abstract

Optical devices are highly attractive for biosensing as they can not only enable quantitative measurements of analytes but also provide information on molecular structures. Unfortunately, typical refractive index-based optical sensors do not have sufficient sensitivity to probe the binding of low-molecular-weight analytes. Non-optical devices such as field-effect transistors can be more sensitive but do not offer some of the significant features of optical devices, particularly molecular fingerprinting. We present optical conductivity-based mid-infrared (mid-IR) biosensors that allow for sensitive and quantitative measurements of low-molecular-weight analytes as well as the enhancement of spectral fingerprints. The sensors employ a hybrid metasurface consisting of monolayer graphene and metallic nano-antennas and combine individual advantages of plasmonic, electronic and spectroscopic approaches. First, the hybrid metasurface sensors can optically detect target molecule-induced carrier doping to graphene, allowing highly sensitive detection of low-molecular-weight analytes despite their small sizes. Second, the resonance shifts caused by changes in graphene optical conductivity is a well-defined function of graphene carrier density, thereby allowing for quantification of the binding of molecules. Third, the sensor performance is highly stable and consistent thanks to its insensitivity to graphene carrier mobility degradation. Finally, the sensors can also act as substrates for surface-enhanced infrared spectroscopy. We demonstrated the measurement of monolayers of sub-nanometer-sized molecules or particles and affinity binding-based quantitative detection of glucose down to 200 pM (36 pg/mL). We also demonstrated enhanced fingerprinting of minute quantities of glucose and polymer molecules.

Introduction

Optical transducers are one of the most promising candidates for label-free, sensitive, and specific on-chip biosensing.^{1–3} Optical devices are unique in their ability to not only detect binding between biochemical molecules but also probe their molecular structures^{4–8}. Most optical sensors, such as those based on surface plasmon

resonance⁹ or optical fibers¹⁰, rely on the detection of changes in refractive indices to quantify analytes or monitor binding processes, but they unfortunately lack the sensitivity to probe the binding involving low-molecular-weight molecules^{11,12}. Spectroscopic methods can reveal molecular fingerprints and quantify analytes according to signal intensities but are limited by noisy background and low sensitivity as well^{13,14}. Non-optical devices such as field effect transistors may provide more sensitive alternatives^{11,15–17} at the expense of molecular spectroscopic information, convenient wireless communication between optical devices, and other features of optical methods. Also, use of these devices may face other complexities in biosensing, such as unstable contact,

Correspondence: Nanfang Yu (ny2214@columbia.edu) or Qiao Lin (qlin@columbia.edu)

¹Department of Mechanical Engineering, Columbia University, New York, NY 10027, USA

²Department of Applied Physics and Applied Math, Columbia University, New York, NY 10027, USA

Full list of author information is available at the end of the article.

These authors contributed equally: Yibo Zhu, Zhaoyi Li

© The Author(s) 2018

Open Access This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons license, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons license, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons license and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this license, visit <http://creativecommons.org/licenses/by/4.0/>.

Fig. 1 Architecture and optical properties of the hybrid metasurface. **a** Schematic of the graphene-metallic metasurface, with small molecules adsorbed on the suspended graphene. **b** Optical micrograph of one device, showing that the graphene monolayer is continuous and uniform over a large device area. **c** Scanning electron microscopy (SEM) image of graphene-coated nanorod antennas. Inset: SEM image of one antenna gap with suspended graphene. **d** Reflectance spectrum of a device showing a primary plasmonic resonance (ω_r) at $\sim 1500\text{ cm}^{-1}$, a resonance dip (ω_{rr}) at 1000 cm^{-1} and a PMMA absorption peak (ω_{pmma}) near 1700 cm^{-1}

electrochemical corrosion, current leakage, and reduced sensitivity in high-ionic-strength media^{18–21}.

Very recently, with highly tunable electrical and optical properties, two-dimensional materials have been employed to construct novel optoelectronic devices^{22–25}. In particular, graphene is very well-suited to develop optical modulators or photodetectors operating in the mid-infrared (mid-IR)^{26–28} range, where spectral fingerprints of many molecular bonds can be found. However, the potential of these new device architectures in bio-sensing has not been widely explored, especially for the detection of low-molecular-weight analytes. While many studies have reported enhancement of the mid-IR fingerprints of protein or polymer layers using metal or graphene plasmons^{29–32}, to date very limited advances have been made in the measurement of the binding of low-molecular-weight analytes in the mid-IR range; such measurements are challenging because of the significant mismatch in size between their sub-nanometer size and micrometer-scale IR wavelengths^{30,33}.

In this paper, we demonstrate optical conductivity-based mid-IR biosensing that allows the sensitive measurement of low-molecular-weight analytes as well as the

enhancement of their spectral fingerprints. The sensors used in this work employ a hybrid metasurface consisting of monolayer graphene and a metallic nano-antenna array. They offer a unique set of advantages that are not simultaneously available from existing electrically or optically based devices. First, the plasmonic resonance frequency of the sensors varies with the carrier doping introduced by target molecules adsorbed on graphene; therefore, the sensor is not limited by the molecular mass and can sensitively detect low-molecular-weight molecules thanks to the high sensitivity of graphene to molecular doping^{34,35}. Meanwhile, the optical measurements are readily amenable to wireless deployment and avoid the complexities involved in electrical measurements that are typically used to measure carrier doping. Second, the sensors can resolve molecules with similar molecular masses by differentiating the resonance shifts caused by the changes in graphene optical conductivity, which is not possible with refractive index-based sensors. Third, the hybrid metasurface features a quality factor that is insensitive to variation in graphene carrier mobility and thus can be more stable than the devices using graphene as their sole functional material, which in general are

affected by the degradation of carrier mobility^{36,37}. Finally, the sensors can act as substrates for surface-enhanced IR absorption (SEIRA) spectroscopy to enhance the fingerprints in the mid-IR range and facilitate the identification of analytes.

We demonstrated the use of these hybrid metasurface sensors for sensitive detection of low-molecular weight analytes. Sub-monolayers of sub-nanometer analytes were detected and differentiated from each other. Glucose from 2 nM to 20 mM was quantified via its reversible binding with boronic acid anchored on graphene. We experimentally and theoretically studied the relationship between the metasurface geometry, electric-field confinement, and sensor sensitivity. This led us to further optimize the sensor design and achieve a detection limit of 200 pM glucose by confining the electric-field on ultra-narrow (10 nm) suspended graphene. Additionally, we demonstrated enhanced fingerprinting of minute quantities of glucose and polymer molecules on the metasurface, which are otherwise unrecognizable without enhancement. This work opens a path to the development of high performance, multifunctional sensors using 2D–3D hybrid structures for biomedical applications.

Results

Sensor response to molecular doping

The hybrid metasurface consisted of an Au nanorod antenna array (length $L = 1.64 \mu\text{m}$) covered by monolayer graphene, atop a Pt mirror with a SiO_2 spacer layer in between (Figs. 1a, b). The size of the gaps separating adjacent nanorod antennas was 30 nm. The SiO_2 layer with the Pt mirror formed an optical cavity. When incident light excites oscillating dipoles (i.e., plasmonic resonance) in the antenna, the optical cavity can introduce imaging dipoles at the same distance underneath the Pt mirror, oscillating in opposite directions. A pair of such dipoles are coupled via near-field interactions and can form a quadrupole to suppress radiation loss in the far field and improve the quality factor of the plasmonic resonance band³⁸. Benefiting from its excellent mechanical strength and flexibility, graphene conformed to the antennas and was suspended at the 30 nm gaps over a large device area (Fig. 1c and Figure S1). Compared to the previously reported devices that loaded graphene underneath the metallic antennas^{26,28,39}, our device structure grants easier access of molecules to the graphene in the nanogap and thus is more suitable for sensing applications. Originating from the coupling between the SiO_2 phonon resonance and the metasurface plasmon resonance, two resonance dips were respectively observed near 1500 cm^{-1} (ω_i) and 1000 cm^{-1} (ω_{tr}) in the mid-IR spectral region where the optical conductivity of graphene is highly tunable^{22,40}. In this work, the peak near 1500 cm^{-1} was used as the primary plasmonic resonance band

for biosensing, as it showed stronger absorption (Fig. 1d), up to 90% at ω_i with a signal to noise ratio exceeding 1000. Such a high signal-to-noise ratio ensures that the absorption bands of other materials, for example, the poly (methyl methacrylate) (PMMA) residue from device fabrication, would not significantly influence the profile of the plasmonic resonance band.

To characterize the plasmonic resonance frequency of the hybrid metasurface as a function of the molecular doping of graphene, we first altered the carrier density of graphene by immobilizing small molecules on graphene. The hybrid metasurfaces with bare graphene were respectively exposed to the organic solvent acetonitrile (ACN) and two pyrene derivatives dissolved in ACN: amino-pyrene (AP, 217.27 g/mol) and boronic acid-pyrene (BAP, 246.07 g/mol). Raman spectroscopy and atomic force microscopy (AFM) indicated that the AP or BAP molecules were immobilized on graphene via π – π interactions (Figure S2a). Measurements of transport characteristics of graphene field effect transistors indicated that the pure ACN bath introduced p-doping to graphene, increasing the hole density (Δn_h) by $2.3 \times 10^{12} \text{ cm}^{-2}$ (Fig. 2a). Immobilization of AP and BAP molecules produced additional increases in hole density of $0.46 \times 10^{12} \text{ cm}^{-2}$ and $1.58 \times 10^{12} \text{ cm}^{-2}$, respectively. Accordingly, ω_i shifted to higher wavenumbers by 23 cm^{-1} , 27 cm^{-1} , and 46 cm^{-1} after treatment with ACN, AP, and BAP, respectively (Fig. 2b). These blue shifts, opposite to the red-shifts found in typical plasmonic sensors, are attributed to the increases in the graphene optical conductivity caused by the molecular doping to graphene, corresponding to a more negative real part of the graphene permittivity²⁶. The amount of the shift, $|\Delta\omega_i|$, as a function of the carrier density doped into graphene by the molecules (Fig. 2c), allowed us to quantify the binding of low-molecular-weight molecules, as demonstrated below.

To further examine the contributions of the changes in optical conductivity and refractive index to the shift of ω_i , a sub-nanometer thick (0.8 nm) Al layer was vapor-phase deposited onto the bare Au metasurface (without graphene) and the hybrid metasurface, respectively. Different from the pyrene derivatives, the Al nanoparticles could be homogeneously coated on the bare sensor surface without graphene so that the effect of the refractive index changes on the shift of ω_i could be examined independently. As the work function of Al (4.28 eV) is lower than that of pristine graphene (4.6 eV), the Al nanoparticles upshifted the graphene Fermi level and donated electrons to graphene. The increase in carrier density caused by the Al nanoparticles was $2 \times 10^{12} \text{ cm}^{-2}$ (Figure S3b), leading to an apparent 24.1 cm^{-1} blue-shift of ω_i in the hybrid metasurface (Fig. 2d). After exclusion of the 2.2 cm^{-1} red-shift observed on the bare Au metasurface, which was due to the change of refractive index by the oxidized Al, the total blue-shift caused by the electron doping of the Al

nanoparticles in the hybrid metasurface was 26.3 cm^{-1} . In other words, the carrier doping-induced shift was ~ 12 times the shift induced by the change of the local refractive index. This confirmed the dominant contribution from graphene optical conductivity in the detection of these sub-nanometer subjects, especially for even smaller analytes, such as the pyrene derivatives (Figure S2b).

Insensitivity of Q factor to graphene mobility

Besides the carrier density (n), the carrier mobility (μ) of graphene usually also influences the performance of graphene-based devices. For example, the intensity and quality factor of the plasmonic resonance excited on graphene nanoribbons vary significantly with graphene carrier mobility^{31,37}. Interestingly, in our hybrid metasurface sensor, the profile of the plasmonic resonance peak did not show appreciable variations. In contrast, the carrier mobility of graphene decreased significantly after chemical treatment, as was evident from the considerably less sharp transfer characteristics (Fig. 2a). The quality factor (Q) of the plasmonic resonance peak remained

approximately constant at 5.5 within a deviation of 5.6% (Fig. 2e), although the hole mobility was lowered by nearly 70%, from $4570 \text{ cm}^2 \text{ V}^{-1} \text{ s}^{-1}$ to $1420 \text{ cm}^2 \text{ V}^{-1} \text{ s}^{-1}$ at 0 V gate bias (see Supplementary Information section 1.2 for details). This is in agreement with the results of the finite-difference time-domain (FDTD) simulation (Figure S4). To better understand the effects of the carrier mobility, we also developed an equivalent circuit model of the hybrid metasurface (Fig. 2f), in which the current I_{tot} represents the reflectance of the metasurface. As expected, the frequency response of I_{tot} highly depended on the graphene carrier density but barely varied with the graphene carrier mobility (Figure S5c). Such weak dependence of ω_r on graphene carrier mobility can be explained by the insensitivity of graphene impedance to the carrier mobility. We note that the graphene impedance referred to here is not equivalent to the electrical resistance but is a function of graphene optical conductivity (see Section 3 in the Supplementary Information for details). The calculated impedance of the graphene load shows approximately 100 times stronger dependence on carrier density than on carrier mobility (Table S1). Therefore, the hybrid

plasmonic sensor exhibits not only high sensitivity to adsorption of small molecules, but also excellent robustness against the degradation of the graphene carrier mobility.

Reversible quantification of glucose

We then conducted measurements of glucose, a monosaccharide that serves as an important source of energy in cellular respiration, to demonstrate the sensitive quantification of small-molecule biomarkers using the hybrid metasurface. Recently, the affinity binding between glucose and boronic acid has attracted much attention in the development of next-generation, enzyme-free, and oxygen independent synthetic chemosensors⁴¹. However, affinity binding-based optical techniques for quantitative measurements of glucose at low concentrations have been scarce, possibly due to the low molecular mass and electrical neutrality of glucose. It has been shown that, upon glucose binding to boronic acid, the BAP could be converted from an electron-withdrawing group to an electron-donating group^{42–45}. This transition of boronic acid is expected to give rise to changes in the graphene optical conductivity and allows for glucose sensing in the hybrid metasurface.

To first examine the possible responses introduced by refractive index variations due to glucose binding, control experiments were conducted on bare Au metasurfaces treated by 4-mercaptophenylboronic acid (4-PBA) (Fig. 3a). X-ray photoelectron spectroscopy was used to confirm the successful immobilization of 4-PBA on the bare Au metasurface (Figure S7). The shift of ω_r of the Au metasurface after 4-PBA immobilization followed by exposure to 20 mM glucose solution was merely 1–2 cm^{-1} (Fig. 3b), which corroborated the inability of a monolayer of small molecules to introduce significant variations in local refractive indices.

We then performed measurements on the hybrid metasurface. Without any intended functionalization of graphene, ω_r did not change after exposure of the device to glucose solutions (Figure S8). In contrast, when the graphene was treated by BAP, ω_r significantly shifted to lower wavenumbers after exposure to glucose solutions (Figs. 3c, d). This was because the boronic acid, with an empty orbital, initially p-doped graphene by withdrawing electrons from graphene, and this electron-withdrawing capability was weakened upon the binding of glucose (Figure S9). Exposure of the device to 2 nM (0.36 ng/mL) glucose led to a clearly resolvable red-shift of $\sim 1.5 \text{ cm}^{-1}$. The detection limit was improved by almost five orders of magnitude compared to the existing affinity binding-based optical glucose sensors that measured the changes in local or bulk refractive indices^{46–49}. We note that the high sensitivity does not rely on a specific biochemical process and can be adapted to sense other small

molecules, such as gas molecules, since graphene is very sensitive to gas adsorption³⁴. Increasing the glucose concentration up to 20 mM monotonically red-shifted the plasmonic resonance. Figure 3e plots the shifts amount of ω_r for different sensor configurations, which clearly shows that the combination of graphene and the metallic metasurface, with BAP functionalization, led to the most significant sensor responses. The dynamic range of the sensor, extracted from this plot, was 2 to 10 mM, over six orders of magnitude. Fitting of the experimental results to the Hill–Langmuir equation gave an estimated dissociation constant of 99.2 μM (Figure S10). Fig. 3f shows the ex-situ measurements after exposure to high glucose concentrations followed by rinsing with fresh buffer solution. The increase and decrease in $|\Delta\omega_r|$, observed after rinsing with glucose solutions and fresh buffer solutions, respectively, demonstrated that the sensor was able to detect both increases and decreases in glucose concentrations thanks to the reversibility of the glucose-boronic acid binding.

Influence of antenna geometry on sensitivity

We further investigated the influence of the antenna's geometry on its biosensing sensitivity. First, with a fixed 30-nm gap size between the neighboring antennas, we conducted studies to compare the sensitivities of rod, disk, and diamond-shaped antennas (Fig. 4a). The dimensions of these antennas were chosen to ensure that the initial resonance frequencies, ω_{r0} , of the metasurfaces (i.e., antenna arrays with bare graphene) were close to each other. The relative shifts of the plasmonic resonance, defined as $\Delta\omega_r$ normalized by ω_{r0} , were used to measure sensitivity. After exposure to 2 mM glucose solution, $\Delta\omega_r/\omega_{r0}$ were 0.54 and 0.38% for disk and diamond-shaped antennas, respectively (Fig. 4a), which were reduced by 50.4 and 65.1%, compared to that achieved using the rod antennas. FDTD simulation was performed based on a larger modulation of graphene carrier density of $8 \times 10^{12} \text{ cm}^{-2}$ to achieve more readily recognizable shifts of $\Delta\omega_r$. The values of $\Delta\omega_r/\omega_{r0}$ in our experiments are in good agreement with the simulation results, which predicted that the sensitivities of the disk and diamond-shaped antennas were decreased by 42.2% and 45.1%, respectively, compared to that of the rod antennas. We then fixed the geometry of nanorod antennas and studied the effect of the gap size g on the sensitivity. When g was increased from 30 to 100 nm and then to 200 nm (Fig. 4b), $\Delta\omega_r/\omega_{r0}$ introduced by 20 mM glucose decreased by nearly 50% and 80%, respectively, qualitatively agreeing with the simulation results.

To explain the dependence of sensitivity on the antenna geometry, we also calculated the sensitivities for different antenna geometries based on perturbation theory. The equation below relates the sensitivity to the electrical field

Fig. 3 Affinity binding-based measurement of glucose. **a** Schematic of a control device without graphene. **b** Reflectance spectra measured with 4-PBA functionalized control device at different glucose concentrations. **c** Schematic of a hybrid metasurface sensor. **d** Spectral measurement of glucose from 2 to 20 mM on the hybrid metasurface. For each concentration, three measurements were conducted on the same device. **e** $|\Delta\omega_r|$ as a function of glucose concentration for different sensor configurations. **f** Reversible measurement of glucose. The data points were achieved from ex-situ measurements after rinsing the device with each solution. The x-axis reflects the measurement sequence and the time of each solution rinsing. For (e)–(f), error bars represent standard deviations calculated from nine measurements (three measurements on each of three independent devices)

energy confined by the metasurface (see the Supplementary Information for a detailed derivation of this equation):

$$\left| \frac{\Delta\omega_r}{\omega_{r0}} \right| \propto |\Delta\epsilon| \cdot \left| \frac{\iiint ds (\vec{E} \cdot \vec{E}_0^*)}{\iiint dv (\epsilon |\vec{E}_0|^2)} \right| = |\Delta\epsilon| \cdot K \quad (1)$$

Here, $\Delta\epsilon$ represents the change of the local permittivity due to the variation of graphene optical conductivity, E_0 and E are, respectively, the electrical fields before and after perturbation (i.e., molecular doping of graphene in this study). S in the numerator integration is the area of the graphene suspended over the nanogap, while V in the denominator covers all the space where E_0 is non-zero

(see Supplementary Information section 5.1 for details). The factor K is a variant representing the relative change in electric-field energy distribution, which depends on the antenna geometry and can be calculated from the near-field distributions (Fig. 4c, d and Figure S11). The trends of the theoretically calculated K (see the Supplementary Information for details) and the simulated $\Delta\omega_r/\omega_{r0}$ agreed nicely with each other for the above five antenna geometries (Fig. 4e). This confirms that the nanorod antennas with 30 nm gaps offer the highest sensitivity. The nanorod antennas show the best performance among the three commonly used antenna geometries because they provide a balance between a strong electric-field enhancement and the volume within which the electric-field can effectively interact with adsorbed molecules on graphene. In contrast, although the diamond antenna shows a higher maximum value of the field enhancement, the ratio of the E field confined at the graphene surface relative to the total E field confined within antenna system was lower. This can also be found from the cross-section views of the near-field distribution and explains the inferior sensitivity of the diamond antenna. On the other hand, with a given antenna shape, increasing the gap between the adjacent nanorod antennas leads to a linear increase in the perturbation volume but an exponential decrease in the

electric-field enhancement and thus will also reduce the sensitivity.

It is noteworthy that the electric-field distribution is changed significantly by adding a monolayer of graphene on the metallic antennas. When graphene is present at the antenna gap, surface plasmon can be excited on graphene surface; thus, standing waves can be formed with multiple maxima located on the suspended graphene, subject to the boundary conditions defined by the metal antenna where the electrical field is zero. In contrast, metallic metasurfaces only support hotspots at the edges of antennas (Fig. 4c). The hotspots distributed along the suspended graphene, despite the slightly reduced maximum enhancement value, create more space for effective light-matter interaction and may particularly benefit measurement at very low concentrations.

The monotonic increase of the sensitivity with decreasing antenna gap size predicts that the sensitivity would continue to improve by bringing the neighboring antennas closer to each other. However, reducing the gap size to below 30 nm would be very challenging for standard lithography and lift-off processes considering the high aspect ratio of the nanostructures (antenna width of ~ 200 nm and thickness of ~ 50 nm). Therefore, a new method was developed to fabricate antenna arrays with a

Fig. 5 Sensitivity improvement in metasurfaces with 10-nm antenna gaps. **a** Fabrication process to create 10-nm antenna gaps. **b** SEM images of one antenna gap at each step of fabrication. **c** Comparison of sensitivities between the cases of $g = 30$ and 10 nm. **d** Detection of 200 pM glucose on a metasurface with $g = 10$ nm

value of 10 nm (Fig. 5a). First, a layer of Al_2O_3 was deposited on the SiO_2 as a protection layer using atomic layer deposition (ALD). Then, Pt antennas with initial gaps of 30 nm were fabricated and an additional 10-nm Pt layer was homogeneously grown using ALD. Finally, a 10-nm Pt layer was anisotropically etched away using inductively coupled plasma (ICP) to create 10-nm gaps between adjacent antennas (Fig. 5b). BAP functionalization of such antenna arrays with 10-nm gaps generated a blue-shift of ω_r of 4.77%, which was 64% larger than that achieved by nanorods with 30-nm gaps (Fig. 5c). Exposure of the BAP-activated metasurface to a 2 mM glucose solution introduced a resonance shift of 1.76%, which was also improved by 60% compared to the case of $g = 30$ nm. A low concentration of 200 pM glucose (36 pg/mL) was detectable on our 10-nm gap metasurface (Fig. 5d), which indicates that the detection limit of the sensor was further improved by approximately one order of magnitude compared to previously demonstrated 30-nm gap metasurfaces. To compare with typical surface plasmon resonance biosensors, we also estimated the detection limit in

weight per unit area (pg/mm^2), which is an important figure of merit for surface-based sensors. In the specific case of glucose sensing, the detection limit was approximately $1.3 \text{ pg}/\text{mm}^2$ when the gap was 30 nm, and further lowered to $0.043 \text{ pg}/\text{mm}^2$ when the gap was 10 nm, which was three times better than that for the state-of-the-art Vernier effect plasmonic resonance biosensors (see Supplementary Information section 5.3 for details on the calculation of the detection limit)⁵⁰. However, these values should not represent the theoretical limit of our method, considering that the optical conductivity change is mainly dependent on the number of charges transferred to graphene rather than the molecular weights of the adsorbed molecules. Therefore, this limit value can be even lower for the detection of stronger dopants with lower molecular weights. It can also be improved by engineering the antenna geometry. Figure 4e implies that the ultimate limit of detection for our approach is defined by quantum effects. Briefly, when the gap size is aggressively reduced to ~ 0.3 nm, direct electron tunneling between the neighboring antenna will cause a sharp

Fig. 6 Enhancement of glucose fingerprints. Optical absorption (ΔR) measured on metallic metasurface without graphene before and after drop-coating. **a** 1 nmol and **b** 100 pmole of glucose on a gold metasurface device with 30-nm antenna gaps

decrease in the field intensity within the gap; thus, the sensitivity will no longer increase^{51,52}.

Enhancement of molecular vibrational fingerprints

The metasurface can also act as a substrate for surface-enhanced IR absorption (SEIRA) spectroscopy. Here, we qualitatively examined the enhancement effect on the glucose fingerprints located within 1300–1500 cm^{-1} , which typically show weaker absorption than glucose OH groups do near 3000 cm^{-1} . We first drop-coated, in total, 1 nmol (2 mM, 0.5 μL) of glucose on a metallic metasurface device. In measurements performed on the bare substrate (SiO_2 -Pt cavity on Si) out of the metasurface region, we found that when the absorption of the OH band was 4.2% (see the Supplementary Information for details), other bands of glucose were hardly recognizable (Fig. 6a). In contrast, for measurements performed within the metasurface ($g = 30 \text{ nm}$), when the intensity of the OH band was close to that measured on bare substrate, the spectrum revealed prominent peaks at 1470 and 1320 cm^{-1} , which were attributed to glucose OCH deformation. The enhancement effect was also observed at 1740 cm^{-1} , which corresponded to the C=O bond of the polymethyl methacrylate (PMMA) residue from the fabrication. Further, when the glucose quantity was reduced to 100 pmol (200 μM , 0.5 μL), the OCH band at 1470 cm^{-1} was still recognizable (Fig. 6b) with a signal-to-noise ratio of 39, even if there was essentially no signal from the OH band as it fell out of the metasurface resonance band. These results demonstrated that the metasurface provided remarkable enhancement of the vibrational fingerprints, which originated from the enhanced light-matter interaction by the strongly confined electrical field at the nanogap and the edges of the metallic antenna. Such fingerprinting enhancement can help identify analytes in biochemical sensing, especially when the analyte quantity is very limited.

Discussion

In summary, the hybrid graphene-metallic metasurface enables ultrasensitive optical measurements of small molecules as well as an enhancement of their fingerprints. The optical conductivity-based sensing shows detection limits orders of magnitude lower than conventional optical sensing techniques based on local refractive index changes. It also allows for differentiation between low-molecular-weight analytes based on different types and/or concentrations of dopants introduced to graphene.

While this paper is focused on demonstrating the principle of this new sensing approach by performing measurements in air, the use of such a sensor is not limited to dry environments and can be further complemented. For example, it would be possible to make the device mid-IR transparent by replacing SiO_2 with CaF_2 and removing the Pt layer, which may allow for integration of the sensor with microfluidic channels for on-chip real-time sensing or monitoring of the binding process³³. To enable the application of the sensor in complex physiological fluids, it is important to develop a functionalization process of graphene to reject non-specific absorption, which is a challenge in general for biosensors employing nanomaterials. In addition, the long-term stability and consistency of sensor output needs to be further studied and improved, possibly by protection of the graphene surface and enhancement of graphene-substrate adhesion^{53,54}.

Materials and methods

Device fabrication

(a) Fabrication of the hybrid graphene-metallic metasurfaces. A 50-nm Pt layer was deposited on a heavily doped silicon substrate, followed by deposition of a 285-nm SiO_2 layer using plasma-enhanced chemical vapor deposition (PECVD). Gold nanorods 50 nm in thickness were then fabricated over a 400 μm by 400 μm area on the SiO_2 layer using electron beam lithography, electron

beam evaporation and lift-off processes. A monolayer graphene film grown on a copper foil via chemical vapor deposition (CVD) was then transferred onto the gold nanorod array using a PMMA-assisted wet transfer method. The PMMA layer was lastly removed by rinsing the device in acetone for over 3 h at room temperature. (b) Fabrication of graphene FETs. A CVD graphene film was transferred onto a thermally oxidized (285-nm SiO₂) silicon substrate, and metallic contacts (2-nm Ti and 45-nm Au) were patterned using electron beam lithography. A monolayer graphene film grown on a copper foil via chemical vapor deposition (CVD) was then transferred onto the gold nanorod array using a PMMA-assisted wet transfer method. The PMMA layer was lastly removed by rinsing the device in acetone for over 3 h at room temperature. (b) Fabrication of graphene FETs. A CVD graphene film was transferred onto a thermally oxidized (285-nm SiO) silicon substrate, and metallic contacts (2-nm Ti and 45-nm Au) were patterned using electron beam lithography, electron beam evaporation and lift-off processes. electron beam evaporation and lift-off processes.

Biochemical functionalization

(a) Immobilization of pyrene derivatives on graphene. 0.5 mM amino-pyrene (AP) or boronic acid-pyrene (BAP) solutions were prepared by dissolving the powders in anhydrous acetonitrile (Sigma-Aldrich). Metasurface devices with as-transferred graphene were immersed in a 2 mL AP or BAP solution for 12 h, and then rinsed with fresh acetonitrile, isopropanol and deionized water baths sequentially to remove molecules that were not immobilized on the graphene. The devices were then gently blow-dried using compressed nitrogen and left in vacuum desiccator for over 30 min at room temperature to completely remove solvents without impacting the properties of graphene. To immobilize boronic acid on gold antennas, devices with bare gold nanorods were immersed in 4-mercaptophenylboronic acid (Fisher Scientific) solution (1 mM in ethanol) for 12 h at room temperature, and then rinsed with fresh ethanol and deionized water baths sequentially. (b) Aluminum nanoparticles were deposited using thermal evaporation. The devices with graphene and the control devices without graphene were secured on a same glass slide for deposition under the same conditions (0.1 Å/s). (c) Glucose solutions were prepared by dissolving D-glucose (Sigma-Aldrich) in 1× phosphate buffered solutions (ThermoFisher Scientific) and diluted to desired concentrations. At each concentration, the device was exposed to glucose solution for 8 min to ensure that the glucose-boronic acid binding reached equilibrium. The devices were then dipped in fresh buffer and ultrapure water (ThermoFisher) and gently blow-dried using purified nitrogen.

Material characterization

The hybrid graphene-metallic metasurfaces and deposited Al nanoparticles were characterized using a scanning electron microscope (SEM, Zeiss Sigma VP). Characterization of the functionalized graphene was conducted using Raman spectroscopy (Renishaw) with an exciting laser at $\lambda = 532$ nm. Functionalization of gold nanorods with 4-PBA was characterized using X-ray photoelectron spectroscopy (XPS, Phi 5500). The thicknesses of the graphene before and after chemical treatment were measured using an atomic force microscope (AFM, Bruker Dimension FastScan).

Measurements setup

(a) Electrical measurements: Transport characteristics of graphene were measured using Keithley 2400 source-meters. (b) Optical measurements: Infrared reflectance spectra of metasurfaces were measured using a Fourier transform infrared (FTIR) microscope (Bruker Vertex 70 v) with light polarized along the antenna rods. A spectrum measured on a gold mirror was used as the reference for reflectance spectral measurements. Measurements at different glucose concentrations were carried out on the same metasurface from the same location with a window size as large as the metasurface area (400 μ m by 400 μ m). All spectral measurements were carried out in ambient atmosphere at room temperature.

Acknowledgements

The work was supported by the National Science Foundation (grants no. ECCS-1509760 and ECCS-1307948), a Defense Advanced Research Projects Agency Young Faculty Award (grant no. D15AP00111), and the Air Force Office of Scientific Research (grants no. FA9550-14-1-0389 and FA9550-16-1-0322). Research was carried out in part at the Center for Functional Nanomaterials, Brookhaven National Laboratory, which is supported by the US Department of Energy, Office of Basic Energy Sciences (contract no. DE-SC0012704).

Author details

¹Department of Mechanical Engineering, Columbia University, New York, NY 10027, USA. ²Department of Applied Physics and Applied Math, Columbia University, New York, NY 10027, USA. ³Department of Chemistry and Biochemistry, University of South Carolina, Columbia, SC 29208, USA. ⁴National Laboratory of Solid State Microstructures, College of Engineering and Applied Sciences, and Collaborative Innovation Center of Advanced Microstructures, Nanjing University, Nanjing 210093, China. ⁵Center for Functional Nanomaterials, Brookhaven National Laboratory, Upton, NY 11973, USA

Author contributions

Y.Z., Z.L., N.Y., and Q.L. conceived and designed the experiments. Y.Z. and Z.L. fabricated the devices with M.L. and A.S.'s assistance. C.D. and Y.H. prepared CVD graphene on Cu foil. Y.Z., Z.L., and Z.H. performed the measurements, characterization, and data analysis, with J.H., N.Y., and Q.L.'s supervision. Y.Z. and Z.L. developed the theoretical model and conducted numerical simulations with N.Y.'s supervision. Y.Z., Z.L., N.Y., and Q.L. wrote the manuscript with input from all co-authors. All authors discussed the results and commented on the manuscript.

Conflict of interest

The authors declare that they have no conflict of interest.

Supplementary information is available for this paper at <https://doi.org/10.1038/s41377-018-0066-1>.

Received: 5 April 2018 Revised: 4 August 2018 Accepted: 17 August 2018
 Accepted article preview online: 31 August 2018
 Published online: 26 September 2018

References

- Cooper, M. A. Optical biosensors in drug discovery. *Nat. Rev. Drug. Discov.* **1**, 515–528 (2002).
- Cetin, A. E. et al. Handheld high-throughput plasmonic biosensor using computational on-chip imaging. *Light Sci. Appl.* **3**, e122 (2014).
- Brolo, A. G. Plasmonics for future biosensors. *Nat. Photonics* **6**, 709–713 (2012).
- Etezadi, D. et al. Nanoplasmonic mid-infrared biosensor for in vitro protein secondary structure detection. *Light Sci. Appl.* **6**, e17029 (2017).
- Laing, S., Jamieson, L. E., Faulds, K. & Graham, D. Surface-enhanced Raman spectroscopy for in vivo biosensing. *Nat. Rev. Chem.* **1**, 0060 (2017).
- Cao, C. et al. Metamaterials-based label-free nanosensor for conformation and affinity biosensing. *ACS Nano* **7**, 7583–7591 (2013).
- Li, Y. L. et al. Graphene plasmon enhanced vibrational sensing of surface-adsorbed layers. *Nano. Lett.* **14**, 1573–1577 (2014).
- Wu, C. H. et al. Fano-resonant asymmetric metamaterials for ultrasensitive spectroscopy and identification of molecular monolayers. *Nat. Mater.* **11**, 69–75 (2012).
- Anker, J. N. et al. Biosensing with plasmonic nanosensors. *Nat. Mater.* **7**, 442–453 (2008).
- Leung, A., Shankar, P. M. & Mutharasan, R. A review of fiber-optic biosensors. *Sens. Actuat B Chem.* **125**, 688–703 (2007).
- Duan, X. X. et al. Quantification of the affinities and kinetics of protein interactions using silicon nanowire biosensors. *Nat. Nanotechnol.* **7**, 401–407 (2012).
- Kabashin, A. V. et al. Plasmonic nanorod metamaterials for biosensing. *Nat. Mater.* **8**, 867–871 (2009).
- Henry, A. I., Sharma, B., Cardinal, M. F., Kurouski, D. & Van Deyne, R. P. Surface-enhanced Raman spectroscopy biosensing: in vivo diagnostics and multimodal imaging. *Anal. Chem.* **88**, 6638–6647 (2016).
- Iverson, N. M. et al. In vivo biosensing via tissue-localizable near-infrared-fluorescent single-walled carbon nanotubes. *Nat. Nanotechnol.* **8**, 873–880 (2013).
- Xu, S. C. et al. Real-time reliable determination of binding kinetics of DNA hybridization using a multi-channel graphene biosensor. *Nat. Commun.* **8**, 14902 (2017).
- Lee, H. et al. A graphene-based electrochemical device with thermoresponsive microneedles for diabetes monitoring and therapy. *Nat. Nanotechnol.* **11**, 566–572 (2016).
- Gao, W. et al. Fully integrated wearable sensor arrays for multiplexed in situ perspiration analysis. *Nature* **529**, 509–514 (2016).
- Wang, C. et al. High-kappa solid-gate transistor configured graphene biosensor with fully integrated structure and enhanced sensitivity. *Adv. Funct. Mater.* **26**, 7668–7678 (2016).
- Fu, W. Y. et al. Graphene transistors are insensitive to pH changes in solution. *Nano. Lett.* **11**, 3597–3600 (2011).
- Gao, N. et al. Specific detection of biomolecules in physiological solutions using graphene transistor biosensors. *Proc. Natl. Acad. Sci. U. S. A.* **113**, 14633–14638 (2016).
- Wang, L. et al. One-dimensional electrical contact to a two-dimensional material. *Science* **342**, 614–617 (2013).
- Wang, F. et al. Gate-variable optical transitions in graphene. *Science* **320**, 206–209 (2008).
- Lopez-Sanchez, O., Lembke, D., Kayci, M., Radenovic, A. & Kis, A. Ultrasensitive photodetectors based on monolayer MoS₂. *Nat. Nanotechnol.* **8**, 497–501 (2013).
- Liu, M. et al. A graphene-based broadband optical modulator. *Nature* **474**, 64–67 (2011).
- Phare, C. T., Lee, Y. H. D., Cardenas, J. & Lipson, M. Graphene electro-optic modulator with 30 GHz bandwidth. *Nat. Photonics* **9**, 511–514 (2015).
- Yao, Y. et al. Broad electrical tuning of graphene-loaded plasmonic antennas. *Nano. Lett.* **13**, 1257–1264 (2013).
- Yan, H. G. et al. Damping pathways of mid-infrared plasmons in graphene nanostructures. *Nat. Photonics* **7**, 394–399 (2013).
- Dabidian, N. et al. Electrical switching of infrared light using graphene integration with plasmonic fano resonant metasurfaces. *ACS Photonics* **2**, 216–227 (2015).
- Farmer, D. B., Avouris, P., Li, Y. L., Heinz, T. F. & Han, S. J. Ultrasensitive plasmonic detection of molecules with graphene. *ACS Photonics* **3**, 553–557 (2016).
- Rodrigo, D. et al. Mid-infrared plasmonic biosensing with graphene. *Science* **349**, 165–168 (2015).
- Hu, H. et al. Far-field nanoscale infrared spectroscopy of vibrational fingerprints of molecules with graphene plasmons. *Nat. Commun.* **7**, 12334 (2016).
- Huck, C. et al. Surface-enhanced infrared spectroscopy using nanometer-sized gaps. *ACS Nano* **8**, 4908–4914 (2014).
- Limaj, O. et al. Infrared plasmonic biosensor for real-time and label-free monitoring of lipid membranes. *Nano. Lett.* **16**, 1502–1508 (2016).
- Schedin, F. et al. Detection of individual gas molecules adsorbed on graphene. *Nat. Mater.* **6**, 652–655 (2007).
- Dontschuk, N. et al. A graphene field-effect transistor as a molecule-specific probe of DNA nucleobases. *Nat. Commun.* **6**, 6563 (2015).
- Dong, X. C., Shi, Y. M., Huang, W., Chen, P. & Li, L. J. Electrical detection of DNA hybridization with single-base specificity using transistors based on CVD-grown graphene sheets. *Adv. Mater.* **22**, 1649–1653 (2010).
- Yu, R. W., Cox, J. D., Saavedra, J. R. M. & De Abajo, F. J. G. Analytical modeling of graphene plasmons. *ACS Photonics* **4**, 3106–3114 (2017).
- Zhu, A. Y., Yi, F., Reed, J. C., Zhu, H. & Cubukcu, E. Optoelectromechanical multimodal biosensor with graphene active region. *Nano. Lett.* **14**, 5641–5649 (2014).
- Yao, Y. et al. Electrically tunable metasurface perfect absorbers for ultrathin mid-infrared optical modulators. *Nano. Lett.* **14**, 6526–6532 (2014).
- Li, Z. Q. et al. Dirac charge dynamics in graphene by infrared spectroscopy. *Nat. Phys.* **4**, 532–535 (2008).
- Wu, X. et al. Selective sensing of saccharides using simple boronic acids and their aggregates. *Chem. Soc. Rev.* **42**, 8032–8048 (2013).
- Shoji, E. & Freund, M. S. Potentiometric saccharide detection based on the pK_a changes of poly (aniline boronic acid). *J. Am. Chem. Soc.* **124**, 12486–12493 (2002).
- Zhu, Y. B. et al. A graphene-based affinity nanosensor for detection of low-charge and low-molecular-weight molecules. *Nanoscale* **8**, 5815–5819 (2016).
- Zhang, Y. et al. Capacitive sensing of glucose in electrolytes using graphene quantum capacitance varactors. *ACS Appl. Mater. Inter.* **9**, 38863–38869 (2017).
- Moore, A. N. J. & Wayner, D. D. M. Redox switching of carbohydrate binding to ferrocene boronic acid. *Can. J. Chem.* **77**, 681–686 (1999).
- Mesch, M., Zhang, C. J., Braun, P. V. & Giessen, H. Functionalized hydrogel on plasmonic nanoantennas for noninvasive glucose sensing. *ACS Photonics* **2**, 475–480 (2015).
- Li, D. C. et al. Affinity based glucose measurement using fiber optic surface plasmon resonance sensor with surface modification by borate polymer. *Sens. Actuat B Chem.* **213**, 295–304 (2015).
- Yetisen, A. K. et al. Reusable, robust, and accurate laser-generated photonic nanosensor. *Nano. Lett.* **14**, 3587–3593 (2014).
- Stephenson-Brown, A. et al. Glucose selective surface plasmon resonance-based bis-boronic acid sensor. *Analyst* **138**, 7140–7145 (2013).
- Azuelos, P. et al. Theoretical investigation of Vernier effect based sensors with hybrid porous silicon-polymer optical waveguides. *J. Appl. Phys.* **121**, 144501 (2017).
- Savage, K. J. et al. Revealing the quantum regime in tunnelling plasmonics. *Nature* **491**, 574–577 (2012).
- Zhu, W. Q. & Crozier, K. B. Quantum mechanical limit to plasmonic enhancement as observed by surface-enhanced Raman scattering. *Nat. Commun.* **5**, 5528 (2014).
- Sarkar, D. et al. MoS₂ field-effect transistor for next-generation label-free biosensors. *ACS Nano* **8**, 3992–4003 (2014).
- Lee, G. H. et al. Highly stable, dual-gated MoS₂ transistors encapsulated by hexagonal boron nitride with gate-controllable contact, resistance, and threshold voltage. *ACS Nano* **9**, 7019–7026 (2015).