REVIEW OPEN

The newborn delivery room of tomorrow: emerging and future technologies

Natalie Batey¹, Caroline Henry^{1,2}, Shalabh Garg³, Michael Wagner⁴, Atul Malhotra⁵, Michel Valstar⁶, Thomas Smith⁶, Don Sharkey^{1,2}✉ and the European Society for Paediatric Research (ESPR) Neonatal Resuscitation Section Writing Group*

© The Author(s) 2022

Advances in neonatal care have resulted in improved outcomes for high-risk newborns with technologies playing a significant part although many were developed for the neonatal intensive care unit. The care provided in the delivery room (DR) during the first few minutes of life can impact short- and long-term neonatal outcomes. Increasingly, technologies have a critical role to play in the DR particularly with monitoring and information provision. However, the DR is a unique environment and has major challenges around the period of foetal to neonatal transition that need to be overcome when developing new technologies. This review focuses on current DR technologies as well as those just emerging and further over the horizon. We identify what key opinion leaders in DR care think of current technologies, what the important DR measures are to them, and which technologies might be useful in the future. We link these with key technologies including respiratory function monitors, electrical impedance tomography, videolaryngoscopy, augmented reality, video recording, eye tracking, artificial intelligence, and contactless monitoring. Encouraging funders and industry to address the unique technological challenges of newborn care in the DR will allow the continued improvement of outcomes of high-risk infants from the moment of birth.

Pediatric Research; <https://doi.org/10.1038/s41390-022-01988-y>

IMPACT:

- Technological advances for newborn delivery room care require consideration of the unique environment, the variable patient characteristics, and disease states, as well as human factor challenges.
- Neonatology as a speciality has embraced technology, allowing its rapid progression and improved outcomes for infants, although innovation in the delivery room often lags behind that in the intensive care unit.
- Investing in new and emerging technologies can support healthcare providers when optimising care and could improve training, safety, and neonatal outcomes.

INTRODUCTION

Delivery room (DR) management of the newborn focuses on ensuring the physiological process of oxygen delivery transitions from the placenta to the lungs. Most infants make this transition independently, aerating the lungs and establishing a functional residual capacity to allow for gas exchange.¹ At birth, 85% of term babies breathe within 10–30 s and require no support, 10% respond to stimulation or airway opening, and approximately 5% of infants will require positive pressure ventilation (PPV), 0.3% chest compressions, and 0.05% adrenaline.²

Advances in DR technology have enhanced management at birth, facilitating care and reducing factors known to affect neonatal outcomes. Neonatal care has embraced technology, leading to the rapid evolution of the speciality. Despite significant advances in monitoring on intensive care units in recent years, much of this technology has not yet translated into everyday DR care. Adoption barriers include the portability of equipment, evidence of benefit, and the training required for utilisation in

high-pressure environments. Furthermore, the DR is a unique setting occurring nowhere else in clinical medicine with factors challenging technology such as fluid-filled lungs, wet skin, 10-fold patient weight differences ranging from <500 g to >5 kg, and a transitioning cardiorespiratory system. It is important to consider the care requirements for the mother and her baby when in the same room as both may need intensive treatment. This requires close communication within teams and with the family. Where possible, technologies should facilitate, rather than impede, family integration and early bonding.

We focus this review on high-income countries where many technologies were designed for, recognising that the greatest burden of morbidity and mortality exists in less well-resourced settings where initiatives such as NEST360³ are driving forward technological innovations to improve care. This review discusses the evolution of newborn technologies in the DR with current recommendations, emerging, and potential future technologies.

¹Nottingham Neonatal Service, Nottingham University Hospitals NHS Trust, Nottingham, UK. ²Centre for Perinatal Research, School of Medicine, University of Nottingham, Nottingham, UK. ³Department of Neonatal Medicine, James Cook University Hospital, Middlesbrough, UK. ⁴Division of Neonatology, Pediatric Intensive Care and Neuropediatrics, Department of Pediatrics, Comprehensive Center for Pediatrics, Medical University of Vienna, Vienna, Austria. ⁵Monash Newborn, Monash Children's Hospital and Department of Paediatrics, Monash University, Melbourne, Australia. ⁶School of Computer Science, University of Nottingham, Nottingham, UK. *A list of authors and their affiliations appears at the end of the paper. ✉email: Don.sharkey@nottingham.ac.uk

Received: 14 September 2021 Revised: 10 January 2022 Accepted: 1 February 2022

Published online: 03 March 2022

Brief history of DR technologies

Paediatric resuscitation has been referenced throughout historical literature with a description of mouth to mouth as far back as 850 BC.⁴ Recognition of establishing respiration during newborn resuscitation, along with the physiological understanding, occurred in the mid-twentieth century. Prior to this, various methods such as intragastric oxygen, respiratory stimulants, hyperbaric oxygen, and rapid hypothermia were all trialled.⁵

Following advances in obstetrics, changes in birth location, and the first neonatal units,⁶ the technology to support newborn care evolved rapidly in the twentieth century. Neonatal physiology research in the 1950s and 1960s^{7,8} improved understanding of acute hypoxia during birth and the management required to alter the pathway of the physiological processes.⁶

Widespread support for positive pressure respiration immediately after birth occurred from the 1950s,⁹ although the method of delivery (face mask vs intubation) remained contentious for some time.¹⁰ The first newborn resuscitator by Vickers Medical in 1965 was introduced into practice.⁶ Transcutaneous oxygen saturation monitoring, developed in the 1970s,¹¹ was not in mainstream DR use until the past 10 years, especially for preterm infants.¹²

Over the past 20 years, technology use in the DR has evolved and impacted the management of newborns. Recommendations for the early use of suction followed by intubation and the use of rebreathing bags with 100% oxygen for resuscitation has transitioned to the use of mask ventilation commencing in the air for term infants and the titration of oxygen for preterm infants.¹³ Some technologies such as end-tidal CO₂ detectors have remained part of current practice,¹⁴ whereas others such as meconium aspirators are less advocated.² Guidelines have mostly been driven by monitoring strategies to support management, this has included the introduction of electrocardiography (ECG) and pulse oximetry monitoring to provide objective measurement in place of more error-prone subjective assessments such as colour or heart rate assessment with a stethoscope.^{13,15}

Current international recommendations

The International Liaison Committee on Resuscitation (ILCOR) publishes a consensus on newborn resuscitation science every 5 years. Current DR design and management strategies are based on these guidelines; this section highlights some of these practice points with a focus on technologies. Key changes to the ILCOR newborn resuscitation guidelines over the past 20 years when technologies started to be integrated into the guidance are outlined in Table 1.

Table 1. Key technologies detailed through the last five iterations of the ILCOR newborn resuscitation guidelines.

Year	Key elements	Technologies
2000	• Suction for meconium	• Stethoscope
	• 100% O ₂ for resuscitation	• Exhaled CO ₂
2005	• Less suction for meconium	• T-piece devices
	• Move to air for term resuscitation	• Plastic bags
2010	• Monitor heart rate and oxygen saturations	• Pulse oximeter
	• Consider CPAP	
2015	• Delayed cord clamping	• Pulse oximetry ± ECG
	• Monitor heart rate and oxygen saturations	• Humidified gases
2020	• Reducing invasive ventilation	• ECG for heart rate
	• Focus on monitoring again	• Pulse oximetry for oxygen saturations

As resuscitation guidance has evolved, there has been a focus on reducing neonatal morbidity and mortality, particularly in preterm infants, with aspects of DR care known to improve outcomes. Some of the most important areas related to temperature management, delayed cord clamping (DCC), and the use of monitoring to minimise invasive or harmful treatments, especially those relating to respiratory outcomes.

Thermoregulation

Temperature management is one of the most important aspects of DR care. ILCOR recommends warming adjuncts (warm blankets, plastic wrapping without drying, cap, thermal mattress) in addition to maintenance of environmental temperature of 23–25 °C for infants <32 weeks gestation, to avoid hypothermia (temperature <36 °C).¹⁵ This particularly applies to infants <28 weeks gestation since hypothermia is an independent risk factor affecting morbidity and mortality, with increased risk for each degree below 36.5 °C.¹⁶ The plastic bag/wrap for very preterm infants is perhaps one of the simplest yet cost-effective ways of reducing hypothermia and, when used in combination with other measures such as humidified gas, is associated with lower mortality and severe brain injury.¹⁷

Delayed cord clamping

With recent evidence supporting its benefits, DCC is recommended for at least 30 s for infants <34 weeks and 60 s for those ≥34 weeks when immediate resuscitation is not required.^{18,19} Several studies are underway looking at resuscitation with the cord intact. With recommendations for optimal cord management translating into practice,⁵ newer resuscitation platforms allowing stabilisation with the cord intact have been developed and undergone clinical evaluation.²⁰

Monitoring in the DR

Several studies have been conducted looking at the effective provision of resuscitative measures (bag and mask ventilation, provision of adequate tidal volume and inspiratory pressure, cardiac compression) as well as monitoring of various parameters to assess the response to resuscitation (chest wall rise, heart rate monitoring, leak assessment, oxygen delivery, and maintenance of saturation targets). Currently, ILCOR recommends ECG for heart rate monitoring²¹ and pulse oximetry for the management of oxygen titration.²² As our understanding of transition at birth increases, it is likely more of these technologies, adapted from the neonatal unit, will evolve further in the DR management of high-risk newborns.

Survey of opinions on DR technologies

As part of this review, we wanted to understand how clinicians and researchers view the current and future role of technologies in the DR care of high-risk infants. We therefore conducted an online survey via two main channels: (1) direct invitation to the European Society for Paediatric Research Neonatal Resuscitation Section members, and (2) via the neonatal theme of the National Institute for Health Research (NIHR) Children and Young People MedTech Cooperative.²³ Details of the survey and who completed it can be found in the supplement. Of the 60 respondents, 82% were from Europe and doctors at a senior level. When asked how advances in DR technologies in the past 5–10 years compared to other areas of clinical medicine, only 37% felt it had increased proportionately or significantly. Similarly, only 9% of respondents felt funding for newborn DR technologies had been good or excellent in their region, 65% said it was poorly funded.

The lack of advancement in DR technologies is perhaps a reflection of the level of funding invested in this niche market with relatively small patient numbers compared to other clinical specialties. Evidence for this lack of investment was demonstrated in paediatric high-risk (class III) devices undergoing FDA approvals between 2008

and 2011.²⁴ Of the 24 approved for use in children, only 3 (16%) had been studied on patients <18 years of age and so approvals for these high-risk devices were mostly based on adult trials only. This is a pattern previously recognised within paediatric drug development and has resulted in the development of paediatric medical device consortia or cooperatives such as those established in the US²⁵ and UK.²⁶ This has also resulted in calls to national funders and policymakers to actively encourage paediatric medical device innovation.²⁷ The use of pulse oximetry to measure heart rate in the DR is a good example of how we cannot simply adopt devices developed for adults and apply them in the unique environment of the DR on patient populations not previously validated with clinical studies.²⁸ Although absolute numbers of babies requiring advanced DR care are small, the need for intervention is not always predictable and therefore a risk at any birth.

There is a multitude of measurements that can be obtained in the DR using technologies. Some, such as temperature, are important when improving outcomes¹⁷ while the utility of others, such as cerebral oxygenation, are being explored in on-going DR studies.²⁹ Our survey found heart rate, oxygen saturations, and temperature to be the most important (Fig. 1). Near infrared spectroscopy (NIRS), for measuring cerebral oxygenation, was ranked the least important reflecting the evolving data on its utility in neonatal care. Respiratory function monitors (RFMs), discussed in more detail later, provide information on ventilation technique and overall was ranked as important by participants.

Emerging technologies

Although not currently part of newborn resuscitation recommendations, there are a number of exciting emerging technologies that could translate into clinical practice in the DR.

Respiratory function monitoring

Healthcare professionals (HCPs) provide PPV via a mask using a T-piece or self-inflating bag during newborn airway support.^{5,12} These devices can generate unknown and variable tidal volumes (V_T) depending on mask seal,^{30,31} airway obstruction,³² and positioning. A recent study showed $V_T > 6$ ml/kg was provided to extremely preterm infants 47.7% of the time.³³ Uncontrolled, large V_T are associated with activation of inflammatory cascades,^{34,35} predisposition to bronchopulmonary dysplasia, and higher rates of severe intraventricular haemorrhage (IVH).³⁶

Given the difficulty of assessing chest wall expansion, especially in preterm infants,^{30,37} and heart rate estimation errors,³⁸ an objective live feedback system of V_T could potentially improve outcomes.^{39,40} RFMs, such as the Florian and Monivent systems, can provide feedback on tidal volumes, pressure waveforms, and gas flow.^{41–43} RFMs can be used to identify when pressure should be titrated with changes in lung compliance,⁴² if there is

suboptimal mask seal⁴⁴ or obstruction,⁴⁰ although device performance can vary.⁴³ RFMs improve HCP performance during simulation^{39,44–46} but can also result in incorrect management with inexperienced users.⁴⁷

Schmölzer et al undertook a small clinical trial with RFMs visible or masked and found a positive impact on ventilation quality when the RFM was visible.⁴⁰ Conversely, Sarrato et al. found no difference in short-term surfactant use or intubation need when using RFM.⁴⁸ The MONITOR randomised control trial, comparing blind versus unblinded RFMs in extremely preterm infants at birth, found no differences in the main outcome measures including ventilation breaths within the V_T range of 4–8 ml/kg.⁴⁹ There were fewer IVHs in the unblinded group although this could be a chance finding and requires additional exploration. These conflicting results raise a number of questions about the usefulness of RFMs and support further evaluation in appropriate clinical trials to understand their usefulness.

Electrical impedance tomography (EIT)

Developed in the 1980s,⁵⁰ EIT is a functional imaging modality that is safe, non-invasive, and not irradiating. It has been used primarily as a research tool and can determine real-time lung aeration during spontaneous or assisted breaths. Electrodes are placed around the chest allowing alternating electrical currents to be channelled between pairs of electrodes and the remaining electrodes measure conductivity.^{51,52} Tomographic algorithms construct frames, up to rates of 50 Hz, of the lungs as air within the lungs has different conductivity to tissue.

EIT has been used to assess lung changes at birth and in other settings including: well infants,⁵³ post surfactant,^{54,55} optimal ventilation strategies,⁵⁶ sustained lung inflations,^{57,58} assessment for extubation,⁵⁹ body positioning,⁶⁰ endotracheal tube placement,^{61,62} effect of suctioning⁶³ and pneumothorax identification.^{64,65} EIT allows real-time observation of over-distension and inhomogeneous ventilation,⁵¹ potentially allowing personalised ventilation strategies.⁶⁶ As with many DR technologies, motion artefact and practicality may limit the usefulness of EIT outside of research settings. Studies using it at birth for preterm infants are also awaited.

Videolaryngoscopy

Securing the newborn airway in an emergency can often be a stressful procedure⁶⁷ but is an essential neonatal skill. Videolaryngoscopy can be a useful aid that facilitates a higher success rate of the first intubation attempt and shortening hypoxic periods.^{68,69} Videolaryngoscopy is possible with infants of all weights, including birth weights around 500 g.⁶⁷

O'Shea et al. demonstrated the main reasons for failed intubation attempts are oesophageal intubation and the failure to recognise anatomical landmarks, something videolaryngoscopy could address.⁷⁰ Videolaryngoscopy could support teaching and training, improve patient safety and minimise unsuccessful intubation attempts. The most important issues of teaching neonatal intubation are visibility, with learners unable to identify the anatomic structures, and the small airway making it is almost impossible for direct supervision. Videolaryngoscopy overcomes these, improving views and communication between trainee and supervisor.^{71,72} Application of videolaryngoscopy in the DR setting does require robust evaluation to establish if it also translates to fewer intubation attempts, and so better outcomes, and its utility in the management of a difficult airway. As with many technologies, the design of the device is crucial as videolaryngoscopy blades differ from conventional blades⁷³ and require a different intubation technique.

Augmented reality (AR)

AR is a technology that displays holograms in the current physical environment. This technology can enhance training experiences

Fig. 1 Survey respondents' ($n = 60$) ranking of the importance of key monitoring measures within the delivery room. SpO₂ oxygen saturations, CO₂ carbon dioxide, O₂ cerebral oxygenation.

as well potentially reduce errors, improving quality of care and patient outcomes, especially in acute settings.⁷⁴ AR might be helpful for training and clinical situations in the DR, as applied in other settings.^{75–77} Dias et al developed and studied the use of AR-assisted videolaryngoscopy for neonatal intubation. They found, in neonatal nurses with minimal intubation experience, that both AR and standard videolaryngoscopy more than doubled intubation success rates, avoided oesophageal intubations, and significantly reduced time to intubate compared with conventional laryngoscopy.⁷⁸ AR could also be used for remote skill training or even real-time support, where a remote specialist can see the first-person view of the provider wearing the AR headset and provides visual instructions remotely.⁷⁹

Clinical decision support

Clinical decision support tools utilise computer-controlled software, allowing access to a broad clinical knowledge database. This can be integrated with patient-specific information and suggest recommendations or assessments specific to the patient, supporting complex decision-making. During simulated newborn resuscitations, a decision support tool that provided visual and auditory prompts reduced algorithm deviations including those related to cardiopulmonary resuscitation.⁸⁰ Simple support tools can also be incorporated into tablet or smartphone apps, for example, to improve the accuracy of heart rate assessment.⁸¹

Video recording

Recording DR resuscitations with video cameras has been performed for a number of years in some centres.^{82–85} Parents find reviewing resuscitations of their babies as acceptable and valuable.⁸⁵ Recording events can help support training, research, and audit,⁸⁶ this could be particularly helpful during complex resuscitations where the accuracy of documentation is often inadequate⁸⁷ and deviations from algorithms are frequent.^{88,89} Retrospective use of video recording in the DR has been demonstrated to be useful. However, the live viewing of video and its utility in supporting care is one area that requires more research to understand its potential with remote and clinical decision support or the guidance of artificial intelligence (AI) linking other measures with team actions.

Telemedicine/remote consulting

Telehealth facilitates access to specialised expertise, education and collaboration, and can also reduce transfers and costs. In the DR, telemedicine has the potential to support high-risk deliveries in non-tertiary centres where in utero transfer is not possible. Retrospective studies have found a higher quality of high-risk newborn resuscitations in community hospitals when using telemedicine⁹⁰ and can reduce inter-hospital transfers.⁹¹ A number of telemedicine neonatal resuscitation simulation studies have been undertaken focussing on team dynamics although not all have shown benefits.^{92,93} Telemedicine in the DR offers an exciting opportunity to enhance training and improve the care of high-risk infants born in non-tertiary centres. Combining telemedicine with AR or eye-tracking warrants exploration to identify additional benefits.

Eye-tracking

The use of eye-tracking glasses enables the analysis of a HCPs visual attention, providing valuable insights into their situational awareness and decision making.⁹⁴ Studies have shown HCPs focus most of their cumulative visual attention time on the infant, followed by monitors, clinical staff, and other physical objects.^{95,96} However, visual attention patterns are different when other technologies or interventions are included (Fig. 2).⁹⁷ Eye tracking studies^{98,99} during simulated and real resuscitations have shown HCPs spend more time looking at the RFM (29 vs 1%), less time looking at the infant (29 vs 46%). Eye-tracking analysis in the DR

Fig. 2 Example heat map after analysis of providers' gaze behaviour in a simulated resuscitation scenario. Visual attention patterns can be analysed allowing different or new technologies to be evaluated.

enables researchers and educators to identify areas of improvement, cognitive (over)load, and the impact of new technologies or monitors. It is also possible to link eye-tracking data with physiological parameters such as heart rate and saturations, facilitating a thorough understanding of HCPs actions and reactions.

AI in the DR

Algorithms for newborn resuscitation are based on international guidelines and allow a structured approach to stabilisation in the first few minutes of life. Retrospective studies after the introduction of standardised resuscitation programmes have found improvements in survival and neurological impairment.¹⁰⁰ However, video analysis of real-life resuscitations highlights that teams are rarely able to undertake the desired assessments or interventions in-line with the algorithm timeline.^{88,89} Using real-life video recordings of resuscitations for training can improve the quality of neonatal resuscitation.⁸⁸

The analysis of video recordings of neonatal resuscitation for research, audit, and medical records can be resource heavy in terms of the time required for event categorisation and documentation. Computer vision and AI approaches could significantly speed this process up and even offer real-time feedback of resuscitation progress to support clinical decision-making and avoid delays with treatments. This could then open up the exciting potential to link these data with long-term outcomes creating a digital database for more efficient, big data analyses.

To explore the potential of AI to analyse complex newborn resuscitation data, Smith et al. used deep-learned superpixels to predict semantic segmentation of videos.¹⁰¹ Superpixels are clusters of image pixels that are grouped together by their

Fig. 3 Process of training a computer system to recognise resuscitation actions using computer vision and artificial intelligence (AI). (1) Image training data set manually labelled, categorising actions for the algorithm. (2) New data set (video) is inputted into the AI algorithm for analysis. (3) The algorithm pixelates items to label and track them. (4) Output allows quantification of important actions, their timing, duration, and frequency. Running in real time has the potential to support stabilisation/resuscitation algorithms to avoid deviations or errors especially in settings lacking experienced support.

similarity and this approach allows training of neural networks on a small data set, Smith utilises 23 segmentation classes from just 50 video frames. They trained recurrent neural networks to find the patterns in the video data that align with newborn resuscitation actions (see Fig. 3). The process is initially time consuming as the algorithm needs to be told what each image represents, in terms of action, before learning about the super-pixel patterns associated with it. The complexity of newborn resuscitation with 20–30 different actions can be hard to train a system and then for it to accurately predict what is going on. As we learn more about the power of AI in neonatal practice, the opportunity to utilise it in the DR is now a near reality and something that could one day offer real-time support and advice in the care we deliver. For example, AI could be linked with videolaryngoscopy to support real-time intubation when more experienced support was not immediately available.

Contactless monitoring

A number of groups have looked at contactless monitoring including using video¹⁰² and radar¹⁰³ in the neonatal unit with mixed results. There are practical issues to translating this into the DR, with the challenging environment and rapidly transitioning newborn, but this perhaps offers an exciting area for future exploration.

Role of human factors in new DR technologies

The analogy of the Formula One pitstop is often used during resuscitation. This team approach is critical in the DR of the high-risk newborn and requires a full appreciation for the role of human factors in achieving the most successful outcome. Understanding human factors, the interaction between humans and elements within their environment, can help optimise performance and reduce errors.¹⁰⁴ In the DR, human factors can be optimised⁹⁹ by decreasing cognitive and technical workload and establishing decision support tools. Teamwork during resuscitation has key success factors, including coordination of tasks, a synergistic working process and constructive conflict management at debriefs.¹⁰⁵ Approaching new technology design and adoption in the DR requires a detailed human factors approach, to avoid cognitive overload and algorithm deviations or errors, although they are rarely considered.¹⁰⁶ An important example by Pickup et al resulted in the identification of previously unrecognised user needs when developing a newborn heart rate monitor.¹⁰⁷ Innovators within the DR space need to incorporate human factors research in their developmental pathway to ensure the device performs its function without introducing new, unforeseen risks and distractions.

Future technologies

We have made great improvements with outcomes of high-risk newborns, in part because of our understanding of transition and

DR stabilisation. Newer practices of DCC and resuscitation/stabilisation with an intact placental circulation introduce new challenges for technologies. The interface between the obstetric and neonatal team is a new setting for technologies and requires careful design and application to avoid harm to the mother, for example through infection during caesarean section, and the infant by avoiding hypothermia while still be able to monitor and stabilise. Technologies play a crucial role in providing the attending team with critical information to manage the care of the newborn.

In our survey, HCPs ranked video analysis and lung aeration as the most important emerging technologies for the next 5–10 years, followed by wireless or contactless vital sign devices (Fig. 4), perhaps reflecting the increasing need to monitor infants during DCC or DR cuddles to avoid trailing wires and the complexity of the stabilisation platform.

To deliver optimal DR care, we need to understand both the clinical status of the newborn and their response to the management we instigate. In the past 20 years, this has become more evident with technology providing us with a wide variety of different measures to guide care. Reducing the complexity of the DR with a whole team circling around a 500 g preterm infant undergoing DCC is going to be important going forward. HCPs ranked wireless monitoring solutions as potentially an important step towards making this happen. There have been some exciting developments with wireless vital signs monitoring in the neonatal unit which offer real potential.¹⁰⁸ These must be suitably adapted to cope with the different demands of the DR with a transitioning infant undergoing DCC. Wireless monitoring simultaneously measuring key information prioritised by clinicians, i.e. heart rate, oxygen saturations, and temperature, has the potential to reduce sensors, device screens, and subsequent cognitive overload for staff. Wireless monitoring also offers the potential to improve family integrated care with kangaroo care an important component. Both parents and staff see device wires as a barrier to kangaroo care¹⁰⁹ and this is likely to be the case when facilitating DR cuddles. Contactless monitoring was also highly ranked as a future DR technology.

Newer systems utilising software algorithms, such as autonomous oxygen titration, or AI were felt to be important but perhaps are further downstream as more data become available from studies underway.

Protecting the brain of the foetus and newborn are at the core of obstetric and neonatal care. In neonatology, this goal has resulted in the evolution of the 'Neuro-NICU' where monitoring of the brain is a critical component.¹¹⁰ NIRS is one of the main modalities currently undergoing intense investigation but even in the Neuro-NICU its role remains unclear but has significant promise. It is therefore not surprising that HCPs did not rank this as an important technology in the DR at this stage. There remain

Fig. 4 Survey ranking by respondents ($n = 60$) for new technology priorities within the delivery room of the future. AI artificial intelligence, NIRS near infrared spectroscopy.

several limitations with neuromonitoring technologies, especially in the DR, with different devices providing different information and uncertainty about translating their use into meaningful outcomes.¹¹¹ Some of these are being addressed, for example targeting cerebral oxygenation in the DR,²⁹ but even here we need to consider the impact of overcomplicating team dynamics with multiple sensors, monitors, and subsequent actions, something that is already being studied.¹¹²

Finally, our survey found extra-uterine systems¹¹³ were ranked as the least important based on advances suggested, this again may reflect the truly 'futuristic' approach to preterm newborn care aiming to mimic the uterine environment and minimise the adverse effects of extreme prematurity. Better monitoring of the breathing and predicting the need for treatments could help reduce some of these adverse effects and technologies may have an important role.^{114,115}

CONCLUSIONS

When introducing any new piece of equipment into a time-critical, acutely stressful DR resuscitation scenario, the benefits must be clear. Increasing monitoring modalities in the DR can result in information overload and affect performance. However, emerging technologies show significant promise for safety and the improvement of outcomes following neonatal stabilisation or resuscitation. The addition of technology to the DR will not only impact an individual infant's care but also enhance the training of HCPs and allow collaborative working across centres. We must now intensify our efforts to deliver new innovative technologies, as other specialties in medicine have done, to improve newborn DR care during the first golden minutes of life, aiming to minimise morbidities that can stay with infants for the rest of their lives.

REFERENCES

- te Pas, A. B., Davis, P. G., Hooper, S. B. & Morley, C. J. From liquid to air: breathing after birth. *J. Pediatr.* **152**, 607–611 (2008).
- Resuscitation Council UK. *Newborn Life Support* 5th edn (Resuscitation Council UK, 2021).
- Home. NEST360. <https://nest360.org/> (2021).
- Bible Gateway passage: 2 Kings 4 - Common English Bible. Bible Gateway. <https://www.biblegateway.com/passage/?search=2%20Kings%204&version=CEB> (2021).
- Resuscitation Council (UK). *Newborn Life Support* 4th edn (Resuscitation Council (UK), 2016).
- Wellcome Collection. Origins of neonatal intensive care in the UK: a Witness Seminar held at the Wellcome Institute for the History of Medicine, London, on 27 April 1999 /Witness Seminar transcript edited by Christie, D. A. and Tansey, E. M.; introduction by Professor Peter Dunn. <https://wellcomecollection.org/works/ug67cgby> (2021).
- Dawes, G. Birth Asphyxia, Resuscitation and Brain Damage. in *Foetal and Neonatal Physiology* 151–159 (Year Book Publisher, 1968).

- Cross, K. W. Resuscitation of the asphyxiated infant. *Br. Med. Bull.* **22**, 73–78 (1966).
- Safar, P., Escarraga, L. A. & Elam, J. O. A comparison of the mouth-to-mouth and mouth-to-airway methods of artificial respiration with the chest-pressure arm-lift methods. *N. Engl. J. Med.* **258**, 671–677 (1958).
- McAdams, R. Learning to breathe: the history of newborn resuscitation, 1929 to 1970. <https://eleanor.lib.gla.ac.uk/record=b2668570> (2021).
- Eberhard, P., Mindt, W., Jann, F. & Hammacher, K. Continuous pO₂ monitoring in the neonate by skin electrodes. *Med. Biol. Eng.* **13**, 436–442 (1975).
- Wyllie, J. et al. Part 7: Neonatal resuscitation: 2015 International Consensus on Cardiopulmonary Resuscitation and Emergency Cardiovascular Care Science with Treatment Recommendations. *Resuscitation* **95**, e169–e201 (2015).
- Perkins, G. D. et al. The International Liaison Committee on Resuscitation—review of the last 25 years and vision for the future. *Resuscitation* **121**, 104–116 (2017).
- American Heart Association. Part 11: Neonatal resuscitation. *Circulation* **102**, i343–i357 (2000).
- Saugstad, O. D., Robertson, N. J. & Vento, M. A critical review of the 2020 International Liaison Committee on Resuscitation treatment recommendations for resuscitating the newly born infant. *Acta Paediatr.* **110**, 1107–1112 (2021).
- Laptook, A. R., Salhab, W. & Bhaskar, B. Admission temperature of low birth weight infants: predictors and associated morbidities. *Pediatrics* **119**, e643–e649 (2007).
- Abiramalatha, T. et al. Delivery room interventions for hypothermia in preterm neonates: a systematic review and network meta-analysis. *JAMA Pediatr.* **175**, e210775 (2021).
- Gomersall, J. et al. Umbilical cord management at term and late preterm birth: a meta-analysis. *Pediatrics* **147**, e2020015404 (2021).
- Seidler, A. L. et al. Umbilical cord management for newborns <34 weeks' gestation: a meta-analysis. *Pediatrics* **147**, e20200576 (2021).
- Katheria, A., Lee, H. C., Knol, R., Irvine, L. & Thomas, S. A review of different resuscitation platforms during delayed cord clamping. *J. Perinatol.* **41**, 1540–1548 (2021).
- Wyckoff, M. H. et al. Neonatal Life Support: 2020 International Consensus on Cardiopulmonary Resuscitation and Emergency Cardiovascular Care Science With Treatment Recommendations. *Circulation* **142**, S185–S221 (2020).
- Soar, J. et al. 2019 International Consensus on Cardiopulmonary Resuscitation and Emergency Cardiovascular Care Science With Treatment Recommendations: Summary From the Basic Life Support; Advanced Life Support; Pediatric Life Support; Neonatal Life Support; Education, Implementation, and Teams; and First Aid Task Forces. *Circulation* **140**, e826–e880 (2019).
- Neonatal Technologies – NIHR CYP MedTech. <https://cypmedtech.nihr.ac.uk/neonatal-technologies/> (2021).
- Hwang, T. J., Kesselheim, A. S. & Bourgeois, F. T. Postmarketing trials and pediatric device approvals. *Pediatrics* **133**, e1197–e1202 (2014).
- Ulrich, L. C., Joseph, F. D., Lewis, D. Y. & Koenig, R. L. FDA's Pediatric Device Consortia: national program fosters pediatric medical device development. *Pediatrics* **131**, 981–985 (2013).
- NIHR CYP MedTech – NIHR Children and Young People MedTech Co-operative. <https://cypmedtech.nihr.ac.uk/> (2021).
- American Academy of Pediatrics. Off-label use of medical devices in children. *Pediatrics* **139**, e20163439 (2017).
- van Vonderen, J. J. et al. Pulse oximetry measures a lower heart rate at birth compared with electrocardiography. *J. Pediatr.* **166**, 49–53 (2015).
- Pichler, G. et al. Cerebral regional tissue Oxygen Saturation to Guide Oxygen Delivery in preterm neonates during immediate transition after birth (COSGOD III): an investigator-initiated, randomized, multi-center, multi-national, clinical trial on additional cerebral tissue oxygen saturation monitoring combined with defined treatment guidelines versus standard monitoring and treatment as usual in premature infants during immediate transition: study protocol for a randomized controlled trial. *Trials* **20**, 178 (2019).
- Schmölzer, G. M. et al. Assessment of tidal volume and gas leak during mask ventilation of preterm infants in the delivery room. *Arch. Dis. Child. Fetal Neonatal Ed.* **95**, F393–F397 (2010).
- Schilleman, K. et al. Leak and obstruction with mask ventilation during simulated neonatal resuscitation. *Arch. Dis. Child. Fetal Neonatal Ed.* **95**, F398–F402 (2010).
- Finer, N. N., Rich, W., Wang, C. & Leone, T. Airway obstruction during mask ventilation of very low birth weight infants during neonatal resuscitation. *Pediatrics* **123**, 865–869 (2009).
- Vaidya, R., Visintainer, P. & Singh, R. Tidal volume measurements in the delivery room in preterm infants requiring positive pressure ventilation via endotracheal tube-feasibility study. *J. Perinatol.* **41**, 1930–1935 (2021).

34. Polglase, G. R. et al. Initiation of resuscitation with high tidal volumes causes cerebral hemodynamic disturbance, brain inflammation and injury in preterm lambs. *PLoS ONE* **7**, e39535 (2012).
35. Hillman, N. H. et al. Brief, large tidal volume ventilation initiates lung injury and a systemic response in fetal sheep. *Am. J. Respir. Crit. Care Med.* **176**, 575–581 (2007).
36. Mian, Q. et al. Impact of delivered tidal volume on the occurrence of intra-ventricular haemorrhage in preterm infants during positive pressure ventilation in the delivery room. *Arch. Dis. Child. Fetal Neonatal Ed.* **104**, F57–F62 (2019).
37. Poulton, D. A., Schmölzer, G. M., Morley, C. J. & Davis, P. G. Assessment of chest rise during mask ventilation of preterm infants in the delivery room. *Resuscitation* **82**, 175–179 (2011).
38. Phillippos, E. et al. Heart rate assessment immediately after birth. *Neonatology* **109**, 130–138 (2016).
39. Bowman, T. A., Paget-Brown, A., Carroll, J., Gurka, M. J. & Kattwinkel, J. Sensing and responding to compliance changes during manual ventilation using a lung model: can we teach healthcare providers to improve? *J. Pediatr.* **160**, 372–376 (2012).
40. Schmölzer, G. M. et al. Respiratory function monitor guidance of mask ventilation in the delivery room: a feasibility study. *J. Pediatr.* **160**, 377–381 (2012).
41. Schmölzer, G. M. et al. Respiratory monitoring of neonatal resuscitation. *Arch. Dis. Child. Fetal Neonatal Ed.* **95**, F295–F303 (2010).
42. Foglia, E. E. & Te Pas, A. B. Effective ventilation: the most critical intervention for successful delivery room resuscitation. *Semin. Fetal Neonatal Med.* **23**, 340–346 (2018).
43. Verbeek, C. et al. Accuracy of currently available neonatal respiratory function monitors for neonatal resuscitation. *Eur. J. Pediatr.* **175**, 1065–1070 (2016).
44. Wood, F. E., Morley, C. J., Dawson, J. A. & Davis, P. G. A respiratory function monitor improves mask ventilation. *Arch. Dis. Child. Fetal Neonatal Ed.* **93**, F380–F381 (2008).
45. O'Curraín, E., Thio, M., Dawson, J. A., Donath, S. M. & Davis, P. G. Respiratory monitors to teach newborn facemask ventilation: a randomised trial. *Arch. Dis. Child. Fetal Neonatal Ed.* **104**, F582–F586 (2019).
46. Schmölzer, G. M., Morley, C. J. & Kamlin, O. C. O. F. Enhanced monitoring during neonatal resuscitation. *Semin. Perinatol.* **43**, 151177 (2019).
47. Milner, A. et al. Evaluation of respiratory function monitoring at the resuscitation of prematurely born infants. *Eur. J. Pediatr.* **174**, 205–208 (2015).
48. Zeballos Sarrato, G. et al. New strategies of pulmonary protection of preterm infants in the delivery room with the respiratory function monitoring. *Am. J. Perinatol.* **36**, 1368–1376 (2019).
49. Foglia, E. A multi-center randomized controlled trial of respiratory function monitoring during stabilization of preterm infants at birth. *clinicaltrials.gov*; 2021. <https://clinicaltrials.gov/ct2/show/NCT03256578> (2021).
50. Brown, B. H., Barber, D. C. & Seagar, A. D. Applied potential tomography: possible clinical applications. *Clin. Phys. Physiol. Meas.* **6**, 109–121 (1985).
51. Frerichs, I. et al. Chest electrical impedance tomography examination, data analysis, terminology, clinical use and recommendations: consensus statement of the TTranslational EIT developmeNt stuDY group. *Thorax* **72**, 83–93 (2017).
52. Masner, A., Blasina, F. & Simini, F. Electrical impedance tomography for neonatal ventilation assessment: a narrative review. *J. Phys. Conf. Ser.* **1272**, 012008 (2019).
53. Tingay, D. G. et al. Imaging the respiratory transition at birth: unraveling the complexities of the first breaths of life. *Am. J. Respir. Crit. Care Med.* **204**, 82–91 (2021).
54. Miedema, M., de Jongh, F. H., Frerichs, I., van Veenendaal, M. B. & van Kaam, A. H. Changes in lung volume and ventilation during surfactant treatment in ventilated preterm infants. *Am. J. Respir. Crit. Care Med.* **184**, 100–105 (2011).
55. van der Burg, P. S., de Jongh, F. H., Miedema, M., Frerichs, I. & van Kaam, A. H. Effect of minimally invasive surfactant therapy on lung volume and ventilation in preterm infants. *J. Pediatr.* **170**, 67–72 (2016).
56. Gómez-Laberge, C. & Wolf, G. K. Toward individualized treatment of acute respiratory distress syndrome: response-guided mechanical ventilation. *Pediatr. Crit. Care Med.* **12**, 690–692 (2011).
57. Tingay, D. G. et al. Gradual aeration at birth is more lung protective than a sustained inflation in preterm lambs. *Am. J. Respir. Crit. Care Med.* **200**, 608–616 (2019).
58. Tingay, D. G. et al. An individualized approach to sustained inflation duration at birth improves outcomes in newborn preterm lambs. *Am. J. Physiol. Lung Cell. Mol. Physiol.* **309**, L1138–L1149 (2015).
59. Rossi, F., de, S., Yagui, A. C. Z., Haddad, L. B., Deutsch, A. D. & Rebello, C. M. Electrical impedance tomography to evaluate air distribution prior to extubation in very-low-birth-weight infants: a feasibility study. *Clinics* **68**, 345–350 (2013).
60. Hough, J. L., Johnston, L., Brauer, S., Woodgate, P. & Schibler, A. Effect of body position on ventilation distribution in ventilated preterm infants. *Pediatr. Crit. Care Med.* **14**, 171–177 (2013).
61. Schmölzer, G. M., Bhatia, R., Davis, P. G. & Tingay, D. G. A comparison of different bedside techniques to determine endotracheal tube position in a neonatal piglet model. *Pediatr. Pulmonol.* **48**, 138–145 (2013).
62. Steinmann, D., Engehausen, M., Stiller, B. & Guttman, J. Electrical impedance tomography for verification of correct endotracheal tube placement in paediatric patients: a feasibility study. *Acta Anaesthesiol. Scand.* **57**, 881–887 (2013).
63. van Veenendaal, M. B. et al. Effect of closed endotracheal suction in high-frequency ventilated premature infants measured with electrical impedance tomography. *Intensive Care Med.* **35**, 2130–2134 (2009).
64. Bhatia, R., Schmölzer, G. M., Davis, P. G. & Tingay, D. G. Electrical impedance tomography can rapidly detect small pneumothoraces in surfactant-depleted piglets. *Intensive Care Med.* **38**, 308–315 (2012).
65. Rahtu, M. et al. Early recognition of pneumothorax in neonatal respiratory distress syndrome with electrical impedance tomography. *Am. J. Respir. Crit. Care Med.* **200**, 1060–1061 (2019).
66. Davies, P. & Silvestre, C. Electrical impedance tomography in clinical use: Unnecessary technology or a unique angle in respiratory monitoring? *Pediatr. Pulmonol.* **55**, 845–846 (2020).
67. Vanderhal, A. L., Berci, G., Simmons, C. F. & Hagiike, M. A videolaryngoscopy technique for the intubation of the newborn: preliminary report. *Pediatrics* **124**, e339–e346 (2009).
68. Lingappan, K., Arnold, J. L., Fernandes, C. J. & Pammi, M. Videolaryngoscopy versus direct laryngoscopy for tracheal intubation in neonates. *Cochrane Database Syst. Rev.* **6**, CD009975 (2018).
69. Moussa, A., Luangxay, Y., Tremblay, S., Lavoie, J. & Lachance, C. Videolaryngoscope for teaching neonatal endotracheal intubation: a randomized controlled trial. *Pediatrics* **137**, e20152156 (2016).
70. O'Shea, J. E., Loganathan, P., Thio, M., Kamlin, C. O. F. & Davis, P. G. Analysis of unsuccessful intubations in neonates using videolaryngoscopy recordings. *Arch. Dis. Child. Fetal Neonatal Ed.* **103**, F408–F412 (2018).
71. O'Shea, J. E. et al. Videolaryngoscopy to teach neonatal intubation: a randomized trial. *Pediatrics* **136**, 912–919 (2015).
72. O'Shea, J. E., Kirolos, S., Thio, M., Kamlin, C. O. F. & Davis, P. G. Neonatal videolaryngoscopy as a teaching aid: the trainees' perspective. *Arch. Dis. Child. Fetal Neonatal Ed.* **106**, 168–171 (2021).
73. Kirolos, S. & O'Shea, J. E. Comparison of conventional and videolaryngoscopy blades in neonates. *Arch. Dis. Child. Fetal Neonatal Ed.* **105**, 94–97 (2020).
74. Munzer, B. W., Khan, M. M., Shipman, B. & Mahajan, P. Augmented reality in emergency medicine: a scoping review. *J. Med. Internet Res.* **21**, e12368 (2019).
75. Yoo, J. S., Patel, D. S., Hrynewycz, N. M., Brundage, T. S. & Singh, K. The utility of virtual reality and augmented reality in spine surgery. *Ann. Transl. Med.* **7**, S171 (2019).
76. Kim, Y., Kim, H. & Kim, Y. O. Virtual reality and augmented reality in plastic surgery: a review. *Arch. Plast. Surg.* **44**, 179–187 (2017).
77. Verhey, J. T., Haglin, J. M., Verhey, E. M. & Hartigan, D. E. Virtual, augmented, and mixed reality applications in orthopedic surgery. *Int. J. Med. Robot.* **16**, e2067 (2020).
78. Dias, P. L., Greenberg, R. G., Goldberg, R. N., Fisher, K. & Tanaka, D. T. Augmented reality-assisted video laryngoscopy and simulated neonatal intubations: a pilot study. *Pediatrics* **147**, e2020005009 (2021).
79. Wang, S. et al. Augmented reality as a telemedicine platform for remote procedural training. *Sensors* **17**, E2294 (2017).
80. Fuerch, J. H., Yamada, N. K., Coelho, P. R., Lee, H. C. & Halamek, L. P. Impact of a novel decision support tool on adherence to Neonatal Resuscitation Program algorithm. *Resuscitation* **88**, 52–56 (2015).
81. Henry, C. et al. Improving newborn heart rate assessment using a simple visual timer. *BMJ Paediatr. Open* **4**, e000638 (2020).
82. Layouni, I. et al. [Video recording of newborn resuscitation in the delivery room: technique and advantages]. *Arch. Pediatr. Organe* **18**, S72–S78 (2011).
83. Carbine, D. N., Finer, N. N., Knodel, E. & Rich, W. Video recording as a means of evaluating neonatal resuscitation performance. *Pediatrics* **106**, 654–658 (2000).
84. Leone, T. A. Using video to assess and improve patient safety during simulated and actual neonatal resuscitation. *Semin. Perinatol.* **43**, 151179 (2019).
85. den Boer, M. C. et al. Reviewing recordings of neonatal resuscitation with parents. *Arch. Dis. Child. Fetal Neonatal Ed.* **106**, 346–351 (2021).
86. Wagner, M. et al. Video-based reflection on neonatal interventions during COVID-19 using eye-tracking glasses: an observational study. *Arch. Dis. Child. Fetal Neonatal Ed.* **0**, F1–F5 (2021).
87. Avila-Alvarez, A., Davis, P. G., Kamlin, C. O. F. & Thio, M. Documentation during neonatal resuscitation: a systematic review. *Arch. Dis. Child. Fetal Neonatal Ed.* **106**, 376–380 (2021).
88. Skåre, C. et al. Video performance-debriefings and ventilation-refreshers improve quality of neonatal resuscitation. *Resuscitation* **132**, 140–146 (2018).

89. Schilleman, K. et al. Auditing resuscitation of preterm infants at birth by recording video and physiological parameters. *Resuscitation* **83**, 1135–1139 (2012).
90. Fang, J. L. et al. The impact of telemedicine on the quality of newborn resuscitation: a retrospective study. *Resuscitation* **125**, 48–55 (2018).
91. Maddox, L. J. et al. Implementation and outcomes of a telehealth neonatology program in a single healthcare system. *Front. Pediatr.* **9**, 648536 (2021).
92. Gross, I. T. et al. Comparison of two telemedicine delivery modes for neonatal resuscitation support: a simulation-based randomized trial. *Neonatology* **117**, 159–166 (2020).
93. Butler, L. et al. The impact of telemedicine on teamwork and workload in pediatric resuscitation: a simulation-based, randomized controlled study. *Telemed. J. E-Health* **25**, 205–212 (2019).
94. Law, B. H. Y. et al. Analysis of neonatal resuscitation using eye tracking: a pilot study. *Arch. Dis. Child. Fetal Neonatal Ed.* **103**, F82–F84 (2018).
95. Weinberg, D. D. et al. Visual attention patterns of team leaders during delivery room resuscitation. *Resuscitation* **147**, 21–25 (2020).
96. Wagner, M. et al. Eye-tracking during simulation-based neonatal airway management. *Pediatr. Res* **87**, 518–522 (2020).
97. Wagner, M. et al. Visual attention during pediatric resuscitation with feedback devices: a randomized simulation study. *Pediatr. Res.* <https://doi.org/10.1038/s41390-021-01653-w> (2021).
98. Herrick, H. et al. Provider visual attention on a respiratory function monitor during neonatal resuscitation. *Arch. Dis. Child. Fetal Neonatal Ed.* **105**, 666–668 (2020).
99. Katz, T. A. et al. Visual attention on a respiratory function monitor during simulated neonatal resuscitation: an eye-tracking study. *Arch. Dis. Child. Fetal Neonatal Ed.* **104**, F259–F264 (2019).
100. Patel, A., Khatib, M. N., Kurhe, K., Bhargava, S. & Bang, A. Impact of neonatal resuscitation trainings on neonatal and perinatal mortality: a systematic review and meta-analysis. *BMJ Paediatr. Open* **1**, e000183 (2017).
101. Smith, T. J., Valstar, M., Sharkey, D. & Crowe, J. Clinical scene segmentation with tiny datasets. In: *2019 IEEE/CVF International Conference on Computer Vision Workshop (ICCVW)*. 1637–1645 (2019).
102. Villarroel, M. et al. Non-contact physiological monitoring of preterm infants in the Neonatal Intensive Care Unit. *Npj Digit. Med.* **2**, 1–18 (2019).
103. Lee, W. H. et al. Feasibility of non-contact cardiorespiratory monitoring using impulse-radio ultra-wideband radar in the neonatal intensive care unit. *PLoS ONE* **15**, e0243939 (2020).
104. Yamada, N. K., Catchpole, K. & Salas, E. The role of human factors in neonatal patient safety. *Semin. Perinatol.* **43**, 151174 (2019).
105. Salih, Z. N. I. & Draucker, C. B. Facilitators of and barriers to successful teamwork during resuscitations in a neonatal intensive care unit. *J. Perinatol.* **39**, 974–982 (2019).
106. Fuerch, J. H., Sanderson, P., Barshi, I. & Liley, H. Developing safe devices for neonatal care. *Semin. Perinatol.* **43**, 151176 (2019).
107. Pickup, L. et al. Development of a clinical interface for a novel newborn resuscitation device: human factors approach to understanding cognitive user requirements. *JMIR Hum. Factors* **6**, e12055 (2019).
108. Chung, H. U. et al. Binodal, wireless epidermal electronic systems with in-sensor analytics for neonatal intensive care. *Science* **363**, eaau0780 (2019).
109. Bonner, O., Beardsall, K., Crilly, N. & Lasenby, J. There were more wires than him: the potential for wireless patient monitoring in neonatal intensive care. *BMJ Innov.* **3**, 12–18 (2017).
110. Austin, T. The development of neonatal neurointensive care. *Pediatr. Res.* <https://doi.org/10.1038/s41390-019-0729-5> (2019).
111. Pichler, G., Cheung, P.-Y., Aziz, K., Urlesberger, B. & Schmölzer, G. M. How to monitor the brain during immediate neonatal transition and resuscitation? A systematic qualitative review of the literature. *Neonatology* **105**, 205–210 (2014).
112. University of Alberta. Effect of additional monitoring devices on healthcare provider mental workload, visual attention, and performance during neonatal resuscitation. [clinicaltrials.gov](https://clinicaltrials.gov/ct2/show/NCT04521725); 2020. <https://clinicaltrials.gov/ct2/show/NCT04521725> (2021).
113. Partridge, E. A. et al. An extra-uterine system to physiologically support the extreme premature lamb. *Nat. Commun.* **8**, 15112 (2017).
114. Ramanathan, A. et al. Assessment of breath sounds at birth using digital stethoscope technology. *Eur. J. Pediatr.* **179**, 781–789 (2020).
115. King, A., Blank, D., Bhatia, R., Marzbanrad, F. & Malhotra, A. Tools to assess lung aeration in neonates with respiratory distress syndrome. *Acta Paediatr.* **109**, 667–678 (2020).

AUTHOR CONTRIBUTIONS

D.S. conceptualised and designed the review. All authors wrote and contributed sections. N.B. and D.S. produced the final version and all authors approved this on behalf of the ESPR Newborn Resuscitation writing group.

FUNDING

A.M. is supported by funding from the National Health and Medical Research Council, Cerebral Palsy Alliance, and Kathleen Tinsley Trust Foundation, Australia. D.S. is supported by the National Institute of Health Research (NIHR) Children and Young People MedTech Co-operative (CYP MedTech), has received funding for technology development from the Medical Research Council, NIHR and Action Medical Research, and is a non-executive director of SurePulse Medical who are developing monitoring solutions for neonatal care. The views expressed are those of the author(s) and not necessarily those of the NHS, the NIHR, or of the Department of Health.

COMPETING INTERESTS

D.S. is supported by NIHR and CYP MedTech and is a non-executive director of SurePulse Medical, see 'Funding information' section. All other authors declare no potential conflict of interest.

ETHICS APPROVAL AND CONSENT TO PARTICIPATE

Ethical approval or patient consent were not required for this review.

ADDITIONAL INFORMATION

Supplementary information The online version contains supplementary material available at <https://doi.org/10.1038/s41390-022-01988-y>.

Correspondence and requests for materials should be addressed to Don Sharkey.

Reprints and permission information is available at <http://www.nature.com/reprints>

Publisher's note Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.

Open Access This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons license, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons license, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons license and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this license, visit <http://creativecommons.org/licenses/by/4.0/>.

© The Author(s) 2022

THE EUROPEAN SOCIETY FOR PAEDIATRIC RESEARCH (ESPR) NEONATAL RESUSCITATION SECTION WRITING GROUP

Mara Niemuth⁷, Helmut Küster⁸, Henry Rozycki⁹, Anne Lee Solevåg^{10,11}, Inmaculada Lara-Cantón¹², Shiraz Badurdeen¹³, Janneke Dekker¹⁴, Peter Davis¹⁵, Calum Roberts¹⁵, Arjan te Pas¹⁴, Máximo Vento¹², Burkhard Simma¹⁶, Marieke den Boer^{17,18}, Heidi Meredith Herrick¹⁹, Mario Rüdiger^{7,20,21,22}, Maxi Kaufmann^{16,21}, Heidi Aichner²², Samir Gupta^{23,24,25}, Willem deBoode²⁶, Charles Christoph Roehr^{27,28}, Britt Nakstad^{29,30}, Stuart Hooper³¹, Natalie Batey³², Caroline Henry^{32,33}, Shalabh Garg³⁴, Michael Wagner³⁵, Atul Malhotra¹⁵, Michel Valstar³⁶, Thomas Smith³⁶ and Don Sharkey³³

⁷Department for Neonatology and Pediatric Intensive Care, Clinic for Pediatric and Adolescence Medicine, Faculty of Medicine, Technische Universität Dresden, Dresden, Germany. ⁸Clinic for Pediatric Cardiology, Intensive Care and Neonatology, University Medical Center Göttingen, Göttingen, Germany. ⁹Division of Neonatal Medicine, Children's Hospital of Richmond, Virginia Commonwealth University, Richmond, VA, USA. ¹⁰The Department of Paediatric and Adolescent Medicine, Oslo University Hospital, Oslo, Norway. ¹¹Neonatal Intensive Care Unit, The Department of Paediatric and Adolescent Medicine, Oslo University Hospital, Oslo, Norway. ¹²Neonatal Research Group, Health Research Institute and University and Polytechnic Hospital La Fe, Valencia, Spain. ¹³Newborn Research Center and Neonatal Services, The Royal Women's Hospital, Melbourne, VIC, Australia. ¹⁴Division of Neonatology, Department of Paediatrics, Leiden University Medical Centre, Leiden, The Netherlands. ¹⁵Department of Paediatrics, Monash University, Clayton, VIC, Australia. ¹⁶Department of Paediatrics, Academic Teaching Hospital, Landeskrankenhaus Feldkirch, Feldkirch, Austria. ¹⁷Division of Neonatology, Leiden University Medical Center, Leiden, The Netherlands. ¹⁸Department of Medical Ethics and Health Law, Leiden University Medical Center, Leiden, The Netherlands. ¹⁹Department of Pediatrics, Division of Neonatology, Children's Hospital of Philadelphia, Philadelphia, PA, USA. ²⁰Saxony Center for Feto-Neonatal Health, Faculty of Medicine, Technische Universität Dresden, Dresden, Germany. ²¹Department of Paediatrics, Division of Neonatology and Paediatric Intensive Care Medicine, Medical Faculty, Technical University Dresden, Dresden, Germany. ²²Saxony Center for Feto-Neonatal Health, Medical Faculty, Technical University Dresden, Dresden, Germany. ²³Department of Engineering, Durham University, Durham, UK. ²⁴Honorary Consultant Neonatologist, University Hospital of North Tees, Stockton-on-Tees, UK. ²⁵Division of Neonatology, Department of Pediatrics, Sidra Medicine, Doha, Qatar. ²⁶Kinderarts-neonatoloog, Radboudumc Amalia Children's Hospital, 6500 HB Nijmegen (804), The Netherlands. ²⁷National Perinatal Epidemiology Unit Clinical Trials Unit, University of Oxford, Oxford, UK. ²⁸Newborn Services, Southmead Hospital, Bristol, UK. ²⁹Department of Pediatrics and Adolescent Health, University of Botswana, Gaborone, Botswana. ³⁰University of Oslo, Faculty Medicine, Oslo, Norway. ³¹Hudson Institute for Medical Research, Department of Obstetrics and Gynaecology, Monash University, Melbourne, VIC, Australia. ³²Nottingham University Hospitals NHS Trust, Nottingham Neonatal Service, Nottingham, UK. ³³Centre for Perinatal Research, School of Medicine, University of Nottingham, Nottingham, UK. ³⁴Department of Neonatal Medicine, James Cook University Hospital, Middlesbrough, UK. ³⁵Division of Neonatology, Pediatric Intensive Care and Neuropediatrics, Department of Pediatrics, Comprehensive Center for Pediatrics, Medical University of Vienna, Vienna, Austria. ³⁶School of Computer Science, University of Nottingham, Nottingham, UK.