AUTHOR CORRECTION

Open Access

Author Correction: Naked-eye point-of-care testing platform based on a pH-responsive superwetting surface: toward the non-invasive detection of glucose

Zhong Feng Gao¹, Ei Ei Sann¹, Xiaoding Lou^{1,2}, Renyi Liu², Jun Dai¹, Xiaolei Zuo³, Fan Xia^{1,2} and Lei Jiang⁴

Correction to: *NPG Asia Materials* <https://doi.org/10.1038/s41427-018-0024-7>; Article published online 11 April 2018

After online publication of this article, the authors noticed an error in the supplementary information section in the manuscript.

The correct supplementary table of this article should have read as below.

“In Table S1, column of the contact angle, the “4.824” of Normal 7 and “6.310” of Normal 8 should revise as “108.7” of Normal 7 and “101.6” of Normal 8, respectively.”

The authors apologize for any inconvenience caused.

Correspondence: Jun Dai (dj_hust1987@sina.com) or Fan Xia (xiafan@cug.edu.cn)

¹Hubei Key Laboratory of Bioinorganic Chemistry & Materia Medica, School of Chemistry and Chemical Engineering, Department of Obstetrics and Gynecology, Tongji Hospital, Tongji Medical College, Huazhong University of Science and Technology, 430074 Wuhan, China

²Engineering Research Center of Nano-Geomaterials of Ministry of Education, Faculty of Materials Science and Chemistry, China University of Geosciences, 388 Lumo Road, 430074 Wuhan, China

³Institute of Molecular Medicine, Shanghai Jiao Tong University, 200240 Shanghai, China

⁴Key Laboratory of Bio-Inspired Smart Interfacial Science and Technology of Ministry of Education, School of Chemistry, Beihang University, 100191 Beijing, China

© The Author(s) 2018

Open Access This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons license, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons license, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons license and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this license, visit <http://creativecommons.org/licenses/by/4.0/>.

Blood specimens	Blood glucose concentration (mM)					
	Contact angle (CA)(°)	Calculated glucose from CA(mM)	Glucometer 1(ACCU-CHEK, Roche)	Glucometer 2 (Contour plus, Bayer)	Glucometer 3(ONETOUCH, Johnson & Johnson)	RSD% (CA vs GM1)
Patient 1	64.0	14.171	13.4	9.8	7.6	3.95%
Patient 2	45.0	19.521	19.7	13.1	10.8	0.65%
Patient 3	60.5	14.903	13.0	8.6	8.4	9.65%
Patient 4	58.4	15.347	18.4	15.8	12.4	12.79%
Patient 5	59.9	15.042	17.5	14.8	23.7	10.68%
Patient 6	60.8	14.837	13.9	10.9	18.8	4.61%
Patient 7	56.8	15.686	15.0	11.2	7.8	3.16%
Patient 8	56.1	15.828	17.8	12.8	32.9	6.49%
Patient 9	55.7	15.912	16.2	14.6	27.7	1.27%
Normal 1	110.2	4.506	5.0	5.7	5.2	7.35%
Normal 2	106.5	5.248	5.9	6.5	6.4	8.27%
Normal 3	106.3	5.326	5.6	6.1	5.9	3.55%
Normal 4	104.2	5.766	6.0	6.4	6.2	2.81%
Normal 5	107.2	5.138	6.0	5.7	6.2	10.94%
Normal 6	107.6	5.054	5.2	4.8	5.7	2.01%
Normal 7	108.7	5.2	5.9	6.0	4.824	5.30%
Normal 8	101.6	6.8	7.3	7.1	6.310	5.29%
Normal 9	105.6	5.472	5.5	6.0	6.5	0.36%