#### CORRESPONDENCE





# Normative data for peripapillary retinal nerve fibre layer OCT scans in patients of African descent

Frederick R. Burgess<sup>1,2</sup> · Kazim Dhalla<sup>3</sup> · Celina F. Mhina<sup>4</sup> · Andrew Blaikie<sup>1</sup>

Received: 23 July 2020 / Revised: 1 September 2020 / Accepted: 3 September 2020 / Published online: 11 September 2020 © Springer Nature Limited 2020

## To the Editor:

We present normative optical coherence tomography (OCT) peripapillary data for healthy adults of African descent (AD) and wish to emphasise the lack of normative control data from all regions of sub-saharan Africa.

We read with interest the recent article by Yashadhana et al. calling for improved diversity in global eye health leadership [1]. Leadership is not the only area of eye health where diversity could be improved. We are currently investigating TEON: a bilateral subacute optic neuropathy of unknown aetiology affecting young people in Tanzania [2]. The condition demonstrates similar clinical and temporal characteristics to Leber's Hereditary Optic Neuropathy (LHON). As part of the TEON Study—a collaborative research project that aims to identify the causes of TEON we sought to compare the OCT findings of TEON and LHON, utilizing a new OCT scanner (Topcon 3D OCT-1 Maestro, Topcon Medical Systems, Oakland, NJ, USA) at Muhimbili National Hospital.

To allow thorough comparison of OCT data, we need a representative normative dataset. Only 20% of the normative dataset of our OCT analysis software contains data from people of AD and only 45% were under the age of 40 [3]. The AD and Glaucoma Evaluation Study, a study that investigated glaucoma in American AD patients, showed significant baseline differences in optic nerve parameters between healthy participants of African and European descent [4].

We therefore conducted a literature search for normative data for OCT peripapillary retinal nerve fibre layer (pRNFL) thickness measurements in healthy adult AD participants, the results of which can be seen in Table 1.

The difference between the averages is in agreement with studies that have previously examined ethnic differences in pRNFL OCT thicknesses—namely that AD patients have thinner temporal peripapillary RNFL thicknesses and thicker RNFL in the other quadrants [5].

It should be noted that many of these RNFL studies concern glaucoma, primarily a disease of the elderly, hence normative data may be skewed by age, as well as differences in scan equipment, scan technique, and optic disc morphology.

The variation in RNFL thicknesses seen in Table 1 shows the difficulties in comparing OCT scan data and emphasises the need for more extensive normative control data from all regions of sub-saharan Africa. This article hopefully highlights the heterogenous nature of the term 'AD' and that there are no published normative datasets for East African patients.

Frederick R. Burgess frb2@st-andrews.ac.uk

<sup>2</sup> Department of Ophthalmology, Princess Alexandra Eye Pavilion, Edinburgh, UK

- <sup>3</sup> Dr Agarwal's Eye Hospital, Dar es Salaam, Tanzania
- <sup>4</sup> Muhimbili University of Health & Allied Sciences, Dar es Salaam, Tanzania

<sup>&</sup>lt;sup>1</sup> Department of Global Health, School of Medicine, University of St Andrews, St Andrews, UK

 
 Table 1 Studies reporting normative pRNFL OCT data for adult participants of African descent.

Study first author	Location	Year	Eyes (n)	Weight <sup>a</sup>	pRNFL	tRNFL	iRNFL	sRNFL	nRNFL
Mashige	South Africa	2017	600	11.88	110	73.63	135.06	131.96	87.24
Girkin	USA	2010	315	6.24	103.69	66.48	135.12	128.83	84.3
Sani	Nigeria	2016	220	4.36	104.2	67.2	129.2	135.3	85.1
Ismail	South Africa	2019	132	2.61	108.7	74.8			77.7
Blumberg	USA	2016	103	2.04	93.98	60.59	122.71	119.16	73.42
Ocansey	Ghana	2020	100	1.98	102.4	66.8	133.6	131.5	76.7
Knight	USA	2012	51	1.01	93.9	57.8	125.2	119.6	73
Rao	USA	2015	50	0.99	97.18		130.04	121.16	
Racette	USA	2005	42	0.83	114.86	84.55	141.69	149.79	82.86
Poon	USA	2018	40	0.79	92.97	62.09	118.65	114.47	76.49
Seraji-Bozorgzad	USA	2016	27	0.53	100.9	68.9	133.7	128.1	72.4
Budenz	USA	2007	27	0.53	101.1				
Alasil	USA	2013	26	0.51	99.2				
Kimbrough	USA	2015	14	0.28	100	67	135	126	73
Weighted average (AD participants)					105.0	69.6	132.6	130.1	82.9
OCT software average (all ethnicities)					104.0	70.6	123.9	116.4	80.6

*pRNFL* average peripapillary RNFL thickness, *tRNFL* average temporal quadrant peripapillary RNFL thickness, *iRNF* average inferior quadrant peripapillary RNFL thickness, *sRNFL* average superior quadrant peripapillary RNFL thickness.

<sup>a</sup>Weight based on sample size relative to the median sample size.

**Funding** This research was funded by Fight for Sight (grant reference: 24RCO5).

#### **Compliance with ethical standards**

**Conflict of interest** We are an international team investigating Tanzanian Endemic Optic Neuropathy (TEON) in Dar es Salaam, Tanzania. FRB, AB, CFM and KD have no competing interests to declare.

**Publisher's note** Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.

## References

1. Yashadhana A, Zhang JH, Yasmin S, Morjaria P, Holland P, Faal H, et al. Action needed to improve equity and diversity in global eye health leadership. Eye. 2020;34:1051–4.

- Plant GT, Mtanda AT, Arden GB, Johnson GJ. An epidemic of optic neuropathy in Tanzania: characterization of the visual disorder and associated peripheral neuropathy. J Neurol Sci. 1997; 145:127–40.
- Chaglasian M, Fingeret M, Gunvant Davey P, Huang W, Leung D, Ng E, et al. The development of a reference database with the Topcon 3D OCT-1 Maestro. Clin Ophthalmol. 2018;12: 849–57.
- Girkin CA, Sample PA, Liebmann JM, Jain S, Bowd C, Becerra LM, et al. African Descent and Glaucoma Evaluation Study (ADAGES). Arch Ophthalmol. 2010;128:541–50.
- Poon LY-C, Antar H, Tsikata E, Guo R, Papadogeorgou G, Freeman M, et al. Effects of age, race, and ethnicity on the optic nerve and peripapillary region using spectral-domain OCT 3D volume scans. Transl Vis Sci Technol. 2018;7:12.