



OPEN

Quantifying neutralising antibody responses against SARS-CoV-2 in dried blood spots (DBS) and paired sera

Kelly J. Roper^{1,4}, Jordan Thomas^{1,4}, Wejdan Albalawi^{1,4}, Emily Maddocks^{1,4}, Susan Dobson^{1,4}, Abdullateef Alshehri^{1,4}, Francesco G. Barone², Murielle Baltazar^{1,4}, Malcolm G. Semple^{1,4,47}, Antonia Ho³, Lance Turtle^{1,4}, ISARIC4C Consortium*, William A. Paxton^{1,4,46} & Georgios Pollakis^{1,4,46}✉

The ongoing SARS-CoV-2 pandemic was initially managed by non-pharmaceutical interventions such as diagnostic testing, isolation of positive cases, physical distancing and lockdowns. The advent of vaccines has provided crucial protection against SARS-CoV-2. Neutralising antibody (nAb) responses are a key correlate of protection, and therefore measuring nAb responses is essential for monitoring vaccine efficacy. Fingerstick dried blood spots (DBS) are ideal for use in large-scale sero-surveillance because they are inexpensive, offer the option of self-collection and can be transported and stored at ambient temperatures. Such advantages also make DBS appealing to use in resource-limited settings and in potential future pandemics. In this study, nAb responses in sera, venous blood and fingerstick blood stored on filter paper were measured. Samples were collected from SARS-CoV-2 acutely infected individuals, SARS-CoV-2 convalescent individuals and SARS-CoV-2 vaccinated individuals. Good agreement was observed between the nAb responses measured in eluted DBS and paired sera. Stability of nAb responses was also observed in sera stored on filter paper at room temperature for 28 days. Overall, this study provides support for the use of filter paper as a viable sample collection method to study nAb responses.

The severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2) pandemic has been one of the most detrimental viral outbreaks of the twenty-first century and is so far responsible for nearly 6.9 million deaths globally¹. The rapid development of SARS-CoV-2 vaccines has facilitated large-scale immunisation programmes that have provided vital protection against the virus²⁻⁴. Measuring antibody responses against SARS-CoV-2 is crucial to understanding population seroprevalence, the longevity of immunity and vaccine efficacy, especially with the emergence of SARS-CoV-2 variants⁵. Humoral immune responses measured from patients infected with SARS-CoV-2 have been found to predominantly target the Spike (S) glycoprotein and Nucleocapsid (N) protein antigens⁶. Many of the antibodies generated against S and N will bind to the viral antigens, but antibodies that can bind and subsequently prevent viral entry, known as neutralising antibodies (nAbs), are the type of antibodies that correlate with protection against future infections^{7,8}. Many assays are used to investigate humoral responses against SARS-CoV-2 but not all assays measure nAbs⁹. Live virus assays are the gold standard for measuring nAb responses, but for SARS-CoV-2 they can only be performed in high-containment facilities. Pseudo-virus particle (PVP) assays overcome the containment issues associated with live virus and nAb levels measured against SARS-CoV-2 with PVP assays have been observed to correlate with the results obtained in live virus assays¹⁰⁻¹³.

¹Department of Clinical Infection, Microbiology and Immunology (CIMI), Institute of Infection, Veterinary and Ecological Sciences (IVES), University of Liverpool, Liverpool L69 7BE, UK. ²Department of Biochemistry and Systems Biology, Institute of Systems, Molecular and Integrative Biology (ISMIB), University of Liverpool, Liverpool L69 3BX, UK. ³MRC-University of Glasgow Centre for Virus Research, 464 Bearsden Road, Glasgow G61 1QH, UK. ⁴Faculty of Health and Life Sciences, Institute of Infection, Veterinary and Ecological Sciences, University of Liverpool, Liverpool, UK. ⁴⁶NIHR Health Protection Research Unit in Emerging and Zoonotic Infections, Liverpool, UK. ⁴⁷Respiratory Medicine, Alder Hey Children's Hospital, Institute in The Park, University of Liverpool, Liverpool, UK. *A list of authors and their affiliations appears at the end of the paper. ✉email: g.pollakis@liverpool.ac.uk

During the pandemic, traditional serology was limited due to physical distancing restrictions and/or lockdowns imposed by many governments. Dried blood spots (DBS) are an ideal solution to overcome these sample collection issues, as they can be collected from fingerstick blood in non-clinical settings and therefore provide the potential for self-sampling¹⁴. To date, DBS are most well known for their use in newborn screening for genetic health conditions, but they have also been widely used in diagnosing and monitoring therapy for human immunodeficiency virus (HIV) and hepatitis virus infections^{15–17}. The practicality and reliability of DBS to measure antibody responses against a range of viruses has also been demonstrated^{17,18}. Many studies have utilised fingerstick DBS (FDBS) to quantify binding antibodies against SARS-CoV-2 S^{19–26} or against both S and N^{27–32}. Some studies have also used FDBS to assess SARS-CoV-2 seroprevalence^{33–38}, however, only a few studies have explored the use of FDBS to measure SARS-CoV-2 nAb responses^{39–41}. We utilised Schleicher & Schuell 903 filter paper cards, previously used to measure HIV viral loads⁴², to store sera, venous blood and fingerstick blood. Samples were utilised from SARS-CoV-2 acutely infected individuals, SARS-CoV-2 convalescent individuals and SARS-CoV-2 vaccinated individuals. By comparing nAb responses in paired sera to eluted filter paper samples we have shown that DBS are a viable sample collection method to study nAb responses against SARS-CoV-2.

Results

Human serum retains neutralising ability against SARS-CoV-2 S after storage on filter paper. To determine if filter paper can be used to store human sera for serological testing, we tested dried serum spots (DSS) eluates in neutralisation assays using single-round infectious PVP expressing the SARS-CoV-2 S glycoprotein. For all filter paper eluate (FPE) tested in a neutralisation assay, a paired serum sample was also assayed (Supplementary Table S1). Additionally, one or two replicates of a known positive control serum were tested in every assay to measure reliability between assays (Supplementary Fig. S1). Due to availability of sera, 4 different positive control sera were used over the course of the experiments. Initial experiments were performed to test 3 different sample elution conditions. When compared to the neutralisation observed for paired serum, we found that elution at 4 °C overnight was an optimal protocol for eluting DSS from filter paper (Supplementary Fig. S2). Following this, we tested 53 DSS eluates to further evaluate if storage on filter paper affected neutralising capacity when compared to the standard storage method (Fig. 1). Three sera and DSS eluates failed to prevent PVP infectivity and were classed as non-neutralising and subsequently excluded from further analysis, though results agreed. For the remaining 50 paired samples, the inhibitory concentrations (IC) that reduced PVP infectivity by 50% (IC50), 70% (IC70) or 90% (IC90) were calculated. One serum IC50 value was found to be above the limit of detection (LOD) and 2 IC70 and 6 IC90 values were determined to be below the LOD. All IC values that were found to be outside the limits of detection for sera were also outside the limits for FPE and were excluded from further analysis.

43/49 (88%) DSS eluates had comparable IC50 values to paired sera (Fig. 1a). One DSS eluate had an IC50 value that was below the LOD for FPE and 5 had IC50 values that were above the LOD for FPE. There was a positive relationship between the IC50 values of DSS eluates and the IC50 values of paired sera (simple linear regression analysis $R^2 = 0.7432$, $p < 0.0001$) (Fig. 1d). 47/48 (98%) IC70 values were measured for DSS eluates, with 1 below the LOD for FPE (Fig. 1b). A positive relationship was also observed between IC70 values for DSS eluates and sera (simple linear regression analysis $R^2 = 0.8381$, $p < 0.0001$) (Fig. 1e). Two DSS eluates IC90 values were below the LOD for FPE meaning 42/44 (95%) IC90 values were measured (Fig. 1c). IC90 values were also positively related between assays (simple linear regression analysis $R^2 = 0.8286$, $p < 0.0001$) (Fig. 1f). Bland–Altman analyses on the absolute differences between the IC50 values (Fig. 1g), IC70 values (Fig. 1h) and IC90 values (Fig. 1i) for DSS eluates and sera were also performed to assess the agreement between storage methods⁴³. Overall, there was high agreement between methods for all IC values measured. The difference between an IC value measured for a DSS eluate and the IC value measured for its paired sera was then calculated for all samples. The differences were then averaged to determine the bias of one assay over or under-estimating IC values compared to the other. The smallest bias of -0.0271 was observed for IC90 values (95% confidence interval [CI] 1.10 to -1.15), followed by IC50 values with a bias of -0.0335 (CI 1.85 to -1.91). IC70 values had the largest bias of -0.0602 (CI 1.30 to -1.41). The small biases observed indicate nAb responses measured for FPE are highly similar to those measured in sera.

Next, we tested if PVP neutralisation was affected by storing DSS at room temperature (RT) (approximately 20 °C) over a period of up to 28 days (Supplementary Fig. S3). Seven serum samples were selected and paired DSS were left at RT for 2, 5, 7, 14 and 21 days. Due to limitations in sera availability, only 4 samples (CCP-UK01, CCP-UK05, CCP-UK31, CCP-UK10b) had a DSS left at RT for 28 days (Supplementary Fig. S4). IC70 and IC90 values for CCP-UK01 serum were excluded as they were below the LOD. Overall, 30/39 (77%) IC50 values, 32/33 (97%) IC70 values and 31/33 (94%) IC90 values were measured for eluates from DSS that were stored at RT (Supplementary Table S2). IC50 values for 1 sample (CCP-UK10b) were below the LOD for the FPE for all 6 of the RT DSS eluates tested. One IC90 value (CCP-UK17 eluate from the 2-day RT DSS) was below the LOD due to an elution issue that resulted in a lower eluate volume. A lower volume of eluted sample meant that the detection limit was reduced, though IC50 and IC70 values were still measured. Wilcoxon paired t-tests were performed to assess the difference between the average IC50 values (Fig. 2a and d), IC70 values (Fig. 2b and e) and IC90 values (Fig. 2c and f) for all eluates of DSS stored at RT compared to paired sera and no significant differences ($p > 0.05$) were found. The results indicate that storage of DSS on filter paper cards at RT for up to 28 days did not affect the measurement of nAb responses in FPE.

Venous whole blood stored on filter paper can act as a surrogate for measuring nAb responses against SARS-CoV-2 S in place of sera. Following the evaluation of DSS, we proceeded to measure venous DBS (VDBS) eluates in neutralisation assays (Supplementary Fig. S5). Initially 15 μ L and 25 μ L VDBS




Figure 1. Serum stored on filter paper retains neutralising capacity against SARS-CoV-2 spike. Pseudo-virus particles (PVP) expressing SARS-CoV-2 spike were used to measure the neutralisation capacity of paired human sera stored in direct aliquots and eluted from dried serum spots (DSS) stored on filter paper. Neutralisation activity was defined as the serum dilution that reduced PVP infectivity by 50%, 70% or 90% (IC₅₀, IC₇₀ or IC₉₀, respectively). (a–c) Slope charts from left to right display on the y-axis PVP neutralisation as (a) IC₅₀, (b) IC₇₀ and (c) IC₉₀. The x-axes show the sample type with sera represented by orange circles and DSS eluates represented by blue circles. IC₅₀, IC₇₀ and IC₉₀ values for paired sample types are connected via grey lines. The dotted lines across the charts represent the lower limits of detection, with the limit for direct sera in orange and DSS in blue (d–f) Scatter plots from left to right show direct sera IC₅₀, IC₇₀ and IC₉₀ plotted against DSS eluates IC₅₀, IC₇₀ and IC₉₀ values. Simple linear regression analysis was performed and found significant positive relationships between sera aliquots and DSS eluates IC₅₀ values [$n = 43$, $R^2 = 0.7432$, $p < 0.0001$], IC₇₀ values [$n = 47$, $R^2 = 0.8381$, $p < 0.0001$] and IC₉₀ values [$n = 42$, $R^2 = 0.8286$, $p < 0.0001$]. The dotted lines represent the upper and lower 95% confidence intervals (CI) of the line of best fit. (g–i) Bland–Altman plots display on the x-axes the average IC₅₀ values ($n = 43$), IC₇₀ values ($n = 47$) and IC₉₀ values ($n = 42$) for sera aliquots and DSS eluates and the difference between the two values on the y-axes. Red lines represent bias, and the dotted lines represent the upper and lower 95% CI.

eluates along with paired sera were screened for 14 samples and IC₇₀ values were calculated. Two samples were excluded because 1 failed to prevent PVP infectivity and so was classed as non-neutralising and the other had an IC₇₀ value below the LOD. For the remaining 12 samples 50 μ L VDBS eluates were also tested. IC₇₀ values were obtained for 12/12 (100%) 25 μ L and 50 μ L VDBS eluates. Only 10/12 (83%) IC₇₀ values were measured for 15 μ L VDBS eluates, with 2 being below the LOD for FPE (Fig. 3). IC₇₀ values for all three volumes of VDBS eluates were found to have a positive relationship with sera (Fig. 3b–d). 50 μ L VDBS eluates had the strongest relationship with sera (simple linear regression analysis $R^2 = 0.8655$, $p < 0.0001$), followed by 25 μ L VDBS eluates (simple linear regression analysis $R^2 = 0.7429$, $p = 0.0003$) and then by 15 μ L VDBS eluates (simple linear regression analysis $R^2 = 0.6780$, $p = 0.0034$). Bland–Altman analyses on the absolute differences between the IC₇₀ values for 15 μ L VDBS eluates (Fig. 3e), 25 μ L VDBS eluates (Fig. 3f) and 50 μ L VDBS eluate (Fig. 3g) and sera were also performed, again to assess the agreement of measuring nAb responses between sample types⁴³. The smallest bias of -0.0417 was observed for 50 μ L VDBS eluates (CI 0.88 to -0.97), followed by a bias of -0.4408




Figure 2. Neutralising capacity of serum against SARS-CoV-2 spike is not significantly affected by storage on filter paper at room temperature for up to 28 days. Single-round infectious pseudo-virus particles (PVP) expressing SARS-CoV-2 spike were used to measure the neutralisation capacity of 7 human sera samples (squares-CCP-UK01, downwards triangles-CCP-UK10a, upwards triangles CCP-UK10b, circles-CCP-UK05, diamonds-CCP-UK16, hexagons-CCP-UK17 and stars-CCP-UK31). Sera were stored on filter paper kept at room temperature (RT) as dried serum spots (DSS) for 0–28 days before elution and compared to sera stored as direct aliquots at -80°C (serum). Due to availability of sera only 4 out of 7 had DSS left at RT for 28 days. Bar charts a–f show neutralisation activity on the y-axes defined as the serum dilution that reduced PVP infectivity by 50% (a and d), 70% (b and e) and by 90% (c and f) (IC₅₀, IC₇₀ or IC₉₀, respectively). Error bars represent standard deviation from the mean. Wilcoxon t-tests were run, and no significant differences (ns) were found between the mean IC values for serum controls and the mean IC values for DSS stored at RT from 2 to 28 days ($p > 0.05$).

for 25 μL VDDBS eluates (CI 0.75 to -1.64). The 15 μL VDDBS eluates had the largest bias of -0.82 (CI 0.29 to -1.93). The 10 samples with complete IC₇₀ values for all four sample types (sera, 15 μL , 25 μL and 50 μL VDDBS eluates) were analysed in a repeated measures one-way ANOVA test and significant differences ($p < 0.0001$) were found between mean IC₇₀ values (Fig. 3a). A multiple comparisons test was performed to identify which of the mean IC₇₀ values for VDDBS eluates were significantly different from the serum controls. The mean IC₇₀ value for 15 μL VDDBS eluates was found to be significantly higher ($p < 0.0001$) than the mean IC₇₀ value for the serum controls. No significant differences were identified between the mean IC₇₀ values for the serum controls and the 25 μL and 50 μL VDDBS eluates ($p > 0.05$).

Fingerstick DBS can be used to reliably estimate nAb responses against SARS-CoV-2 S. Lastly, we tested 12 FDDBS with paired sera to evaluate if capillary blood stored on filter paper could also be used to




Figure 3. Venous whole blood stored on filter paper shows comparable serum neutralisation of SARS-CoV-2 spike to paired sera aliquots. Pseudo-virus particles (PVP) expressing SARS-CoV-2 spike were used to measure the neutralisation capacity of paired human sera with venous blood eluates that were stored on filter paper as dried blood spots (VDBS). Neutralisation activity was defined as the serum dilution that reduced PVP infectivity by 70% (IC_{70}). As whole blood contains approximately 55% serum this was accounted for when calculating IC_{70} values for VDBS eluates. **(a)** Violin plot displays on the y-axis PVP neutralisation as IC_{70} values. The x-axis shows paired sample types ($n = 10$) with sera represented by orange circles and 3 pipetted volumes of VDBS (15 μL , 25 μL & 50 μL) represented by pink circles. The dotted lines across the plot represent the lower limits of detection, with the limit for sera in orange and VDBS eluates in pink. A repeated measures one-way ANOVA test was run to compare the mean IC_{70} value for sera against the mean IC_{70} values for VDBS eluates. A significant difference was found ($n = 10$, $p < 0.0001$) and a Holm–Šidák’s multiple comparisons test identified it to be between the mean sera and mean 15 μL VDBS eluates IC_{70} values ($p < 0.0001$). No significant differences (ns) were found between the mean IC_{70} values for sera and the 25 μL and 50 μL VDBS eluates ($p > 0.05$) **(b–d)** Scatter plots show the IC_{70} values for sera plotted against the IC_{70} values for VDBS eluates **(b)** 15 μL VDBS **(c)** 25 μL and **(d)** 50 μL VDBS. Simple linear regression analysis found significant relationships between IC_{70} values for sera and 15 μL VDBS eluates [$n = 10$, $R^2 = 0.6780$, $p = 0.0034$], 25 μL VDBS eluates [$n = 12$, $R^2 = 0.7429$, $p = 0.0003$] and 50 μL VDBS eluates [$n = 12$, $R^2 = 0.8655$, $p < 0.0001$]. The dotted lines represent the upper and lower 95% confidence intervals (CI) of the line of best fit. **(3e–3g.)** Bland–Altman plots display on the x-axes the average IC_{70} values for sera and VDBS eluates **(e)** 15 μL **(f)** 25 μL and **(g)** 50 μL and the difference between the IC_{70} values on the y-axes. The red lines represent bias, and the dotted lines represent the upper and lower 95% CI.

detect nAb responses against SARS-CoV-2 S PVP (Supplementary Fig. S6). As the exact volume of blood blotted for each FDBS was unknown, known blood spot volumes for VDBS were used to estimate FDBS blood volumes (Supplementary Fig. S7c). From visual comparisons we estimated FDBS to be between 15 and 25 μL of blood. From this estimation, we used two different input volumes of 15 μL and 25 μL to calculate the nAb responses in FDBS. One out of 12 samples tested failed to prevent PVP infectivity and so was classed as non-neutralising and excluded from further analysis. IC_{50} , IC_{70} and IC_{90} values were then calculated for the remaining samples (Fig. 4). 1/11 sera IC_{70} values (AVIS08b) and 3/11 sera IC_{90} values (AVIS10, AVIS07, and AVIS08b) were below the LOD and were excluded. Following this, 8/11 (73%) IC_{50} values, 8/10 (80%) IC_{70} values and 7/8 (88%) IC_{90} values were measured for both the 15 μL and 25 μL estimated FDBS eluates. For 1 FDBS eluate the IC_{50} value could not be measured as a result of a poor infectivity curve (AVIS10). Two other IC_{50} values could not be measured as they were below the LOD for FPE (AVIS08a and AVIS08b) along with 2 IC_{70} values (AVIS07 and AVIS08a) and 1 IC_{90} value (AVIS08a). The mean IC_{50} , IC_{70} and IC_{90} values that were measured for estimated




Figure 4. Dried blood spots obtained via fingerstick show comparable serum neutralisation of SARS-CoV-2 spike to paired serum aliquots. Pseudo-virus particles (PVP) expressing SARS-CoV-2 spike were used to measure the neutralisation capacity of human sera and eluted dried blood spots obtained via fingerstick (FDBS) for 11 participant samples. Neutralisation activity was defined as the serum dilution that reduced PVP infectivity by 50%, 70% or 90% (IC₅₀, IC₇₀ or IC₉₀, respectively). As whole blood contains approximately 55% serum this was accounted for when calculating IC values for FDBS eluate. For each FDBS the exact volumes of blood blotted from the participants' fingers were not measured therefore two volumes were estimated for FDBS volume. The dotted lines across the graph represent the lower limits of detection, with the limit for sera in orange, DSS eluate in blue, $\approx 15 \mu\text{L}$ FDBS eluate in yellow and $\approx 25 \mu\text{L}$ FDBS eluate in brown.

FDBS eluates and paired sera and were analysed in Friedman ANOVA tests (Supplementary Fig. S8). No significant differences were identified between the mean IC₅₀ values ($p=0.079$) (Supplementary Fig. S8a) and IC₉₀ values ($p=0.052$) (Supplementary Fig. S8b). Significant differences were found between the mean IC₇₀ values ($p=0.018$), but a multiple comparisons test did not find these differences to be between the mean IC₇₀ value for the serum controls and the $15 \mu\text{L}$ estimated FDBS eluates ($p=0.16$) or the $25 \mu\text{L}$ estimated FDBS eluates ($p=0.63$) (Supplementary Fig. S8c).

Finally, two participants had all filter paper sample types (DSS, VDBS and FDBS) tested in neutralisation assays (Supplementary Fig. S9). Friedman ANOVA tests found no significant differences ($p>0.05$) between mean IC₅₀ values (Supplementary Fig. S9a), IC₇₀ values (Supplementary Fig. S9b) and IC₉₀ values (Supplementary Fig. S9c) for all sample types tested from the two participants.

Discussion

The application of FDBS sample collection to study humoral responses could provide a practical alternative to blood collection via traditional venepuncture. FDBS are less invasive, low cost, and can be sampled in the field. This study aimed to assess if storing samples on filter paper affected the quantification of nAb responses against SARS-CoV-2. We observed high levels of agreement between the neutralisation capacity of paired serum samples and FPE, supporting the storage of samples on filter paper to study nAb responses. Other studies have also provided support for the use of DBS to measure SARS-CoV-2 nAb responses^{39–41}. Sancilio et al. used a surrogate virus neutralization test (sVNT) that quantified the inhibition of the interaction between SARS-CoV-2 S and human angiotensin-converting enzyme 2 receptor protein and demonstrated that nAb responses in DBS from SARS-CoV-2 PCR positive and negative samples had a concordance correlation of 0.991 to paired sera⁴¹. These findings concur with our results, however, PVP assays have been shown in the literature to correlate more closely with live virus assay than sVNTs¹². Similarly, Danh et al. used a Split-Oligonucleotide Neighbouring Inhibition Assay (SONIA) to measure the ability of nAb to block the spike protein subunit 1-ACE2 interaction³⁹. A sensitivity of 91–97% was observed between SONIA and the live virus assay for measuring nAb in sera. SONIA was then used to successfully measure nAb responses in self-collected FDBS. One limitation of measuring responses against the spike protein subunit 1 is that nAbs can be generated against other regions of the spike⁴⁴. Itell et al. also used a PVP assay and found a strong correlation of 0.99 between nAb responses in VDBS eluate and paired sera⁴⁰. However, their results on FDBS were limited, as only two self-collected FDBS were tested with paired sera.

Many studies have found that binding antibodies against SARS-CoV-2 are not affected by storage of DBS at RT for durations ranging from 7 days to 6 months^{26,30,40,45,46}. Others have utilised DBS stability for large serosurveillance studies and have also found antibodies to be stable after transport at ambient temperature^{33,34}. No other study, within our knowledge, has measured nAb responses in FPE stored at RT for a defined period. Although nAbs have been shown to be stable in self-collected DBS transported at ambient temperatures, the exact durations that DBS were at RT were not specified^{33,40}. We found that DSS eluates retained comparable nAb responses to paired sera after storage on filter paper for 2–28 days at RT. We observed high agreement between nAb responses from FDBS and paired DSS eluates and therefore, we believe that this result, together with previous studies, provides support for the stability of nAb responses against SARS-CoV-2 in DBS stored at RT for up to 28 days. One consideration, however, is that RT varies in different settings and in this UK study, RT was approximately 20 °C. Other countries may experience higher RT conditions. Nevertheless, several studies have stored DBS at temperatures $\geq 37 \text{ }^\circ\text{C}$ and saw limited effects on measurement of binding antibodies^{28,45,46}. The effect of higher temperatures on nAbs would need to be explored further. Overall, it appears the application of DBS sampling within the field may facilitate the removal of high-cost cold-chain handling of samples before analysing a key correlate of immune protection.

Additionally, we demonstrated that VDDBS eluate nAb responses strongly agree with nAb responses measured from paired sera. However, we did observe a loss of sensitivity for the smallest volume of VDDBS, with the mean IC70 value for the 15 μL VDDBS eluate found to be significantly higher than paired sera. The overestimation of nAb responses for 15 μL VDDBS eluates was likely due to cutting out excess filter paper area not saturated with blood when VDDBS were processed (Supplementary Fig. S7c). As a result, the elution buffer was absorbed into the dry paper surrounding the blood spot and caused the 15 μL VDDBS eluates to be more concentrated than estimated. This may explain why the 50 μL VDDBS eluate had the highest level of agreement to paired sera as 50 μL of blood was observed to saturate the whole filter paper cut-out. Sensitivity issues connected to blood spot cut-out size have been described previously¹⁴. One study also noted that blood spot size significantly affected OD values measured⁴⁵. In this study 19 mm sample discs were cut out, but utilising a punch device that cuts out a smaller sample disc size would be one solution to the discussed problem. Another technicality to overcome this issue would be to excise a sub-spot from within the saturated area of the dried sample. This technique has been shown to reliably yield the same volume of serum on average from varying volumes of blotted blood as long as the area within the sub-spot was fully saturated⁴⁷. As smaller volumes of blood are obtained from fingerstick collection, filter paper cut-out size should be considered when using FDBS to improve sensitivity and accuracy.

A limitation of our work was the unknown volume of blood sampled for FDBS. Although we observed that estimated volumes of FDBS were still predictive of nAb responses, we saw decreased sensitivity when measuring IC values (73–88%). The application of quantitative microsampling is one way to overcome the issues associated with unknown sample volume⁴⁸. One example is the neoteryx[®] Mitra[®] device, which uses volumetric absorptive microsampling tips to collect either 10, 20, or 30 μL volumes of fingerstick blood. The Mitra[®] device has successfully been used to measure binding antibodies against SARS-CoV-2 in the field^{49,50}. Another type of micro sampler used to successfully measure binding antibody responses against SARS-CoV-2 is the Capitainer qDBS, which collects an exact sample volume into a pre-cut DBS disc⁵¹. However, a limitation of microsampling devices is an increased cost and another simpler solution to measure FDBS volume could be the use of capillary blood tubes (Supplementary Fig. S10).

To conclude, FDBS overcome many of the logistical issues associated with blood collection via standard venepuncture. FDBS are minimally invasive, low cost, stable at ambient temperatures and provide the possibility for self-collection. With growing evidence to support the use of DBS in SARS-CoV-2 serology, the adoption of DBS sampling would greatly benefit large population-based studies and the collection of samples in resource-limited settings. The acceptance of DBS as a viable sample collection mechanism would also aid in future pandemic preparedness.

Methods

Ethics and study participants. Participant samples (Supplementary Table S1) used in this work were acquired from two study protocols. Paired serum and DSS were obtained from the Clinical Characterisation Protocol UK (CCP-UK) study which is a part of the International Severe Acute Respiratory and Emerging Infections Consortium (ISARIC) supported by the WHO Ethics Review Committee (RPC571 and RPC572, 25 April 2013). 36 paired samples from both acute ($n=27$) and convalescent ($n=9$) hospitalised COVID-19 patients were collected between March 2020 and May 2020. Study participants were confirmed to be SARS-CoV-2 positive by reverse transcription polymerase chain reaction (rtPCR) or were highly suspected cases based on clinical presentation. Acute infection samples were collected within 28 days following the onset of symptoms. Convalescent (CV) samples were collected 28 days or more post symptoms onset. Paired serum, DSS, VDDBS and FDBS were obtained from the Human Immune Responses to Acute Virus Infections study (AVIS) (Rec reference 16/NW/0160) at The University of Liverpool (protocol number UoL001207). 26 paired samples were collected between June 2020 and June 2022 and were categorized into three sample type groups. The first group were CV samples from recovered participants 28 days or more following SARS-CoV-2 infection ($n=4$). The second group were vaccinated samples from participants 28 days or more post-vaccination with either a first, second or third dose of an available SARS-CoV-2 vaccine in the UK (Moderna/Pfizer/AstraZeneca) ($n=7$). The samples in the third group were from CV and vaccinated participants who either had a SARS-CoV-2 infection before vaccination, experienced a breakthrough infection following vaccination or experienced both ($n=15$). SARS-CoV-2 infection was confirmed by positive rtPCR or by self-reported positive lateral flow test. Informed consent was obtained for all participants across the two study cohorts. Where required all experiments were conducted under accordance with the relevant guidelines and regulations, with experimental protocols approved by the University of Liverpool licensing committees.

Sample collection. Sera from ISARIC CCP-UK study was received on dry ice in 2 mL screw cap tubes and immediately stored at $-80\text{ }^{\circ}\text{C}$. For AVIS, sera isolation was achieved by collecting venous blood in silica tubes (BD Vacutainer[®] SST II Advance Tubes). The venous blood was then left at RT (approximately $20\text{ }^{\circ}\text{C}$) for a minimum of 30 min before being centrifuged at 1860 g for 10 min to isolate serum. Serum was then aliquoted into 2 mL tubes and stored at $-80\text{ }^{\circ}\text{C}$ until further use. Ahead of use in the neutralisation assay, all sera were thawed and heat-inactivated (HI) at $56\text{ }^{\circ}\text{C}$ for 30 min. HI sera were then centrifuged at 9600 g for 10 min and transferred into new tubes and stored at $-80\text{ }^{\circ}\text{C}$. Before HI, 50 μL of selected sera were blotted onto Schleicher & Schuell 903 filter paper cards into pre-printed circle guides (Supplementary Fig. S7a) and left to dry for 30 min at RT, before being placed into plastic zip-lock bags and stored at $-80\text{ }^{\circ}\text{C}$ until further use. For a further selected number of samples, DSS cards were left at RT for 2, 5, 7, 14, 21 and 28 days ahead of storage at $-80\text{ }^{\circ}\text{C}$. VDDBS and FDBS were also processed in the same way. For VDDBS, 15 μL , 25 μL , 50 μL and 100 μL of venous blood, collected in heparin tubes (BD Vacutainer[®] Heparin Tubes), were blotted onto Schleicher & Schuell 903 filter paper cards. For FDBS, participant fingers were pierced with a 2.0 mm lancet (UniStik3 extra AT1012) and massaged

to encourage blood flow. The resulting blood excretions were then blotted onto Schleicher & Schuell 903 filter paper cards.

DSS, VDBS and FDBS elution. DSS, VDBS and FDBS were removed from plastic zip-lock storage bags and brought to RT in a microbiology safety cabinet. Dried sample spots were then extracted using a punch device (Supplementary Fig. S7b) that created 19 mm sample discs. The punch device was disinfected with 70% ethanol before and after use. Sample discs were extracted into single wells of a 12-well plate (CytoOne). 300 μ L of Phosphate-buffered saline (PBS) buffer was then added to each well containing a sample disc. 1 mL of PBS was added to any empty wells within the plate to minimise evaporation. Plates were then wrapped in parafilm and placed at 4 °C overnight. The next day each sample eluate was collected into a 2 mL tube and HI for 30 min at 56 °C. HI samples were then centrifuged at 9600 g for 10 min to pellet any debris and the resultant supernatant was transferred to a new tube and stored at –80 °C.

Cell culture. HEK293T (ATCC® CRL-3216™) cells were cultivated in Dulbecco's modified eagle medium (Invitrogen) and supplemented with 10% heat-treated FCS (Sigma), 2 mM/mL L-glutamine (Invitrogen), 100 U/mL penicillin (Invitrogen) and 100 mg/mL streptomycin (Invitrogen), termed complete DMEM (Thermofisher). HEK293T/ACE-2 (Creative Biogene CSC-RO0641) cells were used to monitor single-round infectious PVP infectivity and in performing sample neutralisation assays. All cells were cultured at 37 °C and 5% CO₂.

PVP production and infection. The ancestral SARS-CoV-2 S glycoprotein (Accession MN908947) was cloned into the pCDNA3.1 expression plasmid (produced by GeneArt Gene Synthesis) and was used in generating PVP stocks via a lentiviral system to generate single-cycle infectious viral particles as previously described^{52–54}. HEK293T cells (5.0×10^5 in each well of a 6-well tissue culture flask) (Corning) were grown in 2.0 mL of complete DMEM overnight. Cells were transfected with 750 ng of the lentiviral luciferase reporter construct, pCSFLW, along with 450 ng of the SARS-CoV-2 S expression plasmid and 500 ng of the lentiviral backbone, p8.91, using cationic polymer transfection reagent (Polyethylenimine) (Polysciences) and in the presence of OptiMEM (Invitrogen). OptiMEM/plasmid mix was removed 16 h post-transfection and 2.0 mL complete DMEM was added with the single-cycle infectious SARS-CoV-2 stock harvested 48 h later, passed through a 0.45 μ M filter, aliquoted and stored at –80 °C. PVP infection was monitored on HEK293T/ACE-2 cells by measuring luciferase activity. 100 μ L of virus stock was used to infect 1.5×10^4 cells/well for 6 h in a white 96-well plate (Corning). Following infection 100 μ L DMEM complete medium was added to each well. 48 h post-infection, media was discarded from the wells and the cells were washed with PBS (Thermofisher), lysed with 30 μ L cell lysis buffer (Promega) and luciferase activity determined utilising the commercially available luciferase assay (Promega) and measured using a BMGLabtech FluoroStar Omega luminometer.

Neutralisation assay. SARS-CoV-2 S enveloped PVP was used in neutralisation assays as previously described⁵⁴. Serum samples from participants were serially diluted twofold with complete DMEM, and 28 μ L serum dilution was incubated with 420 μ L diluted SARS-CoV-2 PVP for 30 min at RT. 200 μ L of virus/serum dilution mix was used to infect HEK293T/ACE-2 cells. The same protocol was followed for dried sample spot eluates (DSS, VDBS and FDBS) with a couple of adaptations. As filter paper eluted samples were already diluted in DPBS following elution from filter paper, only 6 twofold serial dilutions of sample eluates were made in complete DMEM. Then when eluate/DMEM dilution mixes were added to PVP an additional dilution was made by adding 38 μ L of neat eluate directly to 420 μ L diluted SARS-CoV-2 PVP. Eluate/PVP mixes were then incubated for 30 min at RT before 200 μ L of each PVP/eluate mix was used to infect HEK293T/ACE-2 cells. Sera dilutions for DSS, VDBS and FDBS eluates were adjusted accordingly to account for initial dilution in PBS, which was performed to elute the sample from filter paper. As whole blood contains approximately 55% of serum this was also adjusted for when calculating sera dilutions for whole blood samples. Luciferase activity readings of neutralised virus were analysed i) by considering 0% inhibition as the infection values of the virus in the absence of participant sera included in each experiment, ii) by considering 0% inhibition as the infection values of two consecutive high dilutions of participant sera not inhibiting virus entry. The neutralisation activity was defined as the sera dilution that reduced viral infectivity by 50%, 70% or 90% (IC50, IC70 or IC90, respectively).

Statistical analysis. Statistical analyses were performed using GraphPad Prism 8.0 software. Paired sample comparisons were conducted for all data and individual Figures state the corresponding statistical test performed. The normality of data was assessed by running a D'Agostino and Pearson test. The tests performed include simple linear regression, non-parametric t-tests (Wilcoxon test), repeated measures one-way ANOVA tests and non-parametric Friedman one-way ANOVA tests. Alpha levels of 0.05 were used for all tests and significant *P* values < 0.05 were depicted by *. Bland–Altman plots were also performed to evaluate if dried sample (sera/whole blood) spots eluted from filter paper can be used to measure neutralising antibody responses equivalent to those measured for paired sera.

Data availability

The authors welcome requests for access to the data used in this study conducted in response to the COVID-19 outbreak which will be made available by the corresponding author upon reasonable request.

Received: 13 April 2023; Accepted: 4 September 2023

Published online: 11 September 2023

References

- World Health Organization. WHO Coronavirus Disease (COVID-19) Dashboard With Vaccination Data | WHO Coronavirus (COVID-19) Dashboard With Vaccination Data. *World Health Organization* 1–5 <https://covid19.who.int/> (2021).
- Polack, F. P. *et al.* Safety and efficacy of the BNT162b2 mRNA Covid-19 vaccine. *N. Engl. J. Med.* **383**, 2603–2615 (2020).
- Baden, L. R. *et al.* Efficacy and Safety of the mRNA-1273 SARS-CoV-2 vaccine. *N. Engl. J. Med.* **384**, 403–416 (2021).
- Sadoff, J. *et al.* Safety and efficacy of single-Dose Ad26.COV2.S Vaccine against Covid-19. *N. Engl. J. Med.* **384**, 2187–2201 (2021).
- Tao, K. *et al.* The biological and clinical significance of emerging SARS-CoV-2 variants. *Nat. Rev. Genet.* **22**, 757–773 (2021).
- Qiu, M. *et al.* Antibody responses to individual proteins of SARS coronavirus and their neutralization activities. *Microbes Infect.* **7**, 882–889 (2005).
- Cromer, D. *et al.* Neutralising antibody titres as predictors of protection against SARS-CoV-2 variants and the impact of boosting: A meta-analysis. *The Lancet Microbe* **3**, e52–e61 (2022).
- Khoury, D. S. *et al.* Neutralizing antibody levels are highly predictive of immune protection from symptomatic SARS-CoV-2 infection. *Nat. Med.* **27**, 1205–1211 (2021).
- Yuying, L. *et al.* Advances in neutralization assays for SARS-CoV-2. *Scandinavian J. Immunol.* <https://doi.org/10.1111/sji.13088> (2021).
- Hyseni, I. *et al.* Characterisation of SARS-CoV-2 Lentiviral pseudotypes and correlation between pseudotype-based neutralisation assays and live virus-based micro neutralisation assays. *Viruses* **12**, 1011 (2020).
- James, J. *et al.* Comparison of serological assays for the detection of SARS-CoV-2 antibodies. *Viruses* **13**, 713 (2021).
- Sholukh, A. M. *et al.* Evaluation of cell-based and surrogate SARS-CoV-2 neutralization assays. *J. Clin. Microbiol.* **59**, 527–548 (2021).
- Schmidt, F. *et al.* Measuring SARS-CoV-2 neutralizing antibody activity using pseudotyped and chimeric viruses. *J. Exp. Med.* <https://doi.org/10.1084/jem.20201181> (2020).
- McDade, T. W., Williams, S. & Snodgrass, J. J. What a drop can do: Dried blood spots as a minimally invasive method for integrating biomarkers into population-based research. *Demography* **44**, 899–925 (2007).
- Johannessen, A., Trøseid, M. & Calmy, A. Dried blood spots can expand access to virological monitoring of HIV treatment in resource-limited settings. *J. Antimicrob. Chemother.* **64**, 1126–1129 (2009).
- Vojnov, L. *et al.* The performance of using dried blood spot specimens for HIV-1 viral load testing: A systematic review and meta-analysis. *PLoS Med.* **19**, e1004076 (2022).
- Snijdewind, I. J. M. *et al.* Current and future applications of dried blood spots in viral disease management. *Antiviral Res.* **93**, 309–321 (2012).
- Amini, F. *et al.* Reliability of dried blood spot (DBS) cards in antibody measurement: A systematic review. *PLoS ONE* **16**(3), e0248218. <https://doi.org/10.1371/journal.pone.0248218> (2021).
- Meyers, E. *et al.* Diagnostic performance of the SARS-CoV-2 S1RBD IgG ELISA (ImmunoDiagnostics) for the quantitative detection of SARS-CoV-2 antibodies on dried blood spots. *J. Clin. Virol.* **155**, 105270 (2022).
- Meyers, E. *et al.* Comparison of dried blood spots and venous blood for the detection of SARS-CoV-2 antibodies in a population of nursing home residents. *Microbiol. Spectr.* <https://doi.org/10.1128/Spectrum.00178-21> (2021).
- Turgeon, C. T. *et al.* Detection of SARS-CoV-2 IgG antibodies in dried blood spots. *Diagn. Microbiol. Infect. Dis.* **101**, 115425 (2021).
- Brinc, D. *et al.* Evaluation of dried blood spot testing for sars-cov-2 serology using a quantitative commercial assay. *Viruses* **13**, 962 (2021).
- Morley, G. L. *et al.* Sensitive detection of SARS-CoV-2-specific antibodies in dried blood spot samples. *Emerg. Infect. Dis.* **26**, 2970–2973 (2020).
- McDade, T. W. *et al.* High seroprevalence for SARS-CoV-2 among household members of essential workers detected using a dried blood spot assay. *PLoS ONE* **15**, 1–8 (2020).
- Catlett, B. *et al.* Evaluation of serological assays for SARS-CoV-2 antibody testing from dried blood spots collected from cohorts with prior SARS-CoV-2 infection: SARS-CoV-2 detection from dried blood spots. *J. Clin. Virol. Plus* **2**, 10093 (2022).
- Sims, M. D. *et al.* Dried blood spots are a valid alternative to venipuncture for COVID-19 antibody testing. *J. Immunol. Methods* **513**, 113420 (2023).
- Turgeon, C. T. *et al.* Validation of a multiplex flow immunoassay for detection of IgG antibodies against SARS-CoV-2 in dried blood spots. *PLoS ONE* **16**, e0252621 (2021).
- Zava, T. T. & Zava, D. T. Validation of dried blood spot sample modifications to two commercially available COVID-19 IgG antibody immunoassays. *Bioanalysis* **13**, 13–28 (2021).
- Nikiforuk, A. M. *et al.* Erratum for Nikiforuk et al., “Performance of immunoglobulin G Serology on Finger prick capillary dried blood spot samples to detect a SARS-CoV-2 antibody response”. *Microbiol. Spectr.* <https://doi.org/10.1128/spectrum.03376-22> (2022).
- Brown, L. *et al.* Self-sampling of capillary blood for SARS-CoV-2 serology. *Sci. Rep.* **11**, 1–6 (2021).
- Karp, D. G. *et al.* A serological assay to detect SARS-CoV-2 antibodies in at-home collected finger-prick dried blood spots. *Sci. Rep.* **10**, 1–7 (2020).
- Weisser, H. *et al.* Evaluation of dried blood spots as alternative sampling material for serological detection of anti-SARS-CoV-2 antibodies using established ELISAs. *Clin. Chem. Lab. Med.* **59**, 979–985 (2021).
- Wong, M. P. *et al.* Development and implementation of dried blood spot-based COVID-19 serological assays for epidemiologic studies. *Microbiol. Spectr.* <https://doi.org/10.1128/spectrum.02471-21> (2022).
- Beyerl, J. *et al.* A dried blood spot protocol for high throughput analysis of SARS-CoV-2 serology based on the Roche Elecsys anti-N assay. *EBioMedicine* **70**, 103502 (2021).
- Carter, R. E. *et al.* Prevalence of SARS-CoV-2 antibodies in a multistate Academic Medical Center. *Mayo Clin. Proc.* **96**, 1165–1174 (2021).
- Khan, M. *et al.* Simple, sensitive, specific self-sampling assay secures SARS-CoV-2 antibody signals in sero-prevalence and post-vaccine studies. *Sci. Rep.* **12**, 1–11 (2022).
- Snyder, T. *et al.* Serological surveys to estimate cumulative incidence of SARS-CoV-2 infection in adults (Sero-MAss study), Massachusetts, July–August 2020: A mail-based cross-sectional study. *BMJ Open* **11**, e051157 (2021).
- Styer, L. M. *et al.* High-throughput multiplex SARS-CoV-2 IgG microsphere immunoassay for dried blood spots: A public health strategy for enhanced serosurvey capacity. *Microbiol. Spectr.* <https://doi.org/10.1128/Spectrum.00134-21> (2021).
- Danh, K. *et al.* Detection of neutralizing antibodies against multiple SARS-CoV-2 strains in dried blood spots using cell-free PCR. *Nat. Commun.* **13**, 4–11 (2022).
- Itell, H. L. *et al.* SARS-CoV-2 Antibody Binding and Neutralization in Dried Blood Spot Eluates and Paired Plasma. *Microbiol. Spectr.* <https://doi.org/10.1128/Spectrum.01298-21> (2021).
- Sanclio, A. E. *et al.* A surrogate virus neutralization test to quantify antibody-mediated inhibition of SARS-CoV-2 in finger stick dried blood spot samples. *Sci. Rep.* **11**, 1–6 (2021).
- Ayele, W. *et al.* Use of dried spots of whole blood, plasma, and mother’s milk collected on filter paper for measurement of human immunodeficiency virus type 1 burden. *J. Clin. Microbiol.* **45**, 891–896 (2007).

43. Martin Bland, J. & Altman, D. Statistical methods for assessing agreement between two methods of clinical measurement. *Lancet* **327**, 307–310 (1986).
44. Gruell, H. *et al.* Antibody-mediated neutralization of SARS-CoV-2. *Immunity* **55**, 925–944 (2022).
45. Moat, S. J. *et al.* Development of a high-throughput SARS-CoV-2 antibody testing pathway using dried blood spot specimens. *Ann. Clin. Biochem. Int. J. Lab. Med.* **58**, 123–131 (2021).
46. Anderson, M. *et al.* Expanding access to SARS-CoV-2 IgG and IgM serologic testing using fingerstick whole blood, plasma, and rapid lateral flow assays. *J. Clin. Virol.* **141**, 104855 (2021).
47. Mei, J. V., Alexander, J. R., Adam, B. W. & Hannon, W. H. Use of filter paper for the collection and analysis of human whole blood specimens. *J. Nutr.* **131**, 1631–1636 (2001).
48. Protti, M., Mandrioli, R. & Mercolini, L. Quantitative microsampling for bioanalytical applications related to the SARS-CoV-2 pandemic: Usefulness, benefits and pitfalls. *J. Pharm. Biomed. Anal.* **191**, 113597 (2020).
49. Campbell, C. *et al.* Quantitative serology for SARS-CoV-2 using self-collected saliva and finger-stick blood. *Sci. Rep.* **12**, 6560 (2022).
50. Garcia-Beltran, W. F. *et al.* Remote fingerstick blood collection for severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2) antibody testing. *Arch. Pathol. Lab. Med.* **145**, 415–418 (2021).
51. Marchand, A., Roulland, L., Semence, F., Beck, O. & Ericsson, M. Use of quantitative dried blood spots to evaluate the post-vaccination level of neutralizing antibodies against sars-cov-2. *Life* **11**, 1125 (2021).
52. Genova, C. *et al.* Production, titration, neutralisation, storage and lyophilisation of severe acute respiratory syndrome coronavirus 2 (sars-cov-2) lentiviral pseudotypes. *bio-protocol* <https://doi.org/10.21769/BioProtoc.4236> (2021).
53. Carnell, G., Grehan, K., Ferrara, E., Molesti, E. & Temperton, N. An optimized method for the production using PEI, titration and neutralization of SARS-CoV spike luciferase pseudotypes. *Bio-Protoc.* **7**, 1–10 (2017).
54. Parker, E. *et al.* SARS-CoV-2 antibody responses associate with sex, age and disease severity in previously uninfected people admitted to hospital with COVID-19: An ISARIC4C prospective study. *Front. Immunol.* **14**, 1189 (2023).

Acknowledgements

We thank the study participants and their families for their contributions. We thank all ISARIC4C Consortium Investigators. We also thank all the frontline NHS clinical and research staff as well as volunteer medical students, who collected the clinical data in challenging circumstances and aided in the collection of the study specimens. This work uses data provided by patients and collected by the NHS as part of their care and support #DataSavesLives. We would also like to thank Dr Andrew Owens for providing the HEK293T/ACE-2 cells. W.A. was funded under a Saudi Arabia grant (1074237684) in association with the Department Clinical Laboratory Sciences, College of Applied Medical Sciences, Sakakah 72388, the University of Aljouf, Saudi Arabia. A.A. was funded under a Saudi Arabia grant (NJU246) in association with the Department of Medical Microbiology, College of Applied Medical Sciences, Najran University, Najran, Saudi Arabia. F.B. was funded by a Wellcome Trust PhD studentship (102172/B/13/Z). L.T. was supported by a Wellcome Trust fellowship (205228/Z/16/Z). L.T. is also supported by the U.S. Food and Drug Administration Medical Countermeasures Initiative contract (75F40120C00085). L.T. and M.S. are supported by the National Institute for Health Research Health Protection Research Unit (HPRU) in Emerging and Zoonotic Infections (NIHR200907) at University of Liverpool in partnership with Public Health England (PHE), in collaboration with Liverpool School of Tropical Medicine and the University of Oxford. L.T. and M.S. are based at University of Liverpool. The views expressed in this article are those of the author(s) and not necessarily those of the NHS, UK Health Security Agency, UKRI, NIHR, or the Department of Health and Social Care. For the purpose of Open Access, the author has applied a CC BY public copyright licence to any Author Accepted Manuscript version arising from this submission. F.B. was funded by a Wellcome Trust PhD studentship (102172/B/13/Z).

Author contributions

G.P. and W.A.P. conceived the concept and the experiments. K.J.R., J.T., E.M., W.A., A.A., F.G.B. and M.B. conducted the experiments. L.T. and M.S. provided participant specimens from the AVIS and ISARIC4C protocols respectively. S.D. administered AVIS participant specimens. S.D. and A.H. curated participant clinical data. KR and GP analysed the results. K.J.R. wrote the first draft of the manuscript and K.J.R., J.T., M.G.S., G.P. and W.A.P. edited and revised the manuscript. All authors contributed to the article and approved the submitted version.

Funding

This work was supported by the University of Liverpool. This research was funded in part by the National Institute for Health and Care Research (CO-CIN-01) and jointly by NIHR and UK Research and Innovation (CV220-169, MC_PC_19059).

Competing interests

The authors declare no competing interests.

Additional information

Supplementary Information The online version contains supplementary material available at <https://doi.org/10.1038/s41598-023-41928-2>.

Correspondence and requests for materials should be addressed to G.P.

Reprints and permissions information is available at www.nature.com/reprints.

Publisher's note Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.



Open Access This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons licence, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons licence, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons licence and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this licence, visit <http://creativecommons.org/licenses/by/4.0/>.

© The Author(s) 2023

ISARIC4C Consortium

Cara Donegan⁴, Rebecca G. Spencer⁴, Shona C. Moore⁴, William A. Paxton^{1,4,46} & Georgios Pollakis^{1,4,46}✉ Hayley Hardwick⁴, Malcolm G. Semple^{1,4,47}, Tom Solomon^{4,48}, Lance C. W. Turtle^{4,49}, Jordan Thomas^{1,4}, Ana da Silva Filipe³, Antonia Ying Wai Ho^{3,50}, Sarah E. McDonald³, Massimo Palmarini³, David L. Robertson³, Janet T. Scott^{3,50}, Emma C. Thomson³, Meera Chand⁵, Richard S. Tedder^{51,6}, Nicholas Price^{7,52}, Gary Leeming⁸, Debby Bogaert⁸, Clark D. Russell⁸, Annemarie B. Docherty^{9,53}, Thomas M. Drake⁹, Cameron J. Fairfield⁹, Ewen M. Harrison⁹, Stephen R. Knight⁹, Kenneth A. Mclean⁹, Derek Murphy⁹, Lisa Norman⁹, Riinu Pius⁹, Catherine A. Shaw⁹, Peter W. Horby¹⁰, Olivia V. Swann¹¹, Kanta Chechi^{12,27}, A. A. Roger Thompson¹³, Graham S. Cooke¹⁴, Shiranee Sriskandan^{14,43}, Charlotte Summers¹⁵, William Greenhalf¹⁶, Vanessa Sancho-Shimizu¹⁷, Saye Khoo¹⁸, Mahdad Noursadeghi¹⁹, David Stuart²⁰, Lucy Norris²¹, Rishi K. Gupta²², Andrew Rambaut²³, Christopher A. Green²⁴, Victoria Shaw²⁵, Seán Keating²⁶, Gail Carson²⁷, Laura Merson²⁷, Louise Sigfrid²⁷, James Lee²⁸, Daniel Plotkin²⁸, Marie Connor²⁹, Jo Dalton²⁹, Chloe Donohue²⁹, Carrol Gamble²⁹, Michelle Girvan²⁹, Sophie Halpin²⁹, Janet Harrison²⁹, Clare Jackson²⁹, Laura Marsh²⁹, Stephanie Roberts²⁹, Egle Saviciute²⁹, Tom Fletcher³⁰, Carlo Palmieri^{31,54}, Alison M. Meynert³², Murray Wham³², Peter J. M. Openshaw^{33,55}, Ryan S. Thwaites³³, Jake Dunning^{34,56}, Maria Zamboni³⁴, Gonçalo dos Santos Correia^{35,57}, Matthew R. Lewis^{35,57}, Lynn Maslen^{35,57}, Caroline J. Sands^{35,57}, Panteleimon Takis^{35,57}, Wei Shen Lim³⁶, Alexander J. Mentzer^{34,37}, Paul Klenerman^{37,58}, Tassos Grammatikopoulos^{38,59}, Susan Knight³⁹, Sarah Tait³⁹, J. Kenneth Baillie^{40,60}, Sara Clohisey⁴⁰, Fiona Griffiths⁴⁰, Ross Hendry⁴⁰, Andrew Law^{40,44}, Wilna Oosthuizen⁴⁰, Beatrice Alex⁴¹, Benjamin Bach⁴¹, James Scott-Brown⁴¹, Petros Andrikopoulos^{42,61}, Marc-Emmanuel Dumas^{42,61}, Julian L. Griffin^{42,61}, Sonia Liggi^{42,61}, Michael Olanipekun^{42,61}, Anthonia Osagie^{42,61}, Zoltan Takats^{42,61}, Wendy S. Barclay⁴², Thushan de Silva⁴³, Samreen Ijaz⁴⁵

⁵Antimicrobial Resistance and Hospital Acquired Infection Department, Public Health England, London, UK. ⁶Blood Borne Virus Unit, Virus Reference Department, National Infection Service, Public Health England, London, UK. ⁷Department of Infectious Diseases, Centre for Clinical Infection and Diagnostics Research, School of Immunology and Microbial Sciences, King's College London, London, UK. ⁸Division of Informatics, Imaging and Data Science, Faculty of Biology, Medicine and Health, Centre for Health Informatics, School of Health Sciences, Manchester Academic Health Science Centre, University of Manchester, Manchester, UK. ⁹Centre for Medical Informatics, The Usher Institute, University of Edinburgh, Edinburgh, UK. ¹⁰Nuffield Department of Medicine, Centre for Tropical Medicine and Global Health, University of Oxford, Old Road Campus, Roosevelt Drive, Oxford, UK. ¹¹Department of Child Life and Health, University of Edinburgh, Edinburgh, UK. ¹²Department of Epidemiology and Biostatistics, Faculty of Medicine, School of Public Health, Imperial College London, London, UK. ¹³Department of Infection, Immunity and Cardiovascular Disease, University of Sheffield, Sheffield, UK. ¹⁴Department of Infectious Disease, Imperial College London, London, UK. ¹⁵Department of Medicine, University of Cambridge, Cambridge, Cambridgeshire, UK. ¹⁶Department of Molecular and Clinical Cancer Medicine, University of Liverpool, Liverpool, UK. ¹⁷Department of Pediatrics and Virology, St Mary's Medical School Bldg, Imperial College London, London, UK. ¹⁸Department of Pharmacology, University of Liverpool, Liverpool, UK. ¹⁹Division of Infection and Immunity, University College London, London, UK. ²⁰Division of Structural Biology, The Wellcome Centre for Human Genetics, University of Oxford, Headington, Oxford OX3 7BN, UK. ²¹EPCC, University of Edinburgh, Edinburgh, UK. ²²Institute for Global Health, University College London, London, UK. ²³Institute of Evolutionary Biology, University of Edinburgh, Edinburgh, UK. ²⁴Institute of Microbiology and Infection, University of Birmingham, Birmingham, UK. ²⁵Institute of Translational Medicine, University of Liverpool, Liverpool, Merseyside, UK. ²⁶Intensive Care Unit, Royal Infirmary Edinburgh, Edinburgh, UK. ²⁷Nuffield Department of Medicine, ISARIC Global Support Centre, Centre for Tropical Medicine and Global Health, University of Oxford, Oxford, UK. ²⁸ISARIC, Global Support Centre, COVID-19 Clinical Research Resources, Epidemic Diseases Research Group, Oxford (ERGO), University of

Oxford, Oxford, UK. ²⁹Liverpool Clinical Trials Centre, University of Liverpool, Liverpool, UK. ³⁰Liverpool School of Tropical Medicine, Liverpool, UK. ³¹Molecular and Clinical Cancer Medicine, Institute of Systems, Molecular and Integrative Biology, University of Liverpool, Liverpool, UK. ³²MRC Human Genetics Unit, MRC Institute of Genetics and Molecular Medicine, University of Edinburgh, Edinburgh, UK. ³³National Heart and Lung Institute, Imperial College London, London, UK. ³⁴National Infection Service, Public Health England, London, UK. ³⁵Department of Metabolism, Digestion and Reproduction, National Phenome Centre, Imperial College London, London W12 0NN, UK. ³⁶Nottingham University Hospitals NHS Trust, Nottingham, UK. ³⁷Nuffield Department of Medicine, John Radcliffe Hospital, Oxford, UK. ³⁸Paediatric Liver, GI & Nutrition Centre and MowatLabs, King's College Hospital, London, UK. ³⁹Public Health Scotland, Edinburgh, UK. ⁴⁰Roslin Institute, University of Edinburgh, Easter Bush, Edinburgh EH25 9RG, UK. ⁴¹School of Informatics, University of Edinburgh, Edinburgh, UK. ⁴²Section of Biomolecular Medicine, Division of Systems Medicine, Department of Metabolism, Digestion and Reproduction, Sir Alexander Fleming Building, Exhibition Rd, London SW7 2AZ, UK. ⁴³Department of Infection, Immunity and Cardiovascular Disease, The Florey Institute for Host-Pathogen Interactions, University of Sheffield, Sheffield, UK. ⁴⁴The Roslin Institute, University of Edinburgh, Edinburgh, UK. ⁴⁵Virology Reference Department, National Infection Service, Public Health England, Colindale Avenue, London, UK. ⁴⁸Walton Centre NHS Foundation Trust, Liverpool, UK. ⁴⁹Tropical and Infectious Disease Unit, Royal Liverpool University Hospital, Liverpool, UK. ⁵⁰Department of Infectious Diseases, Queen Elizabeth University Hospital, Glasgow, UK. ⁵¹NHS Greater Glasgow and Clyde, Glasgow, UK. ⁵²Transfusion Microbiology, National Health Service Blood and Transplant, London, UK. ⁵³Department of Infectious Diseases, Guy's and St Thomas' NHS Foundation Trust, London, UK. ⁵⁴MRC Centre for Molecular Bacteriology and Infection, Imperial College London, London, UK. ⁵⁵Clatterbridge Cancer Centre NHS Foundation Trust, Liverpool L7 8YA, UK. ⁵⁶Imperial College Healthcare NHS Trust: London, London, UK. ⁵⁷Section of Bioanalytical Chemistry, Department of Metabolism, Digestion and Reproduction, Imperial College London, London SW7 2AZ, UK. ⁵⁸Department of Microbiology/Infectious Diseases, John Radcliffe Hospital, Oxford University Hospitals NHS Foundation Trust, Oxford, UK. ⁵⁹Translational Gastroenterology Unit, Nuffield Department of Medicine, University of Oxford, Oxford, UK. ⁶⁰Institute of Liver Studies, King's College London, London, UK. ⁶¹Section of Genomic and Environmental Medicine, Respiratory Division, National Heart and Lung Institute, Guy Scadding Building, Dovehouse St., London SW3 3LY, UK.