ORIGINAL ARTICLE

Comparing handrim biomechanics for treadmill and overground wheelchair propulsion

AM Kwarciak, JT Turner, L Guo and WM Richter

Biomechanics Laboratory, MAX Mobility, LLC, Antioch, TN, USA

Study design: Cross-sectional study.

Objectives: To compare handrim biomechanics recorded during overground propulsion with those recorded during propulsion on a motor-driven treadmill.

Setting: Biomechanics laboratory.

Methods: In all, 28 manual wheelchair users propelled their own wheelchairs, at a self-selected speed, on a low-pile carpet and on a wheelchair accessible treadmill. Handrim biomechanics were recorded with an OptiPush instrumented wheelchair wheel.

Results: Across the two conditions, all handrim biomechanics were found to be similar and highly correlated ($r > 0.85$). Contact angle, peak force, average force and peak axle moment differed by 1.6% or less across the two conditions. Although not significant, power output and cadence tended to be slightly higher for the treadmill condition (3.5 and 3.6%, respectively), owing to limitations in adjusting the treadmill grade.

Conclusion: Based on the results of this study, a motor-driven treadmill can serve as a valid surrogate for overground studies of wheelchair propulsion.

Spinal Cord (2011) 49, 457–462; doi:10.1038/sc.2010.149; published online 2 November 2010

Keywords: wheelchair; biomechanics; propulsion; treadmill; handrim; kinetics

Introduction

Manual wheelchair users are at an increased risk of upper limb pain and pathology owing to the repetitive loads associated with wheelchair propulsion.^{1,2} For this reason, wheelchair propulsion continues to be the focus of a number of research studies.^{3–11} Although these studies share a common goal of helping to reduce the risk of injury to manual wheelchair users, the methods of reaching that goal vary, particularly with respect to the testing environment. In these studies, propulsion testing was conducted overground,^{3–5} on a dynamometer,^{6–8} or on a treadmill.^{9–11} Each of these testing environments offers its own advantage, though with the treadmill and the dynamometer, the edge comes at the cost of typical propulsion. Ideally, all studies of manual wheelchair propulsion would be conducted overground, as it represents the most realistic testing condition. However, overground data collection can be limiting. In order to capture a number of consecutive strokes, a sufficiently long distance is needed. This distance may not be available, or allowable, given the restrictions of the measurement equipment (motion capture, electromyogra-

phy, cardiopulmonary diagnostics, so on). With overground testing, it is also difficult to control experimental conditions such as velocity and power output.¹²

An alternative means of conducting studies of wheelchair propulsion is with a motor-driven treadmill. A treadmill provides a confined and consistent testing environment in which wheel velocity and power output can be controlled. The ability to control parameters that affect testing consistency makes the treadmill useful for studying the effects of propulsion training and other interventions.¹³ To appreciate the similarity of treadmill testing to overground testing, we must study the similarity of the testing results. The only comparative study involving the two conditions, conducted by Stephens and Endsberg,¹⁴ examined recovery patterns during propulsion overground, on a dynamometer and on a pair of treadmills. As most wheelchairs are too wide to fit on a traditional exercise treadmill, the treadmill condition was created by having subjects straddle two independent treadmills running at the same speed. Data from the dynamometer and dual-treadmill conditions were combined into a single condition, making it difficult to conclude what affects, if any, the dual treadmill had on recovery pattern.

The objective of this study was to compare handrim biomechanics recorded during overground propulsion (reference condition) with those recorded for the same subjects, with similar velocity and power output, during treadmill

Correspondence: AM Kwarciak, Biomechanics Laboratory, MAX Mobility, LLC, 5425 Mount View Parkway, Antioch, TN 37013, USA.

E-mail: andrew@max-mobility.com

Received 7 June 2010; revised 3 September 2010; accepted 28 September 2010; published online 2 November 2010

propulsion (experimental condition). We hypothesized that handrim biomechanics on a treadmill would be substantially equivalent to those found while pushing overground.

Materials and methods

Subjects

In all, 31 full-time manual wheelchair users, identified from an internal subject database and by local rehabilitation professionals, were recruited for this study. The criteria for participation included: use of a manual wheelchair as the primary mode of mobility, full function in the upper limbs with no debilitating pain, healthy enough to propel for 5 min at a time and the use of quick-release rear axles. Before enrollment, all subjects provided informed consent. All study documents and test procedures were approved by the Western Institutional Review Board (Olympia, WA, USA).

Instrumentation

Data for this study were collected with an OptiPush instrumented wheelchair wheel (MAX Mobility, LLC, Antioch, TN, USA). The OptiPush measures wheel angle and the three-dimensional forces and moments applied to the handrim during propulsion. Data are sampled at 200 Hz and transmitted via bluetooth to a laptop running the OptiPush software. The software filters forces and moments using a 4th order digital Butterworth filter with a low-pass cutoff frequency of 20 Hz.¹⁵ Filtered kinetics and wheel angle measurements are then used to compute propulsion variables on a stroke-by-stroke basis. Each push stroke (push and recovery phases) is defined by the absolute moment about the axle. The push phase begins when moment about the axle exceeds 1 Nm and the recovery phase begins when moment about the axle drops back within 1 Nm.

Data collection

Data for this study were obtained from a larger study of wheelchair propulsion. The methods listed here are limited to the procedures used to collect the data for this investigation.

The wheels of each subject's wheelchair were removed and replaced by the OptiPush wheel on the right (6.0 kg for a 0.64 m wheel) and an inertia-compensated wheel on the left.

The diameters of the wheels and handrims were chosen to match those used by the subject. Both wheels were fitted with a low-profile pneumatic tire inflated to 100 pounds per square inch. Once the wheels were attached and the OptiPush was initialized, subjects propelled their wheelchairs across a low-pile, commercial carpet with no pad (about 50 m long). Subjects were asked to select a speed that could be maintained for an extended period of time. This was done to help prevent fatigue during treadmill testing. Allowing subjects to propel at a self-selected speed also helped to prevent unwanted changes in normal propulsion technique, which may be affected by speed.¹⁶ Two separate runs were made across the floor, in opposite directions, to account for any surface irregularities. Data from the first 3–4 strokes of each trial (initial acceleration phase¹⁷) and the last 3–4 strokes (deceleration phase) were disregarded. Steady-state data from each trial were combined into a single matrix from which average overground speed, power output and the resulting handrim biomechanics were computed. Table 1 lists the equations for all variables assessed in this study.

Subjects were then loaded onto a wheelchair accessible treadmill (Figure 1). Unlike standard treadmills, this treadmill has a wide belt (1.06 m) and a dynamic safety system. The front of the wheelchair was strapped to the frame of the treadmill using two straps that ride along linear bearings. The straps keep the wheelchair on the treadmill and prevent

Figure 1 Photo of the treadmill test setup.

Table 1 Propulsion variables

Variable	Description (Units)	Equation
Speed	Mean speed during the stroke, (m s^{-1})	$(\theta_{\text{stroke}}/180) \cdot \pi \cdot D/t_{\text{stroke}}$
Cadence	Push frequency, (pushes min^{-1})	$60/t_{\text{stroke}}$
Contact angle	Angle through which wheel rotates when hand is on handrim, (degrees)	Wheel angle at recovery start—wheel angle at push start
Power output	Power generated per stroke, (Watts)	$\Sigma(M_{\text{axle}} \cdot \Delta\theta \cdot \pi/180)/t_{\text{stroke}}$
Peak force	Maximum total force applied to the handrim during the stroke, (Newtons)	$\text{Max}(F_{\text{tot}})$, where $F_{\text{tot}} = \sqrt{(F_x^2 + F_y^2 + F_z^2)}$
Average force	Average total force applied to the handrim during the stroke, (Newtons)	$\text{Mean}(F_{\text{tot}})$
Peak moment	Maximum moment about the axle during the stroke, (Newton-meters)	$\text{Max}(M_{\text{axle}})$

Abbreviations: D , wheel diameter (m); θ , wheel angle (degrees); θ_{stroke} , angle through which the wheel rotates during the stroke (degrees); t_{stroke} , time to complete the stroke (s); $\Delta\theta$, point-to-point change in angle (degrees).

excessive lateral sway; however, they do not impede fore-aft movement of the wheelchair. Subjects were given several minutes to acclimate to pushing on the treadmill. During the acclimation period, the treadmill speed was set to match the subject's average overground speed and a metronome beep was provided to help the subject match treadmill cadence to overground cadence. The grade of the treadmill was increased until the subject's treadmill power output, measured on a stroke-by-stroke basis, matched the average overground power output (The average treadmill grade across all subjects was $0.81 \pm 0.28^\circ$). Once the acclimation period was complete, the treadmill was stopped and subjects were given time to rest. After a few minutes, the treadmill was re-started and subjects were asked to push their wheelchair for approximately 1 min while OptiPush data were recorded from the right side.

Data analysis

Seven propulsion variables, computed from data collected during the overground and treadmill trials, were compared across the two testing conditions (Table 1). These variables were selected because they are common in studies of wheelchair propulsion and have been related to upper limb health.^{1,2,18} For each subject, the variables were averaged across all strokes (minimum of 15 strokes) for each condition. Paired *t*-tests were used to analyze the differences between the mean treadmill value and the mean overground value for each variable, with the null hypothesis that the difference would equal zero. To maximize the sensitivity, no type I error correction was applied and α was set to 0.05. Pearson's product-moment correlation coefficients were also calculated to assess the relationship between the variables calculated from each condition. For visual confirmation, scatter plots of each variable (excluding speed) were created. The plots include a line of equality (slope of 1 and intercept of 0) to evaluate the similarity of the variables. The statistical power of the study was estimated to be 77%, based on an s.d. of 20% and a detectable change of 10%. All statistical analyses were performed using SPSS (SPSS Inc., Chicago, IL, USA) and all plots were made using MATLAB (The MathWorks, Inc., Natick, MA, USA).

We certify that all applicable institutional and governmental regulations concerning the ethical use of human volunteers were followed during the course of this research.

Results

Subjects

A complete set of data were obtained from 28 subjects (24 men, 4 women). Data from three individuals could not be used as the protocol was not followed correctly. This reduced the projected study power to 72%. The remaining subjects ranged in age from 15 to 57 years (34.2 ± 9.5 years) and in duration of injury from 2 to 43 years (14.5 ± 10.4 years). Of the 28 subjects, 20 had paraplegia (T3-L1), 5 had spina bifida (T10-L5), 1 had tetraplegia (C6-7 incomplete), 1 had cerebral palsy and 1 had a spinal lipoma.

Table 2 Comparison of propulsion variables for each condition

Test variable	Overground	Treadmill	Difference (%)	P-value
Speed (m s^{-1})	1.09 (0.23)	1.09 (0.23)	-0.1	0.583
Power output (W)	8.43 (2.32)	8.72 (2.59)	3.5	0.132
Cadence (pushes min^{-1})	50.6 (10.9)	52.4 (11.1)	3.6	0.111
Contact angle (degrees)	86.6 (19.1)	86.2 (18.1)	-0.5	0.803
Peak force (N)	58.1 (11.1)	58.4 (12.4)	0.4	0.853
Average force (N)	38.9 (7.59)	38.9 (8.60)	-0.1	0.959
Peak moment (Nm)	10.9 (2.47)	11.0 (2.98)	1.6	0.460

Values are mean (s.d.).

Table 3 Pearson's correlation coefficients

Test Variable	<i>r</i>	P-value
Speed (m s^{-1})	0.999	<0.001
Power output (W)	0.924	<0.001
Cadence (pushes min^{-1})	0.862	<0.001
Contact angle (deg)	0.883	<0.001
Peak force (N)	0.863	<0.001
Average force (N)	0.851	<0.001
Peak moment (Nm)	0.917	<0.001

Comparison of variables

Table 2 lists the variables calculated for each condition and the percentage differences between them. Most of the variables were nearly identical across the conditions with no greater than 1.6% difference between the averages. Furthermore, none of the differences were found to be statistically significant. Table 3 shows the results of the Pearson's correlation analysis. As expected, the analysis produced high *r*-values and low *P*-values indicating strong correlations between the variables in each condition. The proximity of the variables to each line of equality (Figure 2) also demonstrated the congruencies between the two sets of variables.

Discussion

The results of this study demonstrate the strong similarities in handrim biomechanics between treadmill and overground manual wheelchair propulsion. No statistically significant differences were found between the two conditions. This does not prove equivalence; however, it does not refute it. For this comparison, the most revealing results are the descriptive statistics. Four of the five uncontrolled variables calculated during treadmill propulsion (contact angle, peak force, average force and peak moment) were within 2% of the values recorded during overground propulsion. In addition, every variable was highly correlated between the two conditions ($r > 0.85$) and tightly distributed around the line of equality for the two testing environments (Figure 2).

Compared with previous studies of wheelchair propulsion on a treadmill,^{9,11} the power output measured in this study was 15–65% lower. The lower power output values are attributed to a slower belt speed and/or lesser treadmill

Figure 2 Plots of the variables (top row: cadence, contact angle and peak force; bottom row: average force, peak moment and power output). Each plot includes a line of equality ($y=x$) to show the proximity of the points to being equal across the treadmill and overground conditions.

grade. Given the lower power output values, it is reasonable that the values of contact angle and cadence measured in this study were also lower than in previous studies. On the other hand, average peak force (58.4 N) was within the range of reported values (53.5–58.9 N) for treadmill propulsion.^{9,11}

Although we found no significant differences between the propulsion variables, subtle differences were seen in power output, cadence and peak moment. Power output was a controlled variable that should have been equivalent across the two conditions. Power was adjusted by increasing treadmill grade, which ranged from 0.4° to 1.7° for all subjects, until the power output measured by the OptiPush was similar to the mean power output of the overground trials. The resolution for adjusting grade was 0.1°, which limited our ability to fine tune power output on the treadmill. With greater control over treadmill grade, we are confident that power output could have been adjusted to match the overground condition.

During treadmill propulsion, cadence and peak moment were 3.6 (1.8 pushes min^{-1}) and 1.6% (0.17 Nm) higher, respectively, than during overground propulsion. The increases in both variables were likely due to the increased power output requirement of the treadmill. On average, subjects did not increase their contact angle (−0.5%) or average force (−0.1%), and made a marginal increase in peak force (0.4%). Therefore, to achieve the higher power output on the treadmill, subjects increased their strokes per minute and, to a lesser extent, the peak moment of each stroke. Although the data support this conclusion, we cannot rule

Figure 3 Bland–Altman plot of cadence. The solid line represents the mean difference in cadence between pushing on a treadmill and pushing overground, and the dashed lines define the 95% confidence interval.

out the possibility that subjects altered their technique, particularly cadence, in response to pushing on a treadmill. To help subjects become comfortable on the treadmill we had them push for 2–3 min before data collection. We believe this is an important step, as most wheelchair users have never pushed on a treadmill. Additional time may have helped subjects settle into a lower cadence; however, the acclimation period was purposely limited to avoid subject fatigue.

A supplementary Bland–Altman plot of cadence (Figure 3) was created to further investigate the changes in push frequency. The plot shows a slight trend in the differences between overground and treadmill cadence. Subjects who

pushed more frequently experienced greater changes in cadence across the two conditions. However, given the relatively few number of data points at the extremes, it is difficult to determine whether there is a definitive trend or whether these subjects are outliers. The majority of the subjects pushed at a rate of about 50 strokes min^{-1} . Within this group, there was an even distribution of bias with no apparent trend. The only clear result demonstrated by the plot is a small bias (mean offset from zero) when pushing on the treadmill.

From a mechanics perspective, the two testing conditions should produce similar results. van Ingen Schenau¹⁹ provides an analysis of bipedal locomotion on both treadmill and overground. Assuming that treadmill belt speed remains constant and air resistance is negligible, treadmill locomotion is mechanically similar to overground locomotion. The same should hold true for wheelchair propulsion. The external forces acting on a wheelchair user (gravity and ground reaction force) are the same as those acting on a bipedal walker. By selecting a coordinate system that moves with the belt,¹⁹ wheelchair propulsion on a treadmill can be described with the same model used to describe overground wheelchair propulsion.²⁰

Along with biomechanics, one must consider the advantages and disadvantages of treadmill use. In addition to providing control over speed and power output, treadmills provide the means to study steady-state activity in a limited space. This makes it easier to acquire data from motion capture systems and any tethered or cumbersome devices (for example, a cardiopulmonary monitor). Treadmills also provide unobstructed, pure pushing environments that can be used to develop upper body endurance, as well as proper muscle memory and biomechanics. In this way, treadmills are beneficial to rehabilitation clinics and fitness centers that provide propulsion training and exercise services. On the other hand, treadmills can be expensive, costing over \$10 000. Also, the treadmill cannot adequately represent the changes in surface, speed and direction that occur during typical wheeling. Overground testing represents the most realistic environment for studies of wheelchair propulsion. However, based on the similarities in mechanics and the variables computed in this study, the motor-driven treadmill provides an acceptable substitute when space is limited or when steady-state propulsion is desired. Researchers and clinicians can use these data to support their own use of treadmill testing.

There were several limitations to this study. First, upper body kinematics during overground and treadmill propulsion were not collected, so we could not determine whether there were differences in joint ranges of motion between the conditions. However, based on the strong similarities between the contact angle for the two conditions and the kinematic constraints on the upper extremity during the push, this seems unlikely. And second, we only studied a single power output level. Future work should include additional power output levels, as well as kinematics, metabolics and electromyography to provide a more comprehensive validation of treadmill use for studying wheelchair propulsion.

Conclusions

The motor-driven treadmill provides a controlled environment that is suitable for studying manual wheelchair propulsion. Subjects demonstrated similar handrim biomechanics to those measured during overground propulsion. The values of each propulsion variable were similar and highly correlated across the two testing conditions. To maintain comparable conditions to those overground, proper adjustments must be made to the treadmill to replicate overground speed and power output. Adequate time should also be given to help wheelchair users get accustomed to pushing on a moving belt. With these provisions, the treadmill provides a valid method of measuring wheelchair handrim biomechanics and may be an acceptable surrogate for overground studies when a controlled propulsion environment is desired.

Conflict of interest

We confirm that we have a financial interest in the subject matter and materials discussed in this paper. The authors are all employed by MAX Mobility, LLC, which currently offers a wheelchair-accessible treadmill. By demonstrating the similarity of treadmill propulsion to overground propulsion, we could gain from the potential increase in treadmill sales.

Acknowledgements

We would like to thank Russell Rodriguez for his help with subject recruitment, hardware setup, and data collection. This study was funded by the National Institutes of Health (1 R01 HD053732-01).

References

- 1 Boninger ML, Cooper RA, Baldwin MA, Shimada SD, Koontz AM. Wheelchair pushrim kinetics: body weight and median nerve function. *Arch Phys Med Rehab* 1999; **80**: 910–915.
- 2 Mercer JL, Boninger M, Koontz A, Ren D, Dyson-Hudson T. Shoulder joint kinetics and pathology in manual wheelchair users. *Clin Biomech (Bristol, Avon)* 2006; **21**: 781–789.
- 3 Desroches G, Dumas R, Pradon D, Vaslin P, Lepoutre FX, Cheze L. Upper limb joint dynamics during manual wheelchair propulsion. *Clin Biomech (Bristol, Avon)* 2010; **25**: 299–306.
- 4 Koontz AM, Roche BM, Collinger JL, Cooper RA, Boninger ML. Manual wheelchair propulsion patterns on natural surfaces during start-up propulsion. *Arch Phys Med Rehabil* 2009; **90**: 1916–1923.
- 5 Cowan RE, Nash MS, Collinger JL, Koontz AM, Boninger ML. Impact of surface type, wheelchair weight, and axle position on wheelchair propulsion by novice older adults. *Arch Phys Med Rehabil* 2009; **90**: 1076–1083.
- 6 Rice I, Gagnon D, Gallagher J, Boninger M. Hand rim wheelchair propulsion training using biomechanical real-time visual feedback based on motor learning theory principles. *J Spinal Cord Med* 2010; **33**: 33–42.
- 7 DeGroot KK, Hollingsworth HH, Morgan KA, Morris CL, Gray DB. The influence of verbal training and visual feedback on manual wheelchair. *Disabil Rehabil Assist Technol* 2009; **4**: 86–94.
- 8 Lighthall-Haubert L, Requejo PS, Mulroy SJ, Newsam CJ, Bontrager E, Gronley JK *et al*. Comparison of shoulder muscle electromyographic activity during standard manual wheelchair

- and push-rim activated power assisted wheelchair propulsion in persons with complete tetraplegia. *Arch Phys Med Rehabil* 2009; **90**: 1904–1915.
- 9 Gil-Agudo A, Ama-Espinosa A, Perez-Rizo E, Perez-Nombela S, Crespo-Ruiz B. Shoulder joint kinetics during wheelchair propulsion on a treadmill at two different speeds in spinal cord injury patients. *Spinal Cord* 2010; **48**: 290–296.
 - 10 Bregman DJ, Drongelen SV, Veeger HE. Is effective force application in handrim wheelchair propulsion also efficient? *Clin Biomech (Bristol, Avon.)* 2009; **24**: 13–19.
 - 11 Richter WM, Rodriguez R, Woods KR, Axelson PW. Stroke pattern and handrim biomechanics for level and uphill wheelchair propulsion at self-selected speeds. *Arch Phys Med Rehabil* 2007; **88**: 81–87.
 - 12 Van der Woude LH, Veeger HE, Dallmeijer AJ, Janssen TWJ, Rozendaal LA. Biomechanics and physiology in active manual wheelchair propulsion. *Med Engr Phys* 2001; **23**: 713–733.
 - 13 Haisma JA, Van der Woude LH, Stam HJ, Bergen MP, Sluis TA, Bussmann JB. Physical capacity in wheelchair-dependent persons with a spinal cord injury: a critical review of the literature. *Spinal Cord* 2006; **44**: 642–652.
 - 14 Stephens CL, Engsberg JR. Comparison of overground and treadmill propulsion patterns of manual wheelchair users with tetraplegia. *Disabil Rehabil Assist Technol* 2010; **5**: 420–427.
 - 15 Cooper RA, DiGiovine CP, Boninger ML, Shimada SD, Robertson RN. Frequency Analysis of 3-Dimensional Pushrim Forces and Moments for Manual Wheelchair Propulsion. *Automedica* 1998; **16**: 355–365.
 - 16 De Groot S, Veeger HE, Hollander AP, Van der Woude LH. Effect of wheelchair stroke pattern on mechanical efficiency. *Am J Phys Med Rehabil* 2004; **83**: 640–649.
 - 17 Koontz AM, Cooper RA, Boninger ML, Yang Y, Impink BG, Van der Woude LH. A kinetic analysis of manual wheelchair propulsion during start-up on select indoor and outdoor surfaces. *J Rehabil Res Dev* 2005; **42**: 447–458.
 - 18 Preserving Upper Limb Function in Spinal Cord Injury. *A Clinical Practice Guideline for Health-Care Professionals*. Consortium for Spinal Cord Medicine. Spinal Cord Medicine, Clinical Practice Guideline: Washington, DC, 2005.
 - 19 van Ingen Schenau GJ. Some fundamental aspects of the biomechanics of overground versus treadmill locomotion. *Med Sci Sports Exer* 1980; **12**: 257–261.
 - 20 Cooper RA. A systems approach to the modeling of racing wheelchair propulsion. *J Rehab Res Dev* 1990; **27**: 151–162.