

SCIENTIFIC REPORTS



OPEN

Single-Cell Electrical Phenotyping Enabling the Classification of Mouse Tumor Samples

Yang Zhao^{1,*}, Mei Jiang^{2,*}, Deyong Chen¹, Xiaoting Zhao², Chengcheng Xue¹, Rui Hao¹, Wentao Yue², Junbo Wang¹ & Jian Chen¹

Received: 09 October 2015
Accepted: 14 December 2015
Published: 14 January 2016

Single-cell electrical phenotyping (e.g., specific membrane capacitance (C_m) and cytoplasm conductivity (σ_p)) has long been regarded as potential label-free biophysical markers in tumor status evaluation. However, previous studies only reported the differentiation of tumor cell lines without classifying real tumor samples using cellular electrical properties. In this study, two types of mouse tumor models were constructed by injecting two types of tumor cell lines (A549 and H1299), respectively. Then tumor portions were retrieved for immunohistochemistry studies and single-cell electrical phenotyping based on home-developed microfluidic platforms. Immunohistochemistry results of tumor samples confirmed the adenocarcinoma and large-cell carcinoma characteristics for A549 and H1299 based tumor samples, respectively. Meanwhile, cellular C_m and σ_p were characterized as $2.25 \pm 0.50 \mu\text{F}/\text{cm}^2$ and $0.96 \pm 0.20 \text{ S}/\text{m}$ for A549 based tumor samples ($n_{\text{cell}} = 1336$, Mouse I, II, III) and $1.76 \pm 0.54 \mu\text{F}/\text{cm}^2$ and $1.35 \pm 0.28 \text{ S}/\text{m}$ for H1299 based tumor samples ($n_{\text{cell}} = 1442$, Mouse IV, V, VI). Significant differences in C_m and σ_p were observed between these two types of tumor samples, validating the feasibility of using C_m and σ_p for mouse tumor classification.

Single-cell electrical phenotyping (e.g., the quantification of specific membrane capacitance (C_m) and cytoplasm conductivity (σ_p)) is important for understanding cellular functions and status^{1,2}, enabling the classification of tumor cells^{3–6}, stem cells^{7–10} and blood cells^{11–16}.

In the field of electrical phenotyping of tumor cells, there are mainly two conventional techniques, which are dielectrophoresis and electrorotation^{17,18}. In dielectrophoresis, cells attached to dielectrophoretic electrodes at specific frequencies were counted, and the number of attached cells was then translated to intrinsic electrical properties¹⁹. Based on this technique, electrical property variations of tumor cells with different malignant levels (e.g., normal keratinocytes versus oral squamous cell carcinoma cell lines^{4,20,21}) and different anti-drug capabilities (e.g., K562 cells vs. their multidrug resistant derivatives³, MCF 7 cells vs. their multidrug resistant derivatives³) were reported. However, this technique can only provide electrical properties based on batch testing and cannot quantify electrical properties at the single cell level.

In electrorotation, a rotating electric field is applied to rotate a suspended cell as a result of Maxwell-Wagner polarization^{22,23}. Based on this technique, electrical property variations of tumor cells with different anti-drug capabilities (e.g., K562 cells vs. their multidrug resistant derivatives²⁴) were reported. However, in electrorotation, cell manipulation and positioning in the rotating electric field is time consuming and labor intensive.

With the development of microfluidics featured with typical dimensions in the range of 1–100 μm ^{25,26}, micro electrical impedance spectroscopy was proposed for cell classification at the single cell level^{27–30}. Frazier *et al.* sucked cells on top of electrodes and classified breast³¹ as well as head and neck cells³² with different malignant levels. Sun *et al.* aspirated single cells through a microfluidics based constriction channel (cross sectional area smaller than biological cells) for cellular impedance measurement, enabling classification of breast tumor cells and their multiple drug resistant counterparts³³. However, these classifications can only quantify size-dependent electrical parameters which are heavily dependent on experimental conditions, which cannot indicate intrinsic cellular electrical properties (C_m and σ_p).

¹State Key Laboratory of Transducer Technology, Institute of Electronics, Chinese Academy of Sciences, Beijing, P.R. China, 100190. ²Department of Cellular and Molecular Biology, Beijing Chest Hospital, Capital Medical University, Beijing, P.R. China, 101149. *These authors contributed equally to this work. Correspondence and requests for materials should be addressed to W.Y. (email: yuewt@ccmu.edu.cn) or J.W. (email: jbwang@mail.ie.ac.cn) or J.C. (email: chenjian@mail.ie.ac.cn)




Figure 1. Schematic of this study. (a) mouse tumor formation (subcutaneous injection of lung tumor cells into nude mice). (b) tumor sample retrieval and division into two portions. (c) hematoxylin and eosin staining. (d) sample dissociation and seeding in agar media for purification (removal of fibroblast-like cells). (e) electrical property characterization of single tumor cells with C_m and σ_p quantified.

Recently, we proposed a microfluidic platform enabling the high-throughput quantification of C_m and σ_p by modeling the cellular travelling process within the constriction design^{34,35}. Using this system, electrical variations of paired high- and low-metastatic carcinoma strains, as well as tumor cells with single oncogenes under regulation were reported based on measurement results from hundreds of cells⁶. However, previous studies only reported the differentiation of tumor cell lines without classifying real tumor samples using C_m and σ_p .

To address this issue, in this study, we conducted single-cell electrical phenotyping of two types of mouse tumor samples, which were constructed by injecting two types of tumor cell lines, respectively. Then tumor portions were retrieved for single-cell electrical phenotyping based on the home-developed microfluidic platforms. Significant differences in C_m and σ_p were observed between these two types of tumor samples, validating the feasibility of using C_m and σ_p for mouse tumor classification.

Materials and Methods

Materials. Materials used for isolation of solid tumor samples include CytoSelect™ Clonogenic Tumor Cell Isolation Kit (Cell Biolabs, Inc. San Diego, CA, USA) and collagenase II (Sigma, St. Louis, MO, USA). All cell-culture reagents were purchased from Life Technologies Corporation (Carlsbad, CA, USA) unless otherwise specified. The materials used for the fabrication of microfluidic devices were SU-8 photoresist (MicroChem Corp, Newton, MA, USA) and 184 silicone elastomer (Dow Corning Corp., Midland, MI, USA).

Mouse Tumor Formation, Retrieval, and Colony Formation. Female BALB/c-nude mice were obtained from the Experimental Animal Center of NICPBP (China's National Institute for the Control of Pharmaceutical and Biologic Products). Mice were maintained in a pathogen-free environment, with temperature and humidity constant. All procedures involving animals were approved by the Capital Medical University Animal Care and Use Committee and were carried out in accordance with the approved guidelines. Xenograft assays were performed with 3 animals for each cohort and H1299 or A549 cells (2×10^6 in 0.2 mL PBS) were injected subcutaneously. After seven weeks, mice were sacrificed and the tumor portions were excised and retrieved.

Tumors from mice were formalin-fixed and paraffin-embedded. Paraffin sections were dewaxed in xylene and rehydrated with xylene/alcohol (1:1), 100% alcohol, 95% alcohol, 90% alcohol, 80% alcohol, 70% alcohol and 50% alcohol successively. After stained with hematoxylin and eosin, the sections were hydrated with 95% alcohol, 100% alcohol and xylene successively. Images were taken in a Nikon microscope through a $20 \times$ objective.

For the isolation of solid tumor samples, CytoSelect™ Clonogenic Tumor Cell Isolation Kit (Cell Biolabs) was used. Briefly, xenograft tumors were minced with a razor blade and washed several times in PBS with high doses of penicillin/streptomycin to avoid contamination. Tissue dissociation was carried out by enzymatic digestion (0.8 mg/ml collagenase II, Sigma) at 37°C for 4 hours to allow complete cell dissociation with pipetting every 1 hour. Cells were filtered through $40\ \mu\text{m}$ sterile filter and adjusted to 5×10^6 cells/mL. Then cells were incubated 14 days in a proprietary semisolid agar media. Once these cells formed colonies, they were retrieved from the agar media and separated from the single cells by using a $70\ \mu\text{m}$ sterile filter. The viable cells in these colonies were then dissociated into individual cells and seeded in 6-well plates for single-cell electrical phenotyping.

Microfluidic Device Fabrication, Operation and Data Analysis. The two-layer PDMS device (constriction channel cross-section area of $10\ \mu\text{m} \times 10\ \mu\text{m}$) was replicated from a double-layer SU-8 mold, with detailed fabrication procedures described in previous publications^{6,36}. Briefly, the first layer of SU-8 5 was used to form the constriction channel ($10\ \mu\text{m}$) and the second layer of SU-8 25 was used to form the cell loading channel ($25\ \mu\text{m}$). PDMS prepolymer and curing agent were mixed, degassed, poured on channel masters and baked in an




Figure 2. Images of sacrificed mice with the tumor portions marked by arrows. A549 (tumor type I) and H1299 (Tumor Type II) cells were injected in the flank of 6-week-old female BALB/c-nude mice subcutaneously. Tumors were noticed after 2–3 weeks of tumor cell injection and mice were sacrificed in 7 weeks with the tumor portions retrieved.

oven. PDMS channels were then peeled from the SU-8 masters with reservoir holes punched through and bonded to a glass slide.

The whole detailed operation process and data processing was described in previous publications^{6,36}. Briefly, the cell samples were pipetted to the entrance of the cell loading channel of the microfluidic device where a negative pressure at 1 kPa was applied to aspirate cells continuously through the constriction channel with two-frequency impedance data (1 kHz + 100 kHz) and images recorded (sampling rate: 20 points per sec, experimental duration: 750 sec as one experiment, throughput: ~1 cell per sec). Raw impedance data were translated to impedance data with cells at 1 kHz and 100 kHz, which were used to evaluate the sealing properties of deformed cells with constriction channel walls and equivalent cellular membrane capacitance as well as cytoplasm resistance, respectively. By combining cell elongation length during its traveling process within the constriction channel based on image processing, these impedance data were further translated to C_m and σ_p .




Figure 3. Immunohistochemistry results of xenograft tumor samples. (a–c) Tumor samples from three mice injected with A549 cells are featured with adenocarcinoma characteristics (glandular cavities) since A549 is an adenocarcinoma cell line. (d–f) Tumor samples from three mice injected with H1299 cells are featured with large cell carcinoma characteristics (polygonal-shaped cells) since H1299 is a large cell neuroendocrine cell line.




Figure 4. Time-sequence tumor colony formation, tumor cluster retrieve and seeded in culture plates. Tumor colony formation for A549 based tumor samples (a,b) and H1299 based tumor samples (e,f). (c,g) show the retrieved tumor clusters collected from A549 and H1299 based tumor samples, respectively. These tumor colonies were further dissociated into individual cells and seeded in 6-well plates for single-cell electrical phenotyping (d,h).

Results and Discussion

Currently, the golden-standard approach for tumor classification is the immunohistochemistry of tumor samples where morphology data obtained from hematoxylin and eosin staining were used for tumor status evaluation. However, this is a qualitative approach which requests extensive personnel training to make sound decisions. In this study, we explored the feasibility of using electrical parameters including C_m and σ_p to classify tumor samples, which may provide a quantitative approach for cellular status evaluation.

As shown in Fig. 1, in this study, initially, two human lung tumor cell lines (A549 and H1299) were injected subcutaneously into nude mice, respectively, to form solid tumors (see Fig. 1(a)). Then tumor samples were excised and divided into two portions (see Fig. 1(b)). For the immunohistochemistry assay, tumor samples were formalin-fixed and stained by hematoxylin and eosin (see Fig. 1(c)). In the meanwhile, xenograft tumor samples were dissociated by enzymatic digestion and seeded in agar media to remove fibroblast like cells (see Fig. 1(d)). Purified tumor clones were then retrieved from agar media to form suspended single cells, which were then flushed into the constriction channel based microfluidic platform with C_m and σ_p quantified (see Fig. 1(e)).

After the injection of lung tumor cell lines into nude mice subcutaneously, tumor formation was noticed within two weeks and mice appeared weight loss seven weeks later. Then six mice in total (three mice injected with A549 cells and three mice injected with H1299 cells) were sacrificed with tumor samples retrieved. As shown in Fig. 2, three mice injected with A549 cells and three mice injected with H1299 cells were sacrificed with tumor portions marked with arrows.

For each retrieved tumor sample, it was divided into two portions. One portion of the tumor sample was formalin-fixed and paraffin-embedded. After stained with hematoxylin and eosin, the images were taken for




Figure 5. Impedance amplitude and phase measurement at 1 kHz and 100 kHz, simultaneously as well as cellular images squeezing through the constriction channels for A549 (a) and H1299 (b) based tumor cells, respectively. During the cell squeezing process, there is an amplitude increase and phase decrease for data at both 1 kHz and 100 kHz. Basal impedance amplitudes (no cell entry) at 1 kHz were lower than the values at 100 kHz and there were higher impedance amplitude increases during the cellular squeezing process at 1 kHz than those of 100 kHz. As to the phase data, basal phase values (no cell entry) at 1 kHz were almost zero degree, which were higher than the values at 100 kHz. There were higher phase changes during the cellular squeezing process at 100 kHz than 1 kHz.




Figure 6. C_m and σ_p for A549 (a–c) and H1299 (d–f) based tumor samples, respectively. When a cross line ($C_m = 2.0 \mu\text{F}/\text{cm}^2$ and $\sigma_p = 1.2 \text{ S/m}$) was drawn to split the scatter plots, electrical properties of A549 and H1299 based tumor samples fall within the upper left domain and the lower right domain, respectively. These results confirm the classification of mouse tumor samples based on C_m and σ_p .

tumor type classification. Figure 3(a–f) represent tumor samples from three mice injected with A549 cells or H1299 cells, respectively. Significant differences in the hematoxylin and eosin staining were located, due to the pathological difference of A549 and H1299 cells. A549 is an adenocarcinoma cell line and therefore adenocarcinoma characteristics (i.e., glandular cavities) were observed in corresponding xenograft tumor samples (see Fig. 3(a–c)) while H1299 is a large cell neuroendocrine cell line and therefore large cell carcinoma characteristics (i.e., polygonal-shaped cells) were located in corresponding xenograft tumor samples (see Fig. 3(d–f)). The consistency between the injected tumor cell types and morphologies of xenograft tumor samples confirmed the successful formation of tumor samples. Note that although a significant difference in cellular morphologies between these two types of tumors can be obtained, this approach is qualitative and cannot provide quantitative data.

The second portion for each retrieved tumor sample was dissociated into individual cells and cultured in semisolid agar media to form tumor colonies. This approach is a well-established method, capable of forming pure tumor colonies by removing impurities such as tumor-associated fibroblasts which cannot effectively proliferate in the environment of semisolid agar media (CytoSelect™ Clonogenic Tumor Cell Isolation Kit, Cell Biolabs). As shown in Fig. 4, it is a time-sequence proliferation of tumor cells to form the tumor colonies (A549 based tumor colony formation for Fig. 4(a,b) and H1299 based tumor colony formation for Fig. 4(e,f)). Figure 4(c,g) show the retrieved tumor clusters from A549 and H1299 based tumor samples, respectively. These tumor colonies were further dissociated into individual cells and seeded in 6-well plates for single-cell electrical phenotyping (see Fig. 4(d,h)).

Single-cell electrical phenotyping was realized by a home-developed microfluidic platform where single-cell suspensions were pipetted to the entrance of the microfluidic device and a negative pressure was applied to aspirate cells continuously through the constriction channel (a channel with a cross-sectional area smaller than a cell) with two-frequency impedance data (1 kHz + 100 kHz) and images recorded. Note that 1 kHz impedance data was used to evaluate the cell-channel leakage and 100 kHz impedance data was used to quantify membrane capacitance and cytoplasm resistance.

Figure 5(a,b) show the impedance amplitude and phase measurement at 1 kHz and 100 kHz, simultaneously as well as images of cellular squeezing through the constriction channels for A549 and H1299 based tumor cells, respectively. During the cell squeezing process, there was an amplitude increase and a phase decrease for impedance data at both 1 kHz and 100 kHz. Basal impedance amplitudes (no cell entry) at 1 kHz were lower than the values at 100 kHz and there were higher impedance amplitude increases during the cellular squeezing process at 1 kHz than those of 100 kHz. As to the phase data, basal phase values (no cell entry) at 1 kHz were almost zero degree, which were higher than the values at 100 kHz. There were higher phase changes during the cellular squeezing process at 100 kHz than 1 kHz.

Based on previously developed electrical models^{34,35}, raw impedance data were translated to C_m and σ_p , two size-independent intrinsic electrical parameters of single cells (see Fig. 6). For A549 based tumor samples, C_m and σ_p were quantified as $2.25 \pm 0.54 \mu\text{F}/\text{cm}^2$ and $0.88 \pm 0.17 \text{ S/m}$ ($n_{\text{cell}} = 415$, Mouse I), $2.30 \pm 0.52 \mu\text{F}/\text{cm}^2$ and $0.89 \pm 0.15 \text{ S/m}$ ($n_{\text{cell}} = 440$, Mouse II) as well as $2.22 \pm 0.44 \mu\text{F}/\text{cm}^2$ and $1.11 \pm 0.19 \text{ S/m}$ ($n_{\text{cell}} = 481$, Mouse III). Compared to previous data where C_m and σ_p of pure A549 cell lines were characterized as $2.00 \pm 0.60 \mu\text{F}/\text{cm}^2$ and

0.73 ± 0.17 S/m ($n_{\text{cell}} = 487$)⁶, for retrieved cells from A549 based tumors, both increases in specific membrane capacitance and cytoplasm conductivity were observed. For H1299 based tumor samples, C_m and σ_p were quantified as 1.76 ± 0.51 $\mu\text{F}/\text{cm}^2$ and 1.34 ± 0.30 S/m ($n_{\text{cell}} = 526$, Mouse IV), 1.69 ± 0.53 $\mu\text{F}/\text{cm}^2$ and 1.42 ± 0.27 S/m ($n_{\text{cell}} = 410$, Mouse V) as well as 1.81 ± 0.57 $\mu\text{F}/\text{cm}^2$ and 1.30 ± 0.24 S/m ($n_{\text{cell}} = 506$, Mouse VI). When the data were compared to previous data of pure H1299 cell lines (1.63 ± 0.52 $\mu\text{F}/\text{cm}^2$ and 0.90 ± 0.19 S/m $n_{\text{cell}} = 489$)⁶, only significant increase in cytoplasm conductivity was observed for retrieved H1299 based tumor cells.

Compared to A549 based tumor samples (C_m of 2.25 ± 0.50 $\mu\text{F}/\text{cm}^2$ and σ_p of 0.96 ± 0.20 S/m ($n_{\text{cell}} = 1336$, Mouse I, II, III)), H1299 based tumor samples demonstrated lower C_m and higher σ_p (C_m of 1.76 ± 0.54 $\mu\text{F}/\text{cm}^2$ and σ_p of 1.35 ± 0.28 S/m ($n_{\text{cell}} = 1442$, Mouse IV, V, VI)). When a cross line ($C_m = 2.0$ $\mu\text{F}/\text{cm}^2$ and $\sigma_p = 1.2$ S/m) was drawn to split the scatter plots, electrical properties of A549 and H1299 based tumor samples fall within the upper left domain and the lower right domain, respectively. These results confirm the classification of mouse tumor samples based on C_m and σ_p .

Conclusions

In this paper, electrical property differences were located for two types of tumor samples from three mice injected with A549 cells or H1299 cells, respectively, confirming the feasibility of tumor cell classification based on cellular electrical properties. Future work will focus on electrical property characterization of human tumor samples, with the purpose of investigating the feasibility of human tumor sample classification using C_m and σ_p .

References

- Morgan, H., Sun, T., Holmes, D., Gawad, S. & Green, N. G. Single cell dielectric spectroscopy. *J. Phys. D: Appl. Phys.* **40**, 61–70 (2007).
- Valero, A., Braschler, T. & Renaud, P. A unified approach to dielectric single cell analysis: impedance and dielectrophoretic force spectroscopy. *Lab Chip* **10**, 2216–2225 (2010).
- Labeed, F. H., Coley, H. M., Thomas, H. & Hughes, M. P. Assessment of multidrug resistance reversal using dielectrophoresis and flow cytometry. *Biophys. J.* **85**, 2028–2034 (2003).
- Liang, X., Graham, K. A., Johannessen, A. C., Costea, D. E. & Labeed, F. H. Human oral cancer cells with increasing tumorigenic abilities exhibit higher effective membrane capacitance. *Integr. Biol.* **6**, 545–554 (2014).
- Coley, H. M., Labeed, F. H., Thomas, H. & Hughes, M. P. Biophysical characterization of MDR breast cancer cell lines reveals the cytoplasm is critical in determining drug sensitivity. *Biochim. Biophys. Acta* **1770**, 601–608 (2007).
- Zhao, Y. *et al.* Tumor cell characterization and classification based on cellular specific membrane capacitance and cytoplasm conductivity. *Biosens. Bioelectron.* **57**, 245–253 (2014).
- Song, H. *et al.* A microfluidic impedance flow cytometer for identification of differentiation state of stem cells. *Lab Chip* **13**, 2300–2310 (2013).
- Flanagan, L. A. *et al.* Unique dielectric properties distinguish stem cells and their differentiated progeny. *Stem Cells* **26**, 656–665 (2008).
- Labeed, F. H. *et al.* Biophysical characteristics reveal neural stem cell differentiation potential. *PLoS One* **6**, e25458 (2011).
- Adams, T. N. G., Turner, P. A., Janorkar, A. V., Zhao, F. & Minerick, A. R. Characterizing the dielectric properties of human mesenchymal stem cells and the effects of charged elastin-like polypeptide copolymer treatment. *Biomicrofluidics* **8**, 054109 (2014).
- Watkins, N. N. *et al.* Microfluidic CD4⁺ and CD8⁺ T lymphocyte counters for point-of-care HIV diagnostics using whole blood. *Sci Transl Med* **5**, 214ra170 (2013).
- Han, X., van Berkel, C., Gwyer, J., Capretto, L. & Morgan, H. Microfluidic lysis of human blood for leukocyte analysis using single cell impedance cytometry. *Anal Chem* **84**, 1070–1075 (2012).
- Du, E. *et al.* Electric impedance microflow cytometry for characterization of cell disease states. *Lab Chip* **13**, 3903–3909 (2013).
- Holmes, D. *et al.* Leukocyte analysis and differentiation using high speed microfluidic single cell impedance cytometry. *Lab Chip* **9**, 2881–2889 (2009).
- Holmes, D. & Morgan, H. Single cell impedance cytometry for identification and counting of CD4 T-cells in human blood using impedance labels. *Anal Chem* **82**, 1455–1461 (2010).
- Vykoukal, D. M., Gascoyne, P. R. & Vykoukal, J. Dielectric characterization of complete mononuclear and polymorphonuclear blood cell subpopulations for label-free discrimination. *Integr. Biol.* **1**, 477–484 (2009).
- Gimsa, J. A comprehensive approach to electro-orientation, electrodeformation, dielectrophoresis, and electrorotation of ellipsoidal particles and biological cells. *Bioelectrochemistry* **54**, 23–31 (2001).
- Jones, T. B. Basic theory of dielectrophoresis and electrorotation. *IEEE Eng Med Biol Mag* **22**, 33–42 (2003).
- Broche, L. M., Labeed, F. H. & Hughes, M. P. Extraction of dielectric properties of multiple populations from dielectrophoretic collection spectrum data. *Phys Med Biol* **50**, 2267–2274 (2005).
- Broche, L. M. *et al.* Early detection of oral cancer—Is dielectrophoresis the answer? *Oral Oncol* **43**, 199–203 (2007).
- Mulhall, H. J. *et al.* Cancer, pre-cancer and normal oral cells distinguished by dielectrophoresis. *Anal Bioanal Chem* **401**, 2455–2463 (2011).
- Fuhr, G., Hagedorn, R. & Goring, H. Separation of different cell-types by rotating electric-fields. *Plant Cell Physiol* **26**, 1527–1531 (1985).
- Arnold, W. M. & Zimmermann, U. Electro-rotation-development of a technique for dielectric measurements on individual cells and particles. *J Electrostat* **21**, 151–191 (1988).
- Bahrieh, G., Erdem, M., Ozgur, E., Gunduz, U. & Kulah, H. Assessment of effects of multi drug resistance on dielectric properties of K562 leukemic cells using electrorotation. *RSC Adv* **4**, 44879–44887 (2014).
- Whitesides, G. M. The origins and the future of microfluidics. *Nature* **442**, 368–373 (2006).
- Wootton, R. C. & Demello, A. J. Microfluidics: Exploiting elephants in the room. *Nature* **464**, 839–840 (2010).
- Cheung, K. C. *et al.* Microfluidic impedance-based flow cytometry. *Cytometry A* **77**, 648–666 (2010).
- Sun, T. & Morgan, H. Single-cell microfluidic impedance cytometry: a review. *Microfluid Nanofluidics* **8**, 423–443 (2010).
- Zheng, Y., Nguyen, J., Wei, Y. & Sun, Y. Recent advances in microfluidic techniques for single-cell biophysical characterization. *Lab Chip* **13**, 2464–2483 (2013).
- Chen, J. *et al.* Microfluidic impedance flow cytometry enabling high-throughput single-cell electrical property characterization. *Int. J. Mol. Sci.* **16**, 9804–9830 (2015).
- Han, K. H., Han, A. & Frazier, A. B. Microsystems for isolation and electrophysiological analysis of breast cancer cells from blood. *Biosens. Bioelectron.* **21**, 1907–1914 (2006).
- Cho, Y. *et al.* Whole-cell impedance analysis for highly and poorly metastatic cancer cells. *J. Microelectromech. Syst.* **18**, 808–817, (2009).

33. Chen, J. *et al.* Classification of cell types using a microfluidic device for mechanical and electrical measurement on single cells. *Lab Chip* **11**, 3174–3181 (2011).
34. Zhao, Y. *et al.* A microfluidic system enabling continuous characterization of specific membrane capacitance and cytoplasm conductivity of single cells in suspension. *Biosens. Bioelectron.* **43C**, 304–307 (2013).
35. Zhao, Y. *et al.* A microfluidic system for cell type classification based on cellular size-independent electrical properties. *Lab Chip* **13**, 2272–2277, (2013).
36. Huang, S.-B. *et al.* Classification of cells with membrane staining and/or fixation based on cellular specific membrane capacitance and cytoplasm conductivity. *Micromachines* **6**, 163–171 (2015).

Acknowledgements

The authors would like to acknowledge financial support from National Basic Research Program of China (973 Program, Grant No. 2014CB744600), National Natural Science Foundation of China (Grant No. 61201077, 81261120561 and 61431019), Instrument Development Program of Chinese Academy of Sciences and Beijing NOVA Program of Science and Technology.

Author Contributions

Y.Z. and D.C. designed and performed the experiments; M.J., X.Z. and W.Y. constructed the mouse models and cultured the cells; Y.Z., C.X. and R.H. analysed the data. J.W. and J.C. conceived the experiments and drafted the paper.

Additional Information

Competing financial interests: The authors declare no competing financial interests.

How to cite this article: Zhao, Y. *et al.* Single-Cell Electrical Phenotyping Enabling the Classification of Mouse Tumor Samples. *Sci. Rep.* **6**, 19487; doi: 10.1038/srep19487 (2016).



This work is licensed under a Creative Commons Attribution 4.0 International License. The images or other third party material in this article are included in the article's Creative Commons license, unless indicated otherwise in the credit line; if the material is not included under the Creative Commons license, users will need to obtain permission from the license holder to reproduce the material. To view a copy of this license, visit <http://creativecommons.org/licenses/by/4.0/>