SCIENTIFIC REPORTS

OPEN

Signal-off impedimetric immunosensor for the detection of *Escherichia coli* O157:H7

Jingzhuan Wan¹, Junjie Ai¹, Yonghua Zhang¹, Xiaohui Geng¹, Qiang Gao¹ & Zhiliang Cheng²

Received: 12 June 2015
Accepted: 14 December 2015
Published: 22 January 2016

A signal-off impedimetric immune-biosensor based on gold nanoparticle (AuNP)-mediated electron transfer (ET) across a self-assembled monolayer (SAM) was developed for highly sensitive detection of *Escherichia coli* O157:H7 bacteria. The biosensor was fabricated by covalently grafting an anti-*Escherichia coli* O157:H7 antibody onto SAM-modified gold electrodes. Following bacterial capture, the sensor was further modified by the gold nanoparticles (AuNPs). Due to the strong interaction between AuNPs and *Escherichia coli* O157:H7, AuNPs attached to the surface of the bacteria and acted as ET pathways across the insulating SAMs on the electrode surface, resulting in a significant reduction of the electron transfer resistance (Ret) between the $[\text{Fe}(\text{CN})_6]^{3-/4-}$ redox probe in the solution and the substrate gold surface. Therefore, the attachment AuNPs to captured bacteria significantly enhanced the sensitivity for *Escherichia coli* O157:H7 bacteria detection.

Harmful pathogenic bacteria may be naturally present in food and water and can cause a variety of diseases in humans. One of the most common pathogens is *Escherichia coli* (*E. coli*) O157:H7, which causes hemorrhagic diarrhea, renal failure, anemia and other serious health problems¹. Traditional methods for *E. coli* O157:H7 detection include plate culture², enzyme-linked immunosorbent assays (ELISAs)³, and polymerase chain reaction (PCR)⁴. However, these methods are generally tedious and time-consuming⁵. Therefore, a rapid, sensitive and specific method for the detection of *E. coli* O157:H7 is highly necessary^{6,7}.

Impedimetric biosensors have been proven effective for the detection of pathogenic bacteria⁸. These sensors analyze the impedance of bacterial cells when they are attached to or associated with the electrodes. The capture of bacteria at the sensing surfaces can often increase interface impedance. Currently, most impedimetric biosensors are based on a “signal-on” mechanism (i.e., the signal directly correlates with the analyte concentration)⁹. However, the high background impedance from these sensors may affect their performance in several ways, including reduction in sensitivity. Several signal enhancement strategies have been studied to achieve low detection limits¹⁰. For example, Ruan *et al.*¹¹ reported an impedimetric immunosensor for *E. coli* O157:H7 bacteria detection that used secondary antibodies labeled with horseradish peroxidase (HRP) for signal amplification. Specifically, HRP was applied to precipitate insoluble products on the electrode surface. This biosensor was able to detect *E. coli* O157:H7 bacteria with a detection limit of 6×10^3 cells/mL.

Compared with enzyme-labeled assays, the use of nanomaterial tags, especially gold nanoparticles (AuNPs), can offer several benefits in terms of time, cost and simplicity¹². For impedimetric sensors, AuNPs are often used to modify the electrode to improve substrate performance¹³. As previously reported, AuNPs can act as electron-transfer (ET) mediators across the insulating layer on the electrode surface¹⁴. Moreover, ET at the interface of metal NPs/insulator/metal electrode sandwich structure is more efficient than ET between a metal electrode and redox species in solution by several orders of magnitude¹⁵. In this study, we took advantage of the strong interaction between *E. coli* O157:H7 bacteria and AuNPs to develop a signal-off impedimetric immunosensor for the sensitive detection of *E. coli* O157:H7.

Results and Discussion

The development of the signal-off impedimetric immunosensor is shown in Fig. 1. Specifically, mixed self-assembled monolayers of 11-mercaptoundecanoic acid (MUA) and 1-undecanethiol (UDT) formed on the gold electrode in ethanol (EtOH). To chemically conjugate the antibody, 10 mol % COOH-terminated alkanethiol

¹Key Laboratory of Applied Surface and Colloid Chemistry, Ministry of Education, School of Chemistry and Chemical Engineering, Shaanxi Normal University, Xi'an, 710062, China. ²Department of Bioengineering, University of Pennsylvania, Philadelphia, PA 19104, USA. Correspondence and requests for materials should be addressed to Q.G. (email: gaoqiang@snnu.edu.cn) or Z.C. (email: zcheng@seas.upenn.edu)

Figure 1. Schematic of the biosensor preparation and *E. coli* O157:H7 bacteria detection.

MUA was added during the self-assembled monolayer (SAM) preparation. Anti-*E. coli* O157:H7 antibody immunoglobulin G (IgG) was then conjugated onto COOH-terminated SAM-modified gold electrode based on the 1-ethyl-3-[3-dimethylaminopropyl] carbodiimide hydrochloride (EDC)/N-hydroxysulfosuccinimide (NHS) method. Because SAMs are insulators, they block ET between [Fe(CN)₆]^{3-/4-} in an aqueous solution and the substrate gold electrode. Following antibody conjugation, bacteria were captured onto the IgG-immobilized gold electrode. To improve the detection sensitivity, electrodes covered in surface-bound bacteria were exposed to AuNPs, resulting in AuNP-coated bacteria on the electrode surface. This arrangement significantly increased ET between [Fe(CN)₆]^{3-/4-} and the substrate gold electrode because the attached AuNPs provided an electrical pathway for electron transfer across the insulating layer on the electrode surface. The bacteria were specifically and sensitively detected by measuring the electron transfer resistance (Ret) with electrochemical impedance spectroscopy (EIS).

Prior to coating the electrode surface-bound bacteria with AuNPs, the interaction between AuNPs and bacteria in the aqueous solution was examined. Citrate-capped gold nanoparticles with a mean diameter of approximately 13 nm were used to coat the bacteria. The successful bacterial labeling with AuNPs was confirmed by transmission electron microscopy (TEM). As shown in Fig. 2a, the unlabeled bacteria exhibited a spheroid shape (ca. 2000 nm). However, a strong interaction between AuNPs and bacteria was observed when bacteria were incubated with AuNPs. As shown in Fig. 2b, the bacteria were sufficiently coated with the AuNPs. Moreover, the AuNP-labeled bacteria retained their size and shape. The bacteria labeled with AuNPs were further washed several times with phosphate buffer solution (PBS). As shown in **supp 2c**, PBS washing did not desorb the AuNPs from the bacteria, which indicated that the AuNPs and bacteria strongly interacted. To confirm that the coated AuNPs were presented on the surface of the bacteria, i.e., not inside of the bacteria, HAuCl₄ and NH₂OH were added to sample of AuNP-bacteria complexes. As shown in Fig. 2d, the AuNPs were enlarged on the surface of the bacteria. Au nano-plates were also observed around the complexes¹⁶. These results indicated that AuNPs had mainly adsorbed to surfaces of the bacteria.

To determine the consistency and ensure that the ratio of adsorbed AuNPs to bacteria was similar for every run, TEM was used to further characterize the complexes that formed between AuNPs and different concentrations of bacteria. Under these conditions, excess AuNPs were added to bacteria. As shown in Figure S1, each complex exhibited similar amounts of adsorbed AuNPs, indicating that the bacteria concentration did not affect the ratio of AuNPs to bacteria in the presence of excess AuNPs.

According to the design principle of the signal-off impedimetric sensor, a highly insulating SAM on the electrode surface is critical for the sensitive detection of bacteria. Here, an insulating layer was prepared using binary SAMs consisting of MUA and UDT. EIS was used to characterize the modified processes of the impedimetric sensor, and the results are shown in Fig. 3. The EIS data were analyzed by fitting them to an equivalent electrical circuit model (inset of Fig. 3). The circuit includes the electrolyte resistance (Rs) and Warburg impedance (Zw), which results from the diffusion of ions from the bulk electrolyte to the interface, the double-layer capacitance (Cdl), and the Ret. The semicircle of the impedance spectra in the high-frequency region was attributed to the limited ET process, whereas the linear part in the lower-frequency region was attributed to transport limitations of electroactive species due to diffusion to the reaction interface, which is known as Warburg impedance.

Figure 2. Transmission electron microscopy images of (a) *E. coli* O157:H7 bacteria, (b) AuNP-labeled *E. coli* O157:H7 bacteria without washing, (c) AuNP-labeled *E. coli* O157:H7 bacteria washed with PBS, (d) complex after the addition of HAuCl_4 and NH_2OH to AuNP-labeled *E. coli* O157:H7 bacteria.

Figure 3. Nyquist plots of impedance spectra of (a) bare gold electrode, (b) gold electrode featuring a mixed SMA of MUA and UDT, (c) Antibody-conjugated electrode, (d) electrode coated with captured *E. coli* O157:H7 bacteria, (e) electrode coated with AuNP-labeled *E. coli* O157:H7.

The semicircle diameter reflects the R_{et} , which indicates the blocking effect of the electrode surface for the $[\text{Fe}(\text{CN})_6]^{3-/4-}$ redox probe. Thus, R_{et} was used as a sensor signal. The experimental (dots) and fitted results (lines) closely correlated.

The Nyquist plot obtained using the bare gold electrode was almost a straight line and yielded a R_{et} of $57\ \Omega$, which was characteristic of a diffusion-limiting process (curve a and inset). However, a significant semicircle curve appeared in the high-frequency region for SAM-modified gold electrode (curve b). The R_{et} increased to $245\ \text{k}\Omega$, indicating that electron transfer between $[\text{Fe}(\text{CN})_6]^{3-/4-}$ and the electrode was strongly blocked by the SAM on the electrode. Following antibody conjugation, the R_{et} was $254\ \text{k}\Omega$ (curve c). Little to no change in the R_{et} was observed over one week when the antibody-immobilized electrodes were stored in PBS at a temperature of 4°C . This finding indicated that the sensors were highly stable. To determine the ability of these sensors to detect bacteria, antibody-conjugated electrodes were incubated with $5000\ \text{cfu/mL}$ *E. coli* O157:H7. The capture of *E. coli* O157:H7 to immobilized antibody only resulted in a slight increase in the R_{et} to $268\ \text{k}\Omega$ (curve d). This

Figure 4. (A) Nyquist responses of the biosensor to different concentrations of *E. coli* O157:H7 bacteria. (a) 0, (b) 3×10^2 , (c) 5×10^2 , (d) 1×10^3 , (e) 3×10^3 , (f) 5×10^3 , (g) 1×10^4 , (h) 3×10^4 , (i) 5×10^4 , (j) 1×10^5 , (k) 3×10^5 , (l) 5×10^5 and (m) 1×10^6 cfu/mL. (B) Change in Ret as a function of *E. coli* O157:H7 concentration.

slight increase in the Ret could be due to the high background impedance from the SAM-modified electrode, which may have affected the sensitivity and detection limit for bacteria detection. To improve the performance of these sensors, the AuNPs were subsequently added to an electrode that had captured *E. coli* O157:H7. As expected, the attachment of AuNPs to the captured *E. coli* O157:H7 significantly decreased the Ret. Specifically, the Ret was changed from 268 kΩ to 106 kΩ (curve e). Therefore, the signal-off impedimetric sensor coated with AuNPs-labeled bacteria led to a 60% change in the Ret. This change was mainly due to the reduced electron transfer resistance because of *E. coli* O157:H7-adsorbed AuNPs. For comparison purposes, the signal of only the impedimetric sensor generated 6% change in the Ret when non-AuNP-labeled *E. coli* O157:H7 were captured by the antibody-conjugated electrodes. Thus, the attachment of AuNPs to *E. coli* O157:H7 led to a 10-fold signal change in bacterial detection. We also found that the physical adsorption of AuNPs to sensor surfaces without captured bacteria was very low.

To evaluate the effects of adding the AuNPs on the amount of bacteria captured on the surface, optical images of the sensor surface coated with captured bacteria were acquired before and after the addition of the AuNPs. As shown in Figure S2, the density of *E. coli* O157:H7 on the sensor surface did not significantly differ between bacteria incubated with AuNPs and bacteria that had not been incubated with AuNPs, indicating that adding AuNPs did not significantly affect the amount of bacteria captured on the surface.

As shown in Fig. 4A, a decrease in Ret was observed when the target *E. coli* O157:H7 bacteria concentration was increased from 0 to 1×10^6 cfu/mL, which demonstrated that varying amounts of bacteria could be detected by measuring the Ret change. As shown in Fig. 4B, the linear dynamic range was between 300 and 1×10^5 cfu/mL and characterized by linear correlation coefficient of 0.973. Each point in Fig. 4B represents the average Ret obtained with three different sensors, and error bars indicate the standard deviations of these tests. The obtained detection limit of 100 cfu/mL was based on the signal-to-noise ratio ($S/N = 3$).

The specificity of the biosensor for the detection of *E. coli* O157:H7 was investigated by testing non-target microorganisms, such as *E. coli* DH5α, *E. coli* K12, and *Staphylococcus aureus*. These organisms all produced small changes in the Ret (less than 3%), confirming that IgG-conjugated sensors specifically detected changes in the Ret due to *E. coli* O157:H7.

Conclusion

A signal-off impedimetric immunosensor was developed for the detection of bacteria. Due to the strong interaction between AuNPs and bacteria, AuNPs easily attached to sensors that had captured bacteria and acted as ET pathways across the insulating SAMs on the electrode surface, resulting in a significant signal change. These sensors are expected to be utilized for the highly sensitive detection of other microorganisms.

Methods

Chemicals. 11-mercaptoundecanoic acid (MUA), 1-undecanethiol (UDT), 1-ethyl-3-[3-dimethylaminopropyl] carbodiimide hydrochloride (EDC), N-hydroxysulfosuccinimide (NHS) and $\text{HAuCl}_4 \cdot 3\text{H}_2\text{O}$ were obtained from Sigma-Aldrich. Rabbit anti-*E. coli* O157:H7 IgG was purchased from Halin Biotech. Co. (Shanghai, China). The other chemicals and reagents were commercially available and of analytical grade. The water used was obtained from a Millipore Milli-Q purification system.

Gold nanoparticles. Gold nanoparticles (AuNPs) with a diameter of approximately 13 nm were prepared via the citrate reduction of HAuCl_4 as previously described¹⁷.

Bacterial culture and counting. *E. coli* O157:H7 was purchased from the American Type Culture Collection. *E. coli* DH5α, *E. coli* K12, and *Staphylococcus aureus* were obtained from the China Center of Industrial Culture Collection (Beijing, China). All microbial strains were cultivated following the culturing guidelines of the manufactures and previous reports¹⁸. Typically, *E. coli* or *Staphylococcus aureus* cells were cultured in Luria–Bertani culture medium containing 10 g/L tryptone, 5 g/L yeast extract and 10 g/L NaCl. All strains were

grown in nutrient medium in a shaking incubator (125 rpm) at 37°C. After 24 h of culturing, the bacterial cells were isolated by centrifugation (6000 rpm, 20 min) and then rinsed three times with filtered phosphate buffer solution (10 mM, pH 7.4). The number of bacterial cells was determined using the plate-counting method. The bacteria sample was directly diluted with 10 mM PBS to the desired concentrations.

Fabrication of the biosensor. The gold electrode was sequentially polished with 1.0, 0.3, and 0.05 µm alumina slurry, followed by ultrasonic cleaning in ethanol and ultrapure water. A mixed self-assembled monolayer (SAM) was formed by immersing the electrode into the ethanol solution containing MUA and UDT at a molar ratio of 1:9 for 4 h. After washing with ethanol and buffer, the SAM-modified gold electrodes were subsequently treated with 0.05 M NHS and 0.2 M EDC solution to activate the –COOH on the electrode surface. The antibody was immobilized by dropping the antibody solution (10 µL, 1 mg/mL) onto the surface of the gold electrodes and incubating the mixture overnight at 4°C under a humidified atmosphere. The residual activated surface was completely blocked with ethanolamine hydrochloride (1.0 mM, pH 8.5). Non-covalently bound material was removed by washing with PBS.

Measurement of target *E. coli* O157:H7 bacteria. *E. coli* O157:H7 bacteria were captured by incubating the antibody-conjugated sensor with a bacteria sample for 1 h. After free bacteria were removed by washing with buffer, the coated electrode was then incubated in AuNP solution (8 nM) for another 1 h to attach AuNPs onto bacteria captured on the electrode surface.

Electrochemical impedance spectroscopy (EIS) measurement. EIS was performed on a Zennium electrochemical workstation (Zahner, Germany) with a three-electrode cell. A gold electrode (2 mm in diameter) was used as the working electrode. An Ag/AgCl electrode (saturated with KCl) and a platinum wire were used as the reference and counter electrodes, respectively. All potentials in this study are presented in terms of Ag/AgCl/saturated KCl electrode potentials. The Faradaic impedance measurements were performed in the presence of a 5 mM [Fe(CN)₆]^{3/4–} redox probe (equimolecular mixture in pH 7.0, 10 mM PBS containing 0.1 M KCl). The direct current (DC) potential was set to +0.24 V, which is equivalent to the formal potential of the [Fe(CN)₆]^{3/4–} redox probe. The experimental data, presented as Nyquist plots, were fitted to proper equivalent circuits using the software provided by Zahner.

References

- Nataro, J. P. & Kaper, J. B. Diarrheagenic *Escherichia coli*. *Clin. Microbiol. Rev.* **11**, 142–201 (1998).
- March, S. B. & Ratnam, S. J. Sorbitol-MacConkey medium for detection of *Escherichia coli* O157:H7 associated with hemorrhagic colitis. *J. Clin. Microbiol.* **23**, 869–872 (1986).
- Bessoff, K., Delorey, M., Sun, W. & Hunsperger, E. Comparison of two commercially available dengue virus (DENV) NS1 capture enzyme-linked immunosorbent assays using a single clinical sample for diagnosis of acute DENV infection. *Clin. Vaccine Immunol.* **15**, 1513–1518 (2008).
- Dharmasiri, U. *et al.* Enrichment and detection of *Escherichia coli* O157:H7 from water samples using an antibody modified microfluidic chip. *Anal. Chem.* **82**, 2844–2849 (2010).
- Gracias, K. S. & McKillip, J. L. A review of conventional detection and enumeration methods for pathogenic bacteria in food. *Can. J. Microbiol.* **50**, 883–890 (2004).
- Ma, F. *et al.* Glycosylation of quinone-fused polythiophene for reagentless and label-free detection of *E. coli*. *Anal. Chem.* **87**, 1560–1568 (2015).
- Labib, M. *et al.* Aptamer-based viability impedimetric sensor for bacteria. *Anal. Chem.* **84**, 8966–8969 (2012).
- Yang, L. J. & Bashir, R. Electrical/electrochemical impedance for rapid detection of foodborne pathogenic bacteria. *Biotechnol. Adv.* **26**, 135–150 (2008).
- Barreiros dos Santos, M. *et al.* Highly sensitive detection of pathogen *Escherichia coli* O157:H7 by electrochemical impedance spectroscopy. *Biosens. Bioelectron.* **45**, 174–180 (2013).
- Hassan, A. H. A., de la Escosura-Muñiz, A. & Merkoçi, A. Highly sensitive and rapid determination of *Escherichia coli* O157:H7 in minced beef and water using electrocatalytic gold nanoparticle tags. *Biosens. Bioelectron.* **67**, 511–515 (2015).
- Ruan, C. M., Yang, L. J. & Li, Y. B. Immunobiosensor chips for detection of *Escherichia coli* O157:H7 using electrochemical impedance spectroscopy. *Anal. Chem.* **74**, 4814–4820 (2002).
- Li, F. *et al.* Detection of *Escherichia coli* O157:H7 using gold nanoparticle labeling and inductively coupled plasma mass spectrometry. *Anal. Chem.* **82**, 3399–3403 (2010).
- Wang, Y. X., Ping, J. F., Ye, Z. Z., Wu, J. & Ying, Y. B. Impedimetric immunosensor based on gold nanoparticles modified graphene paper for label-free detection of *Escherichia coli* O157:H7. *Biosens. Bioelectron.* **49**, 492–498 (2013).
- Shein, J. B., Lai, L. M. H., Eggers, P. K., Paddon-Row, M. N. & Gooding, J. J. Formation of efficient electron transfer pathways by adsorbing gold nanoparticles to self-assembled monolayer modified electrodes. *Langmuir* **25**, 11121–11128 (2009).
- Chazalviel, J. N. & Allongue, P. On the origin of the efficient nanoparticle mediated electron transfer across a self-assembled monolayer. *J. Am. Chem. Soc.* **133**, 762–764 (2011).
- Zayats, M., Baron, R., Popov, I. & Willner, I. Biocatalytic growth of Au nanoparticles: from mechanistic aspects to biosensors design. *Nano Lett.* **5**, 21–25 (2005).
- Wang, W. *et al.* Cascade signal amplification for ultra-sensitive impedimetric detection of DNA hybridization using a hairpin DNA as probe. *Electrochim. Acta* **78**, 377–383 (2012).
- Liu, Y., Gilchrist, A., Zhang, J. & Li, X. F. Novel alternatives to antibiotics: bacteriophages, bacterial cell wall hydrolases, and antimicrobial peptides. *Appl. Environ. Microbiol.* **74**, 1502–1507 (2008).

Acknowledgements

The authors gratefully acknowledge the financial support from the National Natural Science Foundation of China (Nos. 21175089 and 91332101), the Program for Changjiang Scholars and Innovative Research Team in University (IRT-14R33), the Program for Innovative Research Team in Shaanxi Province (No. 2014KCT-28), and the Natural Science Basic Research Plan in Shaanxi Province of China (Nos. 2013SZS08-Z01 and 2013SZS08-P01). This work was also supported in part by the National Institutes of Health NCI R01CA175480 (ZC).

Author Contributions

J.W., J.A., Y.Z. and X.G. performed all experiments. Q.G conducted the experiments and analyzed data. Q.G. and Z.C. wrote the paper. All authors reviewed the manuscript.

Additional Information

Supplementary information accompanies this paper at <http://www.nature.com/srep>

Competing financial interests: The authors declare no competing financial interests.

How to cite this article: Wan, J. *et al.* Signal-off impedimetric immunosensor for the detection of *Escherichia coli* O157:H7. *Sci. Rep.* **6**, 19806; doi: 10.1038/srep19806 (2016).

This work is licensed under a Creative Commons Attribution 4.0 International License. The images or other third party material in this article are included in the article's Creative Commons license, unless indicated otherwise in the credit line; if the material is not included under the Creative Commons license, users will need to obtain permission from the license holder to reproduce the material. To view a copy of this license, visit <http://creativecommons.org/licenses/by/4.0/>