SCIENTIFIC REPORTS

OPEN

Aza-heterocyclic Receptors for Direct Electron Transfer Hemoglobin Biosensor

Vinay Kumar^{1,2}, D. M. Nikhila Kashyap², Suraj Hebbar², R. Swetha³, Sujay Prasad³, T. Kamala⁴, S. S. Srikanta⁴, P. R. Krishnaswamy¹ & Navakanta Bhat¹

Received: 12 July 2016

Accepted: 05 January 2017

Published: 07 February 2017

Direct Electron Transfer biosensors, facilitating direct communication between the biomolecule of interest and electrode surface, are preferable compared to enzymatic and mediator based sensors. Although hemoglobin (Hb) contains four redox active iron centres, direct detection is not possible due to inaccessibility of iron centres and formation of dimers, blocking electron transfer. Through the coordination of iron with aza-heterocyclic receptors - pyridine and imidazole - we report a cost effective, highly sensitive and simple electrochemical Hb sensor using cyclic voltammetry and chronoamperometry. The receptor can be either in the form of liquid micro-droplet mixed with blood or dry chemistry embedded in paper membrane on top of screen printed carbon electrodes. We demonstrate excellent linearity and robustness against interference using clinical samples. A truly point of care technology is demonstrated by integrating disposable test strips with handheld reader, enabling finger prick to result in less than a minute.

The point-of-care (POC) glucometer is an excellent example of how a simple and elegant idea can impact millions of lives. Clark and Lyons, through their seminal publication on glucose oxidase electrodes, laid the foundation for modern POC glucose sensors¹. While most of the commercial glucose sensors are second generation, mediator assisted sensors, there has been substantial research in the recent past to enable third generation glucose sensors, by facilitating direct electron transfer between glucose oxidase enzyme and the working electrode^{2–4}. Although the research publications in the field of biosensors have grown exponentially, in the last five decades, the commercialization has alarmingly lagged behind^{5–9}. Many critical reviews and editorials have stressed the need for new strategies for biomolecule recognition, with an emphasis on novel receptor design for robust and reliable sensing^{5–9}.

The receptor in the biosensor should be specific to the biomolecule of interest in the pool of bio-fluid and should be capable of giving a signal through an appropriate transduction mechanism. Antibodies and enzymes have become the first choice as a receptor molecule in POC biosensors^{10,11}. While these biological receptors are highly sensitive and selective, when used with biochemistry analyzers in a pathology laboratory, their use in POC biosensors has significant limitations. By definition, a biosensor is a self-contained device¹² that can be used in POC settings without any special storage and handling requirements. The stability of antibodies and enzymes, after functionalization on disposable strips and cartridges, is still a serious problem in POC biosensors. Moreover, the accuracy is also affected by ambient temperature, humidity and pH variations^{10,13}.

A low cost and robust POC Hb biosensor can have a big impact on public health for the diagnosis of anemia, especially in underdeveloped countries of the world. There can be different reasons for anemia which include iron deficiency, B₁₂ deficiency, nutritional conditions or other chronic conditions such as chronic kidney disease. According to WHO guidelines, Hb as a single biomarker, can be used for the diagnosis of anemia¹⁴. Hb testing is also very important during pregnancy, since it has been associated with an increased risk of preterm delivery¹⁵. The WHO Global database on anaemia for 1993–2005, covering almost half the world's population, estimated the prevalence of anaemia at 25% worldwide while the prevalence is 43% in underdevelopment countries^{16,17}. Anaemia affects 1.62 billion people globally with about 293 million children of preschool age, 56 million pregnant women and 468 million non-pregnant women estimated to be anaemic. Africa and Asia account for more than

¹Centre for Nano Science and Engineering, Indian Institute of Science, Bangalore-560012, India. ²PathShodh Healthcare Pvt Ltd, Bangalore-560012, India. ³Anand Diagnostics Laboratory, Bangalore, India. ⁴Samatvam Diabetes Endocrinology Centre, Bangalore, India. Correspondence and requests for materials should be addressed to N.B. (email: navakant@ece.iisc.ernet.in)

Figure 1. Pyridine and imidazole based concept for electrochemical detection of Hb. (a) Hb after lysing. (b) 300 μ L lysed blood on carbon printed electrode. (c) CV of lysed blood. (d) Structure of Alkaline hematin. (e) 300 μ L of alkaline hematin on carbon printed electrode. (f) CV of alkaline hematin. (g) Pyridine hemichrome. (h) 300 μ L of pyridine hemichrome on carbon printed electrode. (i) CV of pyridine hemichrome. (j) dispersion of imidazole based chemistry on membrane laminated electrode. (k) Ready to use electrode with dried chemistry. (l) Imidazole-metHb hemichrome. (m) CV of imidazole hemichrome.

85% of the absolute anaemia burden in high-risk groups. In India, almost 58% of pregnant women are anaemic and it is estimated that anaemia is the underlying cause for 20–40% of maternal deaths. India contributes to about 80% of the maternal deaths due to anaemia in South Asia¹⁸.

Hb testing was one of the oldest test done in pathology labs, and has evolved in the last several decades. Sahli's acid hematin technique, Hoppe-Seyler's carbon monoxide hemoglobin technique, Lovibond-Drabkin's cyanmethemoglobin technique, Cyanide free SLS technique used in most of the hematology analyzers, are all based on colorimetric principle^{19–22}. Hb has attracted biosensor research community, for the past several years^{23,24}. Due to the presence of iron redox centre in the heme prosthetic group, electrochemical transduction seems to be a natural choice. Although Hb contains four iron (Fe^{+2}) atoms in its structure, the iron center is embedded deep inside the globin chains and it is very difficult for the iron centre to communicate with the electrode surface for any practical sensing applications. Mediators such as ferricyanide, methylene blue, methylene green^{25–27}, have also been employed for Hb biosensor. Ferricyanide, a redox active molecule, is an oxidizing agent for Hb and converts hemoglobin into methemoglobin. Hb is electrochemically assayed using ferricyanide as a mediator between hemoglobin molecule and electrode surface. The current at the electrode resulting from the oxidation of ferrocyanide to ferricyanide would be proportional to the concentration of Hb. Saithip Pakapongpan *et al.* described the electrochemical Hb biosensor based on methylene blue and carbon nanotube. Methylene blue is used as a mediator for Hb detection, where redox current of methylene blue depends on Hb concentration²⁶. Christopher *et al.* proposed electrochemical Hb biosensor based on Poly (methylene blue) as a redox mediator²². Optical techniques such as surface plasmon resonance (SPR) have also been used for the detection of hemoglobin concentration in human blood²⁸. The direct electron transfer of Hb embedded in novel electrode materials such as silver nanoparticles, dendrimers, graphene has been used to sense H_2O_2 and NO ^{29–31}. However, mediator free Hb biosensor, utilizing the direct electron transfer with heme centre has still been elusive.

Results and Discussions

Herein, we demonstrate a novel analytical method for POC electrochemical detection of metalloproteins such as Hb. We propose new receptors based on aza heterocyclic compounds - pyridine and imidazole - to facilitate direct electron transfer between iron centres of Hb and working electrode. Figure 1 illustrates the overall sensing scheme using pyridine based liquid chemistry (a through i) and imidazole based dry chemistry (j through n). Figure 1(c,f) show the cyclic voltammograms (CVs) on carbon screen printed electrodes obtained using lysed blood sample and alkaline hematin respectively from the same blood sample (*see Paragraph-1 in Supporting Information for experimental details*). It is evident from the Fig. 1(c) that the lysed blood (Hb) does not show any electrochemical activity on a carbon printed electrode but alkaline hematin shows an irreversible CV as shown in Fig. 1(f). Although the reduction current of alkaline hematin appears to be a potential candidate for Hb sensing, it is not so in practice. The reduction current peak of alkaline hematin saturates at very low Hb concentration (micro-molar) range which is well below clinically relevant value (*see Paragraph-1 in*

Figure 2. (a) CV of Pyridine hemichrome for different Hb conc. (b) Error bar plot of peak oxidation current Vs. Hb conc. (c) CA plot for Oxidation current Vs. Hb conc. (d) Standard deviations and Coefficients of Variation.

Supporting Information). The reason for this non-linear response is the formation of dimers of hematin in aqueous solution^{32,33}. The dimerization of hematin hinders the diffusion controlled electrochemical reaction, which is highly desirable for biosensing.

The coordination chemistry of iron porphyrins has attracted considerable interest because of the widespread occurrence of the hemeproteins³⁴. The origin of hemichromes was studied and classified in the late 60s. Hemichromes, with relatively high solubility in aqueous solutions, are low spin ferric compounds. They are formed when the 5th, 6th coordination positions of iron covalently attach with a ligand. Reversible hemichrome can be formed with the coordination of internal ligands such as distal histidine with 6th position of metal iron. Reversible hemichrome can convert back to the normal Hb biomolecule. The other types of hemichromes, called irreversible hemichromes, cannot be converted back to normal Hb. Irreversible hemichrome is characterized by the nitrogenous linkage at the 5th and 6th position of iron(III). Hemichromes have been shown to possess good diffusion controlled electrochemical activity without any complicated surface modifications³⁵. This unique property of hemichrome is used here to develop Hb biosensor. We propose to convert Hb, into hemichrome by coordinating with nitrogenous ligands such as pyridine and imidazole. Then we obtain a well behaved reversible CV as opposed to the irreversible CV of alkaline hematin [Fig. 1(i,n)]. Further, nitrogenous ligands enable unhindered electron transfer with electrodes, as will be illustrated later.

Pyridine liquid sensing chemistry based Hb detection. In pyridine based Hb detection, whole blood is lysed using DI water and is converted into alkaline hematin [Fig. 1(d)]. Alkaline hematin is then converted into pyridine hemichrome (pH = 12.9) after adding high concentration of pyridine in alkaline hematin solution [Fig. 1(g,i)]. The addition of DI water and pyridine solution dilutes the Hb concentration in original blood sample. This dilution is compensated after adding an extra Sigma Bovine hemin in the final solution, so that the final Hb concentration of the solution is 25 g/dL (see Paragraph-2 in Supporting Information for experimental details). 300 uL drop of different concentrations of Hb is applied at the carbon printed electrode and CVs are recorded for a wide range of Hb concentrations varying from 0.5 g/dl to 20.8 g/dl. Each concentration of Hb is tested three times on three different electrodes to analyze the repeatability of the assay. As seen in Fig. 2(a,b), the peak oxidation current is linearly proportional to the Hb concentration with $R^2 = 0.97$. Instead of peak oxidation current, if we monitor the oxidation current at a fixed oxidation voltage of -0.06 V, the repeatability and linearity is equally good with $R^2 = 0.96$ (see Fig. S11 in Supporting Information). Typically, Chronoamperometry (CA) is a preferred technique in POC devices. Hence we analyze the efficacy of CA in Hb sensing. As shown in Fig. 2(c), we get an excellent response with $R^2 = 0.98$. (see Fig. S12 in Supporting Information for CA graphs). The formation of pyridine hemichrome, which enables this sensing scheme, is verified using the UV-VIS spectroscopy [see Paragraph-3 in Supporting Information for details]. In pyridine hemichrome experiment, chemicals such

Figure 3. (a) SEM image of paper membrane dispensed with SDS-Imidazole chemistry (Inset EDX spectra of the membrane). (b) XPS wide band spectra of blank paper membrane and SDS-Imidazole paper membrane.

Figure 4. (a) CV of Imidazole hemichrome for different Hb conc. (b) Error bar plot of oxidation current Vs. Hb. (c) CA plot for Oxidation current Vs. Hb conc. (d) Standard deviation and Coefficients of Variation.

as NaOH and pyridine are used to convert Hb from lysed blood into pyridine hemichrome. In addition blood plasma may also contain many redox active molecules. Hence it is necessary to evaluate possible interference. The absence of redox peaks in CVs obtained in pure plasma solution, pyridine, NaOH confirmed the absence of any interference from the constituents of plasma, and other chemicals. The synthetic pyridine hemichrome samples prepared after adding pyridine into alkaline solution of Sigma Bovine hemin, produced redox peaks identical to human blood samples. This further confirmed that the proposed hemichrome conversion of Hb is solely responsible for the fully reversible redox activity [see Paragraph-3 and Fig. S3 in Supporting Information for details]. Figure 2(d) demonstrates that the standard deviation and coefficient of variations are well within the requirements for POC Hb biosensor. Thus the pyridine receptor chemistry is highly accurate and cost effective method for Hb sensing. But it would need some sample preparation protocol, due to its liquid nature.

Imidazole dry sensing chemistry based Hb detection. For a robust POC biosensor, it is desirable to have stable, dry sensing chemistry on disposable carbon printed electrode. In imidazole based dry chemistry, Sodium dodecyl sulphate (SDS) and imidazole are functionalized on a paper membrane. SDS lyses

Figure 5. (a) Multi analyte electrochemical POC reader. (b) Correlation between Hb values of Laboratory Gold standard and SDS-Imidazole Dry chemistry using POC device. (c) Error bar plot for Intraassay variability. (d) Statistics for Intraassay variability.

the red blood cells (RBC) and also converts Hb into methHb by denaturation of protein (see Paragraph-4 in Supporting Information for details). Imidazole coordinates with iron (III) in porphyrin structure [Fig. 1(m)] and forms imidazole-metHb hemichrome complex, which gives the reversible CV [Fig. 1(n)]. A paper membrane, dispensed with SDS and imidazole based sensing chemistry, is laminated on the carbon printed electrodes. (see Paragraph-5 in Supporting Information for details).

The paper membrane, functionalized with SDS-Imidazole sensing chemistry, is characterized by Scanning Electron Microscopy (SEM), Energy-dispersive X-ray spectroscopy (EDX) and X-ray Photoelectron Spectroscopy (XPS). Figure 3(a,b) show SEM, EDX and XPS spectra of the membrane after drying. The presence of sodium (Na) and nitrogen (N) in EDX spectra, clearly indicates signatures of SDS and imidazole. The XPS spectra shows the presence of sulphur (S), sodium (Na) and imidazole (see Paragraph-5 in Supporting Information for details).

300 μ L drop of whole blood sample is applied on the membrane laminated printed carbon electrodes and CVs are recorded for four different Hb concentrations. Each concentration of Hb is repeated three times on three different test strips to estimate intra-assay variation. As seen in the Fig. 4(a,b), the oxidation current is proportional to Hb concentration in actual blood samples with $R^2 = 0.92$. Figure 4(d) shows the standard deviation and coefficient of variation. The efficacy of SDS-imidazole dry chemistry has also been analyzed by CA (Fig. 4(c)). (see Fig. S13 in Supporting Information for CA graphs). From this data an algorithm is developed to map the oxidation current to the actual Hb concentration in blood sample. This algorithm is used to create a novel handheld POC Hb sensor (Fig. 5(a)). (see Paragraph-6 in Supporting Information for details on handheld reader and disposable test strips). We demonstrate excellent sensing capability of proposed POC sensor, through tests performed on 101 clinical samples (Fig. 5(b)). The POC tests on disposable test strips are performed using 70 μ L of whole blood, with a testing time of about 30 seconds. The pathology lab tests are done using photometric method on Sysmex auto-analyzer using SLS-Hb complex. An excellent correlation with $R^2 = 0.917$ is demonstrated between POC device and Sysmex results. The statistical bias was analyzed using Bland-Altman plot (see Fig. S14 in Supporting Information) and the POC device results are within 95% CI interval. (Figure 5(c,d) show the intra assay variability plot and statistics. For this analysis, each blood sample is tested using 3 different test strips. Further the intra-assay variability analysis indicates that the coefficient of variation is in the range of 0.5%

to 7.1%. In summary, this study represents a major step forward in realizing robust and scalable, direct electron transfer POC Hb sensor utilizing aza-heterocyclic receptors.

Methods

Reagents and chemicals. Sodium dodecyl sulfate (SDS) and bovine hemin (>90%) were procured from Sigma-Aldrich. Imidazole, Pyridine and Sodium Hydroxide were procured from Merck. Saline solution was purchased from Baxter (India) Pvt. Limited. All the above chemicals were commercially available and used as received without further purification.

Apparatus and Measurements. Electrochemical experiments were performed on CHI Electrochemical Workstation 660E. SEM images and EDS spectra were captured on Carl Zeiss Ultra 55 FESEM. XPS measurements were done on AXIS ULTRA X-ray Photoelectron Spectroscopy UV-Spectroscopic measurements were performed on Shimadzu UV-Vis-IR Spectrophotometer. Carbon screen printed electrodes were used from PINE Instruments USA and GSI Technologies, USA with carbon material as working and counter and Ag/AgCl (PINE) or carbon (GSI) as reference.

Pyridine hemichrome based sensing. In pyridine based Hb detection, 1.5 ml of whole blood (pH = 7.2) is lysed using 4 ml of cold DI water and is converted into alkaline hematin after adding 1.5 ml of 1 N NaOH. The pH of alkaline hematin solution is 12.6. Alkaline hematin is then converted into pyridine hemichrome (pH = 12.9) after adding 1.5 ml (12 M) of pyridine solution in alkaline hematin solution. The addition of DI water and pyridine solution dilutes the Hb concentration in original blood sample. This dilution is compensated after adding an extra 77 mg of Sigma Bovine hemin in the final solution, so that the final Hb concentration of the solution is 25 g/dL.

Imidazole hemichrome based sensing. The stock solution of 69 mM SDS and 7.4 M imidazole is prepared in DI water and the final pH of the solution was 10.9. Blank paper membrane is laminated on the carbon printed electrodes. A micro droplet of 100 μ L of SDS-imidazole based sensing chemistry is dispersed on paper membrane and dried at 30 °C for 90 minutes with a forced airflow at a velocity of 1.5 m/sec.

References

1. L. Clark, Jr. & C. Lyons. Electrode systems for continuous monitoring in cardiovascular surgery. *Ann. NY Acad. Sci.* **102**, 29 (1962).
2. J. Wang. Electrochemical Glucose Biosensors. *Chem. Rev.* **108**, 814–825 (2008).
3. J. T. Holland, C. Lau, S. Brozik, P. Atanassov & S. Banta. Engineering of glucose oxidase for direct electron transfer via site-specific gold nanoparticle conjugation. *J. Am. Chem. Soc.* **133**, 19262–19265 (2011).
4. K. Cammann *et al.* Chemical Sensors and Biosensors—Principles and Applications, *Angew. Chem. Int. Ed. Engl.* **30**, 516–539 (1991).
5. Jeffrey Kirsch, Christian Siltanen, Qing Zhou, Alexander Revzin & Aleksandr Simonian. Biosensor technology: recent advances in threat agent detection and medicine. *Chem. Soc. Rev.* **42**, 8733 (2013).
6. O. S. Wolfbeis. Probes, Sensors, and Labels: Why is Real Progress Slow? *Angew. Chem. Int. Ed.* **52**, 9864–9865 (2013).
7. M. C. Goldschmidt. Biosensors: Blessing or Bane? *Journal of Rapid Methods and Automation in Microbiology* **2**, 9–15 (1993).
8. R. W. Murray. “Cadmium horses” and glucose. *Anal Chemistry* **76**, 149 A (2004).
9. P. T. Kissinger. Biosensors—a perspective. *Biosensors and Bioelectronics* **20**, 2512–2516 (2005).
10. T. R. J. Holford, F. Davis & S. P. J. Higson. Recent trends in antibody based sensors. *Biosensors and Bioelectronics* **34**, 12–24 (2012).
11. J. D. Newmon & S. J. Setford. Enzymatic biosensors. *Molecular Biotechnology* **32**, 249–268 (2006).
12. D. R. Thevenot, K. Tóth, R. A. Durst & G. S. Wilson. Electrochemical Biosensors: Recommended Definitions and Classification. *Pure Appl. Chem* **71**, 2333–2348 (1999).
13. J. J. O'Malley & R. W. Ulmer. Thermal stability of glucose oxidase and its admixtures with synthetic polymers. *Biotechnology and Bioengineering* **XV**, 917–925 (1973).
14. WHO. Haemoglobin concentrations for the diagnosis of anaemia and assessment of severity. Vitamin and Mineral Nutrition Information System. Geneva, World Health Organization, (WHO/NMH/NHD/MNM/11.1) <http://www.who.int/vmnis/indicators/haemoglobin.pdf> (2011).
15. Theresa O. Scholl *et al.* Anemia, Iron and Pregnancy Outcome. *The Journal Of Nutrition*, American Society for Nutritional Sciences (2000).
16. E. McLean *et al.* Worldwide prevalence of anaemia, WHO Vitamin and Mineral Nutrition Information System, 1993–2005. *Public Health Nutrition*, doi: 10.1017/S1368980008002401, 1–11 (2008).
17. B. de Benoist, E. McLean, I. Egli & M. Cogswell. *Worldwide Prevalence of Anaemia 1993–2005*, ISBN 978 92 4 159665 7.
18. Government of India (GoI), New Delhi. Guidelines for Control of Iron Deficiency Anaemia, Ministry of health and family welfare, http://www.pbnrhm.org/docs/iron_plus_guidelines.pdf (2013).
19. H. Sahli. *Lehrbuch der klinischen untersuchungs-methoden*, Leipsic, 4th edition, 1905, 655664; 5th edition, 1909, 345; Verhandl. Kong. innere Med. 1902, xx, 230; Jahresb. For&h. Thierchem., 1903, xxxii, 222.
20. T. Srivastava, H. Negandhi, S. B. Neogi, J. Sharma & R. Saxena Methods for Hemoglobin Estimation: A Review of “What Works”. *J Hematol Transfus* **2**(3), 1028 (2014).
21. Frieda S. Robscheit. A comparative study of hemoglobin determination by various methods. *The Journal of Biological Chemistry* Vol. XLI, NO-2 (3 Nov, 1919).
22. Christopher M.A. Brett. Poly(methylene blue) modified electrode sensor for haemoglobin. *Analytica Chimica Acta* **385**, 119–123 (1999).
23. F. W. Scheller, N. Bistolas, S. Liu, M. Ja’nchen, M. Katterle & U. Wollenberger. Thirty years of haemoglobin electrochemistry. *Advances in Colloid and Interface Science* **116**, 111–120 (2005).
24. Y. Wu & S. Hu. Biosensors based on direct electron transfer in redox proteins. *Microchim Acta* **159**, 1–17 (2007).
25. M. J. Green & Roger J. Lias. Electrochemical assay for haemoglobin. *US4876205 A* (1989).
26. Saithip Pakapongpana, Rungtiva Palangsantikula & Werasak Surareungchaia. Electrochemical sensors for hemoglobin and myoglobin detection based on methylene blue-multiwalled carbon nanotubes nanohybrid-modified glassy carbon electrode. *Electrochimica Acta* **56**, 6831–6836 (2011).
27. Yimin Zhu & Shaojun Dong. Rapid redox reaction of hemoglobin at methylene green modified platinum electrode. *Electrochim Acta* **35**, 1139 (1990).
28. Anuj K. Sharma. Plasmonic biosensor for detection of hemoglobin concentration in human blood: Design considerations. *J. Appl. Phys.* **114**, 044701 (2013).

29. X. Gan, T. Liu, X. Zhu & G. Li. An electrochemical biosensor for nitric oxide based on silver nanoparticles and hemoglobin. *Analytical Sciences* **20**, 1271 (2004).
30. Yaping He, Qinglin Sheng, Jianbin Zheng, Minzhi Wang & Bin Liu. Magnetite-graphene for the direct electrochemistry of hemoglobin and its biosensing application. *Electrochimica Acta* **56**, 2471–2476 (2011).
31. M. Baccarin, B. C. Janegitz, R. Berté, F. C. Vicentini, C. E. Banks, O. Fatibello-Filho & V. Zucolotto. Direct electrochemistry of hemoglobin and biosensing for hydrogen peroxide using a film containing silver nanoparticles and poly(amidoamine) dendrimerMaterials. *Science and Engineering C* **58**, 97–102 (2015).
32. J. Jordan & T. M. Bednarski. Polarography of Hemin. Evidence for a Two-Electron Transfer. *J. Am. Chem. Soc.* **86**, 5690 (1964).
33. T. M. Bednarski & J. Jordan. Electron-Transfer Characteristics of the Prosthetic Group of Hemoproteins. *J. Am. Chem. Soc.* **89**, 1552–1558 (1967).
34. J. P. Greer, D. A. Arber, B. Grader, A. F. List, R. T. Means j., F. Paraskevas & G. M. Rodgers. Wintrobe's clinical hematology, 13th Edition (Wolters Kluwer) pp 231–232 (Nov, 2013).
35. D. G. Davis & D. J. Orleon. Cyclic Voltammetry of Some Iron Porphyrin Complexes. *Analytical Chemistry* **38**, 179–183 (1966).

Acknowledgements

We thank DeitY and IKP Knowledge Park, BIRAC-DBT for funding support. We thank the technical staff members at Micro and Nano Characterization Facility (MNCF), CeNSE, IISc.

Author Contributions

Vinay Kumar (V.K.) and Navakanta Bhat (N.B.) conceived the idea and outlined the plan of experiments. V.K. and N.B. performed extensive analysis on all the experimental results and wrote the manuscript. V.K. also performed SEM, XPS characterization. Nikhila Kashyap D.M. (N.K.D.M.) along with V.K. performed all the experiments related to electrochemistry, and UV spectroscopy. N.K.D.M. also contributed in creating graphics for various molecules presented in the manuscript. Suraj Hebbar (S.H.) and N.B. conceived and designed the handheld PoC reader and S.H. performed extensive experiments to validate the reader and interface the reader to test strips. P. R. Krishnaswamy (P.R.K.) planned and coordinated the PoC sensor validation study with two clinical centres and designed the experiments on choosing the appropriate clinical samples and statistical data analysis protocol. Swetha R. (S.R.) and Sujay Prasad (S.P.) from Anand Diagnostics Laboratory along with Kamala T. (K.T.) and S. S. Srikantha (S.S.S.) from Samatvam Diabetes Endocrinology Centre performed all the tests on clinical samples and provided their inference for comparing the results obtained by PoC reader and Sysmex auto analyzer. All authors contributed in reviewing the manuscript and provided their valuable inputs.

Additional Information

Supplementary information accompanies this paper at <http://www.nature.com/srep>

Competing financial interests: The project was fully supported by government funding agencies and there was no private funding for the project. Four of the authors (V.K., N.B., N.K.D.M., S.H.) are also part of a start-up, PathShodh Healthcare, incubated at the Indian Institute of Science, Bangalore, to translate this technology to a useful product for the society.

How to cite this article: Kumar, V. *et al.* Aza-heterocyclic Receptors for Direct Electron Transfer Hemoglobin Biosensor. *Sci. Rep.* **7**, 42031; doi: 10.1038/srep42031 (2017).

Publisher's note: Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.

This work is licensed under a Creative Commons Attribution 4.0 International License. The images or other third party material in this article are included in the article's Creative Commons license, unless indicated otherwise in the credit line; if the material is not included under the Creative Commons license, users will need to obtain permission from the license holder to reproduce the material. To view a copy of this license, visit <http://creativecommons.org/licenses/by/4.0/>

© The Author(s) 2017